

US Patent & Trademark Office

Patent Public Search | Text View

United States Patent	12396839
Kind Code	B2
Date of Patent	August 26, 2025
Inventor(s)	Connor; Patrick M. et al.

Method and apparatus for creating a seam-like anatomical low creep attachment of soft tissue to bone

Abstract

Methods and apparatus for creating a seam like array of suture stitches joined in series by sequential anchors in bone. The anchors are transtendon implants positioned through the tendon in the original tendon footprint and each array includes four or more anchors. The seam like array extends across at least a portion of the tendon. The stitches can extend generally perpendicular to the direction of the tendon orientation. In rotator cuff repair, the anchors are positioned in the original footprints of the infraspinatus and/or supraspinatus tendons and the seam like array extends in an anterior to posterior direction across the particular tendons torn and repaired. The entire array of anchors can be positioned in the medial half of the original footprints.

Inventors: Connor; Patrick M. (Charlotte, NC), Harris; Howard W. (Southlake, TX), Labbé; Marc (Spring, TX), Westling; Thomas A. (Orono, MN), Zenz-Olson; Zak (Ham Lake, MN), Van Tran; Nathaniel (Lakeville, MN), Crompton; David M. (St. Paul, MN)

Applicant: Integrity Orthopaedics, Inc. (Orono, MN)

Family ID: 1000008779417

Assignee: Integrity Orthopaedics, Inc. (Maple Plain, MN)

Appl. No.: 17/681430

Filed: February 25, 2022

Prior Publication Data

Document Identifier	Publication Date
US 20220323200 A1	Oct. 13, 2022

Related U.S. Application Data

Publication Classification

Int. Cl.: **A61F2/08** (20060101); **A61B17/04** (20060101)

U.S. Cl.:

CPC **A61F2/0811** (20130101); **A61B17/0401** (20130101); A61B2017/0414 (20130101)

Field of Classification Search

CPC: A61B (17/0401); A61B (2017/0404); A61B (2017/0406); A61B (2017/0408); A61B (2017/0414); A61B (2017/0417); A61B (2017/0419); A61B (2017/0445); A61B (2017/0446); A61B (2017/0458); A61F (2/0811)

References Cited

U.S. PATENT DOCUMENTS

Patent No.	Issued Date	Patentee Name	U.S. Cl.	CPC
4899743	12/1989	Nicholson et al.	N/A	N/A
4968315	12/1989	Gatturna	N/A	N/A
5041129	12/1990	Hayhurst et al.	N/A	N/A
5192303	12/1992	Gatturna et al.	N/A	N/A
5203787	12/1992	Noblitt et al.	N/A	N/A
5207679	12/1992	Li	N/A	N/A
5269809	12/1992	Hayhurst et al.	N/A	N/A
5403348	12/1994	Bonutti	N/A	N/A
5405359	12/1994	Pierce	N/A	N/A
5464426	12/1994	Bonutti	606/232	A61B 17/0401
5468197	12/1994	Loeffler	N/A	N/A
5584835	12/1995	Greenfield	N/A	N/A
5626614	12/1996	Hart	N/A	N/A
5669935	12/1996	Rosenman et al.	N/A	N/A
5683418	12/1996	Luscombe et al.	N/A	N/A
5725557	12/1997	Gatturna et al.	N/A	N/A
5728100	12/1997	Skiba	N/A	N/A
5741300	12/1997	Li	N/A	N/A
5810848	12/1997	Hayhurst	N/A	N/A
5891168	12/1998	Thal	606/232	A61B 17/0401
5921986	12/1998	Bonutti	N/A	N/A
5948002	12/1998	Bonutti	N/A	N/A
5961538	12/1998	Pedlick et al.	N/A	N/A
6056773	12/1999	Bonutti	N/A	N/A
6066160	12/1999	Colvin et al.	N/A	N/A
6117161	12/1999	Li et al.	N/A	N/A

6149669	12/1999	Li	606/68	A61B 17/0401
6270518	12/2000	Pedlick et al.	N/A	N/A
6293961	12/2000	Schwartz et al.	N/A	N/A
6451030	12/2001	Li et al.	N/A	N/A
6500195	12/2001	Bonutti	N/A	N/A
6520980	12/2002	Foerster	N/A	N/A
6527795	12/2002	Lizardi	N/A	N/A
6547800	12/2002	Foerster et al.	N/A	N/A
6592609	12/2002	Bonutti	N/A	N/A
6645227	12/2002	Fallin et al.	N/A	N/A
6660023	12/2002	McDevitt et al.	N/A	N/A
6726707	12/2003	Pedlick et al.	N/A	N/A
6773436	12/2003	Donnelly et al.	N/A	N/A
6843799	12/2004	Bartlett	N/A	N/A
6972027	12/2004	Fallin et al.	N/A	N/A
6986781	12/2005	Smith	N/A	N/A
7041120	12/2005	Li et al.	N/A	N/A
7081126	12/2005	McDevitt et al.	N/A	N/A
7232455	12/2006	Pedlick et al.	N/A	N/A
7320701	12/2007	Haut et al.	N/A	N/A
7556640	12/2008	Foerster	N/A	N/A
7566339	12/2008	Fallin et al.	N/A	N/A
7641672	12/2009	Fallin et al.	N/A	N/A
7645293	12/2009	Martinek et al.	N/A	N/A
7674275	12/2009	Martin et al.	N/A	N/A
7682374	12/2009	Foerster et al.	N/A	N/A
7722644	12/2009	Fallin et al.	N/A	N/A
7806909	12/2009	Fallin et al.	N/A	N/A
7875064	12/2010	Donnelly et al.	N/A	N/A
7909851	12/2010	Stone et al.	N/A	N/A
7963972	12/2010	Foerster et al.	N/A	N/A
8052719	12/2010	Paulos	N/A	N/A
8118835	12/2011	Weisel et al.	N/A	N/A
8298262	12/2011	Stone et al.	N/A	N/A
8348975	12/2012	Dreyfuss	N/A	N/A
8366744	12/2012	Bojarski et al.	N/A	N/A
8419794	12/2012	ElAttrache et al.	N/A	N/A
8425536	12/2012	Foerster et al.	N/A	N/A
8449584	12/2012	Donnelly et al.	N/A	N/A
8454655	12/2012	Yeung et al.	N/A	N/A
8512375	12/2012	Torrie et al.	N/A	N/A
8771314	12/2013	Crombie et al.	N/A	N/A
8777992	12/2013	Yeung et al.	N/A	N/A
8828052	12/2013	Caborn et al.	N/A	N/A
8828054	12/2013	Caborn	606/232	A61B 17/0483
8845699	12/2013	Bonutti	N/A	N/A
8932331	12/2014	Kaiser et al.	N/A	N/A
8951287	12/2014	Green et al.	N/A	N/A
8986346	12/2014	Dreyfuss	N/A	N/A
9072509	12/2014	Stoll, Jr. et al.	N/A	N/A

9173651	12/2014	Stone et al.	N/A	N/A
9192369	12/2014	Bittenson	N/A	N/A
9216036	12/2014	Johnstone	N/A	N/A
9220493	12/2014	Hart et al.	N/A	N/A
9265495	12/2015	Petersen et al.	N/A	N/A
9271714	12/2015	Martin	N/A	N/A
9301756	12/2015	Wardle	N/A	N/A
9307979	12/2015	Bennett et al.	N/A	N/A
9314238	12/2015	Martin	N/A	N/A
9345467	12/2015	Lunn et al.	N/A	N/A
9451945	12/2015	Hawkins	N/A	N/A
9463008	12/2015	Thal	N/A	N/A
9504462	12/2015	Dooney, Jr. et al.	N/A	N/A
9526489	12/2015	Burkhart	N/A	N/A
9532777	12/2016	Kaiser et al.	N/A	N/A
9539000	12/2016	Hendricksen et al.	N/A	N/A
9545251	12/2016	Bojarski et al.	N/A	N/A
9597070	12/2016	Bittenson	N/A	N/A
9655611	12/2016	Green et al.	N/A	N/A
9693765	12/2016	Sullivan et al.	N/A	N/A
9713463	12/2016	Oren et al.	N/A	N/A
9763719	12/2016	Snyder et al.	N/A	N/A
9814565	12/2016	Foerster et al.	N/A	N/A
9872678	12/2017	Spenciner et al.	N/A	N/A
9931150	12/2017	Philippon et al.	N/A	N/A
10130354	12/2017	Dooney, Jr.	N/A	N/A
10172607	12/2018	Burkhart	N/A	N/A
10178989	12/2018	Bennett et al.	N/A	N/A
10285684	12/2018	Spenciner et al.	N/A	N/A
10368856	12/2018	Stone et al.	N/A	N/A
10376260	12/2018	Bojarski et al.	N/A	N/A
10398428	12/2018	Denham et al.	N/A	N/A
10478172	12/2018	Williams et al.	N/A	N/A
10543075	12/2019	Gregoire et al.	N/A	N/A
10575842	12/2019	Lund	N/A	N/A
10588614	12/2019	Gittings et al.	N/A	N/A
10603029	12/2019	Kaiser et al.	N/A	N/A
10667803	12/2019	Lizardi	N/A	N/A
10675015	12/2019	Guo et al.	N/A	N/A
10729421	12/2019	Stone et al.	N/A	N/A
10772622	12/2019	Santangelo et al.	N/A	N/A
10786235	12/2019	Sorensen et al.	N/A	N/A
10863979	12/2019	Sorensen et al.	N/A	N/A
10952719	12/2020	Lombardo et al.	N/A	N/A
10966704	12/2020	Lozier	N/A	A61B 17/0483
10987099	12/2020	Stone et al.	N/A	N/A
2003/0088272	12/2002	Smith	606/232	A61B 17/0401
2006/0178702	12/2005	Pierce et al.	N/A	N/A
2009/0312782	12/2008	Park	N/A	N/A
2013/0190815	12/2012	Mansmann	N/A	N/A

2015/0250470	12/2014	Vargas	606/232	A61B 17/0401
2020/0253715	12/2019	Trenhaile	N/A	N/A
2020/0315775	12/2019	Pilgeram et al.	N/A	N/A

Primary Examiner: Dornbusch; Dianne

Attorney, Agent or Firm: Seager, Tufte & Wickhem LLP

Background/Summary

CROSS REFERENCE TO RELATED APPLICATIONS (1) The present application claims the benefit of and priority to U.S. Prov. Pat. App. No. 63/172,614, filed Apr. 8, 2021, titled METHOD FOR CREATING A TENSIONABLE AND LOCKABLE SUTURE ANCHOR ARRAY FOR ANATOMICAL ATTACHMENT OF TISSUE TO BONE, U.S. Prov. Pat. App. No. 63/281,411, filed Nov. 19, 2021, titled DELIVERY DEVICE FOR IMPLANTING KNOTLESS MICRO-SUTURE ANCHORS AND ANCHOR ARRAYS FOR ATTACHMENT OF SOFT TISSUE TO BONE and is a continuation-in-part of U.S. patent application Ser. No. 17/551,779, filed Dec. 15, 2021, now U.S. Pat. No. 11,382,612, titled Method FOR CREATING A TENSIONABLE AND LOCKABLE SUTURE ANCHOR ARRAY FOR ANATOMICAL ATTACHMENT OF TISSUE TO BONE, the disclosures of which are incorporated herein by reference.

BACKGROUND

(1) Throughout the human body there are many attachments of soft tissue, such as tendons and ligaments, to bone as integral elements of motion in functioning joints such as the shoulder. The shoulder joint includes the humeral head of the upper arm bone in contact with the indentation of the glenoid working in conjunction with the rotator cuff, which is a combination of muscles and tendons forming a capsule that both stabilizes the joint and causes desired motion. Injury to the connection between tendons of the rotator cuff muscles to the humeral head, usually a tear in a tendon, is common. These tears do not self-heal. It is estimated that in the U.S. over 4 million people annually are referred to a surgeon due to shoulder pain and over 500,000 of these referrals result in shoulder surgery to repair the rotator cuff.

(2) Significant effort has been expended over the past 30 years to develop bone and tissue anchor devices and methods to respond to the need for effective rotator cuff repair. Early methods and devices utilized an open surgical technique that required a large incision of 4 to 6 cm and cutting the deltoid muscle, then re-attaching after the rotator cuff repair. This method is still used today for massive tears by some surgeons due to high success rate, however, the procedure is associated with deltoid dysfunction, significant pain during recovery and extensive rehabilitation time. Due to the invasiveness of the open surgery and resulting rehabilitation time, a “mini-open” procedure and associated devices were developed in the early 1990's, wherein the surgeon uses partial arthroscopic techniques followed by an incision and split of the deltoid muscle fibers to access the rotator cuff tendon for repair. By the late 1990's, devices and instruments were further developed to complete the repair of rotator cuff tendon attachment to bone using all-arthroscopic techniques, with further resultant reduction in trauma and recovery time.

(3) Arthroscopic repair of the rotator cuff tendon attachments to the humeral head are the most common technique used today. However, it is recognized that these all-arthroscopic techniques are quite difficult to perform and achieve varying results. The skill of the surgeon with the technology available is a known factor related to the procedure's success. Even with the last 20 years of all-arthroscopic technologic advancement and experience, deficiencies persist as evidenced by studies indicating an overall average rotator cuff repair failure rate of 20% to 40%, with a highly variable

range of 4% to 90% in individual studies. The study results indicate failure rates are much higher for large or massive tendon tears and there are vast variations in failure rates between surgeons, as well as with respect to various patient factors, equipment used, and type of repair completed.

(4) There is significant controversy among professionals as to the reasons for the high incidence of arthroscopic rotator cuff repair failure (i.e. "re-tear of the rotator cuff"). However, studies clearly show there is a need to reduce the failure rate of arthroscopic rotator cuff repair to avoid its effects of patients' lack of mobility, functional deficits, increased pain and/or requiring subsequent and more invasive surgery with the attendant pain and rehabilitation. In particular, there is great concern for patients who have some degree of native tendon or repair tendon failure yet choose to "live with it" rather than going through a first or additional surgery and rehabilitation, thus affecting quality of life and promoting continued joint degradation from lack of use.

(5) The basic device or devices used for repair of a tendon torn from a bone is one or more suture anchors in which a mechanical structure provides an anchor to the bone and a suture or sutures extend therefrom for attachment to the soft tissue or tendon. Many types of anchor technologies have been proposed and used in procedures. A review of the prior art patent literature indicates over a thousand designs for suture anchors, bone anchors, tendon repair systems, delivery devices and methods espousing improved features over the past 25 years, yet repair failure rate is still unacceptable indicating the need for further improvement in the area of arthroscopic reattachment of tendons to bone and in particular in rotator cuff repair.

(6) Applicants have identified at least two major issues with prior suture anchors and methods of repairing rotator cuff tears. First, prior repairs are not anatomical, meaning the tendon is pulled down onto the bone in a different position and way than the original tendon attachment. With prior methods, the re-attached tendon can be offset from the original footprint of prior attachment and can be laid down in a way that include folds and creases not previously present at the interface between tendon and bone. These issues inhibit healing attachment. The non-anatomical repair is due to the fact that existing anchors with sutures extending therefrom are implanted into bone through the hole formed by the torn tendon. After the anchor is implanted, the surgeon uses suture passers to get ends of the suture through the tendon for securing the tendon. However, this requires that the surgeon approximate where to pass the suture through the tendon so that it pulls the tendon to the correct spot on the footprint. This must be done with each suture and results in tendons being pulled down in ways not matching the original anatomical structure and includes creases in the tendon and gaps where the tendon is not compressed to the bone.

(7) Second, the passed sutures must extend from just a few anchor points within the original tendon footprint due to the size of prior anchors and the need to implant through the hole in the tendon because prior anchors are not implanted through a tendon due to difficulty and potential tendon damage. With prior systems the number of points of attachment in the original footprint are few and there is significant distance across each bridge from anchor to anchor. This results in weak compression at limited locations so that the tendon is not firmly or robustly held against the bone and during use the tendon moves relative to the bone (cyclic creep) which interrupts or prevents healing.

Overview

(8) The present inventors have recognized, among other things, that a problem to be solved is the need for new and/or alternative devices and methods for arthroscopically affixing a tendon or other soft tissue to bone such as in rotator cuff repair with low failure rate, preferably under 10% on average, with little variation between surgeons, patient characteristics, and the system/method used for repair. The disclosed devices, systems, and methods, along with a statement of the problem being solved by each element are included in summary form followed by a description of specific claimed structure or methods in the present disclosure. Importantly, the invention is directed to a system and method wherein the implanted array of anchors, with a continuous set of anchor-to-anchor serial single suture stitches therebetween, creates a seam-like attachment of tendon to bone

akin to a seam created by a sewing machine in fabric.

(9) The seam-like attachment is anatomic in that the tendon area that was torn from the humeral head is laid down in continuous contact with the original footprint on the humeral head without creases and folds. Anchors are then implanted through the tendon (transtendon) as properly positioned to maintain flat continuous contact with the bone. Further, as the anchors do not require installation through the hole formed by the torn tendon, there is not the same limit on the number of anchors that can be utilized to get anatomical attachment. Also, there is no suture passing (wherein the tendon is pulled down onto bone based on an approximation of proper position) because the suture is inserted through the tendon into the bone beneath it.

(10) The system also uses a single suture strand (sometimes referred to as the working suture) in each array serially connecting a significantly increased number of anchors rather than multiple strands of suture extending from a few anchors to be attached at angles to distant anchors outside the footprint. This results in stronger, more vertical compression at increased numbers of locations so that the tendon is firmly and robustly held against the bone. Post-surgery, the tendon has little or no cyclic creep (movement relative to the bone) to maintain healing contact at the bone and tendon interface.

(11) The combination of discontinuous tendon contact with the humeral head and micromotion of the tendon on the humeral head during use are believed to be significant sources of failure to heal properly in prior repairs with subsequent re-tear or failure. To address these issues the currently disclosed system and method utilizes six features and/or implantation techniques to assure continuous tendon/bone contact and minimal micromotion. These include: 1. The implants are designed for and implanted transtendon (through the positioned tendon). Each of the anchors is implanted within the original tendon footprint where bone quality is generally better for an individual patient. In rotator cuff repair, the anchors are preferably implanted in the original footprints of the supraspinatus and infraspinatus tendons with preferred location in the medial half of the original footprint. 2. The anchors are implanted in a row across at least a portion of the tendon with serial stitches extending continuously from one anchor to the next to create the seam like array of stitches. In the rotator cuff, the seam generally extends in the anterior to posterior direction across the tendon which can be generally perpendicular to the tendon fiber orientation or direction of tendon function. 3. The anchor insertion size is small so that they can be implanted in a high-density array with significantly greater numbers within the original footprint in close proximity to each other. Preferred distance between anchors is less than 7 millimeters (mm) from edge of bone hole to edge of bone hole or less than 10 mm from center of bone hole to center of adjacent bone hole. With this spacing the anchors are sized to fit within bone holes less than or equal to about 3 mm in diameter. The high density and close proximity of anchors in the original footprint creates short serial stitches between sequential anchors that apply force more vertically to hold the tendon firmly against bone within the original tendon footprint. 4. Each anchor-to-anchor stitch is tensioned at the time of implantation to sequentially compress the tendon flat on the bone. This tension is retained during subsequent anchor implantation until the entire array is implanted and reduces the chance of creating areas of insufficient compression, bunching, or non-contact. 5. The system has a minimum of four anchors forming each serial array with a single common working suture forming stitches between anchors. In preferred systems a single serial array includes a minimum of 4 and up to 8 anchors. Further, multiple arrays can be implanted in larger or massive tears. 6. Each system includes structure to lock or retain tension in the stitches, either individually as each anchor is placed, or as a set once an array of anchors is implanted that is sufficient to handle loads experienced during normal activity.

(12) The present disclosure includes many types of anchors that can be combined with sutures to form implanted arrays accomplishing the above objectives. Further the anchors and sutures can be arranged in pre-strung arrays for easier implantation with mechanisms or means to tighten, hold and/or lock each suture stitch to create that final stable implanted array that retains tension on each

of the suture stitches to hold the tendon in position during movement.

(13) It is recognized in the art that rotator cuff tears are classified into four categories based on tear size and whether a single row or double row repair is completed. Small tears are less than 1 centimeter (cm) in length; medium tears are 1 cm to 3 cm in length; large tears are 3 cm to 5 cm in length and massive tears are greater than 5 cm in length. With prior art devices, surgeons are limited to available large anchors and by the size of the tear as the medial anchors must fit in the tear area that exposes bone. For example, surgeons may use about 1 medial anchor on small tears, 1 or 2 medial anchors on medium tears and 2 or 3 medial anchors on large tears and massive tears. With the high anchor density anatomical repair of the present application, the surgeon is not limited by tear size as the anchors are implanted through the tendon and can use greater than 3 medial anchors on small tears, greater than 5 medial anchors on medium tears, and greater than 6 medial anchors on large tears and massive tears. This can include positioning implants outside the area of a full thickness tear to reinforce areas of partial thickness tears or weaker untorn tendon. For some tears, two rows of anchor may be placed 1-3 cm apart with the present system, if desired. The high-density array of anchors is formed by implantation of the anchors in a chain or row which can be a relatively straight line or curve depending upon the tear to be repaired and the discretion of the surgeon.

(14) One exemplary delivery device system designed for sequential transtendinous implantation of each anchor in the array is disclosed herein as well. The delivery system includes a delivery tool distal portion to be used at the surgical site for implantation of the array, and a proximal portion having a handle and features for managing the anchors and associated sutures and suture lock or other mechanism. The distal portion of the delivery tool includes an anchor delivery tube sized to allow passage of an anchor and associated working suture and suture lock therethrough. The delivery tool is used with a bone punch that is sized as well for passage through the anchor delivery tube. The proximal portion of the delivery tool is configured to allow a physician to introduce an anchor that is pre-strung onto the working suture into the anchor delivery tube. The proximal portion of the delivery tool may include a platform for receiving a magazine carrying a number of cartridges that house the pre-strung anchors individually. The magazine may include a cartridge ejector that allows one cartridge at a time to be removed from the magazine and placed in a slot on the delivery tool. A plunger is used to transfer individual anchors from the cartridge to a lumen at the proximal end of the anchor delivery tube.

(15) In use, the physician places the distal end of the delivery tool at a desired location for introduction of an anchor. Such placement may be performed with the bone punch extending past the distal end of the delivery tool to allow a physician to probe the desired location using the bone punch. The physician then presses the delivery tool distal end against the tendon and applies a force, such as by pounding, against the proximal end of the bone punch to create a path through the tendon and then to create a bone hole. The distal end of the anchor delivery tube, referred to as a nub, may be advanced through the tendon and at least partly into the bone hole as the bone punch is pounded.

(16) The bone punch is then retracted, while the nub is kept in place to maintain registration through the tendon and into the bone hole. With the bone punch retracted, a cartridge is removed from the magazine using the cartridge ejector and transferred to the slot on the proximal portion of the delivery tool, and the plunger is depressed to move an anchor from the cartridge into position for advancement into the anchor delivery lumen. The bone punch is again advanced, this time pressing against the proximal end of the anchor, eventually ejecting the anchor from the anchor delivery tube into the bone hole. As the bone punch pushes the anchor down the anchor delivery tube, the tip of the bone punch can be engaged with a dimple in the anchor, especially if it is a toggle type anchor. If the anchor is a soft anchor, an anchor advancing tool having a slot for receiving the soft anchor may be used rather than the bone punch during advancement of the soft anchor, in some examples.

(17) When the anchor exits the anchor delivery tube within the bone, the anchor must be toggled, oriented or activated so that it is retained in the bone when the working suture is tensioned. Several exemplary anchor types are described below that include differing ways to activate or orient for bone retention. For example, toggle type anchors are rotated (toggled) so their length lays across the punched hole, soft anchors are expanded so they cannot exit the punched hole, and other anchors can include external flanges or other projections that are extended within the bone or bone hole to prevent pullout. The bone punch or anchor advancing tool is advanced so that the tip thereof extends beyond the nub, forcing the anchor into the bone. The bone punch or anchor advancing tool is then retracted into the anchor delivery tube, and the working suture is manipulated to continue orienting or activating the anchor into a preferred retained position. To prevent interference between the anchor and the anchor delivery tube and/or damage to the working suture during orientation of the anchor, the anchor delivery tube may be retracted so that the nub is within the delivery tool. In addition, retracting the anchor delivery tube and/or nub can reduce flossing tension, allowing flossing of the working suture until tensioned and locked. During orienting or activating the anchor, the anchor delivery tube may be pressed against the tendon to prevent or discourage backing out of the anchor and/or fracture of the bone under the applied forces.

(18) In some examples, the first anchor may be affixed to the working suture, and the first anchor omits any locking suture or other means for actively locking the working suture to first anchor. When the first anchor is set in sufficiently strong material inside the bone (which can be harder cancellous bone or may be resting against the under surface of the cortical shell) the delivery device can be set with the punch pin partially extended as it was at the beginning of the procedure and moved for implantation of the next anchor.

(19) With the second and subsequent anchors, both a proximal and a distal suture portions of the working suture extend up through the delivery device. It is recognized that the proximal portion and distal portion of the working suture can be tensioned in some embodiments to aid in rotating and/or seating the anchor in proper position within the bone hole. Prior to such tensioning of the working suture, the nub may be retracted into the outer tube of the delivery tool to prevent interference with or damage to the working suture and/or locking loop (when provided). During setting and orientation of the anchor and subsequent tensioning of the suture, the distal end of an outer tube of the delivery tool may be pressed against the tendon to provide a counterforce against pullout. This is continued until the properly tensioned suture stitch is formed.

(20) This is repeated for a desired number of anchors in the pre-strung chain which is implanted to form a high-density array as described above. As can be understood, the number of suture stitches formed is equal to the number of anchors in the chain implanted minus 1. Further, the string of stitches is serially continuous with each stitch tensioned independently to form a required robust tendon attachment. The continuous string of stitches can form a row or chain of stitches of desired shape. By row or chain, it is meant that the suture stitches extend from one anchor to the next in the sequence of implanted anchors. It is understood that more than one continuous string of stitches can be formed by implanting multiple anchor arrays that together form an overall repair array, especially for large tears.

(21) As previously stated, the distance between ends of a suture stitch (the distance between edges of anchor holes) is preferably less than about 7 mm (less than about 10 mm from center of anchor insertion hole to center of anchor insertion hole) to provide consistent force on the tendon against the bone to reduce creep. One particularly robust array of implanted anchors for rotator cuff repair includes a first array implanted in a medial portion of the original tendon footprint of the supraspinatus and infraspinatus tendons to form a row or line of stitches generally perpendicular to the length or direction of the tendon's forces. A second array can then be implanted laterally nearer the edge of the tear with at least one anchor through the tendon while at least one other anchor is implanted laterally of the tendon edge to reapproximate the tendon properly against the bone. The lateral row can be implanted in a zig zag pattern or other appropriate pattern based on the shape of

the tear. Depending upon tear size and location, multiple patterns can be utilized.

(22) As becomes clear in the above description, the pre-strung array of anchors in combination with the working suture and multiple locking sutures creates a strong need for a delivery system that has components that manage the anchors and their attendant sutures or suture sections to maintain orderly implantation, use and sterility during a procedure. Further, the small size of the anchors necessitates some sort of holder or cartridge for individual anchors. Applicants disclose herein an attachable magazine and multi-cartridge assembly that integrates with the above-described delivery device. The assembly includes a cartridge for each anchor in a given array with the individual cartridges stored and managed in a cartridge magazine in a way that maintains the integrity of the array and allows the surgeon to access and use each anchor in the array sequentially.

(23) A first illustrative and non-limiting example takes the form of a method for repairing a torn rotator cuff tendon by reattachment to a humeral head utilizing a seam-like row of serial stitches extending over incremental portions of the tendon between adjacent implanted bone anchors, the method comprising the steps of: providing a pre-strung plurality of anchors including a first anchor, a plurality of intermediate anchors and a final anchor with each anchor having a bone hole insertion diameter of less than or equal to about 3 mm, wherein the first anchor includes a length of working suture affixed thereto, each intermediate anchor having at least one passage therethrough with the length of working suture slidably and serially threaded through the at least one passage of each intermediate anchor and a final anchor also having at least one passage therethrough with the length of working suture slidably threaded therethrough and further including a means for locking the working suture relative to the final anchor; implanting each anchor in a single row through the torn rotator cuff tendon, the anchors each implanted in a bone hole, with the intermediate anchors placed in bone holes formed in the original footprint of the torn rotator cuff tendon with spacing between adjacent anchors a distance of about 10 mm or less as measured from center of bone hole to center of bone hole; applying tension to the working suture after implantation of each intermediate anchor and the final anchor implantation to form a series of single suture tensioned stitches between anchors; and securing the working suture as tensioned to the final anchor with the means for locking the working suture.

(24) Additionally or alternatively, the plurality of intermediate anchors includes six anchors.

(25) Additionally or alternatively, the means for locking the working suture is a locking loop, the locking loop extending adjacent the final anchor wherein the locking loop encircles a portion of the length of the working suture, the locking loop having a first open position allowing the working suture to slide through the locking loop and a second closed position engaging the working suture and preventing sliding of the working suture within the locking loop.

(26) Additionally or alternatively, the means for locking the working suture is a knot tied in the working suture adjacent the passage in the final anchor after tensioning.

(27) Additionally or alternatively, the means for locking the working suture is a one-way slip knot that allows sliding of the working suture in a first direction but prevents sliding in an opposite direction, wherein the step of applying tension to the working suture is performed by pulling the working suture in the first direction.

(28) Additionally or alternatively, the means for locking the working suture is a set of angled projections in the passages of the anchors that allow working suture movement in a first direction but prevent working suture movement in a direction opposite the first direction, and the step of applying tension to the working suture is performed by pulling the working suture in the first direction.

(29) Additionally or alternatively, the means for locking the working suture is mechanical lock that compresses and secures the working suture to the final anchor when activated.

(30) Additionally or alternatively, each intermediate anchor further includes means for locking the working suture thereto.

(31) Another illustrative and non-limiting example takes the form of a method for securing a

supraspinatus tendon and an infraspinatus tendon to a humeral head, the method comprising the steps of: providing a first anchor having a length of a working suture secured thereto, the anchor having an insertion diameter in a bone hole of less than or equal to about 3 mm; implanting the first anchor through either of the supraspinatus tendon and the infraspinatus tendon into the humeral head within a medial half of an original combined footprint of attachment of the supraspinatus and infraspinatus tendons to the humeral head; providing a plurality of intermediate anchors, each intermediate anchor slidably received on the length of working suture, each intermediate anchor having a passage therethrough wherein each intermediate anchor is configured for insertion in a bone hole having a diameter of less than or equal to about 3 mm; implanting the intermediate anchors through the tendons into bone holes formed in the humeral head in a serial row within the medial half of the original combined footprint, wherein the first of the intermediate anchors is spaced from the first anchor by a distance of less than or equal to about 10 mm measured from center of bone hole to center of bone hole and each subsequently implanted intermediate anchor is spaced from the just previously implanted adjacent intermediate anchor by a distance of less than about 10 mm measured from center of bone hole to center of bone hole; providing a final anchor, the final anchor slidably received on the length of working suture, the final anchor having a passage therethrough wherein the final anchor is configured for insertion in a bone hole having a diameter of less than or equal to about 3 mm wherein the final anchor further includes means for selectively locking the working suture relative to the final anchor; implanting the final anchor through either of the supraspinatus tendon and the infraspinatus tendon into a bone hole formed in the humeral head within the medial half of the original combined footprint, wherein the final anchor is spaced from the last intermediate anchor by a distance of less than or equal to about 10 mm measured from center of bone hole to center of bone hole; tensioning the working suture at locations along its length to form a stitch extending between each anchor in series to hold the supraspinatus tendon and the infraspinatus tendon against the humeral head; and, securing the tensioned working suture to the final anchor utilizing the means for selective locking to maintain desired tension in the formed individual stitches.

(32) Additionally or alternatively, the plurality of intermediate anchors includes six anchors.

(33) Additionally or alternatively, the means for locking the working suture is a locking loop, the locking loop extending adjacent the final anchor wherein the locking loop encircles a portion of the length of the working suture, the locking loop having a first open position allowing the working suture to slide through the locking loop and a second closed position engaging the working suture and preventing sliding of the working suture within the locking loop.

(34) Additionally or alternatively, the means for locking the working suture is a knot tied in the working suture adjacent the passage in the final anchor after tensioning.

(35) Additionally or alternatively, the means for locking the working suture is a one-way slip knot that allows sliding of the working suture in the a first direction but prevents sliding in a direction opposite the first direction, wherein the step of tensioning the working suture includes pulling the working suture in the first direction.

(36) Additionally or alternatively, the means for locking the working suture is a one way passage that includes angled projections that allow working suture movement in a first direction but prevents moving the working suture movement in a direction opposite the first direction wherein the step of tensioning the working suture includes pulling the working suture in the first direction.

(37) Additionally or alternatively, the means for locking the working suture is mechanical lock that compresses and secures the working suture to the final anchor when activated.

(38) Additionally or alternatively, each intermediate anchor further includes means for locking the working suture thereto.

(39) Another illustrative and non-limiting example takes the form of a method for repairing a torn rotator cuff tendon by reattachment to the humeral head utilizing a seam-like row of serial stitches extending over incremental portions of the tendon between adjacent implanted bone anchors in an

original footprint of the tendon on the humeral head, the method comprising the steps of: providing a pre-strung plurality of anchors, including a first anchor, a plurality of intermediate anchors and a final anchor with each anchor having a bone hole insertion diameter of less than or equal to about 3 mm, wherein the first anchor includes a length of working suture affixed thereto, each intermediate anchor having at least one passage therethrough with the length of working suture slidably and serially threaded through the at least one passage of each intermediate anchor and a final anchor also at least one passage therethrough with the length of working suture slidably threaded therethrough and further including a separate locking loop, wherein the separate locking loop encircles a portion of the length of the suture adjacent the final anchor and having a first position allowing the suture to slide through the locking loop and a second position engaging the suture and preventing sliding within the locking loop; implanting the first anchor through the tendon into a bone hole formed in the humeral head within the original footprint; implanting a first of the intermediate anchors of the pre-strung plurality of anchors in a bone hole formed in the original footprint a distance of about 10 mm or less from the first anchor as measured from center of bone hole to center of bone hole, then applying tension to the working suture extending from the first anchor and passing through the first intermediate anchor to form a single suture tensioned stitch between the first anchor and first intermediate anchor; implanting a second of the intermediate anchors in the pre-strung plurality of anchors in the footprint defined by the original insertion of the supraspinatus and infraspinatus tendons a distance of about 10 mm or less from the first intermediate anchor as measured from center of bone hole to center of bone hole, then applying tension to the suture extending from the first intermediate anchor and passing through the second intermediate anchor to form a single suture tensioned stitch between the first and second intermediate anchors; repeating the spacing distances, implanting and tensioning steps for each subsequent serial intermediate anchor and the final anchor; and, activating the locking loop to maintain tension in the created array.

(40) Additionally or alternatively, the plurality of intermediate anchors includes six anchors.

(41) Additionally or alternatively, each intermediate anchor further includes means for locking the working suture thereto.

(42) Additionally or alternatively, at least one intermediate anchor omits any means for locking the working suture thereto.

(43) Another illustrative and non-limiting example takes the form of a method of reattaching a rotator cuff tendon to a humeral head comprising: forming a first bone hole and implanting a first anchor of an anchor system therein, the anchor system comprising: the first anchor; at least one intermediate anchor having an intermediate anchor body with at least one passage therethrough; and a final anchor having a final anchor body with at least one passage therethrough; a working suture secured to the first anchor, passing slidably through the at least one passage of each of the at least one intermediate anchor and through the at least one passage of the final anchor; and a locking means for locking the working suture relative to the final anchor; wherein each of the first anchor, at least one intermediate anchor, and final anchor are configured for placement in bone holes having a diameter of less than or equal to 3 mm; the method further including forming at least one intermediate bone hole and implanting one of the at least one intermediate anchor in the at least one intermediate bone hole; wherein for each intermediate anchor that is implanted the method includes applying tension to the working suture after implanting the intermediate anchor; forming a final bone hole and implanting the final anchor therein; tensioning the working suture after implanting the final anchor; and securing the working suture as tensioned to the final anchor with the means for locking the working suture.

(44) This overview is intended to introduce the subject matter of the present patent application. It is not intended to provide an exclusive or exhaustive explanation. The detailed description is included to provide further information about the present patent application.

Description

BRIEF DESCRIPTION OF THE DRAWINGS

- (1) In the drawings, which are not necessarily drawn to scale, like numerals may describe similar components in different views. Like numerals having different letter suffixes may represent different instances of similar components. The drawings illustrate generally, by way of example, but not by way of limitation, various embodiments discussed in the present document.
- (2) FIG. 1 is an illustration of a representative pattern of continuous tensioned and locked anchor to anchor single suture stitches;
- (3) FIG. 2A is an illustration of a pre-threaded array of toggle type anchors;
- (4) FIG. 2B is an alternative view of FIG. 2A showing the toggle anchors in cross section to illustrate the threading route of the sutures;
- (5) FIG. 2C is an illustration of an alternative array of toggle type anchors;
- (6) FIG. 2D is an alternative view of FIG. 2C showing the toggle anchors in cross section to illustrate the threading route of the sutures;
- (7) FIG. 2E is partial cut-away view of a toggle body with a working suture and locking suture in an open position illustrated;
- (8) FIG. 2F is partial cut-away view of the toggle body of FIG. 2E having a working suture and locking suture in a closed position illustrated;
- (9) FIG. 2G is a schematic illustration of the interaction between the locking suture and the working suture;
- (10) FIG. 2H is a schematic illustration of an alternative interaction between the locking suture and the working suture;
- (11) FIGS. 3A-3C are perspective views of a representative toggle body anchor;
- (12) FIG. 4A illustrates an alternative array of wedge like anchors joined by unidirectional tensionable sutures;
- (13) FIG. 4B illustrates in further detail an anchor of the array of FIG. 4A having a one-way suture passage therethrough;
- (14) FIG. 5A illustrates another alternative anchor incorporating a toggle body with a separate one-way suture passing member;
- (15) FIG. 5B further illustrates the anchor FIG. 5A in its orientation on implantation in bone;
- (16) FIG. 5C provides a cross section view of the one-way locking mechanism within the separate suture passing member of FIG. 5A;
- (17) FIG. 6A-6B illustrate another alternative anchor that can be used in an alternative array incorporating a one-way slip knot;
- (18) FIG. 7A-7B illustrate another alternative anchor incorporating external retention flanges and an internal mechanical suture locking mechanism;
- (19) FIGS. 8A-8B illustrate another alternative anchor having activatable external legs for anchor retention in bone and an internal active suture locking mechanism;
- (20) FIGS. 9A-9B illustrate features of a soft suture anchor that can be utilized in an implant array;
- (21) FIG. 10A is a perspective view of an example anchor delivery device;
- (22) FIGS. 10B-10D are close up views of the distal end of the anchor delivery device of FIG. 10A in various set configurations;
- (23) FIG. 11A is a partial perspective view of an alternative anchor delivery device illustrating the bone punch, anchor cartridges and magazine and plunger for anchor loading;
- (24) FIGS. 11B-11E illustrate a cartridge for holding a toggle anchor; and,
- (25) FIGS. 12A-12G illustrate the steps for implanting exemplary anchors of the current invention.
- ### DETAILED DESCRIPTION
- (26) The present invention includes multiple components, devices and methods to create and use an

overall system for reattaching soft tissue to bone. It is particularly useful to create a robust repair of torn tendons, such as the supraspinatus tendon, in an arthroscopic rotator cuff repair. The implants and delivery devices make possible a faster, easier and lower failure rate anatomical repair. The tendon is securely attached and held with adequate force to its original footprint with very little creep during movement of the joint. This decreases a patient's time in a sling, increases the rate of healing reattachment of tendon to bone and allows early physical therapy to maintain range of motion and strength.

(27) The implanted array of anchors with a continuous set of anchor-to-anchor single suture stitches creates a seam-like attachment akin to a sewing machine construct as illustrated in FIG. 1. FIG. 1 illustrates a rotator cuff tendon **608** attached to the humeral head **602** by a first array of stitches **650**. Each stitch includes a suture portion **610** that extend over a portion of the tendon **608** between two adjacent anchors **612**. Using the methods and devices disclosed herein can create a row of continuous stitches that are closely spaced, individually tensioned and tightened.

(28) As used herein, the direction of a tendon is the direction in which forces are borne by the tendon for functional use; for example, in the rotator cuff tendon **608** shown in FIG. 1, the functional direction of the tendon can be understood by arrow **606**. A preferred pattern includes a row of stitches **650** generally perpendicular to the direction of the tendon as shown in FIG. 1. In a rotator cuff repair these would all be placed in a medial portion of the original tendon footprint, in particular the original tendon footprint of the supraspinatus and the infraspinatus tendons. For illustration, an approximation of the original footprint of the infraspinatus tendon is shown in phantom at **620**, and the original footprint of the supraspinatus tendon is shown roughly in phantom at **622**; the two footprints may be adjacent one another as shown. In the below described methods, reference may be made to the original combined footprint of the supraspinatus and infraspinatus tendons for convenience. It should also be understood that references below to a first anchor and a final anchor may refer to either end of the row of stitches, as the decision of which end to start with will be up to the physician's judgment or preference in view of patient anatomy.

(29) In some preferred embodiments a second row of anchors **652** is also implanted, especially in a large rotator cuff repair. The second row is implanted laterally of the first row and can include a zig zag pattern to put some anchors in the lateral portion of the original footprint and other anchors lateral of the footprint to hold down edges of the torn tendon. Other configurations are also possible depending on the size and shape of the tear. Anchors may also be placed to create stitches over attached portions of the tendon to reinforce the margins/edges of fully or partially torn tendons.

(30) In some examples, the small cross-sectional size of the anchors (sized for placement in a 3 mm diameter bone hole) allows the anchors to be placed in close proximity to one another (less than about 7 mm between adjacent anchors from edge of bone hole to edge of bone hole). This creates an anchor to anchor suture stitch. Combining this concept with the disclosed anchor design allows the suture stitch to be tightened and locked individually or as a group when the adjacent suture anchors are implanted. This can be repeated many times to implant a row of anchors with continuous independently tensioned and locked stitches between adjacent anchors. Also, because the anchors are in a high-density array, the tension force components on the tensioned suture are more vertically applied to the top surface of the tendon (or other connective tissue) to thereby hold the tendon against the footprint of the bone without creep or slippage during joint movement.

(31) In some preferred embodiments, individual anchors, as described in detail below, do not function alone. Instead each is part of a pre-strung or stringable array having a common serially disposed working suture therethrough. FIG. 2A illustrates a pre-strung array **201**. Each anchor **200** can be implanted sequentially within the array, then the working suture section extending from the just implanted anchor to the just previously implanted anchor can be tensioned. In some embodiments as illustrated FIG. 2A, the working suture is then locked at the just implanted anchor so that a suture stitch between the two anchors provides force against the tendon to hold it in place much like a single sewn stitch. With the array, multiple stitches can be formed in a continuous array

similar to a sewn seam.

(32) In FIG. 2A a pre-strung array **201** of individual anchors **200** is depicted. The anchors **200** may be similar in form and function to the anchor **100** described herein. The shown array has four anchors **200** as a representative chain. It is believed chains of 4 to 12 anchors would be useful in tendon repair procedures such as rotator cuff repair. One particular embodiment includes 8 anchors in an array. As depicted in FIG. 2A, the way in which the working suture **115** is pre-threaded through the series of anchors **200** is important to assure that they will toggle in bone as desired and tension to form the stitch when the suture is tightened. The illustration shows the first anchor **202** to be implanted followed by the second anchor **204**, then the third anchor **206** and finally the fourth anchor **208**. With this order of implantation understood, the working suture **115** has been pre-threaded down through the top of the proximal hole **210** and back up through the distal hole **211** of the first anchor **202**. The working suture **115** then continues to the second anchor **204** where it is threaded down through the proximal hole **212** and back up through the distal hole **213** of the second anchor **204**. The working suture **115** then continues to the third anchor **206** where it enters the top of the proximal hole **214** and back up the distal hole **215** of the third anchor **206**. The working suture then continues to the fourth anchor **208** where it enters the top of the proximal hole **216** and passes up through the bottom of distal hole **217** of the fourth anchor **208**. If the array were more than four anchors, the pre-threading would continue as described for each subsequent anchor.

(33) FIG. 2B is a cross sectional view of the array of FIG. 2A which more clearly shows the threading of the working suture **115** within the anchors **200** in the array **201**. The way in which a locking suture **116** is disposed in the middle passage is also shown for each anchor **200** as described above with each locking loop **118** independent for each anchor. The locking suture **116** can have a preferential point of failure so that it can be tightened then broken off above the slidable knot. This can be accomplished by tying a break knot, or making a nick in, in the free tail of the locking loop just above the slidable knot, as further illustrated in FIGS. 2E to 2H, below. In some preferred embodiments the slidable knot is a 4-throw uni knot and the break knot is in the free tail just above the uni knot. The suture lock may be designed to break at a desired tension with the slidable knot in place sufficient to lock the working suture.

(34) To create an implanted serial array of tensioned and independently locked anchor to anchor suture stitches for attaching a tendon to bone, a surgeon would begin with the pre-strung array **201** described in FIGS. 2A and 2B. The first anchor **202** would be implanted through the tendon into a formed bone hole and the working suture locked. Alternatively, the first anchor **202** can be affixed to the first anchor **202** prior to implantation as this anchor does not have suture pulled through it. The second anchor **204** would then be implanted in close proximity to the first anchor **202**, preferably less than 7 mm away (edge of bone hole to edge of bone hole). The second anchor is toggled and the working suture tensioned at the same time by pulling on the working suture **115** that exits the distal hole **213** of the second anchor **204**. Tension at this location not only toggles the second anchor **204** but also tightens the working suture **115** going back to the first anchor **202** to form the tensioned stitch holding the tendon against the footprint. The second anchor **204** is then locked so that the stitch remains tensioned and is isolated or independent of other stitches. The process is repeated for the third anchor **206** and fourth anchor **208** or more. In one preferred array, eight anchors are implanted and 7 tensioned and locked stitches in a continuous row are formed. Further, in a rotator cuff repair, multiple arrays can be implanted such as one array extending across the tendon in the medial portion of the footprint and a second array more lateral to the medial position.

(35) As discussed above, one preferred array can include a plurality of anchors, each having a locking loop associated therewith and a common working suture that connects the anchors serially. When the working suture is tensioned between anchors, the working suture segment tightens and forms a stitch that is compressed against the tendon between the anchor ends. Other alternative anchors and array designs are possible to create the same or a similar seam like array of stitches.

For example, the first anchor in the array of FIG. 2A includes a locking loop. This loop is not necessary and instead the first anchor in any array could simply have the end of the working suture affixed to anchor in any fashion, such as insertion through a passage on the anchor and affixing with a knot. There is no need to have structure that allows flossing or sliding of the working suture relative to the first anchor in an array.

(36) Another alternative array is depicted in FIGS. 2C and 2D. The array is similar to the array of anchors of FIGS. 2A and 2B with like parts numbered the same. The distinction is the varied use of locking loops. In some embodiments, not all anchors of the array include locking loops. As stated above, the working suture can be affixed to the first anchor **202** with a knot (or in the alternative, with adhesive, heat attachment, mechanical fixation such as with a screw, or any other suitable attachment), as shown by the working suture **115** only having one end at the first anchor **202**, with no locking suture on the first anchor **202**. It may be observed that the working suture **115** terminates at a distal portion of the first anchor **202**, which may be useful for aiding in toggling the first anchor **202**. In this example, one or more of the intermediate anchors **204**, **206** in the array may simply have the working suture **115** threaded through the passages in the anchor so that it can be flossed through the anchor as needed to tension the working suture **115** (and also to aid in toggling each anchor). However, as depicted the last anchor **208** in the array does include a locking loop **116** to affix the working suture relative to the last anchor after it is tensioned. It is recognized that in some embodiments none of the intermediate anchors include locking loops while in others some of the intermediate anchors can include locking loops. In some preferred embodiments, the working suture and anchor passage, along with the tortuous path followed by the working suture (such as highlighted in FIG. 3C, below) provide sufficient friction so that the working suture can be tensioned after each anchor implantation and the tension is retained under conditions during surgery where the tendon is not being flexed and loaded/unloaded as during live use. Therefore, the array can be implanted as discussed with respect to FIGS. 2A and 2B with each stitch tensioned after each intermediate anchor implantation, but without a locking suture **116** or other locking structure to secure the working suture at each intermediate anchor **204/206**. The last anchor is then implanted, the working suture tensioned and the locking loop associated with the last anchor is activated. The locking of the working suture on the last anchor essentially locks the entire array in tension so that the implanted array can be exposed to the loads of shoulder movement as needed during healing and rehabilitation. That is, in some examples, only the last anchor of the array may have the locking loop **116**.

(37) FIGS. 2E-2H are a series of illustrations of an exemplary toggle body type anchor showing a single working suture slidably disposed in passages through the anchor and through a locking loop. As discussed above, in some anchors the locking loop is omitted such as in the first anchor which can have the working suture affixed thereto or some intermediate anchors that rely on the last anchor suture lock to hold the array in tension. The locking loop is configured to have an open position allowing movement of the single working suture, and a closed or locked position that prevents movement of the single working suture. In the illustrative example shown in FIGS. 2E to 2H, the free end **121** of the suture lock **116** is configured to break away from the locking loop **118** proximal of the sliding knot **120**. A break knot is illustrated at **122** and is one example of a way of introducing weakness in the suture lock. The break knot **122** is located a distance above the sliding knot **120**, sufficient that when the suture lock **116** breaks away, the sliding knot **120** remains intact and secure. For example, the break knot **122** may be about 3 to 10 mm proximal of the sliding knot, or more or less. Rather than a break knot **122**, a nick, crimp, or other point of weakness may be imparted at the desired or preferential point of failure in the suture lock **116**.

(38) FIGS. 2G and 2H depict the way in which the locking loop **118** pulls the section **117** of the working suture **115** into the oval portion **113** in two different embodiments. The degree to which the section **117** of the working suture **115** enters the slot **113** will be dependent upon how tight the loop is closed, the size of the locking suture and the size of the slotted opening. In preferred

embodiments, at least a portion of the cross section of the working suture **115** is pulled into the slot so that the edge surfaces of the slot walls provide significant friction and aid in locking. In another example, the preferential point of failure is designed to allow the locking loop **118** to be drawn into the slot before the failure occurs.

(39) The locking loop **118** in combination with the design of the middle passage **108** is an assembly for locking a slidable working suture **115** when tensioned in a suture toggle body **100** during tissue fixation to bone. The locking loop **118** encircles a portion of the working suture **115**, wherein collapsing the locking loop **118** compresses the cross section of the working suture **115** to lock the working suture **115** when tensioned. The suture lock **116** is preferably formed of a suture having at least a slidable knot **120** tied therein to form the loop **118** to allow collapsing of the loop **118** when a tightening leg **121** through the second passage **108** is tensioned. The second passage **108** has an upper portion for receiving the slidable knot **120** at least partially therein that terminates in a platform **114** within the toggle body **100** that does not allow passage of the slidable knot. The second passage includes a lower portion having an oval shape for allowing both legs of the locking loop to pass therethrough side by side and out the passage. A particularly preferred knot is a 4-throw uni knot. However, other slidable knots **120** may be used, as desired. Further, the second passage oval portion is sized to allow at least a portion of the working suture **115** to be pulled therein in response to tension on the locking cord. The working suture **115** is preferably a braided multistrand suture having a compressible cross-sectional area that reduces by at least about 25% when the locking loop is tightened during use. The working suture **115** can be a round and/or braided No. 2 suture in some embodiments. Other size, shape and/or configuration sutures may be used.

(40) As also shown in FIGS. 2F and 2H, after the sliding knot **120** is tightened, and the working suture is drawn at least partly into the slot, the preferential point of failure in the locking loop **116** (such as the break knot or nick described above) breaks, leaving free tail at **123** on the locking loop, a distance above the sliding knot, while the rest of the proximal portion of the suture lock **124** can be discarded. In some examples, a more proximal portion of the suture lock is secured to a cartridge, so that a physician may cause the suture lock to break as shown by pulling on the cartridge itself, as further described below. In an example, the preferential point of failure is designed to allow tightening of the locking loop **118** onto the working suture **115**, and/or drawing of the working suture into the slot on the bottom of the anchor before the failure occurs. For example, the locking loop and the preferential point of failure may be configured for breaking under a pull strength in the range of 3-10 pounds of force, more preferably, 5-7 pounds of force, or more or less as desired. The pull strength needed to tighten the locking loop **118** onto the working suture may be less than the pull strength needed for breaking the preferential point of failure in some examples by, for example, an amount in the range of 0.5 to 3 pounds, or 0.75 to 2 pounds, or about 1 pound.

(41) Referring to FIG. 3A, a perspective view of a representative anchor in the form of a toggle body **100** is illustrated. The toggle body **100** can be an elongate body **101** having a length defined by a proximal end **102** and a distal end **104**. The elongate body **101** can be a generally cylindrical body but other shapes are possible. For example, as shown in FIG. 2A, the toggle body **100** is generally cylindrical but the top surface **105** and bottom surface **107** have flat axially-extending surfaces that allow room for sutures when the toggle body **100** is in a round delivery tube. The length of the toggle body **100** is substantially longer than the diameter thereof, allowing the toggle body **100** to be inserted lengthwise or axially into a small bone hole. Once inserted, unlike most anchors used today, the entire body is pivoted or toggled so that it stays within the bone and has substantially its entire length compressed against material inside the bone. That is, the longitudinal axis of the toggle body **100** is rotated or pivoted from the direction used to insert through the bone hole, thereby preventing removal. This approach means that removal would require the anchor itself to fail, rather than simply being released from surrounding tissue, and provides high pullout

strength (greater than 600 N before anchor failure when implanted in the array disclosed herein) from an anchor requiring a very small insertion hole (less than about 3 mm). As previously stated and described in detail below, small insertion holes allow much closer placement of anchors in a high-density array.

(42) The toggle body **100** can have a length of about 6 mm to about 10 mm in some embodiments. This length gives adequate strength while leaving enough room inside the bone for the high number of anchors implanted. The toggle bodies are preferably molded or machined from a polymeric material, preferably a high tensile strength material such as a poly-ether-ether ketone (PEEK) which is highly biocompatible. In applications where MRI imaging would not be an issue, metal can be utilized in part or all of the toggle body.

(43) Referring now also to FIGS. 3B (top view) and 3C (bottom view), it can be seen that the toggle body **100** can include a number of holes or passages through the cross section of the toggle body **100**. As illustrated, the toggle body **100** has a proximal bore or passage **110**, a middle passage **108** and a distal passage **106**. The passages **106**, **108**, **110** extend from the top surface **105** to the bottom surface **107** such that the passages **106**, **108**, **110** extend through the cross section of the elongate body **101**. In other embodiments, the toggle body may have fewer or more bores or passages, such as having a single bore, two bores, or more than three bores. In the illustrated embodiment, the proximal passage **110** and distal passage **106** receive a portion of a common working suture slidable with respect to the toggle body **100** during use. The middle passage **108** receives a locking suture which is independent for each anchor used in an array of anchors.

(44) The distal end **104** of the toggle body **100** has an angled surface. As shown, the angled surface creates a longer upper longitudinal surface **105** than lower longitudinal surface **107**. In other words, the upper surface projects a greater distance distally than the lower surface. This is useful during insertion of the toggle body **100** because the projecting distal surface plows into cancellous spongy bone when implanted to initiate at least partial rotation of the toggle body during insertion. Keeping in mind that the present toggle bodies **100** are preferably implanted through the tendon, it is important that the toggle body **100** toggle every time or it may pull out of the bone hole under tension yet not be visible as it will be under the tendon.

(45) The proximal end **102** of the toggle body **100** can include one or more projecting fins **112**. The illustrated embodiment includes two fins **112**. Each fin **112** projects outward and proximally. Further, in some embodiments, as depicted, the fins **112** project downward as they extend proximally. The bone hole in which the implant will be placed is sized to closely match the general diameter of the toggle body. In contrast, as shown, the fins **112** each project laterally beyond the outer cross section or diameter of the elongate body. During insertion the fins **112** flex inward under compressive force due to contact with the inner diameter of a delivery tube to fit in the bone hole.

(46) Once delivered and released from compressive forces of the delivery tube, the fins **112** relax to a size greater than the bone hole. In some preferred embodiments, each fin tip extends about an additional 0.5 mm beyond the size of the bone hole where that feature is inserted. Such fin tips may also be described as extending about 0.5 mm beyond the maximum outer diameter of the rest of the anchor body, for example, in the range of 0.4 mm to 0.7 mm. This feature provides an added safeguard against the toggle body **100** backing out of the bone hole under tension if the toggle body **100** has not adequately toggled. Further, the fins **112** are positioned so that tension on the toggle body **100** causes the partially toggled anchor to grab cancellous bone and further rotate the anchor.

(47) The top and bottom views of FIGS. 3B and 3C show details of the proximal **110**, middle **108** and distal **106** passages. In particular, the middle hole has a platform **114** formed within the elongate body **101**, part way through the cross section. That is, in this example, the middle passage **108** has a change in size or shape partway along its length, to define a platform **114**. From the bottom view, it can be seen that the middle passage **108** continues from the platform **114** with a slotted or oval shape or portion **111**, while having a circular profile from the top view. The function

of these passages is detailed in the cross-section perspective views of FIGS. 2E and 2F wherein representative cords or sutures **115**, **116** have been pre-strung on the toggle body **100**.

(48) First, there is a single suture, called herein the working suture **115** that extends into the proximal passage **110** from the top surface, and extends out at the bottom surface. The working suture **115** then extends up through the distal passage **106** from the bottom surface and out through the top surface. This leaves a section **117** of the working suture **115** extending past or adjacent the middle passage **108** along the bottom surface. The working suture **115** can be flossed or is slidable through the distal **106** and proximal passage **110**, meaning the toggle body **100** can slide on the working suture **115** when tension is applied. Second there is a locking loop **118** that encircles a portion of the section **117** of the working suture **115** extending adjacent the outer surface of the toggle body **100** between the proximal **110** and distal **106** passages. The locking loop **118** has a first open position as depicted in FIG. 2E wherein the working suture **115** is free to slide through the locking loop **118** and a second closed position depicted in FIG. 2F wherein the locking loop **118** engages the section **117** and prevents it from sliding within the locking loop **118**.

(49) Several examples refer to a suture, cord, or thread, which can be used as the working suture **115** or in the locking loop **118**. These elements may be, for example, made of natural material such as silk and/or synthetic materials such as polyglycolic acid, polylactic acid, and polydioxanone, each of which are known for use as absorbable sutures, and/or nylon and polypropylene, which are typically non-absorbable. Various coatings, including antimicrobial, anti-wicking or lubricious coatings may be applied as well. More broadly, these elements **115**, **118** may include any item that can be used to couple together objects in a surgical environment, such as any sufficiently biocompatible metal, natural material, plastic or other artificial material adapted for use in a surgical procedure. Monofilaments or more complex structures including braids, weaves, windings, twisted threads, coated or multilayer member, etc. may be used.

(50) In the embodiment depicted, the locking loop **118** extends from the bottom surface of the toggle body **100** through the middle passage **108**. The locking loop **118** includes a cord or suture having at least a slidable knot **120** tied therein to allow collapsing of the locking loop **118** when a free end or proximal end **121** of the suture lock **116** extending through the middle passage **108** is tensioned. As shown, the upper portion of the middle passage **108** is sized to receive at least a portion of the slidable knot **120** therein. The slidable knot **120** then contacts the surface of the platform **114** which does not allow the knot to pass through towards the bottom opening. The lower oval portion **113** of the middle passage **108** is a slot or oval which allows both legs of the locking loop **118** to pass therethrough, preferable side by side in the slot direction. The interaction of these components locks the working suture **115** with respect to the toggle body **100**.

(51) As shown, especially seen in FIG. 3C, the bottom of the toggle body **100** includes a channel **125** formed in the bottom surface **107** between the proximal **110** and distal **106** passage. When the working suture **115** is tensioned, it is pulled up into this channel **125** which is sized to make the suture less able to floss or move therethrough by increasing frictional resistance to such movement, but does not lock the suture. Further, the working suture then has two near 90-degree angle turns at the bottom openings of the distal **106** and proximal **110** passage which also make it more difficult to floss, but do not lock the working suture **115**. The locking loop **118** closing around the working suture **115** and pulling it toward and at least partially into the slot or oval portion **113** is the structure that locks the suture so that cumulative friction prevents slippage of the working suture **115**.

(52) The above discussion of arrays, anchors and the working suture forming stitches as implanted focuses on the use of one particular anchor type, the toggle body. This is one preferred anchor design, however it should be recognized that the method of creating a seam like attachment akin to a sewing machine construct can be practiced with any type of suture anchor provided it meets the necessary criteria of being implantable in a bone hole having a diameter of less than or equal to about 3 mm, and can accommodate a working suture therethrough that allows flossing and

tensioning of the working suture to form individual tensioned stitches. Further, the first anchor must have the ability to lock the working suture thereto or be supplied with the working suture affixed as by a knot, and at least the last anchor in the array must have a locking loop or other means for affixing the working suture to the last anchor after tensioning. The locking or affixing of the suture to the first and last anchors at a minimum maintains the stitches in the array in tension against the tendon which is against the bone in healing contact. Intermediate anchors may include, or may omit structures or locking loops for locking the working suture thereto.

(53) Referring now to FIGS. 4A and 4B, one alternative anchor design and array that can be used in accord with the method of the current application is depicted. First, the anchors **450** are of a wedge shape design. Any wedge shape design, as known in the art can be utilized. For implantation the anchors can be sized smaller in one dimension for insertion then rotate like a toggle type anchor for locking in the bone. Alternatively, the anchors can be rolled within a tubular member for insertion into a bone hole where upon insertion the wedge unrolls to the depicted shape and locks within the bone.

(54) As also depicted in FIG. 4A and in more detail in FIG. 4B each anchor can include a passage **452** therethrough for slidably receiving the working suture **115**. In the depicted embodiment each anchor **450** is shown with a plurality of barbs **454** or protrusions projecting inwardly toward the center of each anchor passage **450**. With this feature the working suture **115** becomes slidable in one direction as the barbs are preferably angled to bend in that direction while the working suture can not floss in the opposite direction due the barbs digging into or frictionally engaging the working suture. The array can be implanted as discussed with other embodiments, however, tensioning of the stitch is accomplished by pulling the working suture **115** extending from each anchor in the first direction and the tension is maintained due to the lock accomplished by the barbs **454** or protrusions in the passage. This type of one-way passive lock can be incorporated in any anchors of the present invention if desired. In the example shown, a single working suture **115** may be used. Alternatively, there may be several individual segments of working suture **115**, each having a first end permanently affixed to one anchor (see affixing location **456**), passing through the passage **452** of the subsequent anchor, with a free tail **119** extending therefrom. To tension an anchor to anchor stitch, the free tail **119** of the working suture **115** can be pulled. If, instead, a single working suture is used, the free tails **119** and the affixing location **456** would be omitted, and the suture **115** would simply run from one anchor to another. In some examples, the first anchor **450** may include the affixing location **456** or may otherwise be permanently secured to the working suture **115**. In some examples, the last anchor in the series of anchors may have an additional locking structure associated therewith, such as a locking loop as shown in other examples.

(55) Next referring to FIGS. 5A to 5C, another alternative anchor design is depicted. This exemplary design is a two-piece toggle anchor **560** which includes a toggle body **550** and a separate passive one-way suture locking member **552**. The toggle body **550** can be of similar design to previously discussed toggle type anchors, however the working suture of this embodiment passes through one passage of the toggle body **550**, then passes through the suture locking member **552** before going back through the toggle body **550**. With this two-piece design, the toggle body and anchor locking member can be made with a less than or equal to 3 mm insertion diameter. The toggle body **550** and suture locking member **552** may be aligned lengthwise end to end for insertion in a bone hole. As depicted in FIG. 5B, after implantation in bone, the toggle body can be rotated and pulling on the working suture tails will also cause the anchor locking member to be pulled up against the longitudinal side of the toggle body **550**. Although not shown, the toggle body **550** may include fins (similar to fins **112** shown in FIG. 3A) at a proximal end thereof **556**, if desired, to limit back-out as the anchor is toggled. The toggle body **550** may be toggled by pulling on the more distally passing portion of the working suture, as indicated by direction arrow **558**.

(56) The anchor locking member **552** depicted is preferable a passive suture locking member as

described for the wedge type anchors of FIG. 4A. The working suture passes through a passage in the locking member and can slide relative thereto, however, as best shown in FIG. 5C, the locking member includes internal radial projections that are preferable angled longitudinally and allow the suture to slide one direction while frictionally engaging or digging into the projections in the other direction. It is recognized that the same function could be accomplished with smooth internal surfaces and using a working suture that has projections or barbs, however, this alternative is not preferred due to the need for the working suture to pass through and extend over tendon tissue in forming the seam like row of stitches preferred herein, as any such barbs or projections may damage the tendon as the working suture **115** is tensioned.

(57) Another alternative embodiment of a toggle body and suture arrangement is depicted in FIGS. 6A and 6B. As can be seen in the drawings, the toggle body **651** is the same design as the toggle body **200** (having two bores), described above. In this embodiment the working suture **652** is routed through the passages **653**, **654** to create a one-way knot. The toggle body **651** includes a second hole **654** through the elongated toggle body extending from the first longitudinal surface to the second longitudinal surface at a spaced interval along the elongated toggle body **651** relative to a first hole **653**. The working suture **652** exits the first hole **653** at the second longitudinal surface and wraps around the elongated toggle body and around a portion of the working suture extending from the second hole **654** at the first longitudinal surface of the toggle body **651**, then into the second hole through the second longitudinal surface and out the first longitudinal surface. This forms a one-way locking knot preventing the suture from sliding or increasing friction in one direction but allowing sliding in the other direction. Although not depicted, a locking loop could also encircle the working suture **652**. This anchor design could be included in any of the previously disclosed arrays to achieve the method of forming a seam like repair of anchor to anchor tensioned stitches, and may be used as a first, last, or intermediate toggle body, as desired.

(58) Next referring to FIGS. 7A and 7B, another suture anchor **750** is depicted that can be utilized in methods of forming a seam like array of tensioned stitches. The anchor **750** includes a cylindrical exterior having flanges **751** projecting therefrom for retention of the anchor when inserted in a bone hole. Further, the working suture **752** passes through an opening at a proximal end of the anchor **750**, and can be flossed through the opening, relative to the anchor. With flossing, the anchor could be used as an intermediate anchor in an array. However, the anchor in FIG. 7A is also depicted to include a locking member **754** that can affix the working suture relative to the anchor **750**. As detailed in cross-section in FIG. 7B, the locking member **754** includes a pull suture **755** that extends into the internal core of the anchor and to the distal end where it wraps around a shoulder **756**, before extending back proximally to attach to a locking arm **758**. The working suture passage is at the proximal end of the locking arm **758**. As illustrated, pulling on the pull suture **750** will draw the locking arm **758** distally into an anchor passage **760**. The teeth of the locking arm **758** engage with a pawl **762** as the locking arm **758** advances into the anchor passage **760**. This movement pulls the locking suture in crimped fashion into the anchor body passage **760** and locks it therein. This type of locking anchor could be utilized in any of the disclosed arrays previously discussed.

(59) Next referring to FIGS. 8A and 8B, another alternative suture anchor **850** is illustrated. In this example, retention of the anchor **850** in a bone hole is accomplished with retention arms **852** that are expanded after the anchor is inserted in the bone hole. The anchor **860** includes an outer tubular body **860** that has an internal body **858** slidably disposed therein. The internal body **858** can be moved from a first distally extending position to a more proximal position and locked therein. Retention arms **852** on the outer tubular body **860** stay in a reduced diameter state as they are introduced, and allow one-way movement down into the bone hold. The application of a force to withdraw the retention arms **852** causes each to catch on the surrounding bone, holding the anchor **850** in the bone hole. This prevents back-out as the working suture **854** is flossed to tighten the preceding stitch. It may be noted that the retention arms **852** occupy two sides of the outer body,

allowing the working suture **854** to pass by unimpeded. As flossing is completed, a force is then applied to the locking suture **856** to draw the inner body in an upward direction relative to the bone hole. When the internal body **858** is moved in this way shoulders on the internal body further secure the retention arms **852** in a radially expanded position to lock the anchor in bone. As can be seen in FIG. **8B**, movement of the internal body **858** in a proximal direction, as the suture lock **856** is pulled, causes the working suture **854** to be pulled into the outer tubular body and pinched in locking arrangement between the internal wall of the outer tubular body and the internal body. Protrusions shown on the distal portion of the inner body snap into detents in the outer body to secure the lock. This type of anchor could be utilized in any of the disclosed arrays previously discussed.

(60) FIGS. **9A** and **9B** illustrate an exemplary all-suture anchor that can be used alone or incorporate a locking suture or locking loop configuration relative to the working suture and anchor member. The illustrations also show a single working suture slidably disposed in passages through the anchor member and through a locking loop. The locking loop is configured to have an open position allowing movement of the single working suture, and a closed or locked position that prevents movement of the single working suture relative to the anchor member and/or the locking loop. In some embodiments the locking loop is omitted as was done with other intermediate anchors described above.

(61) Referring to FIG. **9A**, a simplified perspective view of a representative soft anchor **900** is illustrated. The anchor **900** includes two components, namely a working suture **901** that is slidably disposed within the lumen of an anchor member **902**. The anchor member **902** as illustrated can be a length of hollow woven suture material. As depicted the ends **903** of the anchor member **902** flare outward away from the working suture. The anchor member **902** is a soft or flexible member that can easily fold into a linear configuration for insertion in a small bone hole. After insertion, the tails of the working suture **901** can be tensioned which causes the anchor member to migrate toward the bone hole opening. However, the designed ends **903** extending laterally when relaxed will catch on the bone hole and cause the anchor member **902** to expand radially and be locked within the bone hole.

(62) In some examples, the ends **903** of the anchor member **902** are open on one side, as depicted in the figures, while a central portion of the anchor member **902** forms a complete cylinder and is not open on any side thereof, thereby maintaining the anchor member **902** on the working suture **901** while allowing the ends **903** to flare outward as shown. In some examples, the entire anchor member **902** is a soft and/or flexible material. In other examples the ends **903** may have an additional coating or different structure that adds rigidity to the ends and/or central portion of the anchor member **902**. In still another example, the anchor member can be generally stiffer throughout, at least relative to the working suture **901**, to provide additional anchoring strength. For example, the anchor member **902** may be formed of a different type of suture or with a different thickness of suture than the working suture **901**. The anchor member **902** may be, for example, a coated, braided synthetic material having greater stiffness and larger outer diameter, and the working suture may be a relatively thinner, braided (for knot security) natural material of relatively lesser outer diameter. Other combinations may be used, as desired. In an example, the anchor member **902** may be formed of a hollow braid, cut to length, slit to form the ends **903**, and then coated to prevent fraying along the ends, if desired.

(63) FIG. **9B** shows the anchor of FIG. **9A** as implanted and expanded within a bone hole. Tails of the working suture **902** extend from the bone hole for attachment to tendon. At the same time, the projections or flanges **903** of the anchor member **902** are trapped and secured within the bone. In this secured configuration, the working suture **901** is still slidable or flossable relative to the anchor member. A slidable working suture **901** may allow tensioning of individual suture stitches that extend from one anchor to the next anchor in the serial row of the array. Anchors of this exemplary design can be incorporated into previously disclosed exemplary arrays on a working suture. Once

emplaced in the bone hole and with the working suture **901** tensioned to form a stitch, a locking loop **905** can be pulled (similar to other examples having a locking loop) to securely hold the working suture **901** in a fixed position relative to the locking loop and the anchor member **902**. A number of additional example soft anchor may be found in U.S. Prov. Pat. App. No. 63/231,143, filed Aug. 9, 2021, titled TENSIONABLE AND LOCKABLE SOFT SUTURE ANCHORS AND ANCHOR ARRAYS FOR ANATOMICAL ATTACHMENT OF SOFT TISSUE TO BONE, the disclosure of which is incorporated herein by reference.

(64) In some examples a toggle anchor may also include a movable locking structure associated with the toggle body itself. The locking structure in such examples may be a means for locking a working suture, in addition to the various examples shown above. Some illustrative examples are shown in U.S. Prov. Pat. App. No. 63/231,136, filed Aug. 9, 2021, titled TENSIONABLE AND LOCKABLE MICRO SUTURE ANCHORS AND ANCHOR ARRAYS FOR ANATOMICAL ATTACHMENT OF SOFT TISSUE TO BONE, the disclosure of which is incorporated herein by reference.

(65) In some examples, a series of anchors disposed on a single working suture may have multiple different designs. For example, a first anchor may be any of the above designs, but may be unique within the series of anchors in that it is permanently affixed to the working suture. A set of intermediate anchors can then follow, with each intermediate anchor having the same design. A last anchor may have a third design within the array, with the last anchor having a locking loop or other structure to actively secure the working suture once the entire array has been implanted and tensioned. In another example, an array may include a first anchor that is affixed to a working suture, a second anchor that includes an active locking structure, several intermediate anchors lacking a locking structure or only having a passive locking structure, and two final anchors each having an active locking structure, or any other suitable combination of anchors that do not lock the working suture, or which passively lock the working suture with anchors that actively lock the working suture. Illustrative arrays may include any combination of the above anchors, as desired.

(66) One preferred anchor delivery device **300** for transtendinous implantation of individual anchors in an array is depicted in FIG. **10A**. The delivery device **300** is particularly useful to implant anchors disclosed above but especially useful and described in detail with respect to toggle type anchors and arrays as depicted in FIGS. **2A-2H**.

(67) FIG. **10A** is a perspective view of an example anchor delivery device in one exemplary configuration, and FIGS. **10B-10D** are close up views of the distal end of the anchor delivery device corresponding to FIG. **10A** in varied useful configurations. Starting with FIG. **10A**, the delivery device **300** can be a gun-like component that has a proximal housing **310** that includes a pistol grip type handle **311** and trigger **312** that moves from a spring retained released position to an engaged position upon squeezing and holding the trigger (as further illustrated below). The trigger **312** is linked to moveable internal features within the proximal housing **310** to provide desired functions during implantation described below. The delivery device **300** includes an elongate tube **306** extending distally from the proximal housing **310**. The elongate tube **306** includes a longitudinal slot **307** over its length for receiving sutures therethrough as anchors are passed through the central lumen of the tube.

(68) FIG. **10A** also shows that the proximal housing **310** is associated with a bone punch having a distal punch head **322** and a proximal punch head **323**. The proximal punch head **323** has a tapping surface **324** at its proximal side. Combined elements **322** and **323** form a punch head assembly. As illustrated in FIG. **10B**, the bone punch also includes a punch pin **320** having a tapered point **321** adapted for probing through the tendon and/or grabbing the tendon to aid positioning. Positioning may include positioning the tendon in its original footprint, for tendons that are detached. In some examples, positioning as a separate step may be omitted or limited, such as when repairing a partial tear, such as a partial thickness articular side tear or combination of full thickness and articular partial thickness tears. A tendon can be considered positioned at a location for securing to bone

either by virtue of having placed a fully torn or detached tendon at a location, such as its original footprint, where it can be re-attached, or, with a partial tear, when the tendon is located where a physician desires to have it when applying anchors to repair or otherwise address the partial tear. (69) The punch pin **320** and tip are configured for being pounded into bone to create a bone hole. The tapered tip **321** is also used in some methods disclosed herein to engage and push against the proximal end of an anchor. The punch pin **320** extends through the proximal housing **310** and the elongate tube **306**. The punch pin **320** is affixed to the proximal punch head **323** and is slidable within the distal punch head **322**. The distal punch head **322** snap latches to the proximal housing **310** of the delivery device. The proximal punch head **323** and distal punch head **322** are connected by a spring-loaded mechanism that holds the punch pin **320** in a fully extended position when the proximal punch head **323** is pushed against the distal punch head **322** and latched. When the proximal punch head **323** is released from close connection with the distal punch head **322**, the spring loading causes the punch pin **320** to withdraw proximally to a partially retracted position with only a short distal portion of the punch pin **320** extending beyond the elongate tube **306** for use in probing a potential implant site (as shown in FIG. **10B**, for example).

(70) FIG. **10C** shows another configuration of the delivery device **300**. The action of latching together the proximal punch head **323** with the distal punch head **322**, with the distal punch head **322** engaged with the proximal housing **310**, advances the anchor delivery tube **330** distally, and forces a distal portion (referred to as a nub, **332** as highlighted in FIG. **10D**) of the anchor delivery tube **330** past the distal end of the elongate tube **306**. The anchor delivery tube **330** also has a longitudinal slot **331** aligned with the longitudinal slot **307** of the elongate tube for passing a suture therethrough. The fully extended configuration shown in FIG. **10C** may be achieved as the surgeon begins tapping or pounding on the proximal punch head **323**, causing the proximal and distal punch heads **322**, **323** to latch together on the proximal housing.

(71) FIG. **10D** shows a next configuration of the delivery device. Here, the distal punch head **322** is no longer engaged with the proximal housing **310**, and the proximal and distal punch heads **323**, **322**, are not latched together. The disengagement of the distal punch head **322** and housing **310**, and disengagement of the proximal and distal punch heads **323**, **322**, is caused by actuation of the trigger **312**. As described in the method illustration of FIGS. **12A-12G**, below, this configuration would arise after a bone hole is created, and is used to introduce an anchor/suture into the anchor delivery tube for implant. To facilitate such a step in the procedure, a portion of the anchor delivery tube **330** referred to as the nub **332** remains extended from the distal end of the elongate tube **306**. With the bone punch retracted or removed, the anchor delivery tube **330** now defines an open lumen **333** to allow an anchor to be introduced and passed therethrough with the aid of the reinserted bone punch, as detailed below. As also highlighted in FIG. **10D**, optionally, the distal end of the elongate tube **306** may be tapered as shown at **308**. The taper **308**, in some examples, provides the elongate tube **306** with a blunt distal tip that can be used to maintain force against the outside of a tendon during manipulation of an anchor and/or tensioning of a stitch between two anchors.

(72) At a high level, the procedure may be understood as follows. With the anchor delivery device **300** in the configuration shown in FIG. **10B**, the physician may probe the surgical site to identify a location where an anchor is to be implanted. Once the desired location is identified, the physician applies force to the tapping surface **324** of the bone punch to force the bone punch through the tendon and to create a bone hole using the distal tip **321** of the punch pin **320**. As the physician advances the bone punch in this manner, the proximal and distal punch heads **323**, **322** will become latched together to form the configuration as shown in FIG. **10C**. The same action of advancing the bone punch relative to the elongate tube also advances the anchor delivery tube **330** and nub **332** beyond the distal end of the elongate tube **306**. Next, the trigger **312** is actuated to release the bone punch **320**, pulling the bone punch **320** in a proximal direction to create the configuration as shown in FIG. **10D**. The implant tool **300** is held in position, using the nub **332** to maintain registration

with the formed bone hole. In some examples, a portion of the nub **332** will be inserted into the bone hole.

(73) An anchor is then introduced into the anchor delivery tube **330** and passed down the lumen **333** thereof to the distal end, with force applied to advance the anchor using the bone punch assembly. Complete insertion of the anchor can be confirmed by maintaining pressure against the tendon to hold the nub **332** in the desired registration relative to the bone hole, and pushing the proximal punch head **323** in the distal direction until the distal punch head **322** latches with the proximal housing **310** and the proximal punch head **323** latches with the distal punch head **322**. Now the trigger **312** will again be actuated, however, due to mechanisms within the housing, this second actuation of the trigger after insertion of the anchor will apply positive retraction force, along with spring force, to retract the anchor delivery tube **330** and nub **332** into the distal end of the elongate tube **306**, as well as retracting the bone punch.

(74) With the nub retracted, the physician can manipulate toggling of the anchor using the working suture without the nub **332** possibly damaging the working suture, while force can be maintained against the tendon and bone by pressing the distal tip of the elongated tube **306** against the tendon. After toggling the anchor, the delivery tool **300** is pulled back from the implant position and the suture lock is secured by pulling on the suture lock cord. If the anchor is the second or a subsequent anchor in a series, the physician may tighten the working suture to form a stitch while keeping pressure against the tendon with the elongated tube **306** prior to moving the delivery device to a next position. The delivery device is then reset and the configuration of FIG. **10B** is again assumed.

(75) The above procedure may be modified for insertion of a soft suture anchor by, for example, removing the bone punch and inserting into the anchor delivery lumen an anchor delivery tool. The anchor delivery tool may have a slot, for example, for holding such a soft anchor, as disclosed in U.S. Prov. Pat. App. No. 63/231,143, filed Aug. 9, 2021, titled TENSIONABLE AND LOCKABLE SOFT SUTURE ANCHORS AND ANCHOR ARRAYS FOR ANATOMICAL ATTACHMENT OF SOFT TISSUE TO BONE the disclosure of which is incorporated herein by reference.

(76) Referring now to FIG. **11A**, another embodiment of a delivery device is illustrated. The device of FIG. **11A** features a plunger for delivering anchors from individual cartridges **391** to the delivery device and a magazine **390** for holding cartridges on the anchor delivery device. The delivery device is generally shown at **300** with the proximal housing at **310**. On one side of the proximal housing is a receiver **370** into which a plunger **380** can be slidably placed and retained. The top of the receiver includes a slot **372** for receiving a cartridge **392** that carries an anchor to be implanted. The cartridge **392** can be seen to have at least first and second ends of a working suture **393** extending therefrom.

(77) The delivery device is shown relative to a patient **400** having a patient portal **402**, which may be for example a shoulder portal that is formed for performing arthroscopic surgery. In the example shown, the removed cartridge **392** is shown with the working suture **393** extending on either side thereof. The physician may pull the cartridge away from the magazine and the delivery device, as well as the portal **402**, in order to floss the working suture **393** so that an amount of slack is available on either side of the anchor contained in the cartridge **392**. The purpose of this maneuver is to ensure that as the anchor is advanced through the delivery device and into the patient, there will be enough slack to make this passage easy. That is, while it is possible to floss the suture through the anchor during delivery and implantation, it may be preferable to generate slack before the implantation to make advancement of the anchor into position relatively easier. Once the anchor is positioned, the extra slack can be taken out as the physician tensions the working suture to create a stitch between the anchor being toggled and a previously placed anchor.

(78) Opposite the plunger **380** is a magazine **390** that can be releasably secured to the proximal housing **310** and carries a plurality of cartridges **391**. A cartridge ejector is shown at **394** for ejecting cartridges **391/392** one at a time. The magazine is shown with 7 cartridges **391** therein, the 8.sup.th cartridge **392** having already been ejected. In the example shown, at least one additional

cartridge has already been ejected and used, since the working suture **393** can be seen to extend into the elongate tube **306** and into the patient portal **402**. It will be understood as well that the magazine is carried on receiver **398**. Greater detail regarding the magazine and its use can be found in U.S. patent application Ser. No. 17/551,811, filed on Dec. 15, 2021 and titled DELIVERY DEVICE FOR IMPLANTING KNOTLESS MICRO-SUTURE ANCHORS AND ANCHOR ARRAYS FOR ATTACHMENT OF SOFT TISSUE TO BONE, the disclosure of which is incorporated herein by reference.

(79) As illustrated in FIG. **11A**, the plunger itself is shown at **380**, in an extended position relative to a receiver **370**. When a cartridge is placed in the slot **372**, the plunger **380** can be depressed at an appropriate time during an anchor implantation procedure. Doing so laterally transfers the anchor from the cartridge into the bore through the length of the delivery device. The anchor is then ready to be inserted by advancing the bone punch **320** (or anchor insertion tool for a soft anchor) through the proximal housing and down the anchor delivery lumen. The plunger prepares the anchor for delivery by pushing the anchor laterally to the midline of the anchor delivery tube, and holds the anchor in position until the bone punch is advanced to push the anchor down the anchor delivery tube.

(80) In some examples, and as further explained in U.S. patent application Ser. No. 17/551,811, filed on Dec. 15, 2021 and titled DELIVERY DEVICE FOR IMPLANTING KNOTLESS MICRO-SUTURE ANCHORS AND ANCHOR ARRAYS FOR ATTACHMENT OF SOFT TISSUE TO BONE, depressing the plunger **380** modifies the position of various features internal to the housing **310** which determine what happens when the trigger is actuated. In an example, when the plunger is **380** is in the extended position, actuating the trigger will mechanically pull on the bone punch **320** to retract it from a fully extended position (FIG. **10C**), but does not retract the nub and in fact leaves the nub and anchor delivery tool in an extended, locked position, transitioning the distal end of the anchor delivery tool from the configuration of FIG. **10C** to the configuration of FIG. **10D**. The same actuation of the trigger with the plunger **380** in a depressed position, which would occur after the plunger is depressed to move an anchor into the anchor delivery tube and after the bone punch (or anchor insertion tool for a soft anchor) has been fully inserted, will retract both the bone punch (or anchor insertion tool) and the nub, transitioning the distal end of the anchor delivery tool from the configuration of FIG. **10C** to the configuration of FIG. **10B**. Illustrative examples of mechanical features to provide this functionality are disclosed in the Ser. No. 17/551,811 Application, the disclosure of which is incorporated herein by reference.

(81) FIGS. **11B-11E** illustrate a cartridge for holding an anchor. Starting with FIG. **11B**, a cartridge **500** is illustrated with a handle **502** adapted for grasping by the user/physician. An inner holder is shown at **510**, and is surrounded by a cover **520**. The inner holder **510** secures an anchor **100** between an upper anchor support **511** and a boss **512**. In the configuration shown in FIG. **11B**, the cartridge is “closed” in that the anchor **100** cannot be removed.

(82) FIG. **11C** shows the cover **520** raised to an “open” position in which the anchor **100** is no longer secured by the cover **520**. The cover defines two channels at **522**, **524**. First channel **522** provides a path for the working suture out of the cartridge **500**, and second channel **524** provides a path for the suture lock, as will be further detailed below. The cover **520** may be spring biased to the closed position, if desired, to prevent inadvertent removal of the anchor **100** during handling. Alternatively, the cover **520** can include detents to hold the cover in a closed position until pressure is applied during insertion. In addition, the upper anchor support **511** and boss **512** are spaced so that the anchor **100** is held in position against falling out.

(83) An alternative design may have the inner holder **510** open in alignment with slot **527** to allow anchor removal in an axial, rather than lateral direction. For such an alternative, in an example, the upper anchor support **511** and boss **512** would be positioned higher up on the inner holder, such that the anchor **100** would be held at position **516** as shown in FIG. **11C**. Doing so may remove the need to have a plunger on the proximal housing of the anchor delivery tool.

(84) FIG. 11D shows the cartridge 500 in the closed position again with the cover down. The working suture 530 is now shown passing through first channel 522. The suture lock is shown as well, with the free end 540 of the suture lock passing through second channel 524 and the locking loop shown at 542. As can be seen, the boss 512 holds the working suture 530 away from the underside of the anchor 510, making flossing of the working suture easier prior to release of the anchor 510 from the cartridge. That is, because the bottom side of the anchor 510 may include a channel that makes flossing of the working suture therethrough more difficult, keeping the working suture 530 away from the bottom side of the anchor 100 may make flossing easier. Also, when the working suture 530 is pulled close to the bottom side of the anchor 100, the path that must be navigated when flossing includes first and second near ninety degree turns, increasing friction as the working suture 530 is flossed. Therefore, the boss 512 can be seen to make flossing easier in some examples. In other examples, the boss 512 may be designed so that the working suture does not wrap around it, and instead a simple support on the bottom side of the anchor 100 may be provided, with the working suture then resting between the support and the bottom side of the anchor. It may also be noted that having the working suture placed as shown may aid in retaining the anchor in place until it is ejected by the insertion of the plunger in the examples shown above.

(85) FIG. 11E shows the back side of the cartridge 500. Of note here, the free end 540 of the locking loop 542 passes into a channel and then to a spool 514. In an example, the free end 540 is attached to the spool 514, such as by a knot, so that the free end can be pulled a select distance (10 to 20 cm, for example) before reaching a point where it can no longer unspool. When the physician seeks to use the locking loop, the cartridge 500 can be grasped and pulled until the spool runs out. The physician can then pull on the cartridge and therefore on the free end of the locking loop until the locking loop secures the working suture position and, then, breaks at the break knot (or other preferential point of failure), as described below and above. The result is that the physician can manually grasp the cartridge to easily lock the locking loop and break the free end of the locking loop without needing a special tool and/or without needing to attempt to grasp the thin cord of the free end of the locking loop. It can be observed that the spool 514 includes inner features 515 allowing a tool to be inserted and twisted to spool the free end 540 of the locking loop onto the spool 514. As with the steps of toggling and/or tensioning a stitch, the distal end of the anchor delivery tool may be used to apply exterior pressure on the tendon as the locking loop is tightened and the free end is broken off.

(86) Additional details regarding an illustrative magazine and its use may be found in U.S. patent application Ser. No. 17/551,811, filed on Dec. 15, 2021 and titled DELIVERY DEVICE FOR IMPLANTING KNOTLESS MICRO-SUTURE ANCHORS AND ANCHOR ARRAYS FOR ATTACHMENT OF SOFT TISSUE TO BONE, the disclosure of which is incorporated herein by reference.

(87) It should be noted that the illustrative anchor implantation system shown is but one example of how the presently disclosed anchor system may be implanted. For example, a system that fully withdraws the distal end of the bone punch back to the proximal housing as shown may not be necessary. Separate cartridges for each bone anchor are illustrated in the implantation system; in other examples, several anchors may be disposed together in one cartridge in a longitudinal fashion, for example, for sequential loading. Another anchor delivery tool is disclosed, for example, in U.S. Provisional Patent Application Ser. No. 63/172,629, filed Apr. 8, 2021 and titled DELIVERY DEVICE FOR IMPLANTING KNOTLESS MICRO-SUTURE ANCHORS AND ANCHOR ARRAYS FOR ATTACHMENT OF SOFT TISSUE TO BONE, the disclosure of which is incorporated herein by reference. Rather than lateral release of an anchor from a cartridge, an axial release may be used. In some examples, a cartridge can be omitted entirely. Any suitable implantation system may be used, as desired.

(88) In FIGS. 12A through 12G, an exemplary method for implanting individual and an array of anchors is depicted. Referring first to FIG. 12A, a schematic of select parts of the shoulder rotator

cuff **600** is depicted in order to explain the methods of implantation. The illustration includes a portion of the humeral head **602** shown including an outer cortical shell layer **604** and an inner cancellous bone material **606**. A tendon, in this case the supraspinatus tendon **608** is shown overlaying a portion of the humeral head where is attached to the footprint. The method is a transtendinous or through the tendon repair. The tendon **608** is first positioned in a desired location for reattachment to bone in the footprint of original attachment.

(89) The delivery device disclosed herein, or similar is then utilized to implant the type suture anchor through the tendon **608**. To begin the delivery device is set with the distal nub **332** extending from the distal end of the implant delivery tube **330** and elongate tube **306**. The bone punch **320** is fully inserted distally so that it extends beyond the distal end of the nub **332** and is locked in place, as is the nub locked in place. The device as configured is positioned on the tendon at the desired anchor placement and pounded in until the distal end of the outer tubular member is in contact with the tendon as shown in FIG. **12A**. At this point the nub **332** extends through at least a portion of the cortical shell **604** (in thinner bone the nub **332** can extend into the cancellous bone **606**) and the distal end of the bone punch **320** extends deeper into the cancellous bone **606**. To achieve the desired depth of implantation to assure toggling, the bone punch extends beyond the elongate tube **306** distal end a distance of greater than or equal to about 20 mm. Further, to assure nub registration with the bone hole, the nub portion **332** extends beyond the elongate tube **306** distal end a distance of about 6 to about 10 mm.

(90) As depicted in FIG. **12B**, the bone punch **320** is then retracted while maintaining the elongate tube **306** and nub portion **332** in place, with the nub portion **332** providing registration with the formed hole in the bone. Absent such registration with the bone hole by the nub portion **332**, the location under the tendon would be lost and it would be very difficult to feed an anchor through the tendon which would tend to fill the hole through which the bone punch traveled. In some examples, as described above, this step of the method may be performed by depressing a trigger on an implant tool where the implant tool is configured to maintain the nub portion **332** extended under certain circumstances (for example, with the slide stop in place) while applying a positive retraction force to the bone punch **320**.

(91) The first type anchor is transferred or inserted into the proximal portion of the anchor delivery tube inside the elongate tube **306**. As shown in FIG. **12C**, the bone punch **320** is then reinserted into the lumen of the anchor delivery tube and advanced distally. The example shown illustrates a toggle-type anchor with a toggle body **100** being inserted; the skilled person will understand variation in steps that would occur for the other anchor types disclosed above.

(92) As shown in FIG. **12C**, the toggle body **100** of the anchor is pushed out the distal end by the bone punch **320**. The bone punch **320** continues to be advanced in the distal direction to its original depth to push the toggle body **100** into the bone, as shown in FIG. **12D**. It has been found that pushing the proximal end of the anchor deep into the bone with the toggle body **100** having an angled distal end causes or at least initiates rotation of the toggle body **100**. This initial rotation assures continued rotation upon pulling tension on the working suture **115** outside the body.

(93) As shown in FIG. **12E**, the bone punch **320** and nub **332** are then retracted by the application of positive force by the trigger (as discussed above), as well as with spring action. This assures the nub **332** does not cut or fray the working suture. The bone hole remains shown in the drawings. The distal portion of the working suture extending from the distal passage is then pulled to complete the toggling of the anchor as aided by the proximal fins on the toggle body.

(94) With continued tension on the working suture, the toggle body **100** is pulled toward the inside surface of the cortical shell of the bone as shown in FIG. **12F**. To aid this step, the anchor delivery tool distal end may be pressed against the tendon to provide a counterforce against pullout during toggling and/or suture tensioning; that is, as the anchor is toggled and the suture is tensioned, the toggle body **100** may reach and press against the cortical shell. Additional counterforce can be applied in particular in regions of thinner cortical shell, such as at the edges or outside of a tear

footprint, and/or between the greater and lesser tubercles of the humerus. As depicted in FIG. 12G, once the working suture is tensioned, the locking suture is tensioned to close the locking loop **118** around the working suture and fix the working suture relative to the toggle body **100**. In some examples, the locking suture is broken during this step at the knot which is at or inside the central bore of the anchor **100**, thus, FIG. 12G shows only the working suture extending back into the elongate tube **306**.

(95) With implantation of the first anchor, the working suture is simply locked as it cannot be tensioned to form a stitch until the second anchor is implant. In some examples, the first anchor in a chain of anchors can be pre-locked for this purpose; in other examples the surgeon will lock the first anchor suture lock at the time of implant. Therefore, in preferred methods, the second anchor is implanted repeating the above steps, except to the extent that the suture lock is differently engaged. As the working suture is pulled to toggle the anchor, any loose working suture between the first and second anchors is pulled through to form the tensioned stitch. During suture tensioning the distal end of the elongate tube **306** can be maintained against the outer surface of the tendon to prevent pullout or even possible bone fracture at the cortical shell. Once properly tensioned, the second anchor is locked. These steps are repeated for the rest of the anchors in an array. In some examples, one or more intermediate anchors do not include a suture lock, as noted above, though other steps of the method shown in FIGS. 12A-12G may remain the same.

(96) The preceding provides a relatively complete description of the anchor itself, pre-strung anchor arrays, suture lock, cartridge, magazine, and anchor delivery tool. A range of inventions are thus disclosed, and not all components or parts needs to be used together. For example, the delivery tool may be configured to for use with other anchors, cartridges, magazines, etc. Likewise, the anchors may be used in different configurations with other working suture and suture lock arrangements, other cartridges, magazines and delivery tools. Thus the overall combination shown can be modified in a variety of ways.

(97) Additional features and alternative designs for various components, subassemblies and assemblies may be found in the following patent applications, each of which is incorporated herein by reference:

(98) U.S. Prov. Pat. App. No. 63/172,564, filed Apr. 8, 2021, titled KNOTLESS MICRO SUTURE ANCHORS AND ANCHOR ARRAYS FOR ANATOMICAL ATTACHMENT OF SOFT TISSUE TO BONE, and U.S. patent application Ser. No. 17/551,588, filed on Dec. 15, 2021 and titled KNOTLESS MICRO SUTURE ANCHORS AND ANCHOR ARRAYS FOR ANATOMICAL ATTACHMENT OF SOFT TISSUE TO BONE.

(99) U.S. Prov. Pat. App. No. 63/172,565, filed Apr. 8, 2021, titled TENSIONABLE AND LOCKABLE MICRO SUTURE ANCHORS AND ANCHOR ARRAYS FOR

(100) ANATOMICAL ATTACHMENT OF SOFT TISSUE TO BONE, and U.S. patent application Ser. No. 17/551,709, filed on Dec. 15, 2021 and titled TENSIONABLE AND LOCKABLE MICRO SUTURE ANCHORS AND ANCHOR ARRAYS FOR ANATOMICAL ATTACHMENT OF SOFT TISSUE TO BONE.

(101) U.S. Prov. Pat. App. No. 63/172,613, filed Apr. 8, 2021, titled KNOTLESS MICRO SUTURE ANCHOR ARRAY FOR HIGH DENSITY ANATOMICAL ATTACHMENT OF SOFT TISSUE TO BONE and U.S. patent application Ser. No. 17/551,728, filed on Dec. 15, 2021 and titled KNOTLESS MICRO SUTURE ANCHOR ARRAY FOR HIGH DENSITY ANATOMICAL ATTACHMENT OF SOFT TISSUE TO BONE.

(102) U.S. Prov. Pat. App. No. 63/172,629, filed Apr. 8, 2021, titled DELIVERY DEVICE FOR IMPLANTING KNOTLESS MICRO-SUTURE ANCHORS AND ANCHOR ARRAYS FOR ATTACHMENT OF SOFT TISSUE TO BONE, and U.S. Prov. Pat. App. No. 63/281,411, filed Nov. 19, 2021, titled DELIVERY DEVICE FOR IMPLANTING KNOTLESS MICRO-SUTURE ANCHORS AND ANCHOR ARRAYS FOR ATTACHMENT OF SOFT TISSUE TO BONE, and U.S. patent application Ser. No. 17/551,811, filed on Dec. 15, 2021 and titled DELIVERY DEVICE

FOR IMPLANTING KNOTLESS MICRO-SUTURE ANCHORS AND ANCHOR ARRAYS FOR ATTACHMENT OF SOFT TISSUE TO BONE.

(103) U.S. Prov. Pat. App. No. 63/172,624, filed Apr. 8, 2021, titled CARTRIDGE DEVICE FOR SUTURE ANCHOR AND SUTURE MANAGEMENT DURING IMPLANTATION OF A MICRO SUTURE ANCHOR ARRAY, and U.S. patent application Ser. No. 17/551,838, filed on Dec. 15, 2021 and titled CARTRIDGE DEVICE FOR SUTURE ANCHOR AND SUTURE MANAGEMENT DURING IMPLANTATION OF A MICRO SUTURE ANCHOR ARRAY.

(104) U.S. Prov. Pat. App. No. 63/172,568, filed Apr. 8, 2021, titled LOCKING SUTURE CONSTRUCT FOR TENSIONED SUTURE TO SUTURE BRIDGES IN ANCHOR ARRAYS FOR ATTACHING SOFT TISSUE TO BONE and U.S. patent application Ser. No. 17/551,860, filed on Dec. 15, 2021 and titled LOCKING SUTURE CONSTRUCT FOR TENSIONED SUTURE TO SUTURE STITCHES IN ANCHOR ARRAYS FOR ATTACHING SOFT TISSUE TO BONE.

(105) U.S. Prov. Pat. App. No. 63/172,630, filed Apr. 8, 2021, titled METHODS FOR TRANSTENDINOUS IMPLANTATION OF KNOTLESS MICRO SUTURE ANCHORS AND ANCHOR ARRAYS, and U.S. patent application Ser. No. 17/551,885, filed on Dec. 15, 2021 and titled METHODS FOR TRANSTENDINOUS IMPLANTATION OF KNOTLESS MICRO SUTURE ANCHORS AND ANCHOR ARRAYS.

(106) Each of these non-limiting examples can stand on its own or can be combined in various permutations or combinations with one or more of the other examples.

(107) The above detailed description includes references to the accompanying drawings, which form a part of the detailed description. The drawings show, by way of illustration, specific embodiments. These embodiments are also referred to herein as “examples.” Such examples can include elements in addition to those shown or described. However, the present inventors also contemplate examples in which only those elements shown or described are provided. Moreover, the present inventors also contemplate examples using any combination or permutation of those elements shown or described (or one or more aspects thereof), either with respect to a particular example (or one or more aspects thereof), or with respect to other examples (or one or more aspects thereof) shown or described herein.

(108) In the event of inconsistent usages between this document and any documents so incorporated by reference, the usage in this document controls. In this document, the terms “a” or “an” are used, as is common in patent documents, to include one or more than one, independent of any other instances or usages of “at least one” or “one or more.” Moreover, in the claims, the terms “first,” “second,” and “third,” etc. are used merely as labels, and are not intended to impose numerical requirements on their objects.

(109) The above description is intended to be illustrative, and not restrictive. For example, the above-described examples (or one or more aspects thereof) may be used in combination with each other. Other embodiments can be used, such as by one of ordinary skill in the art upon reviewing the above description. The Abstract is provided to comply with 37 C.F.R. § 1.72(b), to allow the reader to quickly ascertain the nature of the technical disclosure. It is submitted with the understanding that it will not be used to interpret or limit the scope or meaning of the claims.

(110) Also, in the above Detailed Description, various features may be grouped together to streamline the disclosure. This should not be interpreted as intending that an unclaimed disclosed feature is essential to any claim. Rather, innovative subject matter may lie in less than all features of a particular disclosed embodiment. Thus, the following claims are hereby incorporated into the Detailed Description as examples or embodiments, with each claim standing on its own as a separate embodiment, and it is contemplated that such embodiments can be combined with each other in various combinations or permutations. The scope of the protection should be determined with reference to the appended claims, along with the full scope of equivalents to which such claims are entitled.

Claims

1. A method for repairing a torn rotator cuff tendon by reattachment to a humeral head utilizing a seam-like row of serial stitches extending over incremental portions of the torn rotator cuff tendon, the method comprising the steps of: providing a pre-strung plurality of anchors including a first anchor, a plurality of intermediate anchors and a final anchor with each of the pre-strung plurality of anchors having a bone hole insertion diameter of less than or equal to about 3 mm, wherein the first anchor includes a length of working suture affixed thereto, each of the plurality of intermediate anchor having at least one passage therethrough with the length of working suture slidably and serially threaded through the at least one passage of each of the plurality of intermediate anchor and the final anchor also having at least one passage therethrough with the length of working suture slidably threaded therethrough and further including a means for locking the length of working suture relative to the final anchor; implanting each of the pre-strung plurality of anchors in a single row through the torn rotator cuff tendon, the pre-strung plurality of anchors each implanted in a bone hole, with the plurality of intermediate anchors placed in bone holes formed in an original footprint of the torn rotator cuff tendon with spacing between adjacent anchors a distance of about 10 mm or less as measured from center of bone hole to center of bone hole; applying tension to the length of working suture after implantation of each plurality of intermediate anchor and the final anchor implantation to form a series of single suture tensioned stitches between the pre-strung plurality of anchors; and securing the length of working suture as tensioned to the final anchor with the means for locking the length of working suture.
2. The method of claim 1, wherein the plurality of intermediate anchors includes six anchors.
3. The method of claim 1, wherein the means for locking the length of working suture is a locking loop, the locking loop extending adjacent the final anchor wherein the locking loop encircles a portion of the length of the working suture, the locking loop having a first open position allowing the length of working suture to slide through the locking loop and a second closed position engaging the length of working suture and preventing sliding of the length of working suture within the locking loop.
4. The method of claim 1, wherein the means for locking the length of working suture is a knot tied in the length of working suture adjacent the passage in the final anchor after tensioning.
5. The method of claim 1, wherein the means for locking the length of working suture is a one-way slip knot that allows sliding of the length of working suture in a first direction but prevents sliding in an opposite direction, wherein the step of applying tension to the length of working suture is performed by pulling the length of working suture in the first direction.
6. The method of claim 1, wherein the means for locking the length of working suture is a set of angled projections in the passages of the pre-strung plurality of anchors that allow the length of working suture movement in a first direction but prevent the length of working suture movement in a direction opposite the first direction, and the step of applying tension to length of working suture is performed by pulling the length of working suture in the first direction.
7. The method of claim 1, wherein the means for locking the length of working suture is mechanical lock that compresses and secures the length of working suture to the final anchor when activated.
8. The method of claim 1, wherein each plurality of intermediate anchor further includes means for locking the length of working suture thereto.
9. A method for securing a supraspinatus tendon and an infraspinatus tendon to a humeral head, the method comprising the steps of: providing a first anchor having a length of a working suture secured thereto, the first anchor having an insertion diameter in a bone hole of less than or equal to about 3 mm; implanting the first anchor through either of the supraspinatus tendon and the infraspinatus tendon into the humeral head within a medial half of an original combined footprint

of attachment of the supraspinatus and infraspinatus tendons to the humeral head; providing a plurality of intermediate anchors, each plurality of intermediate anchor slidably received on the length of working suture, each plurality of intermediate anchor having a passage therethrough wherein each plurality of intermediate anchor is configured for insertion in a bone hole having a diameter of less than or equal to about 3 mm; implanting the plurality of intermediate anchors through the supraspinatus and infraspinatus tendons into bone holes formed in the humeral head in a serial row within the medial half of the original combined footprint, wherein the first of the plurality of intermediate anchors is spaced from the first anchor by a distance of less than or equal to about 10 mm measured from center of bone hole to center of bone hole and each subsequently implanted plurality of intermediate anchors is spaced from the just previously implanted adjacent plurality of intermediate anchors by a distance of less than about 10 mm measured from center of bone hole to center of bone hole; providing a final anchor, the final anchor slidably received on the length of working suture, the final anchor having a passage therethrough wherein the final anchor is configured for insertion in a bone hole having a diameter of less than or equal to about 3 mm wherein the final anchor further includes means for selectively locking the length of working suture relative to the final anchor; implanting the final anchor through either of the supraspinatus tendon and the infraspinatus tendon into a bone hole formed in the humeral head within the medial half of the original combined footprint, wherein the final anchor is spaced from the last plurality of intermediate anchors by a distance of less than or equal to about 10 mm measured from center of bone hole to center of bone hole; tensioning the length of working suture at locations along its length to form a stitch extending between each anchor in series to hold the supraspinatus tendon and the infraspinatus tendon against the humeral head; and, securing the tensioned length of working suture to the final anchor utilizing the means for selective locking to maintain desired tension in the formed individual stitches.

10. The method of claim 9, wherein the plurality of intermediate anchors includes six anchors.

11. The method of claim 9, wherein the means for locking the length of working suture is a locking loop, the locking loop extending adjacent the final anchor wherein the locking loop encircles a portion of the length of the working suture, the locking loop having a first open position allowing the length of working suture to slide through the locking loop and a second closed position engaging the length of working suture and preventing sliding of the length of working suture within the locking loop.

12. The method of claim 9, wherein the means for locking the length of working suture is a knot tied in the length of working suture adjacent the passage in the final anchor after tensioning.

13. The method of claim 9, wherein the means for locking the length of working suture is a one-way slip knot that allows sliding of the length of working suture in a first direction but prevents sliding in a direction opposite the first direction, wherein the step of tensioning the length of working suture includes pulling the length of working suture in the first direction.

14. The method of claim 9, wherein the means for locking the length of working suture is a one way passage through the final anchor that includes angled projections that allow the length of working suture movement in a first direction but prevents moving the length of working suture movement in a direction opposite the first direction wherein the step of tensioning the length of working suture includes pulling the length of working suture in the first direction.

15. The method of claim 9, wherein the means for locking the length of working suture is a mechanical lock that compresses and secures the length of working suture to the final anchor when activated.

16. The method of claim 9, wherein each intermediate anchor further includes means for locking the length of working suture thereto.

17. A method for repairing a torn rotator cuff tendon by reattachment to the humeral head utilizing a seam-like row of serial stitches extending over incremental portions of the torn rotator cuff tendon in an original footprint of the torn rotator cuff tendon on the humeral head, the method

comprising the steps of: providing a pre-strung plurality of anchors, including a first anchor, a plurality of intermediate anchors and a final anchor with each of the pre-strung plurality of anchors having a bone hole insertion diameter of less than or equal to about 3 mm, wherein the first anchor includes a length of working suture affixed thereto, each of the plurality of intermediate anchors having at least one passage therethrough with the length of working suture slidably and serially threaded through the at least one passage of each of the plurality of intermediate anchor and the final anchor also at least one passage therethrough with the length of working suture slidably threaded therethrough and further including a separate locking loop, wherein the separate locking loop encircles a portion of the length of the working suture adjacent the final anchor and having a first position allowing the length of working suture to slide through the locking loop and a second position engaging the length of working suture and preventing sliding within the locking loop; implanting the first anchor through the torn rotator cuff tendon into a bone hole formed in the humeral head within the original footprint; implanting a first of the plurality of intermediate anchors of the pre-strung plurality of anchors in a bone hole formed in the original footprint a distance of about 10 mm or less from the first anchor as measured from center of bone hole to center of bone hole, then applying tension to the length of working suture extending from the first anchor and passing through the first intermediate anchor to form a single suture tensioned stitch between the first anchor and first intermediate anchor; implanting a second of the intermediate anchors in the pre-strung plurality of anchors in the original footprint a distance of about 10 mm or less from the first intermediate anchor as measured from center of bone hole to center of bone hole, then applying tension to the length of working suture extending from the first intermediate anchor and passing through the second intermediate anchor to form a single suture tensioned stitch between the first and second intermediate anchors; repeating the spacing distances, implanting and tensioning steps for each subsequent serial plurality of intermediate anchors and the final anchor; and, activating the locking loop to maintain tension in the created array.

18. The method of claim 17, wherein the plurality of intermediate anchors includes six anchors.

19. The method of claim 17, wherein each plurality of intermediate anchor further includes means for locking the length of working suture thereto.

20. The method of claim 17, wherein at least one of the plurality of intermediate anchor omits any means for locking the length of working suture thereto.
