

US Patent & Trademark Office

Patent Public Search | Text View

United States Patent	12397156
Kind Code	B2
Date of Patent	August 26, 2025
Inventor(s)	Denison; Timothy et al.

Devices, systems, and methods for incontinence control

Abstract

Provided are devices and methods for preventing an episode of incontinence in an individual in need thereof. The devices comprise a sensor and a stimulator electrode that can be implanted into the body of the individual. Once the device is implanted in the individual, the sensor of the device senses a parameter that is associated with a response from the individual that is intended to prevent an episode of incontinence. Then, the device provides an electrical stimulation using the electrode that, together with the response, helps to prevent the episode of incontinence.

Inventors: Denison; Timothy (Oxford, GB), De Wachter; Stefan (Antwerp, BE), Knowles; Charles (London, GB), Crawley; Aidan (London, GB)

Applicant: Amber Therapeutics Ltd (London, GB)

Family ID: 1000008779079

Assignee: Amber Therapeutics Holdings Limited (London, GB)

Appl. No.: 18/734326

Filed: June 05, 2024

Prior Publication Data

Document Identifier	Publication Date
US 20240424296 A1	Dec. 26, 2024

Related U.S. Application Data

continuation parent-doc US 18213225 20230622 US 12064625 child-doc US 18734326
continuation parent-doc US 18068020 20221219 US 12059564 child-doc US 18213225
continuation parent-doc US 17692611 20220311 US 11565109 20230131 child-doc US 18068020
us-provisional-application US 63160322 20210312

Publication Classification

Int. Cl.: A61N1/36 (20060101); A61N1/05 (20060101)

U.S. Cl.:

CPC A61N1/36007 (20130101); A61N1/0551 (20130101); A61N1/36139 (20130101);

Field of Classification Search

CPC: A61N (1/36007)

USPC: 607/41

References Cited

U.S. PATENT DOCUMENTS

Patent No.	Issued Date	Patentee Name	U.S. Cl.	CPC
3942535	12/1975	Schulman	N/A	N/A
5733313	12/1997	Barreras et al.	N/A	N/A
6055456	12/1999	Gerber	N/A	N/A
7295881	12/2006	Cohen et al.	N/A	N/A
7328068	12/2007	Spinelli et al.	N/A	N/A
7415308	12/2007	Gerber et al.	N/A	N/A
7522061	12/2008	Rondoni et al.	N/A	N/A
7647112	12/2009	Tracey et al.	N/A	N/A
7684873	12/2009	Gerber	N/A	N/A
7686760	12/2009	Anderson et al.	N/A	N/A
7702385	12/2009	Moffitt et al.	N/A	N/A
7729772	12/2009	Williams et al.	N/A	N/A
7774069	12/2009	Olson et al.	N/A	N/A
7855653	12/2009	Rondoni et al.	N/A	N/A
7891085	12/2010	Kuzma et al.	N/A	N/A
7894913	12/2010	Boggs et al.	N/A	N/A
7912555	12/2010	Swoyer et al.	N/A	N/A
7974703	12/2010	Goetz et al.	N/A	N/A
7996092	12/2010	Mrva et al.	N/A	N/A
8019423	12/2010	Possover	N/A	N/A
8036756	12/2010	Swoyer et al.	N/A	N/A
8068910	12/2010	Gerber et al.	N/A	N/A
8072338	12/2010	Rondoni et al.	N/A	N/A
8121691	12/2011	Gerber et al.	N/A	N/A
8165695	12/2011	Diubaldi et al.	N/A	N/A
8260436	12/2011	Gerber et al.	N/A	N/A
8260439	12/2011	Diubaldi et al.	N/A	N/A
8295933	12/2011	Gerber et al.	N/A	N/A
8352026	12/2012	Diubaldi	N/A	N/A
8428735	12/2012	Littlewood et al.	N/A	N/A
8435166	12/2012	Burnett et al.	N/A	N/A
8457758	12/2012	Olson et al.	N/A	N/A
8494643	12/2012	Cowan et al.	N/A	N/A
8521292	12/2012	Wei et al.	N/A	N/A
8583256	12/2012	Tracey et al.	N/A	N/A
8612018	12/2012	Gillbe	N/A	N/A
8626314	12/2013	Swoyer et al.	N/A	N/A

8630711	12/2013	Wark et al.	N/A	N/A
8639342	12/2013	Possover	N/A	N/A
8644938	12/2013	Craggs	N/A	N/A
8688238	12/2013	Gerber	N/A	N/A
8706232	12/2013	Su et al.	N/A	N/A
8706233	12/2013	Su et al.	N/A	N/A
8738148	12/2013	Olson et al.	N/A	N/A
8744585	12/2013	Gerber et al.	N/A	N/A
8751003	12/2013	Diubaldi et al.	N/A	N/A
8805508	12/2013	Gerber et al.	N/A	N/A
8825165	12/2013	Possover	N/A	N/A
8886321	12/2013	Rohrer et al.	N/A	N/A
8983612	12/2014	Fang et al.	N/A	N/A
8989861	12/2014	Su et al.	N/A	N/A
9005102	12/2014	Burnett et al.	N/A	N/A
9072886	12/2014	Gaunt et al.	N/A	N/A
9089693	12/2014	Swoyer et al.	N/A	N/A
9089699	12/2014	Su et al.	N/A	N/A
9108025	12/2014	Spenser et al.	N/A	N/A
9114261	12/2014	Yonce	N/A	N/A
9155855	12/2014	Tebbutt et al.	N/A	N/A
9155885	12/2014	Wei et al.	N/A	N/A
9168374	12/2014	Su	N/A	N/A
9173678	12/2014	Siegel et al.	N/A	N/A
9192763	12/2014	Gerber et al.	N/A	N/A
9220895	12/2014	Siff et al.	N/A	N/A
9259587	12/2015	Goetz et al.	N/A	N/A
9265927	12/2015	Yonce et al.	N/A	N/A
9272140	12/2015	Gerber	N/A	N/A
9289607	12/2015	Su et al.	N/A	N/A
9327117	12/2015	Denison et al.	N/A	N/A
9381351	12/2015	Haessler	N/A	N/A
9433783	12/2015	Wei et al.	N/A	N/A
9446235	12/2015	Su et al.	N/A	N/A
9463324	12/2015	Olson et al.	N/A	N/A
9480840	12/2015	Rohrer et al.	N/A	N/A
9498633	12/2015	Laing et al.	N/A	N/A
9555246	12/2016	Jiang et al.	N/A	N/A
9561366	12/2016	Wei et al.	N/A	N/A
9597520	12/2016	Gillbe et al.	N/A	N/A
9623243	12/2016	Chancellor et al.	N/A	N/A
9623253	12/2016	Perryman et al.	N/A	N/A
9630004	12/2016	Rajguru et al.	N/A	N/A
9707390	12/2016	Ahmed	N/A	N/A
9707391	12/2016	Ahmed	N/A	N/A
9707396	12/2016	Su et al.	N/A	N/A
9724509	12/2016	Su et al.	N/A	N/A
9724511	12/2016	Wei et al.	N/A	N/A
9737707	12/2016	Haessler et al.	N/A	N/A
9757584	12/2016	Burnett	N/A	N/A
9764132	12/2016	Rohrer et al.	N/A	N/A
9782583	12/2016	Sharma	N/A	N/A
9789252	12/2016	Gerber et al.	N/A	N/A

9821112	12/2016	Olson et al.	N/A	N/A
9855423	12/2017	Jiang et al.	N/A	N/A
9861547	12/2017	Crunick et al.	N/A	N/A
9950159	12/2017	Beck et al.	N/A	N/A
9956404	12/2017	Brink et al.	N/A	N/A
9993639	12/2017	Gerber	N/A	N/A
9999532	12/2017	Mische	N/A	N/A
9999763	12/2017	Shiroff et al.	N/A	N/A
10004557	12/2017	Gross	N/A	N/A
10016599	12/2017	Lockwood et al.	N/A	N/A
10035020	12/2017	Wang et al.	N/A	N/A
10046164	12/2017	Gerber	N/A	N/A
10076661	12/2017	Wei et al.	N/A	N/A
10086197	12/2017	Lin et al.	N/A	N/A
10143850	12/2017	Cowan et al.	N/A	N/A
10195419	12/2018	Shiroff et al.	N/A	N/A
10201702	12/2018	Bonde et al.	N/A	N/A
10213604	12/2018	Dinsmoor et al.	N/A	N/A
10220205	12/2018	Bhadra et al.	N/A	N/A
10227713	12/2018	Ganatra et al.	N/A	N/A
10232175	12/2018	Possover	N/A	N/A
10238451	12/2018	Sachs et al.	N/A	N/A
10238877	12/2018	Kaula et al.	N/A	N/A
10245439	12/2018	Schwarz et al.	N/A	N/A
10315031	12/2018	Brink et al.	N/A	N/A
10384067	12/2018	Jiang et al.	N/A	N/A
10391313	12/2018	Moffitt et al.	N/A	N/A
10471255	12/2018	Imran	N/A	N/A
10478249	12/2018	Gross et al.	N/A	N/A
10478612	12/2018	Schepis et al.	N/A	N/A
10478619	12/2018	Lee et al.	N/A	N/A
10493270	12/2018	Ahmed	N/A	N/A
10518086	12/2018	Su et al.	N/A	N/A
10525268	12/2019	Torgerson	N/A	N/A
10531944	12/2019	Lamraoui	N/A	N/A
10543366	12/2019	Christopherson et al.	N/A	N/A
10549087	12/2019	Yoo et al.	N/A	N/A
10556104	12/2019	Gerber	N/A	N/A
10561835	12/2019	Gerber	N/A	N/A
10583284	12/2019	Peters et al.	N/A	N/A
10583298	12/2019	Wang et al.	N/A	N/A
10603502	12/2019	Cogan et al.	N/A	N/A
10674963	12/2019	Toth	N/A	N/A
10744260	12/2019	Gerber et al.	N/A	N/A
10751533	12/2019	Edgerton et al.	N/A	N/A
10758723	12/2019	Fang et al.	N/A	N/A
10765355	12/2019	Nelson et al.	N/A	N/A
10828490	12/2019	Sachs et al.	N/A	N/A
10835748	12/2019	Ackermann et al.	N/A	N/A
10994134	12/2020	Thor	N/A	N/A
10994140	12/2020	Yeh et al.	N/A	N/A
11013911	12/2020	Swoyer et al.	N/A	N/A
11045650	12/2020	Brink et al.	N/A	N/A

11077301	12/2020	Creasey et al.	N/A	N/A
11103723	12/2020	Deisseroth et al.	N/A	N/A
11110271	12/2020	Schobben et al.	N/A	N/A
11116985	12/2020	Jiang et al.	N/A	N/A
11147966	12/2020	Tranchina et al.	N/A	N/A
11167138	12/2020	Patel et al.	N/A	N/A
11198006	12/2020	Nijlunsing et al.	N/A	N/A
11213340	12/2021	Su et al.	N/A	N/A
11224748	12/2021	Dinsmoor et al.	N/A	N/A
11229788	12/2021	John et al.	N/A	N/A
11285321	12/2021	Ganguly et al.	N/A	N/A
11305110	12/2021	Toong et al.	N/A	N/A
11318319	12/2021	Schobben et al.	N/A	N/A
11324947	12/2021	Kaula et al.	N/A	N/A
11406822	12/2021	Hvalsøe et al.	N/A	N/A
11413452	12/2021	Shepherd et al.	N/A	N/A
11420052	12/2021	Doskocil et al.	N/A	N/A
11426579	12/2021	John et al.	N/A	N/A
11446490	12/2021	Doskocil	N/A	N/A
11529513	12/2021	John et al.	N/A	N/A
11534200	12/2021	Bonde et al.	N/A	N/A
11565109	12/2022	Denison et al.	N/A	N/A
11571575	12/2022	Su et al.	N/A	N/A
11596792	12/2022	Campean et al.	N/A	N/A
11633593	12/2022	John et al.	N/A	N/A
11638814	12/2022	Doskocil et al.	N/A	N/A
11642540	12/2022	Bornzin et al.	N/A	N/A
11672978	12/2022	Su et al.	N/A	N/A
11712557	12/2022	Toong et al.	N/A	N/A
11730411	12/2022	Jiang et al.	N/A	N/A
11752334	12/2022	Yoo et al.	N/A	N/A
11766561	12/2022	Mishra et al.	N/A	N/A
11779761	12/2022	Druke et al.	N/A	N/A
11794001	12/2022	Imran	N/A	N/A
11794009	12/2022	Tai	N/A	N/A
11826427	12/2022	Garibyan et al.	N/A	N/A
11839762	12/2022	Doskocil et al.	N/A	N/A
11844943	12/2022	Rajguru et al.	N/A	N/A
11850434	12/2022	Knudsen et al.	N/A	N/A
11896734	12/2023	Williams et al.	N/A	N/A
11896824	12/2023	Doskocil	N/A	N/A
11911614	12/2023	Grill et al.	N/A	N/A
11918806	12/2023	Wong et al.	N/A	N/A
11925801	12/2023	Wei et al.	N/A	N/A
11925802	12/2023	Tai	N/A	N/A
11957894	12/2023	Chasensky et al.	N/A	N/A
12059564	12/2023	Denison et al.	N/A	N/A
12064625	12/2023	Denison et al.	N/A	N/A
2002/0055761	12/2001	Mann et al.	N/A	N/A
2003/0045919	12/2002	Swoyer et al.	N/A	N/A
2003/0100930	12/2002	Cohen et al.	N/A	N/A
2005/0216069	12/2004	Cohen et al.	N/A	N/A
2006/0004421	12/2005	Bennett et al.	N/A	N/A

2006/0020297	12/2005	Gerber et al.	N/A	N/A
2006/0035882	12/2005	Koga et al.	N/A	N/A
2006/0122660	12/2005	Boveja et al.	N/A	N/A
2006/0247723	12/2005	Gerber et al.	N/A	N/A
2007/0100387	12/2006	Gerber	N/A	N/A
2007/0150034	12/2006	Rooney et al.	N/A	N/A
2008/0103570	12/2007	Gerber	N/A	N/A
2008/0300649	12/2007	Gerber et al.	N/A	N/A
2009/0036945	12/2008	Chancellor et al.	N/A	N/A
2009/0054950	12/2008	Stephens	N/A	N/A
2009/0254144	12/2008	Bhadra et al.	N/A	N/A
2010/0094372	12/2009	Grill et al.	N/A	N/A
2012/0010680	12/2011	Wei et al.	N/A	N/A
2012/0197336	12/2011	Su	N/A	N/A
2013/0041430	12/2012	Wang et al.	N/A	N/A
2013/0289664	12/2012	Johanek	N/A	N/A
2013/0331905	12/2012	Haessler	N/A	N/A
2014/0058474	12/2013	Haessler	N/A	N/A
2014/0163579	12/2013	Tischendorf et al.	N/A	N/A
2014/0309708	12/2013	Sharma	N/A	N/A
2015/0080906	12/2014	Fjorback et al.	N/A	N/A
2016/0045731	12/2015	Simon et al.	N/A	N/A
2016/0263376	12/2015	Yoo	N/A	A61N 1/36017
2016/0338902	12/2015	St et al.	N/A	N/A
2017/0001003	12/2016	Pivonka et al.	N/A	N/A
2017/0095667	12/2016	Yakovlev et al.	N/A	N/A
2017/0173327	12/2016	Wei et al.	N/A	N/A
2017/0231547	12/2016	Imran	N/A	N/A
2018/0008311	12/2017	Shiroff et al.	N/A	N/A
2018/0078760	12/2017	Lee et al.	N/A	N/A
2018/0214691	12/2017	Famm et al.	N/A	N/A
2018/0289971	12/2017	Yeh et al.	N/A	N/A
2018/0296497	12/2017	Schepis et al.	N/A	N/A
2019/0110732	12/2018	Toth et al.	N/A	N/A
2019/0117974	12/2018	Creasey et al.	N/A	N/A
2019/0143102	12/2018	Romero-Ortega et al.	N/A	N/A
2019/0247664	12/2018	Irazoqui et al.	N/A	N/A
2019/0269924	12/2018	Su et al.	N/A	N/A
2019/0290902	12/2018	Romero-Ortega et al.	N/A	N/A
2019/0336763	12/2018	Spurling et al.	N/A	N/A
2019/0374776	12/2018	Mishra et al.	N/A	N/A
2020/0046552	12/2019	Velis et al.	N/A	N/A
2020/0069939	12/2019	Imran	N/A	N/A
2020/0101294	12/2019	Torgerson	N/A	N/A
2020/0139018	12/2019	Martin et al.	N/A	N/A
2020/0139115	12/2019	Verity	N/A	N/A
2020/0147375	12/2019	Su et al.	N/A	N/A
2020/0171312	12/2019	Dinsmoor et al.	N/A	N/A
2020/0171313	12/2019	Dinsmoor et al.	N/A	N/A
2020/0230406	12/2019	Brink et al.	N/A	N/A
2020/0246617	12/2019	Errico et al.	N/A	N/A
2020/0338267	12/2019	Gerber et al.	N/A	N/A
2020/0376266	12/2019	Toong et al.	N/A	N/A

2020/0397361	12/2019	Nelson et al.	N/A	N/A
2021/0069513	12/2020	Beer et al.	N/A	N/A
2021/0085962	12/2020	Patel	N/A	N/A
2021/0101008	12/2020	Sandgren et al.	N/A	N/A
2021/0113834	12/2020	Wong et al.	N/A	N/A
2021/0175727	12/2020	Schobben et al.	N/A	N/A
2021/0187291	12/2020	Iyer et al.	N/A	N/A
2021/0268278	12/2020	Agahari	N/A	N/A
2021/0275808	12/2020	Jacobson et al.	N/A	N/A
2021/0299442	12/2020	Wei et al.	N/A	N/A
2021/0315638	12/2020	Townley et al.	N/A	N/A
2021/0330981	12/2020	Mishra et al.	N/A	N/A
2021/0346695	12/2020	Grill et al.	N/A	N/A
2021/0361952	12/2020	Bonde et al.	N/A	N/A
2021/0379382	12/2020	John et al.	N/A	N/A
2022/0022951	12/2021	Townley	N/A	N/A
2022/0088373	12/2021	Burnett	N/A	N/A
2022/0105348	12/2021	Brink et al.	N/A	N/A
2022/0111211	12/2021	Li et al.	N/A	N/A
2022/0118245	12/2021	Hvalsoe et al.	N/A	N/A
2022/0134105	12/2021	Hvalsoe et al.	N/A	N/A
2022/0134118	12/2021	Johnson	N/A	N/A
2022/0143394	12/2021	John et al.	N/A	N/A
2022/0143403	12/2021	Offutt et al.	N/A	N/A
2022/0152402	12/2021	Mueller et al.	N/A	N/A
2022/0176128	12/2021	Dinsmoor et al.	N/A	N/A
2022/0193406	12/2021	Imran	N/A	N/A
2022/0193428	12/2021	Martens et al.	N/A	N/A
2022/0218985	12/2021	Schobben et al.	N/A	N/A
2022/0218986	12/2021	Schobben et al.	N/A	N/A
2022/0233848	12/2021	Gad et al.	N/A	N/A
2022/0233860	12/2021	Hamner et al.	N/A	N/A
2022/0241599	12/2021	Fischer et al.	N/A	N/A
2022/0266012	12/2021	Hamner et al.	N/A	N/A
2022/0288388	12/2021	Rondoni et al.	N/A	N/A
2022/0288389	12/2021	Denison et al.	N/A	N/A
2022/0323751	12/2021	Sharma	N/A	N/A
2022/0331584	12/2021	Offutt et al.	N/A	N/A
2022/0331586	12/2021	Offutt et al.	N/A	N/A
2022/0331589	12/2021	Bittner et al.	N/A	N/A
2022/0339443	12/2021	Schobben et al.	N/A	N/A
2022/0347461	12/2021	Campean et al.	N/A	N/A
2022/0361941	12/2021	Townley	N/A	N/A
2022/0395686	12/2021	Tai	N/A	N/A
2022/0401747	12/2021	Possover	N/A	N/A
2022/0409896	12/2021	John et al.	N/A	N/A
2023/0001185	12/2022	Schobben et al.	N/A	N/A
2023/0021801	12/2022	Sridhar et al.	N/A	N/A
2023/0029600	12/2022	Pivonka et al.	N/A	N/A
2023/0042557	12/2022	Bennett et al.	N/A	N/A
2023/0059066	12/2022	Sridhar et al.	N/A	N/A
2023/0074492	12/2022	Offutt et al.	N/A	N/A
2023/0079328	12/2022	Matton et al.	N/A	N/A

2023/0083643	12/2022	Toong et al.	N/A	N/A
2023/0103791	12/2022	Davis et al.	N/A	N/A
2023/0113727	12/2022	Van et al.	N/A	N/A
2023/0121038	12/2022	John et al.	N/A	N/A
2023/0130942	12/2022	Denison et al.	N/A	N/A
2023/0139116	12/2022	Wiygul et al.	N/A	N/A
2023/0149709	12/2022	Romero-Ortega et al.	N/A	N/A
2023/0157604	12/2022	Imran	N/A	N/A
2023/0158306	12/2022	Jiang et al.	N/A	N/A
2023/0166113	12/2022	Nesbitt et al.	N/A	N/A
2023/0166116	12/2022	Crites-Bachert	N/A	N/A
2023/0173274	12/2022	Lee	N/A	N/A
2023/0211158	12/2022	Grill et al.	N/A	N/A
2023/0226356	12/2022	Toong et al.	N/A	N/A
2023/0233853	12/2022	Zirpel et al.	N/A	N/A
2023/0256244	12/2022	John et al.	N/A	N/A
2023/0264019	12/2022	Doskocil et al.	N/A	N/A
2023/0271005	12/2022	Ostroff et al.	N/A	N/A
2023/0277844	12/2022	Errico et al.	N/A	N/A
2023/0302287	12/2022	Jiang et al.	N/A	N/A
2023/0347145	12/2022	Denison et al.	N/A	N/A
2023/0364413	12/2022	Romaniw et al.	N/A	N/A
2023/0381504	12/2022	Yoo et al.	N/A	N/A
2023/0390461	12/2022	Crook et al.	N/A	N/A
2023/0414950	12/2022	Maharbiz et al.	N/A	N/A
2024/0024667	12/2023	Coates et al.	N/A	N/A
2024/0074888	12/2023	Bennett et al.	N/A	N/A
2024/0075295	12/2023	Tesfayesus	N/A	N/A
2024/0087736	12/2023	Errico et al.	N/A	N/A
2024/0091538	12/2023	Fraga et al.	N/A	N/A
2024/0123224	12/2023	Lysy et al.	N/A	N/A
2024/0123225	12/2023	Paralikar et al.	N/A	N/A
2024/0278022	12/2023	Mishra et al.	N/A	N/A
2024/0359007	12/2023	Arora et al.	N/A	N/A

FOREIGN PATENT DOCUMENTS

Patent No.	Application Date	Country	CPC
2019204642	12/2018	AU	N/A
3160396	12/2020	CA	N/A
101460218	12/2008	CN	N/A
1761303	12/2006	EP	N/A
1940510	12/2007	EP	N/A
1967519	12/2007	EP	N/A
1981589	12/2007	EP	N/A
2023819	12/2008	EP	N/A
2155062	12/2009	EP	N/A
2162181	12/2009	EP	N/A
2166950	12/2009	EP	N/A
2314585	12/2010	EP	N/A
2389975	12/2010	EP	N/A
2717961	12/2013	EP	N/A
2861145	12/2014	EP	N/A
2841008	12/2016	EP	N/A

3212162	12/2016	EP	N/A
3242712	12/2016	EP	N/A
3270834	12/2017	EP	N/A
3157620	12/2017	EP	N/A
3310429	12/2017	EP	N/A
3191172	12/2017	EP	N/A
3318304	12/2017	EP	N/A
3331601	12/2017	EP	N/A
3253441	12/2017	EP	N/A
3180072	12/2017	EP	N/A
3364870	12/2018	EP	N/A
3468657	12/2018	EP	N/A
3481494	12/2018	EP	N/A
3532154	12/2018	EP	N/A
3573519	12/2018	EP	N/A
3589248	12/2019	EP	N/A
3319685	12/2019	EP	N/A
3180070	12/2020	EP	N/A
3813647	12/2020	EP	N/A
3866913	12/2020	EP	N/A
3762092	12/2020	EP	N/A
3924035	12/2020	EP	N/A
3969102	12/2021	EP	N/A
3976169	12/2021	EP	N/A
4021518	12/2021	EP	N/A
4069062	12/2021	EP	N/A
4093485	12/2021	EP	N/A
3057652	12/2022	EP	N/A
4240436	12/2022	EP	N/A
3560552	12/2022	EP	N/A
4308219	12/2023	EP	N/A
4326391	12/2023	EP	N/A
2414410	12/2004	GB	N/A
01/83029	12/2000	WO	N/A
2007/074916	12/2006	WO	N/A
2007/082382	12/2006	WO	N/A
2007/130168	12/2006	WO	N/A
2008/147612	12/2007	WO	N/A
2008/150590	12/2007	WO	N/A
2008/153726	12/2007	WO	N/A
2009/134475	12/2008	WO	N/A
2009/137119	12/2008	WO	N/A
2010/111321	12/2009	WO	N/A
2011/156288	12/2010	WO	N/A
2015/059570	12/2014	WO	N/A
2015/196164	12/2014	WO	N/A
2016/039768	12/2015	WO	N/A
2016/127130	12/2015	WO	N/A
2016/154076	12/2015	WO	N/A
2016/203258	12/2015	WO	N/A
2016/209997	12/2015	WO	N/A
2017/070372	12/2016	WO	N/A
2018/038794	12/2017	WO	N/A

2018/053336	12/2017	WO	N/A
2019/173084	12/2018	WO	N/A
2019/209417	12/2018	WO	N/A
2020/006048	12/2019	WO	N/A
2020/033883	12/2019	WO	N/A
2020/165591	12/2019	WO	N/A
2020/231440	12/2019	WO	N/A
2020/243104	12/2019	WO	N/A
2021/042044	12/2020	WO	N/A
2021/176219	12/2020	WO	N/A
2022/189862	12/2021	WO	N/A
2022/200814	12/2021	WO	N/A
2023/161720	12/2022	WO	N/A
2023/177650	12/2022	WO	N/A
2023/230131	12/2022	WO	N/A
2023/242367	12/2022	WO	N/A
2023/250426	12/2022	WO	N/A
2024/003537	12/2023	WO	N/A
2024/057025	12/2023	WO	N/A
2024/057026	12/2023	WO	N/A

OTHER PUBLICATIONS

Fischell, R. E., et al., “A Long-Lived, Reliable, Rechargeable Cardiac Pacemaker”, Engineering in Medicine, vol. 1, 1975, 34 pages. cited by applicant

Peters et al.: A prospective, single-blind, randomized crossover trial of sacral vs pudendal nerve stimulation for interstitial cystitis. BJU Int. 2007;100:835-839 (2007). cited by applicant

Standring, S. et al., Gray's anatomy. 39th Ed. Amsterdam: Elsevier, Churchill Livingstone. 2005. cited by applicant

Young, R. F., et al., “Electrical stimulation of the trigeminal nerve root for the treatment of chronic facial pain”, JNS, vol. 83, No. 1, Jul. 1995, pp. 72-78. cited by applicant

Spinelli et al.: A New Minimally Invasive Procedure for Pudenda! Nerve Stimulation to Treat Neurogenic Bladder: Description of the Method and Preliminary Data. Neurourology and Urodynamics. 24:305-309 (2005). cited by applicant

Sterne et al.: RoB 2: a revised tool for assessing risk of bias in randomised trials. BMJ. 366(14898):1-8 (2019). cited by applicant

Su et al.: Electromyographic Responses Across Different Pulse-Widths of Sacral Neuromodulation in Sheep. Neuromodulation. 22:684-689 (2018). cited by applicant

Tai et al.: Bladder Inhibition or Voiding Induced by Pudenda! Nerve Stimulation in Chronic Spinal Cord Injured Cats. Neurourol Urodyn. 26:570-577 (2007). cited by applicant

Talatt, M., Afferent Impulses in the Nerves Supplying the Urinary Bladder, J Physiol (Lon) 1937, vol. 89, pp. 1-13. cited by applicant

Tanagho et al.: Electrical stimulation in the clinical management of the neurogenic bladder. J Urol. 140:1331-1339 (1988). cited by applicant

Thin et al.: Systematic review of the clinical effectiveness of neuromodulation in the treatment of faecal incontinence. Br J Surg. 100:1430-1447 (2013). cited by applicant

Thomas et al.: A pilot study of chronic pudenda! nerve stimulation for faecal incontinence for those who have failed sacral nerve stimulation. Tech Coloproctol. 18:731-737 (2014). cited by applicant

Thuroff et al.: Chronic stimulation of the sacral roots in dogs. Eur Urol. 9:102-108 (1983). cited by applicant

Thuroff et al.: Functional pattern of sacral root stimulation in dogs. II. Urethral closure. J Urol. 127:1034-1038 (1982). cited by applicant

Thuroff et al.: Mechanisms of urinary continence: an animal model to study urethral responses to stress conditions. J Urol. 127:1202-1206 (1982). cited by applicant

Todd, JK., Afferent Impulses in the Pudenda! Nerves of the Cat, Q J Exp Physiol Cogn Med Sci 1964;49, No.

3, pp. 258-267. cited by applicant

Toth et al.: DyNeuMo Mk-2: An Investigational Circadian-Locked Neuromodulator with Responsive Stimulation for Applied Chronobiology, Conf Proc IEEE Int Conf Syst Man Cybern. 2020:3433-3440 (2020). cited by applicant

Uludag et al.: Effect of sacral neuromodulation on the rectum. Br J Surg. 92:1017-1023 (2005). cited by applicant

Vaganeet et al.: Needle Placement and Position of Electrical Stimulation Inside Sacral Foramen Determines Pelvic Floor Electromyographic Response—Implications for Sacral Neuromodulation. Neuromodulation. 22:1-6 (2019). cited by applicant

Vaganeet et al.: Neural pathway of bellows response during SNM treatment revisited: Conclusive evidence for direct efferent motor response. Neurourol Urodyn. 39:1-8 (2020). cited by applicant

Vaganeet et al.: Pelvic Floor Muscle Electromyography as a Guiding Tool During Lead Placement and (Re) Programming in Sacral Neuromodulation Patients: Validity, Reliability, and Feasibility of the Technique. Neuromodulation. 23:1-8 (2020). cited by applicant

Van Asselt, E. et al., “A comparative study of voiding in rat and guinea pig: simultaneous measurement of flow rate and pressure”, Am J Physiol, 1995, vol. 269, R98-103. cited by applicant

Van Breda et al.: Subject-Controlled, On-demand, Dorsal Genital Nerve Stimulation to Treat Urgency Urinary Incontinence; a Pilot. Front Neurosci. 10(24):1-7 (2016). cited by applicant

Vansteensel et al.: Fully Implanted Brain-Computer Interface in a Locked-In Patient with ALS. The New England Journal of Medicine. 375(21):2060-2066 (2016). cited by applicant

Vodusek et al.: Detrusor Inhibition Induced by Stimulation of Pudenda! Nerve Afferents. Neurourology and Urodynamics. 5:381-389 (1986). cited by applicant

Vodusek et al.: Detrusor inhibition on selective pudenda! nerve stimulation in the perineum. Neurourol Urodyn. 6:389-393 (1988). cited by applicant

Wang et al.: Efficacy of Electrical Pudenda! Nerve Stimulation in Treating Female Stress Incontinence. Urology. pp. 1-10 (2016). dx.doi.org/doi: 10.1016/j.urology.2016.02.027. cited by applicant

Wang et al.: Efficacy of Electrical Pudenda! Nerve Stimulation versus Transvaginal Electrical Stimulation in Treating Female Idiopathic Urgency Urinary Incontinence. J Urol. 1-17 (2017) doi:10.1016/j.juro.2017.01.065. cited by applicant

Wang et al.: Long-term efficacy of electrical pudenda! nerve stimulation for urgency-frequency syndrome in women. Int Urogynecol J. 25:397-402 (2014). cited by applicant

Wang et al.: Simultaneous perinea! ultrasound and vaginal pressure measurement prove the action of electrical pudenda! nerve stimulation in treating female stress incontinence. BJU Int. 110:1338-1343 (2012). cited by applicant

Wei et al.: Functional Electrical Stimulation as a Neuroprosthetic Methodology for Enabling Closed-loop Urinary Incontinence Treatment, Proceedings of the 5th International IEEE Embs Conference on Neural Engineering. pp. 650-654 (2011). cited by applicant

Wenzel et al.: Closed Loop Electrical Control of Urinary Continence. The Journal of Urology. 175:1559-1563 (2006). cited by applicant

Wenzel et al.: Detecting the onset of hyper-reflexive bladder contractions from the electrical activity of the pudenda! nerve. IEEE Trans Neural Syst Rehabil Eng. 13(3):428-435 (2005). cited by applicant

Wenzel et al.: Detection of neurogenic detrusor contractions from the activity of the external anal sphincter in cat and human. Neurourol Urodyn. 25:140-147 (2005). cited by applicant

Woock et al.: Activation and inhibition of the micturition reflex by penile afferents in the cat. Am J Physiol Regul Integr Comp Physiol. 294:R1880-1889 (2008). cited by applicant

Woock et al.: Mechanisms of Reflex Bladder Activation by Pudenda! Afferents. Am J Physiol Regul Integr Comp Physiol. 300:R398-407 (2011). cited by applicant

Woock, J.P., “Intraurethral stimulation evokes bladder responses via 2 distinct reflex pathways”, J Urol, 2009, vol. 182, pp. 366-373. cited by applicant

Wunderlich et al.: The overlapping innervation of the two sides of the external anal sphincter by the pudenda! nerves. J Neural Sci. 59:97-109 (1983). cited by applicant

Wyndaele et al. Mechanisms of pelvic organ crosstalk: 1. Peripheral modulation of bladder inhibition by colorectal distention in rats. J Urol. 190:765-71 (2013). cited by applicant

Wyndaele et al.: Detrusor overactivity. Does it represent a difference if patients feel the involuntary contractions?. *The Journal of Urology*. 172:1915-1918 (2004). cited by applicant

Yoo et al.: Bladder Activation by Selective Stimulation of Pudenda! Nerve Afferents in the Cat. *Exp Neural*. 212 (1):218-225 (2008). cited by applicant

Herschorn et al.: Could Reduced Fluid Intake Cause the Placebo Effect Seen in Overactive Bladder Clinical Trials? Analysis of a Large Solifenacin Integrated Database *Urology*. 106:55-59 (2017). cited by applicant

Hokanson et al.: Stimulation of the sensory pudenda! nerve increases bladder capacity in the rat. *Am J Physiol Renal Physiol*. 314:F543-F550 (2018). cited by applicant

Horvath et al.: Conditional and continuous electrical stimulation increase cystometric capacity in persons with spinal cord injury *Neurourol Urodyn*. 29(3):401-407 (2010). cited by applicant

Huang et al.: Preservation of pudenda! afferents in sacral rhizotomies. *Neurosurgery*. 41:411-415 (1997). cited by applicant

Ioannidis et al.: Better reporting of harms in randomized trials: an extension of the CONSORT statement, *Ann Intern Med*, 2004;141:781-788. cited by applicant

Irwin et al.: Population-based survey of urinary incontinence, overactive bladder, and other lower urinary tract symptoms in five countries: results of the EPIC study. *Eur Urol*. 50:1306-1314; discussion 1314-5 (2006). cited by applicant

Ishigooka et al.: Electrical pelvic floor stimulation in the management of urinary incontinence due to neuropathic overactive bladder. *Front Med Biol Eng*. 5(1):1-10 (1993). cited by applicant

Ishigooka et al.: Modulation of the urethral pressure by high-frequency block stimulus in dogs. *Eur Urol*. 25:334-337 (1994). cited by applicant

Jairam et al.: Onset of Action of Sacral Neuromodulation in Lower Urinary Tract Dysfunction—What is the Optimal Duration of Test Stimulation? *The Journal of Urology*. 199:1584-1590 (2018). cited by applicant

Janneck, C., “Electric stimulation of the bladder and the anal sphincter—a new way to treat the neurogenic bladder”, *Anorectal Malformations and Associated Diseases, Progress in Pediatric Surgery*, 1976, vol. pp. 119-139. cited by applicant

Janssen et al.: Measurement properties of the EQ-5D-5L compared to the EQ-5D-3L across eight patient groups: a multi-country study. *Qual Life Res*. 22:1717-1727 (2013). cited by applicant

Jiang, HH. et al. “Electrical stimulation of the pudenda! nerve promotes neuroregeneration and functional recovery from stress urinary incontinence in a rat model” *Am J Physiol Renal Physiol*, 2018, vol. 315, F1555-F1564. cited by applicant

Jiang, HH. et al., “Effects of acute selective pudenda! nerve electrical stimulation after simulated childbirth injury”, *Am J Physiol Renal Physiol*, 2013, vol. 304, F239-F247. cited by applicant

Jimenez-Shahed, J., “Device profile of the percept PC deep brain stimulation system for the treatment of Parkinson's disease and related disorders”, *Expert Rev Med Devices*, 2021, vol. 18, pp. 319-332. cited by applicant

Jones et al.: A Joint Mechanism of Action for Sacral Neuromodulation for Bladder and Bowel Dysfunction? *Urology*. 97:13-19 (2016). cited by applicant

Jottard et al.: Endoscopic trans gluteal minimal-invasive approach for nerve liberation (ENTRAMI technique) in case of pudenda! and/or cluneal neuralgia by entrapment: One-year follow-up. *Neurourol Urodyn*. 39:1-5 (2020). cited by applicant

Jottard et al.: Pilot study: pudenda! neuromodulation combined with pudenda! nerve release in case of chronic perinea! pain syndrome. The ENTRAMI technique: early results, *Int Urogynecol J*. pp. 1-6 (2020). cited by applicant

Jottard et al.: The ENTRAMI technique: Endoscopic transgluteal minimal invasive technique for implantation of a pudenda! electrode under full visual control: A cadaver study. *Neurourol Urodyn*. 38:130-134 (2019). cited by applicant

Juenemann et al.: Clinical significance of sacral and pudenda! nerve anatomy. *J Urol*. 139:74-80 (1988). cited by applicant

Jung, SY, et al., “Urethral afferent nerve activity affects the micturition reflex; implication for the relationship between stress incontinence and detrusor instability”, *J Urol*, 1999;162:204-212. cited by applicant

Kamm et al.: Rectal mucosal electrosensory testing—evidence for a rectal sensory neuropathy in idiopathic constipation. *Dis Colon Rectum*. 33:419-423 (1990). cited by applicant

Kiff et al.: Evidence of pudenda! neuropathy in patients with perineal descent and chronic straining at stool. *Gut*. 25:1279-1282 (1984). cited by applicant

Kiff et al.: Slowed conduction in the pudenda! nerves in idiopathic (neurogenic) faecal incontinence. *BrJ Surg*. 71:614-616 (1984). cited by applicant

Knight et al.: Conditional neuromodulation of neurogenic detrusor overactivity using transrectal stimulation in patients with spinal cord injury: A proof of principle study. *Neurourol Urodyn*. 37:385-393 (2018). cited by applicant

Knowles et al.: New concepts in the pathophysiology of fecal incontinence. *Annals of Laparoscopic and Endoscopic Surgery*. 7(15):1-19 (2022). cited by applicant

Knowles et al.: Prospective randomized double-blind study of temporary sacral nerve stimulation in patients with rectal evacuatory dysfunction and rectal hyposensitivity. *Ann Surg*. 255:643-649 (2012). cited by applicant

Knowles et al.: Sensory and autonomic neuropathy in patients with idiopathic slow-transit constipation. *Br J Surg*. 86:54-60 (1999). cited by applicant

Knowles et al.: Surgery for constipation: systematic review and clinical guidance: Paper 1: Introduction & Methods *Colorectal Dis*. 19(Suppl 3):5-16 (2017). cited by applicant

Knowles et al.: The science behind programming algorithms for sacral neuromodulation. *Colorectal Dis*. 23:592-602 (2021). cited by applicant

Krier et al.: Physiological, morphological, and histochemical properties of cat external anal sphincter. *Am J Physiol*. 255:G772-G778 (1988). cited by applicant

Kruse et al.: Modulation of the spinobulbospinal micturition reflex pathway in cats. *Am J Physiol*. 262:R478-R484 (1992). cited by applicant

Landin et al.: Technology Integration Methods for Bi-directional Brain-computer Interfaces and XR-based Interventions. *Conf Proc IEEE Int Conf Syst Man Cybern. Europe PMC Funders Group*. 3695-3701 (2020). cited by applicant

Le Feser et al.: Pudenda! nerve stimulation induces urethral contraction and relaxation. *Am J Physiol*. 277:R1368-R1375 (1999). cited by applicant

Lee et al.: Semiconditional electrical stimulation of pudenda! nerve afferents stimulation to manage neurogenic detrusor overactivity in patients with spinal cord injury. *Ann Rehabil Med*. 35:605-612 (2011). cited by applicant

Lee, S., et al., “A meta-analysis of the placebo response in antimuscarinic drug trials for overactive bladder” *BMC Med Res Methodol*, 2009; vol. 9, No. 55., pp. 1-2. cited by applicant

Li et al.: Short-term Clinical Efficacy of Electric Pudenda! Nerve Stimulation on Neurogenic Lower Urinary Tract Disease: A Pilot Research. *Urology*. pp. 1-22 doi.org/10.1016/j. urology.2017.10.047 (2017). cited by applicant

Lindstrom et al.: The neurophysiological basis of bladder inhibition in response to intravaginal electrical stimulation. *J Urol*. 129:405-410 (1983). cited by applicant

Loukas et al.: Anatomical and surgical considerations of the sacrotuberous ligament and its relevance in pudenda! nerve entrapment syndrome. *Surg Radial Anat*. 28:163-169 (2006). cited by applicant

Luber, K.M., “The definition, Prevalence, and Risk Factors for Stress Urinary Incontinence”, *Reviews in Urology*, 2004, vol. 6, Suppl 3, pp. S3-S9. cited by applicant

Lubowski et al.: Faecal incontinence associated with reduced pelvic sensation. *Br J Surg*. 75:1086-1088 (1988). cited by applicant

Lyon et al.: Pudenda! but not tibial nerve stimulation inhibits bladder contractions induced by stimulation of pontine micturition center in cats. *Am J Physiol Regul Integr Comp Physiol*. 310:R366-R374 (2016). cited by applicant

Madbouly, K.M. et al., “Temporary sacral nerve stimulation in patients with fecal incontinence owing to rectal hyposensitivity: a prospective, double-blind study”, *Surgery*, 2015, vol. 157, pp. 56-63. cited by applicant

Martens et al.: Minimal invasive electrode implantation for conditional stimulation of the dorsal genital nerve in neurogenic detrusor overactivity. *Spinal Cord*. 49:566-572 (2011). cited by applicant

Martens et al.: Surgical access for electrical stimulation of the pudenda! and dorsal genital nerves in the overactive bladder: a review. *J Urol*. 186:798-804 (2011). cited by applicant

Matharu G, et al., “Relationship Between Urinary Symptoms Reported in a Postal Questionnaire and

Myodynamic Diagnosis”, *Neurourology and Urodynamics*, 2005, vol. 24, pp. 100-105. cited by applicant

Matzel et al.: Electrical stimulation of sacral spinal nerves for treatment of faecal incontinence. *Lancet*. 346:1124-1127 (1995). cited by applicant

Matzel et al.: Neuroanatomy of the striated muscular anal continence mechanism. Implications for the use of neurostimulation. *Dis Colon Rectum*. 33:666-673 (1990). cited by applicant

Matzel et al.: Sacral Neuromodulation: Standardized Electrode Placement Technique. *Neuromodulation*. 20:816-824 (2017). cited by applicant

McCallum et al.: Gastric Pacing Improves Emptying and Symptoms in Patients With Gastroparesis. *Gastroenterology*. 114:456-461 (1998). cited by applicant

McCrery et al.: Treatment of Urinary Urgency Incontinence Using a Rechargeable SNM System: 6-Month Results of the ARTISAN-SNM Study, *The Journal of Urology*. 1-15 (2019) doi:10.1097/JU.0000000000000458. cited by applicant

Abrams, P. et al., “The standardisation of terminology in lower urinary tract function: report from the standardisation sub-committee of the International Continence Society”, *Urology*, 2003;61 :37-49. cited by applicant

Altomare et al.: Long-term outcomes of sacral nerve stimulation for faecal incontinence *Br J Surg*. (102):407-415 (2015). cited by applicant

Amundsen et al.: The Refractory Overactive Bladder: Sacral NEuromodulation vs. BoTulinum Toxin Assessment: ROSETTA trial, *Contemp Clin Trials*. 37(2):272-283 (2014). cited by applicant

Arakalitis et al.: Healthcare professional's choice for surgical management of stress urinary incontinence in a U.K. tertiary hospital, *Eur J Obstet Gynecol Reprod Biol*. 263:7-14 (2021). cited by applicant

Auchincloss et al.: The reliability of surface EMG recorded from the pelvic floor muscles, *Journal of Neuroscience Methods*. 182:85-96 (2009). cited by applicant

Avery et al.: ICIQ: A Brief and Robust Measure for Evaluating the Symptoms and Impact of Urinary Incontinence. *Neurourol Urodyn*. 2004;23:322-330. cited by applicant

Barber et al.: Innervation of the female levator ani muscles. *Am J Obstet Gynecol*. 187:64-71 (2002). cited by applicant

Barrington, F.J.F., “The component reflexes of micturition in the cat. Part III”, *Brain* 1941;54:239-243. cited by applicant

Barrington, F.J.F., “The component reflexes of micturition in the cat. Parts I and II”, *Brain* 1931;54:177-188. cited by applicant

Benson, K., et al., “One-year outcomes of the ARTISAN-SNM study with the Axonics System for the treatment of urinary urgency incontinence”, *Neurourol Urodyn* 2020;39:1482-1488. cited by applicant

Bhadra, N. et al., “High frequency electrical conduction block of the pudenda! nerve”, *J Neural Eng*, 2006, vol. 3, No. 2, pp. 180-187. cited by applicant

Binnie, NR et al., “Use of the pudenda-anal reflex in the treatment of neurogenic faecal incontinence”, *Gut*, 1990, vol. 31, pp. 1051-1055. cited by applicant

Bock, S. et al., “First experiences with pudenda! nerve stimulation in fecal incontinence: a technical report”, *Tech Coloproctol*, 2010, vol. 14, pp. 41-44. cited by applicant

Boggs, J et al., “Frequency-dependent selection of reflexes by pudendal afferents in the cat”, *J Physiol*, 2006, vol. 577, pp. 115-126. cited by applicant

Bosch, JL., “The bion device: a minimally invasive implantable ministimulator for pudendal nerve neuromodulation in patients with detrusor overactivity incontinence”, *Urol Clin North Am*, 2005;32:109-112. cited by applicant

Bosch, RJ, et al., “Perinea! pudenda! neurotomy versus selective neurotomy of the S2 somatic contribution to the pudenda! nerve. Effects on sacral-root-stimulated bladder and urethral responses in the dog”, *Urol Int*, 1992;48:48-52. cited by applicant

Brink, TS, et al., “A Chronic, Conscious Large Animal Platform to Quantify Therapeutic Effects of Sacral Neuromodulation on Bladder Function”, *J Urol*, 2015, vol. 194, pp. 252-258. cited by applicant

Chan, AW, et al., “SPIRIT 2013 statement: defining standard protocol items for clinical trials”, *Ann Intern Med*, 2013;158:200-7. cited by applicant

Chan, C.L. et al., “Rectal sensorimotor dysfunction in patients with urge faecal incontinence: evidence from prolonged manometric studies”, *Gut*, 2005, vol. 54, pp. 1263-1272. cited by applicant

Chartier-Kastler, E, et al., “Neuromodulation sacree avec le systeme InterStim: resultats du registre national francais”, *Progres en urologie*, 2011;21:209-17. cited by applicant

Cong et al.: Effects of Acute Sacral Neuromodulation at Different Pulse Widths on Bladder Overactivity in Pigs. *Int Neurourol J*. 23:109-115 (2019). cited by applicant

Craggs, M. et al., “Neuromodulation of the Lower Urinary Tract”, *Exp Physiol*, 1999, vol. 84, pp. 149-160. cited by applicant

Cumpston et al.: Updated guidance for trusted systematic reviews: a new edition of the Cochrane Handbook for Systematic Reviews of Interventions, *Cochrane Database Syst Rev* 10:ED000142 (2019). cited by applicant

Dalmose, A.L., et al., “Conditional stimulation of the dorsal penile/clitoral nerve may increase cystometric capacity in patients with spinal cord injury”, *Neurourol Urodyn*, 2003;22:130-137. cited by applicant

Desprez, C, et al., “Ten-year Evaluation of a Large Retrospective Cohort Treated by Sacral Nerve Modulation for Fecal Incontinence: Results of a French Multicenter Study”, *Ann Surg*, 2022, vol. 275, No. 4, pp. 735-742. cited by applicant

Evers et al.: Reversal of sensory deficit through sacral neuromodulation in an animal model of fecal incontinence. *Neurogastroenterol Motil*. 28:665-673 (2016). cited by applicant

Evers, J, et al., “Systematic Review of Animal Models Used in Research of Origins and Treatments of Fecal Incontinence”, *Dis Colon Rectum*, 2017; vol. 60, No. 6, pp. 614-626. cited by applicant

Farthing, M.J.G. et al., “Sensibility of Rectum to Distension and Anorectal Distension Reflex in Ulcerative-Colitis” *Gut* 1978, vol. 19, pp. 64-69. cited by applicant

Fassov, J.L. et al., “A Randomized, Controlled, Crossover Study of Sacral Nerve Stimulation for Irritable Bowel Syndrome”, *Ann Surg*, 2014, vol. 260, No. 1, pp. 31-36. cited by applicant

Fjorback, M.V., et al., “Event driven electrical stimulation of the dorsal penile/clitoral nerve for management of neurogenic detrusor overactivity in multiple sclerosis”, *Neurourol Urodyn*, 2006;25:349-355. cited by applicant

Flisser et al.: Urodynamic Classification of Patients with Symptoms of Overactive Bladder. *The Journal of Urology*. 169:529-533; discussion 533-534 (2003). cited by applicant

Fowler et al.: The neural control of micturition. *Nat Rev Neurosci*. 9:453-66 (2008). cited by applicant

Freedman et al.: Combined bladder neck suspension and augmentation cystoplasty for neuropathic incontinence in female patients. *British Journal of Urology*. 73:621-624 (1994). cited by applicant

Garry, R.C. et al., “Reflexes Involving the External Urethral Sphincter in the Cat”, *J Physiol*, 1959, vol. 149, pp. 653-665. cited by applicant

George et al.: A new minimally invasive technique for pudenda! nerve stimulation. *Colorectal Dis*. 14:98-103 (2011). cited by applicant

George et al.: Pudenda! nerve stimulation for bowel dysfunction in complete cauda equina syndrome. *Ann Surg*. 259(3):502-507 (2014). cited by applicant

Godec, C., et al., “Electrical stimulation for incontinence. Technique, selection, and results”, *Urology*, 1976; vol. 7, No. 4, pp. 388-397. cited by applicant

Goldman et al.: Long Term Clinical Results on Treatment of Urinary Urgency Incontinence with the Axonics Rechargeable Sacral Neuromodulation System. *Journal of Urology*. 203(45):E554-E555 (2020). cited by applicant

Goldman, H.B. et al., “Dorsal Genital Nerve Stimulation for the Treatment of Overactive Bladder Symptoms”, *Neurourol Urodyn* 2008, vol. 27, No. 6, pp. 499-503. cited by applicant

Gonzalez E. et al, “Sensory pudendal nerve stimulation increases bladder capacity through sympathetic mechanisms in cyclophosphamide-induced cystitis rats”, *Neurourol Urodyn* 2019, vol. 38, pp. 135-143. cited by applicant

Grigorescu, B.A. et al., “Innervation of the levator ani muscles: description of the nerve branches to the pubococcygeus, iliococcygeus, and puborectalis muscles”, *Int Urogynecol J Pelvic Floor Dysfunct*, 2008;19:107-116. cited by applicant

Groen, J, et al., “Chronic pudenda! nerve neuromodulation in women with idiopathic refractory detrusor overactivity incontinence: results of a pilot study with a novel minimally invasive implantable mini-stimulator”, *Neurourol Urodyn* 2005;24:226-30. cited by applicant

Grover M, et al., “Survey of geriatricians on the effect of fecal incontinence on nursing home referral”, *J Am Geriatr Soc*, 2010;58:1058-62. cited by applicant

Gustafsson, K.J, et al., “Fascicular anatomy and surgical access of the human pudenda! nerve”, *World J Urol*, 2005;23:411-418. cited by applicant

Hannestad, Y., et al., “A community-based epidemiological survey of female urinary incontinence: the Norwegian EPINCONT study”, *The Journal of Clinical Epidemiology*, 2000, vol. 53, pp. 1150-1157. cited by applicant

Hansen, J. et al., “Treatment of neurogenic detrusor overactivity in spinal cord injured patients by conditional electrical stimulation”, *J Urol*, 2005, vol. 173, pp. 2035-2039. cited by applicant

Heinze, K. et al., “Comparative pilot study of implantation techniques for pudenda! neuromodulation: technical and clinical outcome in first 20 patients with chronic pelvic pain”, *World J Urol*, 2015, vol. 33, pp. 289-294. cited by applicant

Heitmann et al.: Understanding the physiology of human defaecation and disorders of continence and evacuation. *Nat Rev Gastroenterol Hepatol*. 18:751-769 (2021). cited by applicant

Henry et al.: Assessment of pelvic-floor disorders and incontinence by electrophysiological recording of the anal reflex. *Lancet*. 1:1290-1291 (1978). cited by applicant

Herdman et al.: Development and preliminary testing of the new five-level version of EQ-5D (EQ-5D-5L). *Qual Life Res*. 20:1727-1736 (2011). cited by applicant

McGee et al.: Selective co-stimulation of pudenda! afferents enhances bladder activation and improves voiding efficiency. *Neurourol Urodyn*. 33(8):1272-1278 (2014). cited by applicant

Meyer et al.: Stimulated pressure profile at rest: a noninvasive method for assessing urethral sphincter function. *Urology*. 52(4):679-684 (1998). cited by applicant

Minagawa et al.: Mechanisms of pelvic organ cross-talk: 2. Impact of colorectal distention on afferent nerve activity of the rat bladder. *J Urol*. 190:1123-1130 (2013). cited by applicant

Moher et al.: Preferred reporting items for systematic reviews and meta-analyses: the PRISMA statement. *BMJ*. 339 (b2535):1-8 (2009). cited by applicant

Nissenkorn et al.: Patient-adjusted intermittent electrostimulation for treating stress and urge urinary incontinence. *BJU International*. 94:105-109 (2004). cited by applicant

No Author, “Excellence in Continence Care: Practical guidance for commissioners, and leaders in health and social care”, NHS England, retrieved from internet <https://www.england.nhs.uk/publication/excellence-in-continence-care/>, retrieved on Mar. 9, 2022, pp. 1-24. cited by applicant

No Author, “MRC/DH/MHRA Joint Project: Risk-adapted Approaches to the Management of Clinical Trials of Investigational Medicinal Products”, MHRA, retrieved from internet https://assets.publishing.service.gov.uk/government/uploads/system/uploads/attachment_data/file/343677/Risk-adapted_approaches_to_the_management_of_clinical_trials_of_investigational_medicinal_products.pdf. Version: Oct. 10, 2011, pp. 1-31. cited by applicant

No Author, “The economic impact of incontinence in Australia”, Deloitte Access Economics, Continence Foundation of Australia, 2011, pp. 1-76. cited by applicant

O'Bichere et al.: New, simple approach for maximal pudenda! nerve exposure: anomalies and prospects for functional reconstruction. *Dis Colon Rectum*. 43(7):956-960 (2000). cited by applicant

Ohlsson et al.: Effects of external and direct pudenda! nerve maximal electrical stimulation in the treatment of the uninhibited overactive bladder. *Br J Urol*. 64:374-80 (1989). cited by applicant

Oliver et al.: Measuring the sensations of urge and bladder filling during cystometry in urge incontinence and the effects of neuromodulation. *Neurourol Urodyn*. 22:7-16 (2003). cited by applicant

Opisso et al.: Patient controlled versus automatic stimulation of pudenda! nerve afferents to treat neurogenic detrusor overactivity. *J Urol*. 180:1403-1408 (2008). cited by applicant

Opisso et al.: Subject-controlled stimulation of dorsal genital nerve to treat neurogenic detrusor overactivity at home. *Neurourol Urodyn*. 32:1004-1009 (2013). cited by applicant

Opisso et al.: Urethral sphincter EMG-controlled dorsal penile/clitoral nerve stimulation to treat neurogenic detrusor overactivity. *J Neural Eng*. 8:036001 pp. 1-9 (2011). cited by applicant

Parks, AG et al, “Sphincter denervation in anorectal incontinence and rectal prolapse”, *Gut*, 1977, vol. 18:656-665. cited by applicant

Peeters et al.: Long-term follow-up of sacral neuromodulation for lower urinary tract dysfunction. *BJU Int*. 113:789-794 (2014). cited by applicant

Peng et al.: Improved Bladder Emptying in Urinary Retention by Electrical Stimulation of Pudenda! Afferents.

J Neural Eng. 5:144-154 (2008). cited by applicant

Peng et al.: Pudenda! neuromodulation with a closed-loop control strategy to improve bladder functions in the animal study. 35th Annual International Conferences of the IEEE EMBS 201. 3626-3629 (2013). cited by applicant

Peng et al.: Role of pudenda! afferents in voiding efficiency in the rat. *Am J Physiol Regul Integr Comp Physiol.* 294: R660-R672 (2008). cited by applicant

Percy et al.: Electrophysiological study of motor nerve supply of pelvic floor. *Lancet.* 1:16-17 (1981). cited by applicant

Perry et al.: An epidemiological study to establish the prevalence of urinary symptoms and felt need in the community: the Leicestershire MRC Incontinence Study. Leicestershire MRC Incontinence Study Team. *J Public Health Med,* 22:427-434 (2000). cited by applicant

Peters et al. Randomized trial of percutaneous tibial nerve stimulation versus Sham efficacy in the treatment of overactive bladder syndrome: results from the SUmIT trial. *J Urol* 183:1438-43 (2010). cited by applicant

Peters et al.: Pilot Study Exploring Chronic Pudenda! Neuromodulation as a Treatment Option for Pain Associated with Pudenda! Neuralgia. *Low Urin Tract Symptoms.* 7:138-142 (2015). cited by applicant

Peters et al.: Sacral versus pudenda! nerve stimulation for voiding dysfunction: a prospective, single-blinded, randomized, crossover trial. *Neurourol Urodyn.* 24:643-647 (2005). cited by applicant

Peters, KM, et al., "Chronic pudendal neuromodulation: expanding available treatment options for refractory urologic symptoms", *Neurourol Urodyn*, 2010;29:1267-71. cited by applicant

Peters, KM., "Alternative approaches to sacral nerve stimulation", *Int Urogynecol J*, 2010;21:1559-1563. cited by applicant

Peyronnet et al.: A Comprehensive Review of Overactive Bladder Pathophysiology: On the Way to Tailored Treatment. *Eur Urol.* 75:988-1000 (2019). cited by applicant

Pezzella et al.: Two-year outcomes of the ARTISAN-SNM study for the treatment of urinary urgency incontinence using the Axonics rechargeable sacral neuromodulation system. *Neurourol Urodyn.* 40:714-721 (2021). cited by applicant

Possover, M., "A new technique of laparoscopic implantation of stimulation electrode to the pudenda! nerve for treatment of refractory fecal incontinence and/or overactive bladder with urinary incontinence" *J Minim Invasive Gynecol*, 2014;21:729. cited by applicant

Possover, M., "A novel implantation technique for pudendal nerve stimulation for treatment of overactive bladder and urgency incontinence", *J Minim Invasive Gynecol*, 2014, pp. 1-20, doi:10.1016/j.jmig.2014.03.026. cited by applicant

Previnaire et al.: Is there a place for pudendal nerve maximal electrical stimulation for the treatment of detrusor hyperreflexia in spinal cord injury patients?. *Spinal Cord.* 36:100-103 (1998). cited by applicant

Previnaire et al.: Short-term effect of pudenda! nerve electrical stimulation on detrusor hyperreflexia in spinal cord injury patients: importance of current strength. *Paraplegia.* 34:95-99 (1996). cited by applicant

Reitz, A, et al., "Afferent fibers of the pudendal nerve modulate sympathetic neurons controlling the bladder neck", *Neurourol Urodyn*, 2003;22:597-601. cited by applicant

Richardson et al.: Anticholinergic drugs and risk of dementia: case-control study. *BMJ.* 361:k1315 (2018). cited by applicant

Rogers et al.: Disposable pudenda! nerve stimulator: evaluation of the standard instrument and new device. *Gut.* 29:1131-1133 (1988). cited by applicant

Rosen et al.: Sacral nerve stimulation as a treatment for fecal incontinence. *Gastroenterology.* 121:536-541 (2001). cited by applicant

Rubin et al.: States Worse Than Death Among Hospitalized Patients With Serious Illnesses. *JAMA Intern Med.* 176:1557-1559 (2016). cited by applicant

Russo et al.: Effective Relief of Pain and Associated Symptoms With Closed-Loop Spinal Cord Stimulation System: Preliminary Results of the Avalon Study. *Neuromodulation.* 21(1):38-47 (2017). cited by applicant

Russo et al.: Sustained Long-Term Outcomes With Closed-Loop Spinal Cord Stimulation: 12-Month Results of the Prospective, Multicenter, Open-Label Avalon Study. *Neurosurgery.* 87:E485-E495 (2020). cited by applicant

Sand et al.: Pelvic floor stimulation: effect on urethral closure pressure profile In: Ostergard DR aBA, ed. *Urodynamics and the Evaluation of Female Incontinence, A Practical Guide.* London: Springer. 128-129

(1995). cited by applicant

Schmidt et al.: Micturition and the male genitourinary response to sacral root stimulation. *Invest Urol.* 17(2):125-129 (1979). cited by applicant

Schmidt et al.: Sacral Nerve Stimulation for Treatment of Refractory Urinary Urge Incontinence. *The Journal of Urology.* 162:352-357 (1999). cited by applicant

Schmidt, R.A., "Technique of Pudenda! Nerve Localization for Block or Stimulation", *The Journal of Urology*, 1989, vol. 142, pp. 1528-1531. cited by applicant

Schultz-Lampel et al.: Experimental results on mechanisms of action of electrical neuromodulation in chronic urinary retention. *World J Urol.* 16:301-304 (1998). cited by applicant

Shafik, A., "Perinea! nerve stimulation for urinary sphincter control. Experimental study", *Urol Res*, 1994;22:151-155. cited by applicant

Siegel et al.: Five-Year Follow up Results of a Prospective, Multicenter Study of Patients with Overactive Bladder Treated with Sacral Neuromodulation. *The Journal of Urology.* 199:229-236 (2018). cited by applicant

Siegel et al.: Results of a prospective, randomized, multicenter study evaluating sacral neuromodulation with InterStim therapy compared to standard medical therapy at 6-months in subjects with mild symptoms of overactive bladder. *Neurourol Urodyn.* 34:224-30 (2015). cited by applicant

Snellings et al.: Effects of stimulation site and stimulation parameters on bladder inhibition by electrical nerve stimulation. *BJU Int.* 110:136-143 (2012). cited by applicant

Snooks et al.: Damage to the innervation of the voluntary anal and periurethral sphincter musculature in incontinence: an electrophysiological study. *J Neural Neurosurg Psychiatry.* 47:1269-1273 (1984). cited by applicant

Snooks et al.: The innervation of the muscles of continence. *Ann R Coll Surg Engl.* 68:45-49 (1986). cited by applicant

Afshar, P., "A translational platform for prototyping closed-loop neuromodulation systems", *Front. Neural Circuits*, vol. 6, Jan. 22, 2013, 15 pages. cited by applicant

Afshar, P., et al., "Advancing neuromodulation using a dynamic control framework.", 2011 Annual International Conference of the IEEE Engineering in Medicine and Biology, 2011. cited by applicant

Anjum MF et al., "Multi-night cortico-basal recordings reveal mechanisms of NREM slow-wave suppression and spontaneous awakenings in Parkinson's disease." *Nature communications.*, Feb. 27, 2024, vol. 15, Issue 1, pp. 1793. cited by applicant

Benson. K. et al., "One-year outcomes of the ARTISAN-SNM study with the Axonics System for the treatment of urinary urgency incontinence," *Neurourology and urodynamics*, vol. 39, Issue 5, 2020, pp. 1482-1488. cited by applicant

Biardeau, X., "Optimization of electrical stimulation/modulation therapies in the management of neurogenic and non-neurogenic lower urinary tract dysfunction.", Diss. University of Antwerp, 2024, pp. 239. cited by applicant

Birder. L.A. et al., "Urethral sensation: basic mechanisms and clinical expressions," *International Journal of Urology*, vol. 21, 2014, pp. 13-16. cited by applicant

Blok, B., et al., "Two-year safety and efficacy outcomes for the treatment of overactive bladder using a long-lived rechargeable sacral neuromodulation system.", *Neurourology and Urodynamics*, vol. 39, Issue. 4, 2020, pp. 1108-1114. cited by applicant

Blok. B. et al., "Programming settings and recharge interval in a prospective study of a rechargeable sacral neuromodulation system for the treatment of overactive bladder," *Neurourology and Urodynamics*, vol. 37, Issue S2, 2018, pp. S17-S22. cited by applicant

Blok. B. et al., "Three month clinical results with a rechargeable sacral neuromodulation system for the treatment of overactive bladder," *Neurourology and urodynamics*, vol. 37, Issue S2, 2018, pp. S9-S16. cited by applicant

Borton D A et al., "Developing collaborative platforms to advance neurotechnology and its translation." *Neuron.*, Oct. 28, 2020, vol. 108, Issue 2, pp. 286-301. cited by applicant

Botha. C. et al., "Preliminary report: modulation of parasympathetic nervous system tone influences oesophageal pain hypersensitivity," *Gut*, vol. 64, Issue 4, 2015, pp. 611-617. cited by applicant

Carrington, E. V. et al., "Sa2032 Anal Evoked Potentials At 'Twitch' Threshold: A New Method for Exploration of Anal Sensory Function", *Neurogastroenterology & Motility*, vol. 62, No. Suppl 1, Jun. 2013, 1

page. cited by applicant

Cui et al., "Acute to long-term characteristics of impedance recordings during neurostimulation in humans." *Journal of neural engineering.*, Apr. 3, 2024, vol. 21, Issue 2, pp. 026022. cited by applicant

D'Ancona. C. et al., "The International Continence Society (ICS) report on the terminology for adult male lower urinary tract and pelvic floor symptoms and dysfunction," *Neurourology and urodynamics*, vol. 38, Issue 2, 2019, pp. 433-477. cited by applicant

De Bock et al., "Influence of nerve transections and combined bladder filling on intravesical electrostimulation-induced bladder contraction in the rat," *European urology*, vol. 56, Issue 3, 2009, pp. 527-533. cited by applicant

De Wacher, et al., "Can a faked cystometry deceive patients in their perception of filling sensations? A study on the reliability of spontaneously reported cystometric filling sensations in patients with non-neurogenic lower urinary tract dysfunction," *Neurourology and Urodynamics: Official Journal of the International Continence Society*, vol. 27, Issue 5, 2008, pp. 395-398. cited by applicant

De Wachter et al., "New technologies and applications in sacral neuromodulation: an update," *Advances in therapy*, vol. 37, 2020, pp. 637-643. cited by applicant

De Wachter et al., "Sacral neuromodulation: mechanism of action," *European urology focus*, vol. 6, Issue 5, 2020, pp. 823-825. cited by applicant

Deli A, et al., "Bioelectronic Zeitgebers: targeted neuromodulation to re-establish circadian rhythms." In 2023 IEEE International Conference on Systems, Man, and Cybernetics (SMC) Oct. 1, 2023, IEEE., pp. 2301-2308. cited by applicant

Denison, T., et al., "A 2 μ W 100 nV/rtHz Chopper-Stabilized Instrumentation Amplifier for Chronic Measurement of Neural Field Potentials", *IEEE Journal of Solid-State Circuits*, vol. 42, No. 12, Dec. 2007, pp. 2934-2945. cited by applicant

Devane, L. A. et al., "Acute lumbosacral nerve stimulation does not affect anorectal motor function in a rodent model", *Neurogastroenterology & Motility*, vol. 28, No. 3, Mar. 2016, pp. 358-363. cited by applicant

Dudding, T. C. et al., "Sacral Nerve Stimulation With Interstim® For The Treatment Of Faecal Incontinence Refractory To Conservative Treatment: A Cost-Utility Analysis", *European Society of Coloproctology Dublin*, HTA Consulting, Sep. 23-25, 2015, 1 page. cited by applicant

Dudding. T.C. et al., "Reprogramming sacral neuromodulation for sub-optimal outcomes: evidence and recommendations for clinical practice," *Neuromodulation: Technology at the Neural Interface*, vol. 24, Issue 7, 2021, pp. 1247-1257. cited by applicant

Eggermont. M. et al., "Innervation of the Epithelium and Lamina Propria of the Urethra of the Female Rat," *The Anatomical Record*, vol. 302, Issue 2, 2019, pp. 201-214. cited by applicant

Eggermont. M. et al., "Regional structural and functional specializations in the urethra of the female rat: evidence for complex physiological control systems," *The Anatomical Record*, vol. 301, Issue 7, 2018, pp. 1276-1289. cited by applicant

Feldmann L K, et al., "Toward therapeutic electrophysiology: beta-band suppression as a biomarker in chronic local field potential recordings." *npj Parkinson's Disease*. Apr. 19, 2022, vol. 8, Issue 1, pp. 44. cited by applicant

Fleming J E, et al., "Embedding digital chronotherapy into bioelectronic medicines." *Iscience*. Apr. 15, 2022, vol. 25, Issue 4, pp. 1-13. cited by applicant

Fleming J E, et al., "From dawn till dusk: Time-adaptive bayesian optimization for neurostimulation." *PLOS Computational Biology.*, Dec. 13, 2023, vol. 19, Issue 12, pp. e1011674. cited by applicant

Gillespie, J. I., et al., "On the origins of the sensory output from the bladder: the concept of afferent noise.", *BJU international*, vol. 103, No. 10, 2009, pp. 1324-1333. cited by applicant

Goldman. H.B. et al., "International Continence Society best practice statement for use of sacral neuromodulation," *Neurourology and urodynamics*, vol. 37, Issue 5, 2018, pp. 1823-1848. cited by applicant

Grossi, U. et al., "Clinical utility of anorectal manometry for the diagnosis of dyssynergic defaecation: blinded assessment reveals that existing criteria do not differentiate between patients and controls", *Neurogastroenterology & Motility*, vol. 65, Sep. 2014, 11 pages. cited by applicant

Grossi. U. et al., "Diagnostic accuracy study of anorectal manometry for diagnosis of dyssynergic defecation," *Gut*, vol. 65, Issue 3, 2016, pp. 447-455. cited by applicant

Gunduz A et al., "Adding wisdom to 'smart' bioelectronic systems: a design framework for physiologic control

including practical examples.” Bioelectronics in medicine., Mar. 1, 2019, vol. 2, Issue 1, pp. 29-41. cited by applicant

Heesakkers. J. et al., “A prospective study to assess the effectiveness and safety of the BlueWind System in the treatment of patients diagnosed with urgency urinary incontinence,” Neurourology and Urodynamics, vol. 43, Issue 7, 2024, pp. 1491-1503. cited by applicant

Heesakkers. J. et al., “LBA01-05 Pivotal Study of a Novel Wirelessly Powered, Patient Tailored Programmed, Tibial Neurostimulator for the Treatment of Patients With Overactive Bladder,” Journal of Urology, vol. 209, Supplement 4, 2023, pp. e1178. cited by applicant

Herroelen, S., et al., “Pudendal nerve stimulation for treatment of lower urinary tract symptoms: A systematic review of safety, technical feasibility and clinical efficacy.”, Continence, 2024. cited by applicant

Horrocks, E. et al., “PTU-051 Post-Partum Defecatory Dysfunction: A Complex Problem Involving Co-Existing Faecal Incontinence and Rectal Evacuatory Disorder”, GI physiology, vol. 62, No. Suppl 1, Jun. 2013, 1 page. cited by applicant

International Search Report and Written Opinion received for PCT Patent Application No. PCT/GB24/052215, mailed on Jan. 13, 2025, 19 pages. cited by applicant

Invitation to Pay Additional Fees received for PCT Patent Application No. PCT/GB2024/052215, mailed on Nov. 22, 2024, 10 pages. cited by applicant

Jottard, K., “Surgical innovations in the treatment of pudendal and/or cluneal nerve entrapment syndrome.”, Diss. University of Antwerp, 2022. cited by applicant

Jottard. K. et al., “Diagnosis and treatment of pudendal and inferior cluneal nerve entrapment syndrome: a narrative review,” Acta Chirurgica Belgica, vol. 122, Issue 6, 2022, pp. 379-389. cited by applicant

Kavoosi A, et al., “MorpheusNet: Resource efficient sleep stage classifier for embedded on-line systems.” In2023 IEEE International Conference on Systems, Man, and Cybernetics (SMC), Oct. 1, 2023, IEEE., pp. 2315-2320. cited by applicant

Knowles. C.H. et al., “The science behind programming algorithms for sacral neuromodulation,” Colorectal Disease, vol. 23, Issue 3, 2021, pp. 592-602. cited by applicant

Kremen V et al., “Continuous active probing and modulation of neural networks with a wireless implantable system.” In2017 IEEE biomedical circuits and systems conference (BioCAS), Oct. 19, 2017, IEEE., pp. 1-4. cited by applicant

Lehur. P. et al., “Programming algorithms for sacral neuromodulation: clinical practice and evidence—recommendations for day-to-day practice,” Neuromodulation: Technology at the Neural Interface, vol. 23, Issue 8, 2020, pp. 1121-1129. cited by applicant

Leong. R.K. et al., “Current information on sacral neuromodulation and botulinum toxin treatment for refractory idiopathic overactive bladder syndrome: a review,” Urologia internationalis, vol. 84, Issue 3, 2010, pp. 245-253. cited by applicant

Leong. R.K. et al., “PNE versus 1st stage timed lead procedure: a direct comparison to select the most sensitive test method to identify patients suitable for sacral neuromodulation therapy,” Neurourology and urodynamics, vol. 30, Issue 7, 2011, pp. 1249-1252. cited by applicant

Maeda, Y. et al., “Sacral nerve stimulation for faecal incontinence and constipation: a European consensus statement”, Colorectal Dis., vol. 17, No. 4, Apr. 2015, pp. 74-87. cited by applicant

Marcelissen. T. et al., “Psychological and psychiatric factors as predictors for success in sacral neuromodulation treatment,” BJU International—British Journal of Urology, vol. 108, Issue 11, 2011, pp. 1834-1837. cited by applicant

Marcelissen. T. et al., “Sacral neuromodulation as a treatment for neuropathic clitoral pain after abdominal hysterectomy,” International urogynecology journal, vol. 21, 2010, pp. 1305-1307. cited by applicant

Mehnert. U. et al., “The management of urine storage dysfunction in the neurological patient,” SN Comprehensive Clinical Medicine, vol. 1, 2019, pp. 160-182. cited by applicant

Naimi. H. et al., “An innovative, non-invasive sensation meter allows for a more comprehensive understanding of bladder sensation events: a prospective study in participants with normal bladder function,” Neurourology and urodynamics, vol. 38, Issue 1, 2019, pp. 208-214. cited by applicant

Neels, H., “Pelvic floor dysfunction in women: tackling barriers.”, University of Antwerp, 2018. cited by applicant

Okun et al., “Definition of implanted neurological device abandonment: a systematic review and consensus

statement.” JAMA network open., Apr. 1, 2024, vol. 7, Issue 4, pp. e248654-e248654. cited by applicant

Peterken F et al., “Adapting the Finetech-Brindley Sacral Anterior Root Stimulator for Bioelectronic Medicine.” In 2021 43rd Annual International Conference of the IEEE Engineering in Medicine & Biology Society (EMBC), Nov. 1, 2021, IEEE., pp. 6406-6411. cited by applicant

Petkos K, et al., A high-performance 8 nV/v Hz 8-channel wearable and wireless system for real-time monitoring of bioelectrical signals. Journal of neuroengineering and rehabilitation. Dec. 2019, vol. 16, pp. 1-24. cited by applicant

Pezella. A. et al., “Two-year outcomes of the ARTISAN-SNM study for the treatment of urinary urgency incontinence using the Axonics rechargeable sacral neuromodulation system,” Neurourology and urodynamics, vol. 40, Issue 2, 2021, pp. 714-721. cited by applicant

Pochopien, M. et al., “PMD85—Cost-Utility Analysis of Sacral Nerve Stimulation for The Treatment of Fecal Incontinence Refractory to Conservative Treatment: Looking for Optimal Treatment Path”, Value in Health, ISPOR, vol. 18, No. 7, Nov. 2015, 359 pages. cited by applicant

Puyuelo, O., et al., “Real-life after sacral nerve modulation implantation: Rate, reasons, and risk factors for mid-term follow-up discontinuation.”, Progres en Urologie, vol. 33, Issue. 17, 2023, pp. 1047-1061. cited by applicant

Quaghebeur, J. et al., “A brief physiology and pathophysiology of the bladder,” Annals of Translational Medicine, vol. 12, Issue 2, 2024, 8 pages. cited by applicant

Quaghebeur, J. et al., “A critical analysis of integral theory statements concerning pathogenesis of LUTS and chronic pelvic pain,” Pelviperineology, vol. 43, Issue 1, 2024, pp. 30-39. cited by applicant

Quaghebeur, J. et al., “Peripheral muscle force balance/imbalance in neurological binary control of bladder function and dysfunction,” Annals of Translational Medicine, vol. 12, Issue 2, 2024, 10 pages. cited by applicant

Quaghebeur. J. et al, “The integral theory, pelvic floor biomechanics, and binary innervation,” International Neurourology Journal, vol. 28, Issue 3, 2024, pp. 181-184. cited by applicant

Roarr B N, et al., “OMNI: Open Mind Neuromodulation Interface for accelerated research and discovery.” In 2021 10th International IEEE/EMBS Conference on Neural Engineering (NER), May 4, 2021, IEEE., pp. 985-988. cited by applicant

Sakakibara, R., et al., “Lower urinary tract dysfunction in uncommon neurological diseases, Part III: A report of the Neuro-urology Promotion Committee of the International Continence Society.”, Continence, vol. 11, 2024. cited by applicant

Sermon et al., “1: 2 entrainment is not a device-induced artefact, except when it is.” Brain stimulation., Feb. 6, 2024, vol. 17, Issue 2, pp. 149. cited by applicant

Seth, J. et al., “Single Centre Randomised Pilot Study Of Two Regimens (30 Mins Daily Or 30 Mins Weekly For 12 Weeks) Of Transcutaneous Tibial Nerve Stimulation Using A Novel Device For Treating Multiple Sclerosis-Related Overactive Bladder Symptoms”, Journal of Neurology, Neurosurgery and Psychiatry, vol. 86, No. 11, 2015, 1 page. cited by applicant

Sinha, S., et al., “Lower urinary tract dysfunction in uncommon neurological diseases, Part II: A report of the Neuro-urology Promotion Committee of the International Continence Society.”, Continence, vol. 7, 2023. cited by applicant

Sladky V, et al., “Integrated human-machine interface for closed-loop stimulation using implanted and wearable devices.” In 2022 IEEE International Conference on Systems, Man, and Cybernetics (SMC), Oct. 9, 2022, IEEE., pp. 1730-1735. cited by applicant

Sorkhabi M M et al., “A neurostimulator system for real, sham, and multi-target transcranial magnetic stimulation.” Journal of neural engineering., Apr. 12, 2022, vol. 19, Issue 2, pp. 026035. cited by applicant

Sorkhabi M M, et al., “Design analysis and circuit topology optimization for programmable magnetic neurostimulator.” In 2021 43rd Annual International Conference of the IEEE Engineering in Medicine & Biology Society (EMBC), Nov. 1, 2021, IEEE., pp. 6384-6389. cited by applicant

Sorkhabi M, et al., “Estimation of the Motor Threshold for Near-Rectangular Stimuli Using the Hodgkin-Huxley Model.” Computational intelligence and neuroscience., 2021, vol. 2021, Issue 1, pp. 4716161. cited by applicant

Sorkhabi MM, et al., “Numerical modeling of plasticity induced by quadri-pulse stimulation.” IEEE Access., Feb. 8, 2021, vol. 9, pp. 26484-26490. cited by applicant

Tilborghs. S. et al., "Sacral neuromodulation for the treatment of overactive bladder: systematic review and future prospects," Expert review of medical devices, vol. 19, Issue 2, 2022, pp. 161-187. cited by applicant

Toth. R et al., "DyNeuMo Mk-2: an investigational circadian-locked neuromodulator with responsive stimulation for applied chronobiology," 2020 IEEE International Conference on Systems, Man, and Cybernetics (SMC). IEEE, 2020, pp. 3433-3440. cited by applicant

Tu. L.M. et al., "Initial clinical experience with selective bladder denervation for refractory overactive bladder," Neurourology and Urodynamics, vol. 38, Issue 2, 2019, pp. 644-652. cited by applicant

Vaganee. D. et al., "Needle placement and position of electrical stimulation inside sacral foramen determines pelvic floor electromyographic response-implications for sacral neuromodulation," Neuromodulation: Technology at the Neural Interface, vol. 22, Issue 6, 2019, pp. 709-715. cited by applicant

Vaganee. D. et al., "Repeatability of tools to assist in the follow up and troubleshooting of sacral neuromodulation patients using the sensory response," Neurourology and Urodynamics, vol. 38, Issue 2, 2019, pp. 801-808. cited by applicant

Van Meel et al., "Repeated ice water tests and electrical perception threshold determination to detect a neurologic cause of detrusor overactivity," Urology, vol. 70, Issue 4, 2007, pp. 772-776. cited by applicant

Voorham. J.C. et al., "The effect of EMG biofeedback assisted pelvic floor muscle therapy on symptoms of the overactive bladder syndrome in women: a randomized controlled trial," Neurourology and urodynamics, vol. 36, Issue 7, 2017, pp. 1796-1803. cited by applicant

Wendt K, et al., "Physiologically informed neuromodulation." Journal of the Neurological Sciences., Mar. 15, 2022, vol. 434, pp. 120121. cited by applicant

Wiest C, et al., "Evoked resonant neural activity in subthalamic local field potentials reflects basal ganglia network dynamics." Neurobiology of disease. Mar. 1, 2023, vol. 178, pp. 106019. cited by applicant

Zamora M, et al., "DyNeuMo Mk-1: Design and pilot validation of an investigational motion-adaptive neurostimulator with integrated chronotherapy." Experimental neurology. May 1, 2022, vol. 351, pp. 113977. cited by applicant

Primary Examiner: Mahmood; Nadia A

Attorney, Agent or Firm: Knobbe, Martens, Olson & Bear, LLP

Background/Summary

CROSS-REFERENCE (1) This application is a continuation of U.S. application Ser. No. 18/213,225, filed Jun. 22, 2023, which is a continuation of U.S. application Ser. No. 18/068,020 filed Dec. 19, 2022, which is a continuation of U.S. application Ser. No. 17/692,611, filed Mar. 11, 2022, now U.S. Pat. No. 11,565,109, issued Jan. 31, 2023, which claims the benefit of U.S. Provisional Application No. 63/160,322, filed Mar. 12, 2021, which are incorporated herein by reference in their entirety.

BACKGROUND

(1) Incontinence, or a lack of control over micturition or bowel movements, has many causes but commonly involves injury or weakness of the pelvic floor muscles and the nerves that innervate these muscles and involved organs. Electrical stimulation has been used to treat incontinence by stimulating the muscle directly or the sacral nerve to improve control over micturition and defecation.

SUMMARY

(2) A lack of voluntary control over micturition, defecation, incontinence, or any combination thereof is a problem that can impact quality of life and cause social embarrassment. Urinary incontinence, or loss of bladder control, fecal incontinence, loss of control of bowel movements, or any combination thereof often relate to neurological issues.

(3) Electrical stimulation has been used to treat incontinence by stimulating the muscle directly, sacral nerve, or other pelvic nerves to improve control over micturition and bowel movements. However, electrical nerve stimulation approaches usually only deliver a set stimulation protocol and are not capable of adapting to the condition of the subject. This can result in overstimulating or under

stimulating the target area, resulting in inadequate control over micturition or bowel defecation. Therefore, it would be beneficial to deliver an electrical stimulation that accounts for the condition of the subject and adapts the stimulation level accordingly.

(4) Additionally, traditional approaches to treating incontinence overlook that an innate human response (i.e. reflex) to prevent an incontinent event may be insufficient rather than absent. For some individuals who exhibit incontinence events, it may be that an innate mechanism for preventing incontinence is not entirely absent but rather is insufficient. For example, some individuals who suffer from incontinence may experience insufficient preventative response, such as a muscle contraction of at least one pelvic floor muscle to prevent a leakage event in response to an increased intra-abdominal pressure. In some instances, the leakage event may be urine, or alternatively gas or stool. Additionally, some individuals who suffer from stress incontinence may exhibit a delayed response to preventing an incontinence event in response to a stress event. Alternatively, some individuals may suffer from stress incontinence related to urethral hypermobility (i.e., insufficient support) that may lead to an increased pressure transmitted to the bladder and subsequently an incontinence event.

(5) Aspects of the disclosure herein provide a method for preventing an episode of incontinence in an individual in need thereof, the method comprising: (a) implanting a sensor and an stimulator electrode within a body of the individual; (b) sensing, with the sensor electrode, a parameter that is associated with a response from the individual that is intended to prevent an episode of incontinence; and (c) providing an electrical stimulation with the stimulation electrode, that together with the response, prevents the episode of incontinence. In some embodiments, the episode of incontinence comprises urinary incontinence. In some embodiments, the episode of incontinence comprises fecal incontinence. In some embodiments, the sensor is configured to sense a contraction of a muscle of the individual that results in a partial contraction of a sphincter or a pelvic floor muscle that controls bladder or bowel voiding. In some embodiments, the sensor electrode is positioned within the pelvis of the individual. In some embodiments, the stimulator electrode provides an electrical stimulation to the pudendal nerve. In some embodiments, the sensor and the stimulator electrode are located on a single lead. In some embodiments, the method comprises a step of providing a constant electrical stimulation at a lower intensity level than the electrical stimulation provided in step (c). In some embodiments, the episode of incontinence is urinary incontinence and is urge incontinence type. In some embodiments, the intensity or duration of the electrical stimulation provided in step (c) varies according to the response that is sense in step (b). In some embodiments, the response that is sensed in step (b) is insufficient on its own to prevent the episode of incontinence and the electrical stimulation provided in step (b) adds just enough, together with the response, to prevent the episode of incontinence. In some embodiments, the response that is sense in step (b) is insufficient on its own to prevent the episode of incontinence and the electric stimulation provided in step (b), together with the response, prevent the episode on incontinence. In some embodiments, the method comprises a step of implanting a second stimulator electrode, wherein a first stimulator electrode stimulates one pudendal nerve, and the second stimulator electrode stimulates the a different pudendal nerve. In some embodiments, the stimulator electrode stimulates one pudendal nerve, and the different stimulator electrode stimulates another spatially independent region of the same pudendal nerve. In some embodiments, the method comprises a step of implanting a different stimulator electrode, wherein the stimulator electrode stimulates the main trunk of the pudendal nerve and the different stimulator electrode stimulates the distal nerve of the pudendal nerve. In some embodiments, the distal nerve of the pudendal nerve comprises branches thereof the distal pudendal nerve. In some embodiments, the method comprises a step of implanting a different stimulator electrode, wherein the electrode stimulates the trunk of the pudendal nerve the different stimulator electrode stimulates a main branch of the pudendal nerve e.g., dorsal genital nerve. In some embodiments, the method comprises a step of implanting a different stimulator electrode, wherein the first stimulator electrode stimulates a pudendal nerve, and the different stimulator electrode stimulates a sacral spinal nerve. In some embodiments, the individual suffers from urinary incontinence of varying classification. In some cases, the various classification may comprise stress incontinence, urge incontinence, overflow incontinence, or any combination thereof (i.e., mixed incontinence). In some embodiments, the sensor and stimulator electrodes are operatively coupled to a processor and a non-transitory computer readable medium that includes software. In some embodiments, the sensor electrode is calibrated by the individual using an

external input device that interfaces with the software. In some embodiments, the software is configured to record a signal from the sensor electrode. In some embodiments, the software is configured to adjust the sensor electrode in response to signal.

(6) In some embodiments, the sensor electrode is configured to sense an EMG signal. In some embodiments, the EMG signal determines that a contraction of at least one pelvic muscle has occurred. In some embodiments, the strength of the EMG signal determines that a contraction of at least one pelvic muscle has occurred. In some embodiments, the strength of the EMG signal is proportional to the strength of the contraction of at least one pelvic muscle.

(7) Aspects of the disclosure herein provide a system for preventing an episode of incontinence in an individual in need thereof, the apparatus comprising: (a) a sensor electrode configured to sense a parameter that is associated with a response from the individual that is intended to prevent the episode of incontinence; (b) an stimulator electrode configured to provide electrical stimulation; (c) a processor operably coupled to the sensor and stimulator electrodes; (d) a non-transitory computer readable storage medium including software configured to cause the processor to: (i) receive the parameter that is associated with the response from the individual that is intended to prevent the episode of incontinence; (ii) analyze the parameter that is associated with the response from the individual that is intended to prevent the episode of incontinence; and (iii) cause the stimulator electrode to provide the electrical stimulation to the individual so that the electrical stimulation together with the response from the individual that is intended to prevent the episode of incontinence prevents the episode of incontinence. In some embodiments, the software comprises analyzing a global positioning system (GPS) location of the individual that in combination with the parameter associated with the response from the individual is intended to prevent the episode of incontinence. In some embodiments, the episode of incontinence comprises urinary incontinence. In some embodiments, the episode of incontinence comprises fecal incontinence. In some embodiments, the episode of incontinence comprises a combination of urinary incontinence and fecal incontinence. In some embodiments, the episode of incontinence comprises urinary stress incontinence. In some embodiments, the sensor electrode is configured to sense a contraction of a muscle of the individual that results in a partial contraction of a sphincter that controls bladder or bowel voiding. In some embodiments, the sensor electrode is positioned within the pelvis of the individual. In some embodiments, the stimulator electrode provides the electrical stimulation to the pudendal nerve. In some embodiments, the sensor and stimulator electrodes are located on a single lead. In some embodiments, the stimulator electrode is configured to provide a constant electrical stimulation at a lower intensity level than the electrical stimulation provided to prevent an episode of incontinence. In some embodiments, the episode of incontinence is urinary incontinence and is urge incontinence type. In some embodiments, an intensity or duration of the electrical stimulation varies according to the response that is sensed. In some embodiments, the response that is sensed by the sensor electrode is insufficient on its own to prevent the episode of incontinence and the electrical stimulation provided adds together with the response, to prevent the episode of incontinence. In some embodiments, the system provides a lower intensity stimulation to the subject, whereby the lower intensity stimulation increased or improves the subject's tolerance to the stimulation signal. In some embodiments, the increase or improvement in the subject's tolerance provides tolerance of higher intensity stimulation signal. In some embodiments, the sensor electrode is configured to sense an EMG signal. In some embodiments, the EMG signal determines that a contraction of at least one pelvic muscle has occurred. In some embodiments, the strength of the EMG signal is proportional to the strength of the contraction of at least one pelvic muscle. In some embodiments, the stimulator electrode comprises a first stimulator electrode and a second stimulator electrode, wherein the first stimulator electrode stimulates a first pudendal nerve, and the second stimulator electrode stimulates a second pudendal nerve. In some embodiments, the individual suffers from urinary incontinence of a mixed type. In some embodiments, the sensor electrode is calibrated by the individual using an external input device that interfaces with the software. In some embodiments, the software is configured to further cause the processor to record a signal from the sensor electrode. In some embodiments, the software is configured to adjust the sensor electrode in response to the signal.

(8) Aspects of the disclosure herein provide a non-transitory computer readable storage medium including software for preventing an episode of incontinence in an individual in need thereof, configured

to cause a processor to: (i) receive a parameter by a sensor electrode that is associated with a response from the individual intended to prevent the episode of incontinence; (ii) analyze the parameter that is associated with the response from the individual that is intended to prevent the episode of incontinence; and (iii) cause an stimulator electrode to provide an electrical stimulation to the individual so that the electrical stimulation together with the response from the individual that is intended to prevent the episode of incontinence prevents the episode of incontinence. In some embodiments, the software comprises analyzing a global positioning system (GPS) location of the individual that in combination with the parameter associated with the response from the individual is intended to prevent the episode of incontinence. In some embodiments, the episode of incontinence comprises urinary incontinence. In some embodiments, the episode of incontinence comprises fecal incontinence. In some embodiments, the episode of incontinence comprises urinary stress incontinence. In some embodiments, the sensor electrode is configured to sense a contraction of a muscle of the individual that results in a partial contraction of a sphincter that controls bladder or bowel voiding. In some embodiments, the sensor electrodes, stimulation electrodes, or any combination thereof is positioned within, adjacent, or around the pelvis of the individual. In some embodiments, the stimulator electrode provides the electrical stimulation to the pudendal nerve. In some embodiments, the sensor and stimulator electrodes are located on a single lead. In some embodiments, the stimulator electrode is configured to provide a constant electrical stimulation at a lower intensity level than the electrical stimulation provided to prevent an episode of incontinence. In some embodiments, the episode of incontinence is urinary incontinence and is urge incontinence type. In some embodiments, an intensity or duration of the electrical stimulation varies according to the response that is sensed. In some embodiments, the response that is sensed by the sensor electrode may be insufficient on its own to prevent the episode of incontinence and the electrical stimulation provided adds just enough, together with the response, to prevent the episode of incontinence. In some embodiments, the response that is sensed by the sensor electrode may be insufficient on its own to prevent the episode of incontinence and the electrical stimulation provided, together with the response, prevents the episode of incontinence. In some embodiments, the sensor electrode may be configured to sense an EMG signal. In some embodiments, the EMG signal may determine that a contraction of at least one pelvic muscle has occurred. In some embodiments, the strength of the EMG signal may be proportional to the strength of the contraction of at least one pelvic muscle. In some embodiments, the stimulator electrode may comprise a first stimulator electrode and a second stimulator electrode, wherein the first stimulator electrode may stimulate a first pudendal nerve and the second stimulator electrode may stimulate another portion of the pudendal nerve. In some embodiments, the first stimulator electrode may stimulate one pudendal nerve, and the second stimulator electrode may stimulate another spatially independent region of the same pudendal nerve. In some embodiments, the first stimulator electrode may stimulate the main trunk of the pudendal nerve and the second stimulator electrode may stimulate the distal nerve of the pudendal nerve. In some embodiments, the distal nerve of the pudendal nerve may comprise branches thereof of the distal pudendal nerve. In some embodiments, the first stimulator electrode may stimulate the trunk of the pudendal nerve and the second stimulator electrode may stimulate a main branch of the pudendal nerve e.g., dorsal genital nerve. In some embodiments, the individual suffers from urinary incontinence of a mixed type. In some embodiments, the sensor electrode is calibrated by the individual using an external input device that interfaces with the software. In some embodiments, the software is configured to further cause the processor to record a signal from the sensor electrode. In some embodiments, the software is configured to adjust the sensor electrode in response to the signal.

(9) Aspects of the disclosure herein provide a device for preventing an episode of incontinence in an individual in need thereof, the device comprising: (a) a sensor electrode configured to sense a parameter that is associated with a response from the individual that is intended to prevent the episode of incontinence; a stimulator electrode configured to provide electrical stimulation; and a processor operatively coupled to the sensor and stimulator electrode. In some embodiments, the episode of incontinence comprises urinary incontinence. In some embodiments, the episode of incontinence comprises fecal incontinence. In some embodiments, the episode of incontinence comprises urinary stress incontinence. In some embodiments, the sensor electrode is configured to sense a contraction of a muscle of the individual that results in a partial contraction of a sphincter that controls bladder or bowel voiding. In some embodiments, the sensor electrode is positioned within the pelvis of the individual. In

some embodiments, the stimulator electrode provides the electrical stimulation to the pudendal nerve. In some embodiments, the stimulator electrode is configured to provide a constant electrical stimulation at a lower intensity level than the electrical stimulation provided to prevent an episode of incontinence. In some embodiments, the episode of incontinence is urinary incontinence and is urge incontinence type. In some embodiments, the intensity or duration of the electrical stimulation varies according to the response that is sensed. In some embodiments, the response that is sensed is insufficient on its own to prevent the episode of incontinence and the electrical stimulation provided adds just enough, together with the response, to prevent the episode of incontinence. In some embodiments, the response that is sensed is insufficient on its own to prevent the episode of incontinence and the electrical stimulation provided, together with the response, prevents the episode of incontinence. In some embodiments, the sensor electrode is configured to sense an EMG signal, ENG signal, or any combination thereof. In some embodiments, the EMG signal determines that a contraction of at least one pelvic muscle has occurred. In some embodiments, the strength of the EMG signal is proportional to the strength of the contraction of at least one pelvic muscle. In some embodiments, the stimulator electrode comprises a first stimulator electrode and a second stimulator electrode, wherein the first stimulator electrode stimulates a first pudendal nerve and the second stimulator electrode stimulates a second pudendal nerve. In some embodiments, the individual suffers from urinary incontinence of a mixed type. In some embodiments, the device further comprises a non-transitory computer readable medium that includes software. In some embodiments, the sensor electrode is calibrated by the individual using an external input device that interfaces with the software. In some embodiments, the software is configured to record a signal from the sensor electrode. In some embodiments, the software is configured to adjust the sensor electrode in response to the signal.

(10) Aspects of the disclosure herein provide a method of data processing, said method comprising: (i) receiving a measurement of a parameter previously measured by a sensor electrode, which parameter is predictive of an episode of incontinence in an individual; (ii) analyzing the parameter; and (iii) synthesizing an electrical stimulation signal for the individual, so that when the electrical stimulation signal is provided by a stimulator electrode to the individual, the electrical stimulation signal, together with an effort from the individual that is intended to prevent an episode of incontinence, prevents the episode of incontinence. In some embodiments, the parameter is associated with a response from the individual intended to prevent an episode of incontinence. In some embodiments, the parameter is associated with the individual's effort in trying to prevent an episode of incontinence, and wherein the electrical stimulation signal is synthesized so as to supplement the individual's effort with an electrical stimulation pattern that will, together with the effort from the individual, be sufficient to prevent an episode of incontinence. In some embodiments, the response from the individual is insufficient on its own to prevent the episode of incontinence and the electrical stimulation signal is such that, when applied, it adds enough, together with the response, to prevent the episode of incontinence. In some embodiments, the episode of incontinence comprises urinary incontinence. In some embodiments, the episode of incontinence comprises fecal incontinence. In some embodiments, the episode of incontinence comprises urinary stress incontinence. In some embodiments, the parameter is a signal from a sensor electrode that is configured to sense a contraction of a muscle of the individual related to a partial contraction of a sphincter that controls bladder or bowel voiding. In some embodiments, the electrical stimulation signal is for the electrical stimulation of the pudendal nerve. In some embodiments, the electrical stimulation signal is synthesized to include a constant electrical stimulation component and a measurement parameter specific component. In some embodiments, the episode of incontinence is urinary incontinence and is urge incontinence type. In some embodiments, an intensity or duration of the electrical stimulation that will be provided by the electrical stimulation signal varies according to the value of the parameter that is received. In some embodiments, the parameter is an EMG signal. In some embodiments, the EMG signal determines that a contraction of at least one pelvic muscle has occurred. In some embodiments, a strength of the EMG signal is proportional to the strength of the contraction of at least one pelvic muscle. In some embodiments, the electrical stimulation signal comprises first and second signals for the stimulation of a first pudendal nerve and a second pudendal nerve respectively. In some embodiments, the method further comprises recording the signal previously measured by the

sensor electrode. In some embodiments, the method further comprises synthesizing an adjustment signal to adjust the sensor electrode in response to the recorded signal.

Description

BRIEF DESCRIPTION OF THE DRAWINGS

(1) The novel features of the invention are set forth with particularity in the appended claims. A better understanding of the features and advantages of the present invention will be obtained by reference to the following detailed description that sets forth illustrative embodiments, in which the principles of the invention are utilized, and the accompanying drawings of which:

(2) FIG. 1 shows an exemplary embodiment of an open-loop bioelectronic system comprising an implantable pulse generator. The system delivers a predefined stimulation protocol and does not receive input from the subject.

(3) FIG. 2 shows an exemplary embodiment of a closed-loop bioelectronic system comprising a two-way neural interface comprising a sensor electrode that captures the neural response of the subject and a processing module that can interpret the neural response. The system is capable of delivering an adapted stimulation based on the neural response.

(4) FIG. 3 shows an exemplary embodiment of the devices and methods described herein targeting the pudendal nerve by placing the sensor and stimulator electrodes near the pudendal nerve. The sensor electrode captures the bio-signal to classify any stress events, and the stimulator electrodes deliver the stress electrical stimulation (“bio-simulator”), which is adapted from the base stimulation to account for the stress event, to act on the target sphincter muscle.

(5) FIG. 4 shows exemplary embodiments of depicting a bio-signal and a bio-stimulator electrical stimulation. The bio-signal captured by the sensor electrode is analyzed and classified to identify any stress events, such as a cough or a sudden movement, from an electromyography (EMG) signal. The stimulator electrodes deliver the bio-simulator, an electrical stimulation that is adapted from the base stimulation to account for the stress event.

(6) FIG. 5 shows an exemplary embodiment of a system block diagram for the devices and methods described herein configured to implement slow- and fast-adapting algorithms. API is an application programming interface; MICS is the Medical Information and Communication band.

(7) FIGS. 6A-6B show a flowchart of a method of closed-loop operation of the device of the disclosure.

(8) FIGS. 7A-7B show an example embodiment of the patient controller module, as described in some embodiments herein.

(9) FIG. 8 shows a flow diagram for purposeful patient contraction and manual operation of the devices and systems, as described in some embodiments herein.

(10) FIG. 9 shows a flow diagram for the closed loop system operation, as described in some embodiments herein.

(11) FIG. 10 shows a flow diagram for signal processing and threshold detection of incontinent events, as described in some embodiments herein.

(12) FIGS. 11A-11E show patient purposeful muscle contraction, Valsalva maneuver, and coughing EMG data acquired and processed with the methods of the disclosure, as described in some embodiments herein.

(13) FIGS. 12A-12B show flow diagrams for detecting a purposeful or intent based contraction by the patient and providing electrical stimulation to prevent an incontinent event, as described in some embodiments herein.

(14) FIG. 13 shows a flow diagram for training on-board machine learning classifier of the devices and systems of the disclosure, as described in some embodiments herein.

DETAILED DESCRIPTION

(15) A lack of voluntary control over micturition or defecation, known as urinary and fecal incontinence, respectively, is a problem that can impact quality of life and cause social embarrassment. Urinary and fecal incontinence may affect individuals of all ages. In some cases, older individuals may exhibit a greater probability of incontinence with varied pathophysiology. Both urinary incontinence and fecal

incontinence may involve injury, weakness, or overactivity of the pelvic floor muscles, including but not limited to the urethral and anal sphincter, and the nerves that innervate these muscles and involved organs, such as the bladder, rectum, or anus.

(16) Urinary incontinence may be categorized into one of four main types: urge incontinence, stress incontinence, overflow incontinence, and mixed incontinence. Urge incontinence is often due to an overactive bladder. Individuals with urge incontinence have a strong and sudden need to urinate immediately, often leaving them with insufficient time to reach a bathroom. Stress incontinence is usually due to a poorly functioning urethral sphincter muscle or hypermobility of the urethra or bladder neck. An individual may experience stress incontinence during activities such as coughing, sneezing, laughing, or exercise. Overflow incontinence may typically be due to poor bladder contraction or blockage of the urethra. Mixed incontinence may involve features of stress and urge incontinence. Incontinence often involves neurological issues, including but not limited to impaired nerve conduction between the brain and/or the affected muscles, and nervous system conditions or injuries (e.g., multiple sclerosis or stroke), or mental confusion. Other causes of incontinence include but are not limited to weakness of pelvic or urethral muscles and pelvic prolapse.

(17) Fecal incontinence, also referred to as bowel incontinence, is the loss of bowel control, causing an individual to pass stool unexpectedly from the rectum. Fecal incontinence may be categorized into two main types: urge incontinence, and passive incontinence, or a combination thereof. Urge incontinence is due to an overactive bowel. Individuals with urge incontinence have a strong and sudden need to defecate immediately, often leaving them with insufficient time to reach a bathroom. Passive incontinence is when an individual feels no urge to open their bowel although the rectum is full and ready to be voided. Individuals suffering from passive incontinence cannot consciously control their bowel movements and stool can pass without their knowledge.

(18) Incontinence often involves neurological issues, including but not limited to impaired nerve conduction between the brain and/or the affected muscles, and nervous system conditions or injuries (e.g., multiple sclerosis or stroke), or mental confusion. Causes of fecal incontinence include but are not limited to nerve damage, anal sphincter muscle damage, constipation, diarrhea, hemorrhoids, surgery, loss of rectum storage capacity, rectal prolapse, and rectocele.

(19) Electro galvanic stimulation of muscles has been used to treat incontinence by training the pelvic floor muscles thereby improving strength and function and/or by stimulating the sacral nerve to improve control over urination and defecation. However, these electrical nerve stimulation approaches are not able to adapt to the condition and circumstances of the subject and are only capable of delivering a predefined stimulation protocol. This can result in overstimulating or under stimulating the target area, resulting in inadequate control over muscles involved in urination or bowel movements. As such, it would be beneficial to be able to deliver an electrical stimulation that accounts for the condition of the subject and adapts the stimulation level accordingly.

(20) Provided herein are devices, systems, and methods for preventing, and/or reducing the severity of an episode of incontinence in an individual in need thereof. In some instances, the episode of incontinence may comprise urinary incontinence, fecal incontinence, or any combination thereof. The devices, systems and methods disclosed herein may treat a sub-type of incontinence. In some cases, the sub-type of incontinence may comprise urge incontinence, stress incontinence, overflow incontinence, or mixed incontinence.

(21) In some instances, the devices, systems, and methods described herein may increase bladder capacity. Bladder capacity may be measured as the maximum liquid volume a bladder may contain without inducing pain or micturition. In some cases, bladder capacity of a healthy individual as a baseline value may comprise a liquid volume of about 400 milliliters (mL) to about 600 mL. In some cases, bladder capacity may be measured by ultrasound calculations of a still ultrasound image of a subject's bladder. The bladder's volume may be measured by multiplying the length, width, and height of the measured bladder on the still ultrasound image. Alternatively, or in addition to, bladder capacity may be measured by cytometry where a catheter is inserted into the subject's urethra to fill and may measure the volume of urine. In some cases, a subject's bladder capacity may be reduced by disease. In some cases, the devices and systems described herein may increase bladder capacity by a percent increase from baseline bladder capacity. In some instances, bladder capacity may be increased by providing to the

patient, an electrical stimulation by the stimulator electrode leads, described herein. In some cases, the stimulator may provide a frequency or pulse width of electrical stimulation, described elsewhere herein. (22) In some cases, bladder capacity may be increased utilizing either open or closed-loop system, described elsewhere herein. In some instances, bladder capacity may be increased by about 5% to about 75%. In some instances, bladder capacity may be increased by about 5% to about 10%, about 5% to about 15%, about 5% to about 20%, about 5% to about 25%, about 5% to about 30%, about 5% to about 35%, about 5% to about 40%, about 5% to about 45%, about 5% to about 50%, about 5% to about 70%, about 5% to about 75%, about 10% to about 15%, about 10% to about 20%, about 10% to about 25%, about 10% to about 30%, about 10% to about 35%, about 10% to about 40%, about 10% to about 45%, about 10% to about 50%, about 10% to about 70%, about 10% to about 75%, about 15% to about 20%, about 15% to about 25%, about 15% to about 30%, about 15% to about 35%, about 15% to about 40%, about 15% to about 45%, about 15% to about 50%, about 15% to about 70%, about 15% to about 75%, about 20% to about 25%, about 20% to about 30%, about 20% to about 35%, about 20% to about 40%, about 20% to about 45%, about 20% to about 50%, about 20% to about 70%, about 20% to about 75%, about 25% to about 30%, about 25% to about 35%, about 25% to about 40%, about 25% to about 45%, about 25% to about 50%, about 25% to about 70%, about 25% to about 75%, about 30% to about 35%, about 30% to about 40%, about 30% to about 45%, about 30% to about 50%, about 30% to about 70%, about 30% to about 75%, about 35% to about 40%, about 35% to about 45%, about 35% to about 50%, about 35% to about 70%, about 35% to about 75%, about 40% to about 45%, about 40% to about 50%, about 40% to about 70%, about 40% to about 75%, about 45% to about 50%, about 45% to about 70%, about 45% to about 75%, about 50% to about 70%, about 50% to about 75%, or about 70% to about 75%. In some instances, bladder capacity may be increased by about 5%, about 10%, about 15%, about 20%, about 25%, about 30%, about 35%, about 40%, about 45%, about 50%, about 70%, or about 75%. In some instances, bladder capacity may be increased by at least about 5%, about 10%, about 15%, about 20%, about 25%, about 30%, about 35%, about 40%, about 45%, about 50%, or about 70%. In some instances, bladder capacity may be increased by at most about 10%, about 15%, about 20%, about 25%, about 30%, about 35%, about 40%, about 45%, about 50%, about 70%, or about 75%.

(23) In some cases, the devices, systems, and methods described herein may increase urethral closure pressures. In some cases the urethral closure pressure may be expressed by a profile of pressures. Urethral closure pressure, in some instances, may be described as measurement of integrated pressure curves from the entire length of a subject's urethra. In some instances, urethral closure pressures may be measured by the slow withdrawal of a pressure-measuring catheter through the subject's urethra at a constant rate. In some instances, urethral closure pressures may be measured by a trans-urethral pressure profile catheter. In some cases, transurethral pressure may be measured in units of centimeters of water (cm H.sub.2O). In some instances, the devices and systems described herein may increase urethral closure pressure by stimulating one or more nerves of the subject. In some instances, the one or more nerves may comprise the pudendal nerve, sacral nerve, or any combinations or branches thereof. In some cases, the stimulator described elsewhere herein, may provide an electrical stimulation at the subject's nerve, via a stimulator electrode lead at a frequency and/or frequency ranges, described elsewhere herein, to increase urethral closure pressure. In some instances, the urethral closure pressure may comprise about 10 centimeters of water (cm H.sub.2O) to about 75 cm H.sub.2O. In some instances, the urethral closure pressure may comprise about 10 cm H.sub.2O to about 15 cm H.sub.2O, about 10 cm H.sub.2O to about 20 cm H.sub.2O, about 10 cm H.sub.2O to about 25 cm H.sub.2O, about 10 cm H.sub.2O to about 30 cm H.sub.2O, about 10 cm H.sub.2O to about 35 cm H.sub.2O, about 10 cm H.sub.2O to about 40 cm H.sub.2O, about 10 cm H.sub.2O to about 45 cm H.sub.2O, about 10 cm H.sub.2O to about 50 cm H.sub.2O, about 10 cm H.sub.2O to about 60 cm H.sub.2O, about 10 cm H.sub.2O to about 70 cm H.sub.2O, about 10 cm H.sub.2O to about 75 cm H.sub.2O, about 15 cm H.sub.2O to about 20 cm H.sub.2O, about 15 cm H.sub.2O to about 25 cm H.sub.2O, about 15 cm H.sub.2O to about 30 cm H.sub.2O, about 15 cm H.sub.2O to about 35 cm H.sub.2O, about 15 cm H.sub.2O to about 40 cm H.sub.2O, about 15 cm H.sub.2O to about 45 cm H.sub.2O, about 15 cm H.sub.2O to about 50 cm H.sub.2O, about 15 cm H.sub.2O to about 60 cm H.sub.2O, about 15 cm H.sub.2O to about 70 cm H.sub.2O, about 15 cm H.sub.2O to about 75 cm H.sub.2O, about 20 cm H.sub.2O to about 25 cm H.sub.2O, about 20 cm H.sub.2O to about 30 cm H.sub.2O, about 20 cm H.sub.2O to about 35 cm

H.sub.2O, about 20 cm H.sub.2O to about 40 cm H.sub.2O, about 20 cm H.sub.2O to about 45 cm H.sub.2O, about 20 cm H.sub.2O to about 50 cm H.sub.2O, about 20 cm H.sub.2O to about 60 cm H.sub.2O, about 20 cm H.sub.2O to about 70 cm H.sub.2O, about 20 cm H.sub.2O to about 75 cm H.sub.2O, about 25 cm H.sub.2O to about 30 cm H.sub.2O, about 25 cm H.sub.2O to about 35 cm H.sub.2O, about 25 cm H.sub.2O to about 40 cm H.sub.2O, about 25 cm H.sub.2O to about 45 cm H.sub.2O, about 25 cm H.sub.2O to about 50 cm H.sub.2O, about 25 cm H.sub.2O to about 60 cm H.sub.2O, about 25 cm H.sub.2O to about 70 cm H.sub.2O, about 25 cm H.sub.2O to about 75 cm H.sub.2O, about 30 cm H.sub.2O to about 35 cm H.sub.2O, about 30 cm H.sub.2O to about 40 cm H.sub.2O, about 30 cm H.sub.2O to about 45 cm H.sub.2O, about 30 cm H.sub.2O to about 50 cm H.sub.2O, about 30 cm H.sub.2O to about 60 cm H.sub.2O, about 30 cm H.sub.2O to about 70 cm H.sub.2O, about 30 cm H.sub.2O to about 75 cm H.sub.2O, about 35 cm H.sub.2O to about 40 cm H.sub.2O, about 35 cm H.sub.2O to about 45 cm H.sub.2O, about 35 cm H.sub.2O to about 50 cm H.sub.2O, about 35 cm H.sub.2O to about 60 cm H.sub.2O, about 35 cm H.sub.2O to about 70 cm H.sub.2O, about 35 cm H.sub.2O to about 75 cm H.sub.2O, about 40 cm H.sub.2O to about 45 cm H.sub.2O, about 40 cm H.sub.2O to about 50 cm H.sub.2O, about 40 cm H.sub.2O to about 60 cm H.sub.2O, about 40 cm H.sub.2O to about 70 cm H.sub.2O, about 40 cm H.sub.2O to about 75 cm H.sub.2O, about 45 cm H.sub.2O to about 50 cm H.sub.2O, about 45 cm H.sub.2O to about 60 cm H.sub.2O, about 45 cm H.sub.2O to about 70 cm H.sub.2O, about 45 cm H.sub.2O to about 75 cm H.sub.2O, about 50 cm H.sub.2O to about 60 cm H.sub.2O, about 50 cm H.sub.2O to about 70 cm H.sub.2O, about 50 cm H.sub.2O to about 75 cm H.sub.2O, about 60 cm H.sub.2O to about 70 cm H.sub.2O, about 60 cm H.sub.2O to about 75 cm H.sub.2O, or about 70 cm H.sub.2O to about 75 cm H.sub.2O. In some instances, the urethral closure pressure may comprise about 10 cm H.sub.2O, about 15 cm H.sub.2O, about 20 cm H.sub.2O, about 25 cm H.sub.2O, about 30 cm H.sub.2O, about 35 cm H.sub.2O, about 40 cm H.sub.2O, about 45 cm H.sub.2O, about 50 cm H.sub.2O, about 60 cm H.sub.2O, about 70 cm H.sub.2O, or about 75 cm H.sub.2O. In some instances, the urethral closure pressure may comprise at least about 10 cm H.sub.2O, about 15 cm H.sub.2O, about 20 cm H.sub.2O, about 25 cm H.sub.2O, about 30 cm H.sub.2O, about 35 cm H.sub.2O, about 40 cm H.sub.2O, about 45 cm H.sub.2O, about 50 cm H.sub.2O, about 60 cm H.sub.2O, or about 70 cm H.sub.2O. In some instances, the urethral closure pressure may comprise at most about 15 cm H.sub.2O, about 20 cm H.sub.2O, about 25 cm H.sub.2O, about 30 cm H.sub.2O, about 35 cm H.sub.2O, about 40 cm H.sub.2O, about 45 cm H.sub.2O, about 50 cm H.sub.2O, about 60 cm H.sub.2O, about 70 cm H.sub.2O, or about 75 cm H.sub.2O.

(24) The devices and systems described herein may comprise one or more sensor electrodes, one or more stimulator electrodes, a processor, a power source, or any combination thereof. In some instances, the one or more stimulator electrodes, one or more sensor electrodes, processor, power source, or any combination thereof may be implanted into the body of the individual. In some instances, the one or more stimulator electrodes, one or more sensor electrodes, processor, power source, or any combination thereof may be placed superficially on the body of the individual. In some instances, the processor, power source, or any combination thereof may be implanted into the body of the individual. In some instances, the device may comprise a wireless transmission module capable of transmitting and receiving wireless data wirelessly by and between a remote device (e.g., a mobile phone, a tablet, a computer, etc.) and the incontinence prevention device described herein. In some cases, the device may be implantable into an individual. Alternatively, the device may comprise a hermetically sealed connector placed on the individual's skin superficially that may electrically couple to a remote device by a cable. In some instances, the individual may manually modify or change the one or more sensor electrode or one or more stimulator electrode parameters through a graphical user interface. In some instances, the device may automatically modify or change the one or more sensor electrode or one or more stimulator electrode parameters.

(25) In some cases, an implanted sensor electrode of the device may be configured to sense one or more parameters that may be associated with a response of an episode of incontinence or an attempt of the individual in trying to prevent an episode of incontinence. In some cases, the parameter may comprise an electromyography (EMG) signal that may be predictive of an episode of incontinence. In some instances, EMG signals may comprise electrical activity of muscles, specifically from action potentials in muscle

fibers. In some cases, the parameter may comprise an electroneurogram (ENG). In some cases, an electroneurogram may comprise electrical activity from one or more neurons and typically refers to recordings made from bundles of axons in peripheral nerves. In some cases, the parameter may comprise a change of electric impedance caused by physical deformation of the sensor electrode material. In some instances, the physical deformation may comprise stretching, compression, or any combination thereof. Alternatively, the parameter may comprise a change in pressure, velocity, acceleration, or 3-D spatial direction. In some instances, 3-D spatial direction may be determined by GPS signal. The GPS signal, in some cases, may indicate and recognize when the individual is in proximity to locations such as the individual's home. In some instances, the GPS signal may be configured to modulate the stimulator, described elsewhere herein, based on the GPS coordinates and/or GPS location of the subject. In some instances, modulating the stimulator may comprise adjusting a detection parameter (e.g., signal intensity threshold) of the classifier, described elsewhere herein. In some cases, the GPS signal of a location and/or GPS coordinates may indicate locations and/or regions traveled to the subject that may impose a higher risk of an incontinent event. In some cases, 3-D spatial direction may measure the posture of an individual. In some instances, changes in pressure may be measured by a pressure sensor. In some cases, the pressure sensor may comprise a differential pressure sensor, absolute pressure sensor, or any combination thereof. In some cases, changes in velocity, acceleration, or changes in 3-D spatial direction may be measured by an accelerometer, gyroscope, magnetometer, or any combination thereof. In some instances, the device may provide an electrical stimulation using the one or more implanted stimulator electrodes that, together with the individual's preventative response, may prevent the episode of incontinence.

(26) The devices, systems, and methods described herein may be capable of preventing and/or reducing the severity of an incontinent event by detecting an individual's incontinence preventative parameter and adjusting an electrical stimulation by one or more stimulator electrodes based on the incontinence preventative parameter's characteristics. In some cases, the incontinence preventative parameter characteristics may comprise the changes in the amplitude, frequency, phase, or any combination thereof one or more EMG, ENG, accelerometer, gyroscope, magnetometer, pressure sensor signals, or any combination thereof. In some instances, an incontinence preventative parameter may comprise physical movement of the individual, one or more EMG signals, or any combination thereof. By providing adaptive electrical stimulation in combination to the individual's response, the devices, systems, and methods described herein may prevent incontinence episodes.

(27) Described herein are devices, systems, and methods for preventing and/or reducing the severity of an episode of incontinence in an individual by providing adaptive electrical stimulation based on the individual's incontinence preventative parameter. The device may be implanted in the individual in proximity to the pudendal nerve, sacral nerve, or branches thereof. In some cases, the one or more stimulator electrodes of the device may be placed at or near the pudendal nerve or the sacral nerve. In some cases, the one or more stimulator electrodes or one or more sensor electrodes may be implanted in proximity to the pudendal nerve unilaterally or bilaterally. In some cases, the one or more stimulator electrodes may be implanted to stimulate motor nerve fibers (e.g., to the pelvic floor and sphincters). Alternatively, the one or more stimulator electrodes may be implanted within, proximate, or adjacent to the muscle (e.g., pelvic floor and sphincter muscles). The sensor electrode of the implanted device may sense a signal that indicates that an individual may exhibit an episode of incontinence or is trying to prevent an episode of incontinence. The device may analyze the signal and classify the signal as a real-time or prospective episode of incontinence. The device may generate an electrical stimulation that is modulated based on the classified episode of incontinence and may deliver the modulated electrical stimulation using one or more stimulator electrodes to the targeted nerve. In some instances, the modulation of the one or more stimulator electrodes electrical stimulation may comprise changing the frequency of electric stimulation, amplitude of electrical stimulation, pulse width of electric stimulation, stimulator electrode configuration, or any combination thereof. In some cases, the one or more electrodes may comprise one or more stimulation electrodes and one or more sensing electrodes. The one or more stimulation electrodes and one or more sensing electrodes may be configured to switch between stimulating and sensing operations on demand, programmatically, user controlled, medical personal control, or any combination thereof. The electrical stimulation may be modulated to improve the muscle

and/or nerve response to prevent the episode of incontinence. The modulated electrical stimulation may result in improved muscle response, as measured by response time, muscle function, or other markers of incontinence prevention, to prevent the potential episode of incontinence. The electrical stimulation may be modulated to reduce a severity and/or duration of an episode of incontinence. For example, a subject who exhibits an urge incontinence event (e.g., a cough) may receive an electrical stimulation provided by the devices, systems, and methods provided herein, in response to the incontinence episode. In some cases, the electrical stimulation may reduce the volume of uncontrolled micturition and/or defecation. In some cases, the electrical stimulation may reduce the volume of uncontrolled micturition and/or defecation without preventing the episode of incontinence entirely. In some cases, the electrical stimulation may reduce the duration of the incontinent event that would occur without the electrical stimulation. In some cases, the electrical stimulation may be configured to inhibit reflex incontinence. In some cases, the electrical stimulation may inhibit a bladder detrusor muscle contraction experienced during reflex incontinence.

(28) The electrical stimulation may comprise an electrical stimulation provided over a period of time. In some cases, the period of time may comprise the period of time a subject may provide an intent based or purposeful muscle contraction to trigger an EMG threshold detection.

(29) In some cases, the period of time for the electrical stimulation may comprise about 1 second to about 30 seconds. In some cases, the period of time for the electrical stimulation may comprise about 1 second to about 5 seconds, about 1 second to about 10 seconds, about 1 second to about 15 seconds, about 1 second to about 20 seconds, about 1 second to about 25 seconds, about 1 second to about 30 seconds, about 5 seconds to about 10 seconds, about 5 seconds to about 15 seconds, about 5 seconds to about 20 seconds, about 5 seconds to about 25 seconds, about 5 seconds to about 30 seconds, about 10 seconds to about 15 seconds, about 10 seconds to about 20 seconds, about 10 seconds to about 25 seconds, about 10 seconds to about 30 seconds, about 15 seconds to about 20 seconds, about 15 seconds to about 25 seconds, about 15 seconds to about 30 seconds, about 20 seconds to about 25 seconds, about 20 seconds to about 30 seconds, or about 25 seconds to about 30 seconds. In some cases, the period of time may comprise about 1 second, about 5 seconds, about 10 seconds, about 15 seconds, about 20 seconds, about 25 seconds, or about 30 seconds. In some cases, the period of time for the electrical stimulation may comprise at least about 1 second, about 5 seconds, about 10 seconds, about 15 seconds, about 20 seconds, or about 25 seconds. In some cases, the period of time for the electrical stimulation may comprise at most about 5 seconds, about 10 seconds, about 15 seconds, about 20 seconds, about 25 seconds, or about 30 seconds.

(30) Embodiments of a Device for Treating and Preventing Incontinence

(31) The disclosure describes devices for preventing an episode of incontinence comprising one or more sensor electrodes, one or more stimulator electrodes, an electric stimulator, or any combination thereof. In some cases, the electric stimulator may comprise a processor, memory, a user interface, a power source, or any combination thereof for preventing an episode of incontinence. The devices may be implantable. The surgical procedure to implant the device may be completed under awake sedation, general anesthesia, local anesthesia, twilight anesthesia, or any combination thereof. The devices may be implanted wholly or partly in an individual's pelvic region. In some cases, the devices may be implanted by one or more surgical instruments. In some instances, surgical instruments may comprise introducers, sheaths, directable probes, wires, needles, or any combination thereof.

(32) In some cases, the device may further comprise a transmitter electrically coupled to a processor capable of wirelessly transmitting and receiving data from a remote device, such as a mobile phone, a tablet, or a computer. In some cases, the device may be configured for open-loop configuration. In some cases, the device may be configured for a close-loop or feedback-controlled configuration. The devices described herein may be used to prevent an episode of urinary incontinence. In some instances, urinary incontinence may comprise urge incontinence, stress incontinence, overflow incontinence, mixed incontinence, or any combination thereof. In some embodiments, the episode of incontinence is urinary incontinence and is urge incontinence type. In some cases, the devices described herein may be used to prevent an episode of fecal incontinence.

(33) Open Loop Configuration

(34) FIG. 1 shows an open loop configuration of the device described herein configured to prevent an

episode of incontinence for an individual **110**. The device in an open loop configuration may comprise an implantable pulse generator **102** comprising one or more stimulator electrodes **106**, and a power source. In some cases, the implantable pulse generator **102** may comprise a processor, and a wireless transmission module configured to execute software to administer an electrical stimulation pattern **104**. In some cases, the power source may comprise a battery. In some instances, the battery may be a lithium polymer ion battery, lithium iodine, lithium manganese dioxide, lithium carbon monofluoride, or any combination thereof. In some cases, the battery may be wirelessly charged by an inductive charger. Alternatively, the battery power source may be a single use.

(35) The implantable pulse generator **102** may deliver a predefined electrical stimulation pattern **104** that has been set by a healthcare provider on a remote device **100** (e.g., a mobile phone, a tablet, a computer, etc.) and transmitted wirelessly **105** to the implantable pulse generator **102**. Alternatively, the implantable pulse generator **102** may deliver a predefined electrical stimulation pattern **104** that has been set by a healthcare provider on a remote computing device **100** (e.g., a mobile phone, a tablet, a computer, etc.) and transmitted via a wired communication **101** to the implantable pulse generator **102**. In some instances, the healthcare provider may set the predefined electrical stimulation parameters through a graphical user interface on the remote device. In some cases, an individual **110** with an implantable pulse generator **102** may modify or set electrical stimulation parameters via an external input device **103** (e.g., a mobile phone, a tablet, a computer, etc.) via a wireless communication **105** of the implantable pulse generator **102**. Alternatively, an individual **110** with an implantable pulse generator **102** may modify or set electrical stimulation parameters via an external input device **103** via a wired communication **107** of the implantable pulse generator **102**. In some instances, an individual **110** with an implantable pulse generator **102** may adjust the electrical stimulation pattern **104** using a graphical user interface on the external input device **103**. In some cases, the electrical stimulation pattern **104** parameters that may be adjusted comprise frequency, amplitude, pulse width, or any combination thereof.

(36) Closed Loop Configuration

(37) FIG. 2 shows a closed loop configuration of the device described herein configured to prevent an episode of incontinence for an individual **126**. The device in a closed loop configuration may comprise an implantable pulse generator **118**, one or more stimulator electrodes **122**, one or more sensor electrodes **120**, and a power source. In some cases, the implantable pulse generator **118** may comprise a processor, and a wireless transmission module configured to execute software to detect, analyze myoelectric electromyograph (EMG) signals via one or more sensor electrodes **120** and administer an electrical stimulation pattern **124**. In some instances, the power source may comprise a battery. The battery may be rechargeable or single use. In some cases, the battery may be charged through inductive charging. In some instances, the device configured in a closed loop configuration may measure EMG signals via one or more sensor electrodes **120** to detect the onset of a stress incontinence event or a level of innate myoelectric electrical activity. In some instances the device configured in a closed loop configuration may measure inertial signals such as rapid acceleration, shock, posture-orientation, or any combination thereof via the one or more sensor electrodes **120** to detect the onset of a stress incontinence event or a level of innate myoelectric electrical activity. In some cases, the stress incontinence event may comprise actions such as coughing, sneezing, laughing, or exercise. Once detected the implantable pulse generator **118** may provide an electrical stimulation pattern **124** to prevent micturition or uncontrolled defecation. In some cases, the threshold level for detecting a stress incontinence event may be modified and adjusted by the individual **126** via graphical user interface on an external input device **114** either via wireless communication **113** or a wired connection **115**.

(38) In some embodiments, a closed-loop configuration of the systems and methods described herein, may comprise a fully synchronized system, as shown in FIG. 9. In some cases, an electrode **902**, comprising a sensor and/or stimulator electrode may detect an EMG, ENG, and/or pressure signal through circuitry **904** (e.g., analog to digital circuitry) and then may pass the detected signal to a classifier algorithm **906**. Once the classifier algorithm **906** has classified the detected EMG, ENG, and/or pressure signal as an incontinent event, the classifier may enable an electric stimulation **908** to be delivered to the patient. As shown in FIG. 9, the electrical stimulation may be provided by the same electrode **902** that sensed the EMG and/or ENG signal initially.

(39) In some instances, a closed loop configuration of the device described herein may be configured to detect an individual's effort in trying to prevent an episode of incontinence. In some instances, the methods and systems described here may supplement the patient's effort with an electrical stimulation pattern via one or more stimulator electrodes **122** sufficient to prevent urinary or fecal defecation. In some cases, an individual's effort may comprise an EMG, ENG, pressure, acceleration, gyroscope, magnetometer, 3-D spatial, or any combination thereof signal at a threshold. In some instances, the threshold myoelectric signal may be detected by one or more sensor electrodes **120**. In some cases, the supplemental excitation, may comprise an excitation signal provided by the stimulator electrodes, described elsewhere herein, with parameters e.g., frequency, pulse width, and/or amplitude such that in combination with the detected individual's effort may prevent an episode of incontinence (e.g., urinary, and/or fecal defecation). In some instances, the supplemental excitation may comprise an excitation signal provided by the stimulator electrodes with parameters e.g., frequency, pulse width, and/or amplitude to prevent an episode of incontinence in response to detecting a stress incontinence event. In some instances, the stress incontinence event may comprise actions such as coughing, sneezing, laughing, or exercise, detectable by e.g., a gyroscope, accelerometer, and/or magnetometer, described elsewhere herein. In some instances the one or more parameters of the supplemental excitation, may be determined on an individual subject basis and/or on a large-scale population of subjects with similar presentation of incontinence. For example, a given subject's supplemental excitation one or more parameters may be tuned and/or determined by whether such supplemental excitation signal prevented an episode of incontinence in real-time or after the incontinent event through a user interface of the device and systems, described elsewhere herein. In some instances, a given subject's supplemental excitation one or more parameters may be tuned to values and/or parameters found to prevent incontinence events in subjects with similar clinical presentation (e.g., age, type of incontinence, frequency of incontinence events, other subject clinical meta data, etc.).

(40) In some instances, the system and methods described herein may comprise system and methods configured to provide an electrical stimulation to prevent an incontinent event based on a subject and/or patient's purposeful muscle contraction and/or movement, as seen in FIG. **8**. In some cases, the subject and/or patient **814** may, upon realizing that they may exhibit an incontinent event, induce movement and/or contraction of one or more muscles or muscle groups to trigger an EMG, ENG, pressure, acceleration, gyroscope, magnetometer, 3-D spatial, or any combination thereof signal **818**. In some instances, the induced movement and/or contraction of one or more muscle groups may be amplified **808**, classified (by a classifier) **806**, passed through a control logic algorithm **804**, and used as a trigger **820** to enable the flow of therapy and respective basal **802** and/or active **801**, described elsewhere herein, stimulation pattern parameters through the stimulator **810** and neural interface **812** to prevent an incontinent event. In some cases, the classifier may comprise a machine learning classifier. In some instances, the classifier may comprise an intensity threshold classifier, described elsewhere herein. In some cases, the patient and/or subject **814** may manually **816** enable the delivery of electrical stimulation via a button on the patient controller module **156**, described elsewhere herein.

(41) In some cases, the detection threshold of the implantable pulse generator **118** (i.e., stimulator) may be modified and tuned via a graphical user interface on an external input device **114**. The external input device **114** may be able to modify and tune the threshold of the implantable pulse generator **118** via a wireless communication **113** or a wired connection **115**. In some cases, the implantable pulse generator threshold may be tuned via a healthcare provider on a remote computing device **112** (e.g., a mobile phone, a tablet, a computer, etc.) and transmitted via a wired communication **116** to the implantable pulse generator **118**. Alternatively, the implantable pulse generator threshold may be tuned via a healthcare provider on a remote computing device **112** (e.g., a mobile phone, a tablet, a computer, etc.) and transmitted via wireless communication **113** to the implantable pulse generator **118**. In some cases, the implantable pulse generator threshold may be tuned via the individual **110** on a remote computing device **112** (e.g., a mobile phone, a tablet, a computer, etc.) and transmitted via a wired communication **116** to the implantable pulse generator **118**. Alternatively, the implantable pulse generator threshold may be tuned via the individual **110** on a remote computing device **112** (e.g., a mobile phone, a tablet, a computer, etc.) and transmitted via wireless communication **113** to the implantable pulse generator **118**.

(42) In some cases, the electrical stimulation pattern **124** provided via one or more stimulator electrodes

122 may be tuned and adjusted by the detected threshold level to supplement an individual's effort to prevent an episode of incontinence. In some instances, the adjusted electrical stimulation pattern **124** may be determined by mapping a detectable physiological signal representing effort and a provided electrical stimulation pattern **124** by, piecewise linear mapping, linear mapping, sigmoidal mapping, or any variations thereof. The electric stimulation pattern **124** may be tuned by modifying or changing the electrical stimulation pattern parameters comprising frequency, pulse-width, and amplitude. In some cases, the electrical stimulation pattern parameters of the implantable pulse generator **118** may be modified and tuned via a graphical user interface on an external input device **114**. The external input device **114** may be able to modify and tune the electrical stimulation pattern parameters of implantable pulse generator **118** via a wireless communication **113** or a wired connection **115**. In some cases, the electrical stimulation pattern parameters of implantable pulse generator **118** may be tuned via a healthcare provider on a remote computing device **112** (e.g., a mobile phone, a tablet, a computer, etc.) and transmitted via a wired communication **116** to the implantable pulse generator **118**. In some cases, the electrical stimulation pattern parameters of implantable pulse generator **118** may be tuned via a healthcare provider on a remote computing device **112** (e.g., a mobile phone, a tablet, a computer, etc.) and transmitted via wireless communication **113** to the implantable pulse generator **118**. In some cases, the electrical stimulation pattern parameters of implantable pulse generator **118** may be tuned by a machine learning model executed by the processor of the implantable pulse generator **118** based on input from the individual **126**. In some cases, the machine learning model may be configured to determine a whether or not a subject is at risk of unwanted micturition and/or fecal defecation based on muscle EMG signals detected by the one or more sensor electrodes. In some cases, the machine learning model may be trained to determine the presence or lack thereof an individual's effort, described elsewhere herein. For example, the machine learning model may be trained with one or more EMG signals characteristic of a subject's muscle contractions in particular EMG signals that lead to unwanted micturition and/or fecal defecation.

(43) In some embodiments, the electrical stimulation pattern parameters, as described elsewhere herein (e.g., frequency, amplitude, etc.), may be set or determined by a stimulation machine learning model. In some cases, the stimulation machine learning model may comprise a Bayesian optimization model. In some instances, the stimulation machine learning model may be trained with stimulation patterns that users of the devices and systems, described elsewhere herein, indicate as inhibiting incontinent events for particular types and/or subtypes of incontinence. In some cases, the stimulation machine learning model may be trained with the patient's type of incontinence. The stimulation machine learning algorithms may be trained in a cloud computing network and/or server in communication with the implantable and user-devices, described elsewhere herein, and re-distributed or downloaded to one or more users and/or patients. In effect users and/or patients may update and/or download new updates to the software of the devices and systems described herein. This aspect of the invention described herein provides an unexpected result of a patient specific and optimized electrical stimulation signal that would otherwise not be achievable with a traditional stimulator.

(44) One or more machine learning algorithms may be used to construct the machine learning model, such as support vector machines that deploy stepwise backwards parameter selection and/or graphical models, both of which may have advantages of inferring interactions between parameters. For example, machine learning algorithms or other statistical algorithms may be used such as alternating decision trees (ADTree), decision stumps, functional trees (FT), logistic model trees (LMT), logistic regression, random forests (rf), receiver operational characteristic curves (ROC), linear regression, extreme gradient boosting (xgb), classification and regression trees, support vector machines (SVM), generalized additive model using splines (e.g., gamSpline), glmnet, multivariate adaptive regression splint (earth), neural network, k-means clustering, or any machine learning algorithm or statistical algorithm known in the art. One or more algorithms may be used together to generate an ensemble method, wherein the ensemble method may be optimized using a machine learning ensemble meta-algorithm such as boosting (e.g., AdaBoost, LPBoost, TotalBoost, BrownBoost, MadaBoost, LogitBoost, etc.) to reduce bias and/or variance.

(45) In some embodiments, the machine learning algorithm may comprise a constrained machine learning algorithm configured to run on micro-processors. In some cases, the machine learning algorithm

may comprise a machine learning algorithm operating on within a TinyML framework. In some instances, the machine learning algorithm, described elsewhere herein may be trained offline. The offline training may be completed on a server, cloud, or other dedicated computing clusters. In some cases, the offline trained machine learning algorithm could then be downloaded, deployed, and/or imported into the device to iteratively improve upon the device and system performance in preventing incontinent events.

(46) One of ordinary skill would realize that such an off-line training structure is feasible and realized in related but different implementation. For example, such a machine learning architecture may be utilized in performing “wake up-word” text classification that are commonly seen in smartphone devices (e.g., “hey siri”, “okay google”, etc.). In some cases, the machine learning algorithms described herein may operate within a framework similar to the “wake up-words” speech machine learning classifier. In some cases, the machine learning algorithms described herein may operate on processing power and memory allocation determined to be sufficient for “wake up” speech machine learning classifiers.

(47) In some cases, the software, described elsewhere herein, may be executed by a processor located on the implanted stimulator. In some cases, the software located on the implanted stimulator may utilize a TinyML constrained machine learning model to accommodate the processing and memory parameters of the implanted stimulator. In some instances, the software may be executed offline on a cloud based computing and/or dedicated computing cluster(s). In some cases, the offline processing workflow may include a high-speed (Bluetooth, Wi-Fi, medical implant communication systems, etc.) data transfer between the implanted stimulator and a local personal computing device (smartphone, tablet, laptop, etc.). The personal processing device may then communicate the implanted stimulator data to a one or more cloud and/or computer clusters that will then send back a resulting output, command, and/or notification to the device. In some cases, the command and/or notification may comprise a warning, alert, initiation of electrical stimulation, or any combination thereof. In some cases, the command may comprise the output of a machine learning classifier configured to determine a threshold intensity of EMG and/or ENG signals indicative of an incontinence event.

(48) The machine learning models may be trained on one or more datasets. In some instances, the one or more datasets may comprise data generated by a user and/or subject, or data generated by a population or segment thereof. In some cases, the data generated by subject and/or the data generated by a population may comprise effort signals, excitation signals, that indicated an incontinent event, and prevented an incontinent event, respectively. In some cases, the devices, systems, and corresponding methods described herein, may record user data and/or input of the user when interacting with the systems and devices described elsewhere herein. In some cases, the data may comprise user labeled EMG, ENG, accelerometer, gyroscope, or any combination thereof sensors, as described elsewhere herein, that lead to an incontinent event. In some cases, these signals may be obtained from the device **1302**, and used in characterizing **1304** and training a machine learning classifier **1306**, as seen in FIG. 13. In some instances, the trained machine learning classifiers trained on or more datasets may then be downloaded to each patient's device **1308** to further improve the machine learning classifier's accuracy. In some cases, the machine learning models may be configured to sense subject effort and/or providing sufficient excitation based on e.g., parameters of frequency, amplitude, and/or pulse-width as described elsewhere herein. In some instances, the datasets of one or more individuals may be pooled together as a training dataset where the subjects show characteristics of similarity between clinical presentation and parameters of excitatory/sensory input. In some cases, clinical presentation may comprise clinical incontinence type, subject clinical meta data, e.g., gender, age, past medical history, current medications taken, past surgical intervention, etc. In some cases, a pooled training datasets may be utilized for an individual during the initial period of training a device implanted into a subject.

(49) In some cases, the one or more machine learning models, described elsewhere herein, may be trained on raw and/or processed signals measured by the devices, sensors, and systems, described elsewhere herein. In some cases, the processed signals may comprise original raw signals that have been filtered to optimize the signal-to-noise ratio of the raw signal. In some cases, the filter may comprise a high-pass, low-pass, band-pass, notch, or any combination thereof filters. In some instances, the one or more machine learning models may alternatively or in addition to, be trained on user feedback regarding prior excitation signal parameters and whether or not such excitation signal parameters prevented an incontinent event.

(50) In some aspects, the disclosure provides a method of processing detected signals **1001**, described elsewhere herein to determine incontinent event precursor signal intensity thresholds, as seen in FIG. **10**. In some cases, the EMG, ENG, accelerometer, gyroscope, magnetometer, pressure sensor signals, or any combination thereof signals **1000** may be detected through an amplifier circuit **1002**. In some cases, the amplifier circuit may comprise an operational amplifier circuit.

(51) In some instances, the amplifier circuit may be configured to amplify signals from about 10 microvolt (μV) to about 1,000 μV . In some instances, the amplifier circuit may be configured to amplify signals from about 10 μV to about 50 μV , about 10 μV to about 100 μV , about 10 μV to about 150 μV , about 10 μV to about 300 μV , about 10 μV to about 500 μV , about 10 μV to about 700 μV , about 10 μV to about 900 μV , about 10 μV to about 1,000 μV , about 50 μV to about 100 μV , about 50 μV to about 150 μV , about 50 μV to about 300 μV , about 50 μV to about 500 μV , about 50 μV to about 700 μV , about 50 μV to about 900 μV , about 50 μV to about 1,000 μV , about 100 μV to about 150 μV , about 100 μV to about 300 μV , about 100 μV to about 500 μV , about 100 μV to about 700 μV , about 100 μV to about 900 μV , about 100 μV to about 1,000 μV , about 150 μV to about 300 μV , about 150 μV to about 500 μV , about 150 μV to about 700 μV , about 150 μV to about 900 μV , about 150 μV to about 1,000 μV , about 300 μV to about 500 μV , about 300 μV to about 700 μV , about 300 μV to about 900 μV , about 300 μV to about 1,000 μV , about 500 μV to about 700 μV , about 500 μV to about 900 μV , about 500 μV to about 1,000 μV , about 700 μV to about 900 μV , about 700 μV to about 1,000 μV , or about 900 μV to about 1,000 μV . In some instances, the amplifier circuit may be configured to amplify signals from about 10 μV , about 50 μV , about 100 μV , about 150 μV , about 300 μV , about 500 μV , about 700 μV , about 900 μV , or about 1,000 μV . In some instances, the amplifier circuit may be configured to amplify signals from at least about 10 μV , about 50 μV , about 100 μV , about 150 μV , about 300 μV , about 500 μV , about 700 μV , or about 900 μV . In some instances, the amplifier circuit may be configured to amplify signals from at most about 50 μV , about 100 μV , about 150 μV , about 300 μV , about 500 μV , about 700 μV , about 900 μV , or about 1,000 μV .

(52) In some cases, the amplifier circuit may be configured to amplify signals with a frequency of about 1 Hz to about 1,500 Hz. In some cases, the amplifier circuit may be configured to amplify signals with a frequency of about 1 Hz to about 20 Hz, about 1 Hz to about 40 Hz, about 1 Hz to about 80 Hz, about 1 Hz to about 100 Hz, about 1 Hz to about 150 Hz, about 1 Hz to about 200 Hz, about 1 Hz to about 250 Hz, about 1 Hz to about 500 Hz, about 1 Hz to about 750 Hz, about 1 Hz to about 1,000 Hz, about 1 Hz to about 1,500 Hz, about 20 Hz to about 40 Hz, about 20 Hz to about 80 Hz, about 20 Hz to about 100 Hz, about 20 Hz to about 150 Hz, about 20 Hz to about 200 Hz, about 20 Hz to about 250 Hz, about 20 Hz to about 500 Hz, about 20 Hz to about 750 Hz, about 20 Hz to about 1,000 Hz, about 20 Hz to about 1,500 Hz, about 40 Hz to about 80 Hz, about 40 Hz to about 100 Hz, about 40 Hz to about 150 Hz, about 40 Hz to about 200 Hz, about 40 Hz to about 250 Hz, about 40 Hz to about 500 Hz, about 40 Hz to about 750 Hz, about 40 Hz to about 1,000 Hz, about 40 Hz to about 1,500 Hz, about 80 Hz to about 100 Hz, about 80 Hz to about 150 Hz, about 80 Hz to about 200 Hz, about 80 Hz to about 250 Hz, about 80 Hz to about 500 Hz, about 80 Hz to about 750 Hz, about 80 Hz to about 1,000 Hz, about 80 Hz to about 1,500 Hz, about 100 Hz to about 150 Hz, about 100 Hz to about 200 Hz, about 100 Hz to about 250 Hz, about 100 Hz to about 500 Hz, about 100 Hz to about 750 Hz, about 100 Hz to about 1,000 Hz, about 100 Hz to about 1,500 Hz, about 150 Hz to about 200 Hz, about 150 Hz to about 250 Hz, about 150 Hz to about 500 Hz, about 150 Hz to about 750 Hz, about 150 Hz to about 1,000 Hz, about 150 Hz to about 1,500 Hz, about 200 Hz to about 250 Hz, about 200 Hz to about 500 Hz, about 200 Hz to about 750 Hz, about 200 Hz to about 1,000 Hz, about 200 Hz to about 1,500 Hz, about 250 Hz to about 500 Hz, about 250 Hz to about 750 Hz, about 250 Hz to about 1,000 Hz, about 250 Hz to about 1,500 Hz, about 500 Hz to about 750 Hz, about 500 Hz to about 1,000 Hz, about 500 Hz to about 1,500 Hz, about 750 Hz to about 1,000 Hz, about 750 Hz to about 1,500 Hz, or about 1,000 Hz to about 1,500 Hz. In some cases, the amplifier circuit may be configured to amplify signals with a frequency of about 1 Hz, about 20 Hz, about 40 Hz, about 80 Hz, about 100 Hz, about 150 Hz, about 200 Hz, about 250 Hz, about 500 Hz, about 750 Hz, about 1,000 Hz, or about 1,500 Hz. In some cases, the amplifier circuit may be configured to amplify signals with a frequency of at least about 1 Hz, about 20 Hz, about 40 Hz, about 80 Hz, about 100 Hz, about 150 Hz, about 200 Hz, about 250 Hz, about 500 Hz, about 750 Hz, or about 1,000 Hz. In some cases, the amplifier circuit may be configured to amplify signals with a frequency of at most about 20 Hz, about 40 Hz, about 80 Hz, about 100 Hz, about 150 Hz, about 200 Hz, about 250 Hz, about 500

Hz, about 750 Hz, about 1,000 Hz, or about 1,500 Hz.

(53) In some cases, the amplified signal may then be passed to a filter **1004**. In some cases, the filter may comprise a low-pass, high-pass, band-pass, notch, or any combination thereof filter.

(54) In some cases, the filter may be configured to filter the frequency band of about 1 Hz to about 70 Hz. In some cases, the filter may be configured to filter the frequency band of about 1 Hz to about 5 Hz, about 1 Hz to about 10 Hz, about 1 Hz to about 15 Hz, about 1 Hz to about 20 Hz, about 1 Hz to about 25 Hz, about 1 Hz to about 40 Hz, about 1 Hz to about 50 Hz, about 1 Hz to about 60 Hz, about 1 Hz to about 70 Hz, about 5 Hz to about 10 Hz, about 5 Hz to about 15 Hz, about 5 Hz to about 20 Hz, about 5 Hz to about 25 Hz, about 5 Hz to about 40 Hz, about 5 Hz to about 50 Hz, about 5 Hz to about 60 Hz, about 5 Hz to about 70 Hz, about 10 Hz to about 15 Hz, about 10 Hz to about 20 Hz, about 10 Hz to about 25 Hz, about 10 Hz to about 40 Hz, about 10 Hz to about 50 Hz, about 10 Hz to about 60 Hz, about 10 Hz to about 70 Hz, about 15 Hz to about 20 Hz, about 15 Hz to about 25 Hz, about 15 Hz to about 40 Hz, about 15 Hz to about 50 Hz, about 15 Hz to about 60 Hz, about 15 Hz to about 70 Hz, about 20 Hz to about 25 Hz, about 20 Hz to about 40 Hz, about 20 Hz to about 50 Hz, about 20 Hz to about 60 Hz, about 20 Hz to about 70 Hz, about 25 Hz to about 40 Hz, about 25 Hz to about 50 Hz, about 25 Hz to about 60 Hz, about 25 Hz to about 70 Hz, about 40 Hz to about 50 Hz, about 40 Hz to about 60 Hz, about 40 Hz to about 70 Hz, about 50 Hz to about 60 Hz, about 50 Hz to about 70 Hz, or about 60 Hz to about 70 Hz. In some cases, the filter may be configured to filter the frequency band of about 1 Hz, about 5 Hz, about 10 Hz, about 15 Hz, about 20 Hz, about 25 Hz, about 40 Hz, about 50 Hz, about 60 Hz, or about 70 Hz. In some cases, the filter may be configured to filter the frequency band of at least about 1 Hz, about 5 Hz, about 10 Hz, about 15 Hz, about 20 Hz, about 25 Hz, about 40 Hz, about 50 Hz, or about 60 Hz. In some cases, the filter may be configured to filter the frequency band of at most about 5 Hz, about 10 Hz, about 15 Hz, about 20 Hz, about 25 Hz, about 40 Hz, about 50 Hz, about 60 Hz, or about 70 Hz.

(55) After passing the signal through the filter, the systems and methods described herein, may rectify **1006** the filtered signal. In the process of rectifying the signal, as understood by one of ordinary skill in the art, the signal will convert the alternating current detected signal to a direct current signal. The rectified signal may be additionally filtered with a low pass filter **1007** that may smooth the rectified signal. The signal may be subjected to a threshold detector **1008**, where the threshold detector determines the onset of an incontinent event from a threshold intensity value of the rectified and smooth processed signals **1000** of EMG, ENG, accelerometer, gyroscope, magnetometer, pressure sensor signals, or any combination thereof. If an incontinent event is determined by the threshold detector **1008**, the system may enable the delivery of electrical stimulation **1010**, as described elsewhere herein.

(56) In some cases, training may be supervised training. Alternatively, training may be unsupervised training. In some instances, the data set may be a retrospective data set. Alternatively, the data set may be prospectively developed dataset, and the machine learning model may be iteratively improved over time.

(57) In some aspects, the disclosure provided herein may comprise a method to train a machine learning model with a data set that comprises sensed signal profiles and excitation signals that have and have not prevented incontinent events. The method may comprise the steps of: preprocessing, training, and predicting.

(58) The method may extract training data from a database, or intake new data, described elsewhere herein. The preprocessing step may apply one or more transformations to standardize the training data or new data for the training step or the prediction step. The preprocessed training data may be passed to the training step, which may construct a machine learning model based on training data. The training step may further comprise a validation step, configured to validate the trained machine learning model using any appropriate validation algorithm (e.g., Stratified K-fold cross-validation). In some cases, the k-fold cross-validation may comprise at least 1-fold, 2, folds, 3 folds, 4 folds, 5 folds, 6 folds, 7 folds, 8 folds, 9 folds, or 10 folds. In some cases, the k-fold cross-validation may comprise up to 1-fold, 2 folds, 3 folds, 4 folds, 5 folds, 6 folds, 7 folds, 8 folds, 9 folds, or 10 folds.

(59) The preprocessing step may apply one or more transformations to the training data to clean and normalize the data. The preprocessing step may be configured to discard parameters which contain spurious data or contain very few observations. The preprocessing module can be further configured to standardize the encoding of parameter values. The preprocessing step may recognize the encoding

variation for the same value and standardize the dataset to have a uniform encoding for a given parameter value. The processing step may thus reduce irregularities in the input data for the training and prediction steps, thereby improving the robustness of the training and prediction steps.

(60) The training step may utilize a machine learning algorithm or other algorithm to construct and train a machine learning model to be used in the association of an excitation stimulation, sensed signal profile, and the presence or lack thereof an incontinent event. A machine learning model may be constructed to capture, based on the training data, the statistical relationship, if any, between excitation stimulation parameters, sensed signal profiles, and the presence or lack thereof an incontinent event.

(61) The machine learning algorithm may have an accuracy greater than about 60%, 70%, 80%, 85%, 90%, 95%, or 99%. The machine learning algorithm may have a positive predictive value greater than about 60%, 70%, 75%, 80%, 85%, 90%, 95%, or 99%. The machine learning algorithm may have a negative predictive value greater than about 60%, 70%, 80%, 90%, 95%, or 99%.

(62) Machine learning data analysis, machine learning model training, or any combination thereof may be performed using one or more of many programming languages and platforms known in the art, such as R, Weka, Python, and/or MATLAB, for example.

(63) The use of such closed-loop bioelectronic systems may potentially provide improved electrical stimulation devices to prevent or reduce stress incontinence. In some cases, the use of such closed-loop bioelectronic systems may provide a more precise approach to prevent or reduce or treat urge incontinence and mixed incontinence (combination of urge and stress incontinence). In some instances, individuals having the implanted device may provide feedback to the parameters provided positive outcomes, negative outcomes, or neutral outcomes. In some cases, positive outcomes may comprise preventing an incontinence event. In some instances, negative outcomes may comprise not preventing an incontinence event, producing pain, or any combination thereof. In some cases, neutral outcomes may comprise not preventing an incontinence event, not producing pain, or any combination thereof. In some cases, the positive outcomes, negative outcomes, neutral outcomes, sensor data, or any combination thereof from a plurality of individuals having the implanted device may be used in tuning algorithms to suggest changes to sensor electrode thresholds and electrical stimulation patterns **124**.

(64) Electrical Simulation Pattern

(65) In some cases, the electric stimulation pattern provided by an implantable pulse generator and one or more stimulator electrodes may be modified or changed to suit the needs of the individual in need thereof preventing an episode of incontinence. In some cases, the one or more stimulator electrodes **122** may output an electric stimulation pattern **124** in response to what is detected by the one or more sensor electrodes **120**. In some cases, the electric stimulation pattern **124** may comprise one or more electrical signals. For example, the electric stimulation pattern **124** may comprise a continuous wave signal (e.g., an electric stimulation signal with a constant frequency in time) and a burst or beating signal superimposed onto the continuous wave signal. In some cases, the burst or beating signal may only be enabled for a short duration of time compared to the continual temporal aspect of the continuous wave signal. In some instances, the combination of the continuous wave signal and/or a burst or beating signal may increase a pain threshold of a subject allowing the stimulator to provide higher amplitude electric stimulation burst pattern to prevent incontinent events. In some instances, the frequency of the electric stimulation pattern may be modified or changed. The frequency pattern may comprise a constant profile, swept profile, beating profile, burst profile, chirped profile, monophasic profile, biphasic profile, or any combination thereof. In some instances, the constant profile is comprised of excitation values at a constant amplitude with a frequency value of 0 Hz. In some cases, a swept profile may comprise a signal with time varying frequency of excitation. In some instances, a beating profile may comprise a combination of one or more excitation signals of varying frequency. In some cases, the burst profile may comprise a signal with a constant frequency that is enveloped by a square, delta, sine, or any combination thereof envelope functions. In some instances, a monophasic profile may comprise an excitation signal with only positive or negative amplitude (e.g., signal with values from 0 to -5V or 0 to 5V only) with a constant frequency. In some cases, the beating or burst profile may comprise an electric stimulation pattern that is provided to a patient and/or subject for during an on-state for a first period of time and is not provided to a patient and/or subject during an off-state for a second period of time. In some cases, beating or burst profiles may provide an excitation signal that may provide for a lengthier

period of muscle excitation without suffering muscle fatigue. In some cases, the on-state and/or off-state may comprise about 0.1 second(s) to about 6.5 s. In some cases, the on-state and/or off-state may comprise about 0.1 s to about 0.5 s, about 0.1 s to about 1 s, about 0.1 s to about 1.2 s, about 0.1 s to about 1.5 s, about 0.1 s to about 2 s, about 0.1 s to about 2.5 s, about 0.1 s to about 3 s, about 0.1 s to about 3.5 s, about 0.1 s to about 4 s, about 0.1 s to about 5 s, about 0.1 s to about 6.5 s, about 0.5 s to about 1 s, about 0.5 s to about 1.2 s, about 0.5 s to about 1.5 s, about 0.5 s to about 2 s, about 0.5 s to about 2.5 s, about 0.5 s to about 3 s, about 0.5 s to about 3.5 s, about 0.5 s to about 4 s, about 0.5 s to about 5 s, about 0.5 s to about 6.5 s, about 1 s to about 1.2 s, about 1 s to about 1.5 s, about 1 s to about 2 s, about 1 s to about 2.5 s, about 1 s to about 3 s, about 1 s to about 3.5 s, about 1 s to about 4 s, about 1 s to about 5 s, about 1 s to about 6.5 s, about 1.2 s to about 1.5 s, about 1.2 s to about 2 s, about 1.2 s to about 2.5 s, about 1.2 s to about 3 s, about 1.2 s to about 3.5 s, about 1.2 s to about 4 s, about 1.2 s to about 5 s, about 1.2 s to about 6.5 s, about 1.5 s to about 2 s, about 1.5 s to about 2.5 s, about 1.5 s to about 3 s, about 1.5 s to about 3.5 s, about 1.5 s to about 4 s, about 1.5 s to about 5 s, about 1.5 s to about 6.5 s, about 2 s to about 2.5 s, about 2 s to about 3 s, about 2 s to about 3.5 s, about 2 s to about 4 s, about 2 s to about 5 s, about 2 s to about 6.5 s, about 2.5 s to about 3 s, about 2.5 s to about 3.5 s, about 2.5 s to about 4 s, about 2.5 s to about 5 s, about 2.5 s to about 6.5 s, about 3 s to about 3.5 s, about 3 s to about 4 s, about 3 s to about 5 s, about 3 s to about 6.5 s, about 3.5 s to about 4 s, about 3.5 s to about 5 s, about 3.5 s to about 6.5 s, about 4 s to about 5 s, about 4 s to about 6.5 s, or about 5 s to about 6.5 s. In some cases, the on-state and/or off-state may comprise about 0.1 s, about 0.5 s, about 1 s, about 1.2 s, about 1.5 s, about 2 s, about 2.5 s, about 3 s, about 3.5 s, about 4 s, about 5 s, or about 6.5 s. In some cases, the on-state and/or off-state may comprise at least about 0.1 s, about 0.5 s, about 1 s, about 1.2 s, about 1.5 s, about 2 s, about 2.5 s, about 3 s, about 3.5 s, about 4 s, or about 5 s. In some cases, the on-state and/or off-state may comprise at most about 0.5 s, about 1 s, about 1.2 s, about 1.5 s, about 2 s, about 2.5 s, about 3 s, about 3.5 s, about 4 s, about 5 s, or about 6.5 s.

(66) In some instances, an electrical stimulation pattern provided during an on-state may comprise an oscillating electric stimulation pattern. In some cases, the oscillating electric stimulation pattern may comprise one or more frequencies. In some cases, the frequency of the oscillating electrical stimulation may be chosen based upon prior knowledge of how similar subjects respond with a particular frequency or range of frequencies of the oscillating electrical stimulation pattern. In some cases, a low frequency (e.g., 2-15 Hz) may induces reductions in bladder contractility and thus prevent voiding or a potential incontinent event. In some instances, a higher frequency (e.g., 20-50 Hz) may lead to potentiation of bladder contractions and voiding. In some instances, the frequency of the electrical stimulation pattern may be about 1 Hz to about 60 Hz. In some instances, the frequency of the electrical stimulation pattern may be about 1 Hz to about 5 Hz, about 1 Hz to about 10 Hz, about 1 Hz to about 15 Hz, about 1 Hz to about 25 Hz, about 1 Hz to about 30 Hz, about 1 Hz to about 35 Hz, about 1 Hz to about 40 Hz, about 1 Hz to about 45 Hz, about 1 Hz to about 50 Hz, about 1 Hz to about 55 Hz, about 1 Hz to about 60 Hz, about 5 Hz to about 10 Hz, about 5 Hz to about 15 Hz, about 5 Hz to about 25 Hz, about 5 Hz to about 30 Hz, about 5 Hz to about 35 Hz, about 5 Hz to about 40 Hz, about 5 Hz to about 45 Hz, about 5 Hz to about 50 Hz, about 5 Hz to about 55 Hz, about 5 Hz to about 60 Hz, about 10 Hz to about 15 Hz, about 10 Hz to about 25 Hz, about 10 Hz to about 30 Hz, about 10 Hz to about 35 Hz, about 10 Hz to about 40 Hz, about 10 Hz to about 45 Hz, about 10 Hz to about 50 Hz, about 10 Hz to about 55 Hz, about 10 Hz to about 60 Hz, about 15 Hz to about 25 Hz, about 15 Hz to about 30 Hz, about 15 Hz to about 35 Hz, about 15 Hz to about 40 Hz, about 15 Hz to about 45 Hz, about 15 Hz to about 50 Hz, about 15 Hz to about 55 Hz, about 15 Hz to about 60 Hz, about 25 Hz to about 30 Hz, about 25 Hz to about 35 Hz, about 25 Hz to about 40 Hz, about 25 Hz to about 45 Hz, about 25 Hz to about 50 Hz, about 25 Hz to about 55 Hz, about 25 Hz to about 60 Hz, about 30 Hz to about 35 Hz, about 30 Hz to about 40 Hz, about 30 Hz to about 45 Hz, about 30 Hz to about 50 Hz, about 30 Hz to about 55 Hz, about 30 Hz to about 60 Hz, about 35 Hz to about 40 Hz, about 35 Hz to about 45 Hz, about 35 Hz to about 50 Hz, about 35 Hz to about 55 Hz, about 35 Hz to about 60 Hz, about 40 Hz to about 45 Hz, about 40 Hz to about 50 Hz, about 40 Hz to about 55 Hz, about 40 Hz to about 60 Hz, about 45 Hz to about 50 Hz, about 45 Hz to about 55 Hz, about 45 Hz to about 60 Hz, about 50 Hz to about 55 Hz, about 50 Hz to about 60 Hz, or about 55 Hz to about 60 Hz. In some instances, the frequency of the electrical stimulation pattern may be about 1 Hz, about 5 Hz, about 10 Hz, about 15 Hz, about 25 Hz, about 30 Hz,

about 35 Hz, about 40 Hz, about 45 Hz, about 50 Hz, or about 55 Hz. In some instances, the frequency of the electrical stimulation pattern may be at least about 1 Hz, about 5 Hz, about 10 Hz, about 15 Hz, about 25 Hz, about 30 Hz, about 35 Hz, about 40 Hz, about 45 Hz, about 50 Hz, or about 55 Hz. In some instances, the frequency of the electrical stimulation pattern may be at most about 5 Hz, about 10 Hz, about 15 Hz, about 25 Hz, about 30 Hz, about 35 Hz, about 40 Hz, about 45 Hz, about 50 Hz, about 55 Hz, or about 60 Hz. In some embodiments, the frequency refers to mean frequency of the electrical stimulation pattern. In some embodiments, the frequency refers to the median frequency. In some embodiments, the frequency refers to the maximum frequency.

(67) In some cases, a burst signal with an on-state and/or an off-state, described elsewhere herein may be provided to the subject to prevent fatigue on the one or more muscles innervated by the subject's pudendal nerve. In some cases, the burst electrical stimulation pattern may comprise one or more frequencies described elsewhere herein.

(68) In some instances, the amplitude of the electric stimulation pattern may be modified or changed. In some instances, the amplitude of the electrical stimulation pattern may be about 1 volt (V) to about 15 V. In some instances, the amplitude of the electrical stimulation pattern may be about 1 V to about 2 V, about 1 V to about 3 V, about 1 V to about 4 V, about 1 V to about 5 V, about 1 V to about 6 V, about 1 V to about 7 V, about 1 V to about 8 V, about 1 V to about 9 V, about 1 V to about 10 V, about 1 V to about 12 V, about 1 V to about 15 V, about 2 V to about 3 V, about 2 V to about 4 V, about 2 V to about 5 V, about 2 V to about 6 V, about 2 V to about 7 V, about 2 V to about 8 V, about 2 V to about 9 V, about 2 V to about 10 V, about 2 V to about 12 V, about 2 V to about 15 V, about 3 V to about 4 V, about 3 V to about 5 V, about 3 V to about 6 V, about 3 V to about 7 V, about 3 V to about 8 V, about 3 V to about 9 V, about 3 V to about 10 V, about 3 V to about 12 V, about 3 V to about 15 V, about 4 V to about 5 V, about 4 V to about 6 V, about 4 V to about 7 V, about 4 V to about 8 V, about 4 V to about 9 V, about 4 V to about 10 V, about 4 V to about 12 V, about 4 V to about 15 V, about 5 V to about 6 V, about 5 V to about 7 V, about 5 V to about 8 V, about 5 V to about 9 V, about 5 V to about 10 V, about 5 V to about 12 V, about 5 V to about 15 V, about 6 V to about 7 V, about 6 V to about 8 V, about 6 V to about 9 V, about 6 V to about 10 V, about 6 V to about 12 V, about 6 V to about 15 V, about 7 V to about 8 V, about 7 V to about 9 V, about 7 V to about 10 V, about 7 V to about 12 V, about 7 V to about 15 V, about 8 V to about 9 V, about 8 V to about 10 V, about 8 V to about 12 V, about 8 V to about 15 V, about 9 V to about 10 V, about 9 V to about 12 V, about 9 V to about 15 V, about 10 V to about 12 V, about 10 V to about 15 V, or about 12 V to about 15 V. In some instances, the amplitude of the electrical stimulation pattern may be about 1 V, about 2 V, about 3 V, about 4 V, about 5 V, about 6 V, about 7 V, about 8 V, about 9 V, about 10 V, about 12 V, or about 15 V. In some instances, the amplitude of the electrical stimulation pattern may be at least about 1 V, about 2 V, about 3 V, about 4 V, about 5 V, about 6 V, about 7 V, about 8 V, about 9 V, about 10 V, or about 12 V. In some instances, the amplitude of the electrical stimulation pattern may be at most about 2 V, about 3 V, about 4 V, about 5 V, about 6 V, about 7 V, about 8 V, about 9 V, about 10 V, about 12 V, or about 15 V. In some embodiments, the amplitude refers to the mean amplitude. In some embodiments, the amplitude refers to the median amplitude. In some embodiments, the amplitude refers to the maximum amplitude. In some embodiments, the amplitude refers to peak to peak amplitude.

(69) In some instances, the amplitude of the electrical stimulation pattern may be about 0.05 milliamperes (mA) to about 10 mA. In some instances, the amplitude of the electrical stimulation pattern may be about 0.05 mA to about 1 mA, about 0.05 mA to about 2 mA, about 0.05 mA to about 3 mA, about 0.05 mA to about 4 mA, about 0.05 mA to about 5 mA, about 0.05 mA to about 6 mA, about 0.05 mA to about 7 mA, about 0.05 mA to about 8 mA, about 0.05 mA to about 9 mA, about 0.05 mA to about 10 mA, about 1 mA to about 2 mA, about 1 mA to about 3 mA, about 1 mA to about 4 mA, about 1 mA to about 5 mA, about 1 mA to about 6 mA, about 1 mA to about 7 mA, about 1 mA to about 8 mA, about 1 mA to about 9 mA, about 1 mA to about 10 mA, about 2 mA to about 3 mA, about 2 mA to about 4 mA, about 2 mA to about 5 mA, about 2 mA to about 6 mA, about 2 mA to about 7 mA, about 2 mA to about 8 mA, about 2 mA to about 9 mA, about 2 mA to about 10 mA, about 3 mA to about 4 mA, about 3 mA to about 5 mA, about 3 mA to about 6 mA, about 3 mA to about 7 mA, about 3 mA to about 8 mA, about 3 mA to about 9 mA, about 3 mA to about 10 mA, about 4 mA to about 5 mA, about 4 mA to about 6 mA, about 4 mA to about 7 mA, about 4 mA to about 8 mA, about 4 mA to about 9 mA, about 4 mA to about 10 mA, about 5 mA to about 6 mA, about 5 mA to about 7 mA, about 5 mA to about 8 mA, about 5 mA to

about 9 mA, about 5 mA to about 10 mA, about 6 mA to about 7 mA, about 6 mA to about 8 mA, about 6 mA to about 9 mA, about 6 mA to about 10 mA, about 7 mA to about 8 mA, about 7 mA to about 9 mA, about 7 mA to about 10 mA, about 8 mA to about 9 mA, about 8 mA to about 10 mA, or about 9 mA to about 10 mA. In some instances, the amplitude of the electrical stimulation pattern may be about 0.05 mA, about 1 mA, about 2 mA, about 3 mA, about 4 mA, about 5 mA, about 6 mA, about 7 mA, about 8 mA, about 9 mA, or about 10 mA. In some instances, the amplitude of the electrical stimulation pattern may be at least about 0.05 mA, about 1 mA, about 2 mA, about 3 mA, about 4 mA, about 5 mA, about 6 mA, about 7 mA, about 8 mA, or about 9 mA. In some instances, the amplitude of the electrical stimulation pattern may be at most about 1 mA, about 2 mA, about 3 mA, about 4 mA, about 5 mA, about 6 mA, about 7 mA, about 8 mA, about 9 mA, or about 10 mA. In some embodiments, the amplitude refers to the mean amplitude. In some embodiments, the amplitude refers to the median amplitude. In some embodiments, the amplitude refers to the maximum amplitude.

(70) In some instances, the pulse width of the electric stimulation pattern may be modified or changed. In some instances, the pulse width of the electrical stimulation pattern may be about 60 μ s to about 390 μ s. In some instances, the pulse width of the electrical stimulation pattern may be about 60 μ s to about 90 μ s, about 60 μ s to about 120 μ s, about 60 μ s to about 150 μ s, about 60 μ s to about 180 μ s, about 60 μ s to about 210 μ s, about 60 μ s to about 240 μ s, about 60 μ s to about 270 μ s, about 60 μ s to about 300 μ s, about 60 μ s to about 330 μ s, about 60 μ s to about 360 μ s, about 60 μ s to about 390 μ s, about 90 μ s to about 120 μ s, about 90 μ s to about 150 μ s, about 90 μ s to about 180 μ s, about 90 μ s to about 210 μ s, about 90 μ s to about 240 μ s, about 90 μ s to about 270 μ s, about 90 μ s to about 300 μ s, about 90 μ s to about 330 μ s, about 90 μ s to about 360 μ s, about 90 μ s to about 390 μ s, about 120 μ s to about 150 μ s, about 120 μ s to about 180 μ s, about 120 μ s to about 210 μ s, about 120 μ s to about 240 μ s, about 120 μ s to about 270 μ s, about 120 μ s to about 300 μ s, about 120 μ s to about 330 μ s, about 120 μ s to about 360 μ s, about 120 μ s to about 390 μ s, about 150 μ s to about 180 μ s, about 150 μ s to about 210 μ s, about 150 μ s to about 240 μ s, about 150 μ s to about 270 μ s, about 150 μ s to about 300 μ s, about 150 μ s to about 330 μ s, about 150 μ s to about 360 μ s, about 150 μ s to about 390 μ s, about 180 μ s to about 210 μ s, about 180 μ s to about 240 μ s, about 180 μ s to about 270 μ s, about 180 μ s to about 300 μ s, about 180 μ s to about 330 μ s, about 180 μ s to about 360 μ s, about 180 μ s to about 390 μ s, about 210 μ s to about 240 μ s, about 210 μ s to about 270 μ s, about 210 μ s to about 300 μ s, about 210 μ s to about 330 μ s, about 210 μ s to about 360 μ s, about 210 μ s to about 390 μ s, about 240 μ s to about 270 μ s, about 240 μ s to about 300 μ s, about 240 μ s to about 330 μ s, about 240 μ s to about 360 μ s, about 240 μ s to about 390 μ s, about 270 μ s to about 300 μ s, about 270 μ s to about 330 μ s, about 270 μ s to about 360 μ s, about 270 μ s to about 390 μ s, about 300 μ s to about 330 μ s, about 300 μ s to about 360 μ s, about 300 μ s to about 390 μ s, about 330 μ s to about 360 μ s, about 330 μ s to about 390 μ s, or about 360 μ s to about 390 μ s. In some instances, the pulse width of the electrical stimulation pattern may be about 60 μ s, about 90 μ s, about 120 μ s, about 150 μ s, about 180 μ s, about 210 μ s, about 240 μ s, about 270 μ s, about 300 μ s, about 330 μ s, about 360 μ s, or about 390 μ s. In some instances, the pulse width of the electrical stimulation pattern may be at least about 60 μ s, about 90 μ s, about 120 μ s, about 150 μ s, about 180 μ s, about 210 μ s, about 240 μ s, about 270 μ s, about 300 μ s, about 330 μ s, or about 360 μ s. In some instances, the pulse width of the electrical stimulation pattern may be at most about 90 μ s, about 120 μ s, about 150 μ s, about 180 μ s, about 210 μ s, about 240 μ s, about 270 μ s, about 300 μ s, about 330 μ s, about 360 μ s, or about 390 μ s. In some embodiments, the pulse width refers to the mean pulse width. In some embodiments, the pulse width refers to the median pulse width. In some embodiments, the pulse width refers to the maximum pulse width.

(71) Sensors for Use with Embodiments of the Device Described Herein

(72) In some instances, a sensor electrode for use with embodiments of the device described herein may be configured to sense a parameter associated with a response of an individual intending to prevent an episode of incontinence or that indicates the individual may have an episode of incontinence. In some cases, the sensor electrode may be configured to sense a contraction of a muscle of the individual that results in a partial contraction of a sphincter that controls bladder or bowel voiding. In some cases, a sensor electrode may be configured to sense bulk motion or anatomical stress of an individual. In some instances, the sensor electrode may comprise a sensor configured to detect electromyography (EMG) signals. In some cases, the sensor electrode may be configured to sense myoelectric activity. In some

instances, an EMG signal threshold may determine that a contraction of at least one pelvic muscle has occurred. In some cases, the strength of the EMG signal may be proportional to the strength of the contraction of at least one pelvic muscle. In some instances, the sensor electrode detects an action potential signal. In some cases, the device comprises an amplifier to amplify the signal obtained by the sensor electrode, to facilitate analysis and classification of the signal by the processor.

(73) In some cases, the sensor electrode may be implanted within the pelvis of the individual. In some instances, the sensor electrode may be implanted at or adjacent to the pudendal nerve. In some cases, the sensor electrode may be implanted at or adjacent to the sacral nerve. In some cases, the sensor electrode may be implanted within or adjacent to one or more of the pelvic muscles. In some instances, the sensor electrode may be implanted at or adjacent to the pelvic floor. In some cases, the sensor electrode may be implanted at or adjacent to the urethral sphincter. In some instances, the sensor electrode may be implanted at or adjacent to one or more of the urethra, ureter, and bladder. In some cases, the sensor electrode may be implanted at or adjacent to the anal sphincter. In some instances, the sensor electrode may be implanted at or adjacent to one or more of the anus, rectum, and bowel. In some instances, the devices described herein may comprise a plurality of sensor electrodes. In some cases, the devices described herein may comprise one or more sensor electrodes. In some instances, the devices described herein may comprise a different sensor electrode, or a second sensor electrode. In some cases, the sensor electrode may detect the signal from a muscle area innervated by a first pudendal nerve and the different or second sensor electrode may detect signal from a muscle area innervated by a second pudendal nerve. In some instances, the sensor electrode may detect a signal from a muscle area innervated by a first sacral nerve and the different or second sensor electrode detects a signal from a muscle area innervated by a second sacral nerve.

(74) In some instances, the sensor electrode may comprise a casing and a lead. In some instances, the casing may be made of titanium, titanium alloy, tantalum, or any combination thereof. In some cases, the lead may be made of a metal alloy. In some embodiments, a sensor electrode and a stimulator electrode may be positioned on a single lead. In some instances, the lead may be electrically coupled to one or more sensor electrodes or one or more stimulator electrodes. In some embodiments, a sensor electrode and a stimulator electrode are positioned on separate leads. In some cases, the one or more sensor electrodes may comprise bioelectrical sensor electrodes.

(75) Electrodes for Use with Embodiments of the Device Described Herein

(76) In some cases, the sensor electrode and the stimulator electrode may be located on a single lead. In some instances, one or more sensor electrodes and one or more stimulator electrodes may be located on a single lead. In some cases, one sensor electrode and one stimulator electrode may be located on a single lead. In some instances, the sensor electrode and the stimulator electrode may be located on separate leads. In some cases, the sensor electrode and the stimulator electrode may each be located on its own lead. In some cases, the one or more sensor electrodes and one or more stimulator electrodes may be in a linear geometry, triangular geometry, square geometry, hexagonal geometry, or a general polygonal geometry. In some cases, the electrodes located on a single lead may be spaced by a distance. In some cases, the spacing provides the capability to stimulate multiple locations along the length of the nerve. In some cases, the electrodes may be separated by a distance of at least about 1 mm, about 1.5 mm, about 2 mm, about 2.5 mm, about 5 mm, about 10 mm, about 20 mm, about 30 mm, about 40 mm, about 50 mm, or about 60 mm. In some cases, the electrodes may be separated by a distance of at most about 1.5 mm, about 2 mm, about 2.5 mm, about 5 mm, about 10 mm, about 20 mm, about 30 mm, about 40 mm, about 50 mm, about 60 mm, or about 80 mm.

(77) In some instances, the device may comprise one or more leads. In some cases, the device may comprise at least two leads. In some instances, the device may comprise at least 1, 2, 3, 4, 5, 6, 7, 8, 9, or 10 leads. In some cases, the device may comprise at most 1, 2, 3, 4, 5, 6, 7, 8, 9, or 10 leads. In some instances, the device may comprise 1, 2, 3, 4, 5, 6, 7, 8, 9, or 10 leads. In some cases, each lead may comprise one or more electrodes. In some cases, the one or more electrodes may comprise one or more sensor and/or stimulation electrodes. In some cases, the one or more electrodes on a lead may comprise about 1 electrode to about 10 electrodes. In some cases, the one or more electrodes on a lead may comprise about 1 electrode to about 2 electrodes, about 1 electrode to about 3 electrodes, about 1 electrode to about 4 electrodes, about 1 electrode to about 5 electrodes, about 1 electrode to about 6

electrodes, about 1 electrode to about 7 electrodes, about 1 electrode to about 8 electrodes, about 1 electrode to about 9 electrodes, about 1 electrode to about 10 electrodes, about 2 electrodes to about 3 electrodes, about 2 electrodes to about 4 electrodes, about 2 electrodes to about 5 electrodes, about 2 electrodes to about 6 electrodes, about 2 electrodes to about 7 electrodes, about 2 electrodes to about 8 electrodes, about 2 electrodes to about 9 electrodes, about 2 electrodes to about 10 electrodes, about 3 electrodes to about 4 electrodes, about 3 electrodes to about 5 electrodes, about 3 electrodes to about 6 electrodes, about 3 electrodes to about 7 electrodes, about 3 electrodes to about 8 electrodes, about 3 electrodes to about 9 electrodes, about 3 electrodes to about 10 electrodes, about 4 electrodes to about 5 electrodes, about 4 electrodes to about 6 electrodes, about 4 electrodes to about 7 electrodes, about 4 electrodes to about 8 electrodes, about 4 electrodes to about 9 electrodes, about 4 electrodes to about 10 electrodes, about 5 electrodes to about 6 electrodes, about 5 electrodes to about 7 electrodes, about 5 electrodes to about 8 electrodes, about 5 electrodes to about 9 electrodes, about 5 electrodes to about 10 electrodes, about 6 electrodes to about 7 electrodes, about 6 electrodes to about 8 electrodes, about 6 electrodes to about 9 electrodes, about 6 electrodes to about 10 electrodes, about 7 electrodes to about 8 electrodes, about 7 electrodes to about 9 electrodes, about 7 electrodes to about 10 electrodes, about 8 electrodes to about 9 electrodes, about 8 electrodes to about 10 electrodes, or about 9 electrodes to about 10 electrodes. In some cases, the one or more electrodes on a lead may comprise about 1 electrode, about 2 electrodes, about 3 electrodes, about 4 electrodes, about 5 electrodes, about 6 electrodes, about 7 electrodes, about 8 electrodes, about 9 electrodes, or about 10 electrodes. In some cases, the one or more electrodes on a lead may comprise at least about 1 electrode, about 2 electrodes, about 3 electrodes, about 4 electrodes, about 5 electrodes, about 6 electrodes, about 7 electrodes, about 8 electrodes, or about 9 electrodes. In some cases, the one or more electrodes on a lead may comprise at most about 2 electrodes, about 3 electrodes, about 4 electrodes, about 5 electrodes, about 6 electrodes, about 7 electrodes, about 8 electrodes, about 9 electrodes, or about 10 electrodes.

(78) In some cases, each electrode may comprise a length, whereby the length may provide localized excitation of one or more nerves of the sacral or pudendal nerve. In some cases, each electrode may comprise a length of about 0.1 millimeter (mm) to about 2 mm. In some cases, each electrode may comprise a length of about 0.1 mm to about 0.3 mm, about 0.1 mm to about 0.5 mm, about 0.1 mm to about 0.7 mm, about 0.1 mm to about 0.8 mm, about 0.1 mm to about 1 mm, about 0.1 mm to about 1.2 mm, about 0.1 mm to about 1.4 mm, about 0.1 mm to about 1.5 mm, about 0.1 mm to about 2 mm, about 0.3 mm to about 0.5 mm, about 0.3 mm to about 0.7 mm, about 0.3 mm to about 0.8 mm, about 0.3 mm to about 1 mm, about 0.3 mm to about 1.2 mm, about 0.3 mm to about 1.4 mm, about 0.3 mm to about 1.5 mm, about 0.3 mm to about 2 mm, about 0.5 mm to about 0.7 mm, about 0.5 mm to about 0.8 mm, about 0.5 mm to about 1 mm, about 0.5 mm to about 1.2 mm, about 0.5 mm to about 1.4 mm, about 0.5 mm to about 1.5 mm, about 0.5 mm to about 2 mm, about 0.7 mm to about 0.8 mm, about 0.7 mm to about 1 mm, about 0.7 mm to about 1.2 mm, about 0.7 mm to about 1.4 mm, about 0.7 mm to about 1.5 mm, about 0.7 mm to about 2 mm, about 0.8 mm to about 1 mm, about 0.8 mm to about 1.2 mm, about 0.8 mm to about 1.4 mm, about 0.8 mm to about 1.5 mm, about 0.8 mm to about 2 mm, about 1 mm to about 1.2 mm, about 1 mm to about 1.4 mm, about 1 mm to about 1.5 mm, about 1 mm to about 2 mm, about 1.2 mm to about 1.4 mm, about 1.2 mm to about 1.5 mm, about 1.2 mm to about 2 mm, about 1.4 mm to about 1.5 mm, about 1.4 mm to about 2 mm, or about 1.5 mm to about 2 mm. In some cases, each electrode may comprise a length of about 0.1 mm, about 0.3 mm, about 0.5 mm, about 0.7 mm, about 0.8 mm, about 1 mm, about 1.2 mm, about 1.4 mm, about 1.5 mm, or about 2 mm. In some cases, each electrode may comprise a length of at least about 0.1 mm, about 0.3 mm, about 0.5 mm, about 0.7 mm, about 0.8 mm, about 1 mm, about 1.2 mm, about 1.4 mm, or about 1.5 mm. In some cases, each electrode may comprise a length of at most about 0.3 mm, about 0.5 mm, about 0.7 mm, about 0.8 mm, about 1 mm, about 1.2 mm, about 1.4 mm, about 1.5 mm, or about 2 mm.

(79) In some cases, the one or more leads may comprise a length between the proximal tip of the lead to the distal end electrically coupled to the stimulator. In some cases, the length of the lead may vary based upon the anatomy of the individual or subject receiving the implanted device and leads. In some cases, the lead length may comprise a length, whereby the lead electrodes may be placed near and/or adjacent to the sacral and/or pudendal nerve yet reach the placement of the stimulator in buttock fat pockets of the subject. In some instances, the one or more leads may comprise a length of about 20 centimeters (cm) to

about 50 cm. In some instances, the one or more leads may comprise a length of about 20 cm to about 25 cm, about 20 cm to about 30 cm, about 20 cm to about 35 cm, about 20 cm to about 40 cm, about 20 cm to about 45 cm, about 20 cm to about 50 cm, about 25 cm to about 30 cm, about 25 cm to about 35 cm, about 25 cm to about 40 cm, about 25 cm to about 45 cm, about 25 cm to about 50 cm, about 30 cm to about 35 cm, about 30 cm to about 40 cm, about 30 cm to about 45 cm, about 30 cm to about 50 cm, about 35 cm to about 40 cm, about 35 cm to about 45 cm, about 35 cm to about 50 cm, about 40 cm to about 45 cm, about 40 cm to about 50 cm, or about 45 cm to about 50 cm. In some instances, the one or more leads may comprise a length of about 20 cm, about 25 cm, about 30 cm, about 35 cm, about 40 cm, about 45 cm, or about 50 cm. In some instances, the one or more leads may comprise a length of at least about 20 cm, about 25 cm, about 30 cm, about 35 cm, about 40 cm, or about 45 cm. In some instances, the one or more leads may comprise a length of at most about 25 cm, about 30 cm, about 35 cm, about 40 cm, about 45 cm, or about 50 cm.

(80) In some cases, the one or more leads may comprise a diameter. In some cases, the diameter of the lead may vary based upon the anatomy of the individual or subject receiving the implanted device and leads. In some cases, the lead diameter may comprise a diameter, whereby the diameter provides a form factor for minimally invasive placement of the lead in the subject. In some cases, the diameter of the lead may comprise a diameter at which the lead will resist breakage. In some cases, the one or more leads may have an outer diameter of about 0.1 mm to about 2 mm. In some cases, the one or more leads may have an outer diameter of about 0.1 mm to about 0.2 mm, about 0.1 mm to about 0.3 mm, about 0.1 mm to about 0.5 mm, about 0.1 mm to about 0.8 mm, about 0.1 mm to about 1 mm, about 0.1 mm to about 1.2 mm, about 0.1 mm to about 1.4 mm, about 0.1 mm to about 1.5 mm, about 0.1 mm to about 2 mm, about 0.2 mm to about 0.3 mm, about 0.2 mm to about 0.5 mm, about 0.2 mm to about 0.8 mm, about 0.2 mm to about 1 mm, about 0.2 mm to about 1.2 mm, about 0.2 mm to about 1.4 mm, about 0.2 mm to about 1.5 mm, about 0.2 mm to about 2 mm, about 0.3 mm to about 0.5 mm, about 0.3 mm to about 0.8 mm, about 0.3 mm to about 1 mm, about 0.3 mm to about 1.2 mm, about 0.3 mm to about 1.4 mm, about 0.3 mm to about 1.5 mm, about 0.3 mm to about 2 mm, about 0.5 mm to about 0.8 mm, about 0.5 mm to about 1 mm, about 0.5 mm to about 1.2 mm, about 0.5 mm to about 1.4 mm, about 0.5 mm to about 1.5 mm, about 0.5 mm to about 2 mm, about 0.8 mm to about 1 mm, about 0.8 mm to about 1.2 mm, about 0.8 mm to about 1.4 mm, about 0.8 mm to about 1.5 mm, about 0.8 mm to about 2 mm, about 1 mm to about 1.2 mm, about 1 mm to about 1.4 mm, about 1 mm to about 1.5 mm, about 1 mm to about 2 mm, about 1.2 mm to about 1.4 mm, about 1.2 mm to about 1.5 mm, about 1.2 mm to about 2 mm, about 1.4 mm to about 1.5 mm, about 1.4 mm to about 2 mm, or about 1.5 mm to about 2 mm. In some cases, the one or more leads may have an outer diameter of about 0.1 mm, about 0.2 mm, about 0.3 mm, about 0.5 mm, about 0.8 mm, about 1 mm, about 1.2 mm, about 1.4 mm, about 1.5 mm, or about 2 mm. In some cases, the one or more leads may have an outer diameter of at least about 0.1 mm, about 0.2 mm, about 0.3 mm, about 0.5 mm, about 0.8 mm, about 1 mm, about 1.2 mm, about 1.4 mm, or about 1.5 mm. In some cases, the one or more leads may have an outer diameter of at most about 0.2 mm, about 0.3 mm, about 0.5 mm, about 0.8 mm, about 1 mm, about 1.2 mm, about 1.4 mm, about 1.5 mm, or about 2 mm.

(81) In some cases, the one or more leads described elsewhere herein may comprise an internal stylet and/or mandrel configured to provide rigidity to the lead for implantation and/or insertion to a subject. In some cases, the internal stylet may be removed from the lead once the lead has been inserted and implanted. In some instances, the one or more leads described elsewhere herein may be sterilizable with conventional methods of sterilization used in the medical field, e.g., gas sterilization, steam sterilization, UV sterilization, etc.

(82) In some instances, the one or more leads, described elsewhere herein, may be electrically coupled to the electric stimulator, described elsewhere herein. In some instances, the one or more leads may be coupled to the stimulator and may at a later point in time be uncoupled from the stimulator. In some instances, the one or more leads may couple to the electric stimulator with a quick release electrical coupling. In some cases, the one or more leads may be coupled to the electric stimulator by a set screw fastener, whereby a lead is inserted into a hollow cylindrical geometry in electrical communication with the electric stimulator internal circuitry. The lead may then be fastened i.e., held in tension against the inner wall of the hollow cylindrical geometry, to the conductive hollow cylindrical geometry with a non-conductive machine set screw. The one or more leads may be placed into the electric stimulator prior to

or during the surgical implantation procedure.

(83) In some instances, the sensor electrode and the stimulator electrode may be operatively coupled to a processor and a non-transitory computer readable medium that includes software. In some cases, the sensor electrode may be calibrated by the individual using an external input device that interfaces with the software. In some instances, the software may be configured to record a signal from the sensor electrode. In some cases, the software may be configured to adjust the sensor electrode in response to the signal.

(84) In some instances, the stimulator electrode may provide an electrical stimulation to the pudendal nerve. In some cases, the stimulator electrode may provide an electrical stimulation to the sacral nerve. In some instances, the stimulator electrode may provide an electrical stimulation to one or more of the nerves innervating the pelvic muscles. In some cases, the electrode may comprise one or more electrodes or leads e.g., a first and second electrode. In some instances, the electrode may comprise one or more stimulator electrodes. In some cases, an electrode may comprise a first stimulator electrode configured to stimulate one pudendal nerve, and a second stimulator electrode configured to stimulate another spatially independent region of the same pudendal nerve. In some cases, the first stimulator electrode may stimulate the main trunk of the pudendal nerve and the second stimulator electrode may stimulate the distal nerve of the pudendal nerve. In some instances, the distal nerve of the pudendal nerve comprises branches thereof the distal pudendal nerve. In some cases, the first stimulator electrode may stimulate the trunk of the pudendal nerve and the second stimulator electrode may stimulate a main branch of the pudendal nerve e.g., dorsal genital nerve. In some cases, the stimulator electrode may provide an electrical stimulation to one or more of the nerves innervating the urethral sphincter. In some instances, the stimulator electrode may provide an electrical stimulation to one or more of the nerves innervating the urethra. In some cases, the stimulator electrode may provide an electrical stimulation to one or more of the nerves innervating the bladder. In some instances, the stimulator electrode may provide an electrical stimulation to one or more of the nerves innervating the ureter. In some cases, the stimulator electrode may provide an electrical stimulation to one or more of the nerves innervating the anal sphincter. In some cases, the stimulator electrode may provide an electrical stimulation to one or more of the nerves innervating the anus. In some instances, the stimulator electrode may provide an electrical stimulation to one or more of the nerves innervating the rectum. In some cases, the stimulator electrode may provide an electrical stimulation to one or more of the nerves innervating the bowel. In some instances, the devices described herein may comprise a plurality of stimulator electrodes. In some instances, the devices described herein may comprise one or more stimulator electrodes. In some instances, the devices described herein may comprise a different stimulator electrode, or a second stimulator electrode. In some cases, the stimulator electrode may stimulate a first pudendal nerve and the different or second stimulator electrode stimulates a second pudendal nerve. In some cases, the stimulator electrode may stimulate a first sacral nerve and the different or second stimulator electrode stimulates a second sacral nerve.

(85) In some instances, the stimulator electrode of the device may provide a constant electrical stimulation. In some cases, the stimulator electrode of the device may provide a constant electrical stimulation at a lower intensity level than the electrical stimulation provided to prevent an episode of incontinence. In some cases, constant electrical stimulation may comprise constant frequency, amplitude, current, or any combination thereof. In some instances, the intensity or duration of the electrical stimulation provided may prevent an episode of incontinence varies according to the individual's response to a possible episode of incontinence that is sensed by the sensor electrode. In some cases, the individual's response to prevent a possible episode of incontinence that is sensed by the sensor electrode may be insufficient on its own to prevent the episode of incontinence and the electrical stimulation delivered by the stimulator electrode of the device provides sufficient stimulation, together with the response, to prevent the episode of incontinence. In some cases, the individual's response to prevent a possible episode of incontinence that may be sensed by the sensor electrode combined with the electrical stimulation delivered by the stimulator electrode of the device provides sufficient stimulation to trigger the action potential of the muscles responsible for the incontinence, resulting in contraction of the muscle. In some cases, the combined stimulation from the individual and the device may result in contraction of a muscle. In some instances, the muscle may comprise a urethral sphincter. In some cases,

the muscle may comprise an anal sphincter. In some cases, the muscle may comprise one or more of the pelvic floor muscles.

(86) In some instances, the sensor electrode may comprise casing and a lead. In some instances, the casing may be made of titanium or a titanium alloy. In some cases, the lead may be made of a metal alloy.

(87) Device Anchors

(88) In some cases, the devices may be anchored when implanted by one the one or more surgical device. In some instances, anchoring of the devices may be achieved by sliding the device over the lead, described elsewhere herein, and then compressing it onto the lead using ligatures such that it is immobile. These ligatures may be used to fix the anchoring device to native adjacent tissue such as ligament or periosteum. In some instances, the device may comprise grooves for the purpose of aligning compression ligatures. In some instances, the anchoring device may comprise a torque system to compress the device onto the lead such that it is immobile. In some instances, the device may be compressed at a single point onto the lead. In some instances, the device is compressed at two or more points onto the lead. The electric stimulator may comprise radio-opaque markers that may permit visualization of the electric stimulator under x-ray e.g., fluoroscopy during and/or after implantation. In some instances, fluoroscopy may be used alone or in combination with EMG sensor electrode readings of pelvic floor muscles to verify placement or to adjustment placement of sensor electrode leads, stimulator electrode leads, and/or the stimulator.

(89) Electrical Signal

(90) FIG. 4 depicts the sensed myoelectric EMG signal and a corresponding electrical stimulation (“Bio-stimulator”) provided by the devices disclosed herein. In some cases, the myoelectric EMG signal (“Bio-signal”) **140** may comprise electrical fluctuations **142** corresponding to contractile activity of the muscle near the sensor electrode. In some cases, the electrical fluctuations **142** may represent a cough, sudden movement, change in inertia, or any combination thereof from an electromyography (EMG) reading. In some instances, the bio-signal captured by the sensor electrode may be analyzed and classified **144** to identify any stress events, such as a cough or a sudden movement, from an EMG reading. In some cases, the identified stress event may initiate a process **146** by which the implantable electric stimulator may deliver an electric stimulation pattern **148** to via one or more stimulator electrodes to a one or more pudendal nerve to prevent an episode of incontinence. In some cases, the stimulator electrode may deliver an electrical stimulation by the one or more stimulator electrodes **122** that is configured to supplement the innate reflex detected by the one or more sensor electrodes **120** to account for the stress event otherwise leading to an incontinence event.

(91) In some cases, the electric stimulator may comprise a width and length to be readily implantable in a patient. In some cases, the electric stimulator may comprise a width and length to accommodate circuitry and/or other system level components, described elsewhere herein.

(92) In some instances, the electric stimulator may comprise a width and length to provide sufficient space for a battery, where the battery may comprise a lifetime after which the battery may be replaced. In some cases, the battery lifetime may comprise about 5 years to about 15 years. In some cases, the battery lifetime may comprise about 5 years to about 6 years, about 5 years to about 7 years, about 5 years to about 8 years, about 5 years to about 9 years, about 5 years to about 10 years, about 5 years to about 11 years, about 5 years to about 12 years, about 5 years to about 13 years, about 5 years to about 14 years, about 5 years to about 15 years, about 6 years to about 7 years, about 6 years to about 8 years, about 6 years to about 9 years, about 6 years to about 10 years, about 6 years to about 11 years, about 6 years to about 12 years, about 6 years to about 13 years, about 6 years to about 14 years, about 6 years to about 15 years, about 7 years to about 8 years, about 7 years to about 9 years, about 7 years to about 10 years, about 7 years to about 11 years, about 7 years to about 12 years, about 7 years to about 13 years, about 7 years to about 14 years, about 7 years to about 15 years, about 8 years to about 9 years, about 8 years to about 10 years, about 8 years to about 11 years, about 8 years to about 12 years, about 8 years to about 13 years, about 8 years to about 14 years, about 8 years to about 15 years, about 9 years to about 10 years, about 9 years to about 11 years, about 9 years to about 12 years, about 9 years to about 13 years, about 9 years to about 14 years, about 9 years to about 15 years, about 10 years to about 11 years, about 10 years to about 12 years, about 10 years to about 13 years, about 10 years to about 14 years, about 10

years to about 15 years, about 11 years to about 12 years, about 11 years to about 13 years, about 11 years to about 14 years, about 11 years to about 15 years, about 12 years to about 13 years, about 12 years to about 14 years, about 12 years to about 15 years, about 13 years to about 14 years, about 13 years to about 15 years, or about 14 years to about 15 years. In some cases, the battery lifetime may comprise about 5 years, about 6 years, about 7 years, about 8 years, about 9 years, about 10 years, about 11 years, about 12 years, about 13 years, about 14 years, or about 15 years. In some cases, the battery lifetime may comprise at least about 5 years, about 6 years, about 7 years, about 8 years, about 9 years, about 10 years, about 11 years, about 12 years, about 13 years, or about 14 years. In some cases, the battery lifetime may comprise at most about 6 years, about 7 years, about 8 years, about 9 years, about 10 years, about 11 years, about 12 years, about 13 years, about 14 years, or about 15 years.

(93) In some cases, the electric stimulator battery may require charging once in about 1 day to about 12 days. In some cases, the electric stimulator battery may require charging once in about 1 day to about 2 days, about 1 day to about 3 days, about 1 day to about 4 days, about 1 day to about 5 days, about 1 day to about 6 days, about 1 day to about 7 days, about 1 day to about 8 days, about 1 day to about 9 days, about 1 day to about 10 days, about 1 day to about 11 days, about 1 day to about 12 days, about 2 days to about 3 days, about 2 days to about 4 days, about 2 days to about 5 days, about 2 days to about 6 days, about 2 days to about 7 days, about 2 days to about 8 days, about 2 days to about 9 days, about 2 days to about 10 days, about 2 days to about 11 days, about 2 days to about 12 days, about 3 days to about 4 days, about 3 days to about 5 days, about 3 days to about 6 days, about 3 days to about 7 days, about 3 days to about 8 days, about 3 days to about 9 days, about 3 days to about 10 days, about 3 days to about 11 days, about 3 days to about 12 days, about 4 days to about 5 days, about 4 days to about 6 days, about 4 days to about 7 days, about 4 days to about 8 days, about 4 days to about 9 days, about 4 days to about 10 days, about 4 days to about 11 days, about 4 days to about 12 days, about 5 days to about 6 days, about 5 days to about 7 days, about 5 days to about 8 days, about 5 days to about 9 days, about 5 days to about 10 days, about 5 days to about 11 days, about 5 days to about 12 days, about 6 days to about 7 days, about 6 days to about 8 days, about 6 days to about 9 days, about 6 days to about 10 days, about 6 days to about 11 days, about 6 days to about 12 days, about 7 days to about 8 days, about 7 days to about 9 days, about 7 days to about 10 days, about 7 days to about 11 days, about 7 days to about 12 days, about 8 days to about 9 days, about 8 days to about 10 days, about 8 days to about 11 days, about 8 days to about 12 days, about 9 days to about 10 days, about 9 days to about 11 days, about 9 days to about 12 days, about 10 days to about 11 days, about 10 days to about 12 days, or about 11 days to about 12 days. In some cases, the electric stimulator battery may require charging once in about 1 day, about 2 days, about 3 days, about 4 days, about 5 days, about 6 days, about 7 days, about 8 days, about 9 days, about 10 days, about 11 days, or about 12 days. In some cases, the electric stimulator battery may require charging once in at least about 1 day, about 2 days, about 3 days, about 4 days, about 5 days, about 6 days, about 7 days, about 8 days, about 9 days, about 10 days, or about 11 days. In some cases, the electric stimulator battery may require charging once in at most about 2 days, about 3 days, about 4 days, about 5 days, about 6 days, about 7 days, about 8 days, about 9 days, about 10 days, about 11 days, or about 12 days.

(94) In some instances, the electric stimulator may comprise a width of about 1 mm to about 50 mm. In some instances, the electric stimulator may comprise a width of about 1 mm to about 5 mm, about 1 mm to about 10 mm, about 1 mm to about 15 mm, about 1 mm to about 20 mm, about 1 mm to about 25 mm, about 1 mm to about 30 mm, about 1 mm to about 35 mm, about 1 mm to about 40 mm, about 1 mm to about 45 mm, about 1 mm to about 50 mm, about 5 mm to about 10 mm, about 5 mm to about 15 mm, about 5 mm to about 20 mm, about 5 mm to about 25 mm, about 5 mm to about 30 mm, about 5 mm to about 35 mm, about 5 mm to about 40 mm, about 5 mm to about 45 mm, about 5 mm to about 50 mm, about 10 mm to about 15 mm, about 10 mm to about 20 mm, about 10 mm to about 25 mm, about 10 mm to about 30 mm, about 10 mm to about 35 mm, about 10 mm to about 40 mm, about 10 mm to about 45 mm, about 10 mm to about 50 mm, about 15 mm to about 20 mm, about 15 mm to about 25 mm, about 15 mm to about 30 mm, about 15 mm to about 35 mm, about 15 mm to about 40 mm, about 15 mm to about 45 mm, about 15 mm to about 50 mm, about 20 mm to about 25 mm, about 20 mm to about 30 mm, about 20 mm to about 35 mm, about 20 mm to about 40 mm, about 20 mm to about 45 mm, about 20 mm to about 50 mm, about 25 mm to about 30 mm, about 25 mm to about 35 mm, about 25

5 mm, about 4.5 mm to about 5.5 mm, or about 5 mm to about 5.5 mm. In some instances, the electric stimulator may comprise a height of about 0.5 mm, about 1 mm, about 1.5 mm, about 2 mm, about 2.5 mm, about 3 mm, about 3.5 mm, about 4 mm, about 4.5 mm, about 5 mm, or about 5.5 mm. In some instances, the electric stimulator may comprise a height of at least about 0.5 mm, about 1 mm, about 1.5 mm, about 2 mm, about 2.5 mm, about 3 mm, about 3.5 mm, about 4 mm, about 4.5 mm, or about 5 mm. In some instances, the electric stimulator may comprise a height of at most about 1 mm, about 1.5 mm, about 2 mm, about 2.5 mm, about 3 mm, about 3.5 mm, about 4 mm, about 4.5 mm, about 5 mm, or about 5.5 mm.

(97) In some instances, the electric stimulator may comprise a mass of about 1 g to about 18 g. In some instances, the electric stimulator may comprise a mass of about 1 g to about 3 g, about 1 g to about 6 g, about 1 g to about 8 g, about 1 g to about 10 g, about 1 g to about 12 g, about 1 g to about 14 g, about 1 g to about 16 g, about 1 g to about 17 g, about 1 g to about 18 g, about 3 g to about 6 g, about 3 g to about 8 g, about 3 g to about 10 g, about 3 g to about 12 g, about 3 g to about 14 g, about 3 g to about 16 g, about 3 g to about 17 g, about 3 g to about 18 g, about 6 g to about 8 g, about 6 g to about 10 g, about 6 g to about 12 g, about 6 g to about 14 g, about 6 g to about 16 g, about 6 g to about 17 g, about 6 g to about 18 g, about 8 g to about 10 g, about 8 g to about 12 g, about 8 g to about 14 g, about 8 g to about 16 g, about 8 g to about 17 g, about 8 g to about 18 g, about 10 g to about 12 g, about 10 g to about 14 g, about 10 g to about 16 g, about 10 g to about 17 g, about 10 g to about 18 g, about 12 g to about 14 g, about 12 g to about 16 g, about 12 g to about 17 g, about 12 g to about 18 g, about 14 g to about 16 g, about 14 g to about 17 g, about 14 g to about 18 g, about 16 g to about 17 g, about 16 g to about 18 g, or about 17 g to about 18 g. In some instances, the electric stimulator may comprise a mass of about 1 g, about 3 g, about 6 g, about 8 g, about 10 g, about 12 g, about 14 g, about 16 g, about 17 g, or about 18 g. In some instances, the electric stimulator may comprise a mass of at least about 1 g, about 3 g, about 6 g, about 8 g, about 10 g, about 12 g, about 14 g, about 16 g, or about 17 g. In some instances, the electric stimulator may comprise a mass of at most about 3 g, about 6 g, about 8 g, about 10 g, about 12 g, about 14 g, about 16 g, about 17 g, or about 18 g.

(98) In some instances, the electric stimulator may be sterilizable with conventional methods of sterilization used in the medical field, e.g., gas sterilization, steam sterilization, UV sterilization, etc.

(99) In some cases, the one or more stimulator electrodes may provide a constant electrical stimulation, or base stimulation **146** and **150** for an individual experiencing urge incontinence. In some cases, constant electrical stimulation may comprise constant frequency, amplitude, current, or any combination thereof. In some instances, the stimulator electrode may provide a temporary electrical stimulation lasting the duration of a stress incontinence **148** episode for an individual experiencing stress incontinence. In some cases, the stimulator electrode may provide a constant electrical stimulation (i.e., basal stimulation pattern) with a temporary activated stimulation (i.e., activation stimulation pattern) lasting the duration of a stress incontinence episode for an individual experiencing mixed incontinence.

(100) In some instances, the device disclosed herein may comprise a non-transitory computer readable medium that includes software. In some cases, the software may be configured to record a signal from the one or more sensor electrodes. In some cases, the software may be configured to process the recorded signal from the one or more sensor electrodes to determine whether an electrical stimulation pattern may need to be delivered to the one or more stimulator electrodes innervating the one or more pudendal nerves. In some instances, the software may be configured to adjust parameters of the sensor electrode in response to the observed signal. In some cases, the recorded signal of the sensor electrode by the software may observe a signal that saturates the sensor electrode's dynamic range. In some cases, the gain of the sensor electrode may be adjusted by the software to allow for sufficient monitoring and thresholding of the myoelectric EMG signals of the individual.

(101) Nerve Stimulation Using the Device

(102) FIG. 3 shows an exemplary embodiment of the devices and methods described herein implanted in an individual. In some cases, the implanted device may target the pudendal nerve **133** by placing the sensor electrodes **134** and **136** and stimulation electrodes near pudendal nerve **130** and **132**. The sensor electrode **134** and **136** may capture the bio-signal to classify any stress events, and the stimulator electrodes **130** and **132** may deliver the stress electrical stimulation (“bio-simulator”), which is adapted from the base stimulation to account for the stress event, to act on the target sphincter muscle.

(103) In some cases, providing electrical stimulation to the pudendal nerve instead of the sacral nerve may provide a greater precision as the pudendal nerve or branches thereof is lower than the sacral nerve and closer to the organs and tissues involved in incontinence than the sacral nerves. The sensor electrodes of the device capture the bio-signal to classify any stress events, and the stimulator electrodes of the device deliver the stress electrical stimulation (“bio-simulator”), which is adapted from the base stimulation to account for the stress event, to act on the target sphincter muscle.

(104) System for Nerve Stimulation for Managing Incontinence

(105) FIG. 5 shows an exemplary embodiment of a system block diagram for the devices and methods described herein with the slow- and fast-adapting algorithms. In some cases, the systems disclosed herein may comprise a plurality of sub-modules. In some instances, the sub-modules may comprise: an offline analysis module **152**, a clinician control module **154**, a patient controller module **156**, an implantable module **160**, or any combination thereof.

(106) In some cases, the offline analysis module **152** may comprise a data repository, an analysis software a visualization software, or any combination thereof. In some instances, the offline analysis module may be used to retrospectively analyze and graphically visualize an individual's implant performance to prevent episodes of incontinence. The off-line analysis module **152** may be programmatically coupled to the clinician control module through an application programming interface (API). In some instances, the clinician control module may comprise software that may tune or change the electrical stimulation patterns of the individual's electrical stimulation implant.

(107) In some cases, the clinical control module **154** may comprise stimulation management and device monitor software, stimulation programming map, classification configuration dashboard, streaming data collection dashboard, or any combination thereof. In some cases, a health care personnel may assist an individual with an electrical implant by updating or modifying their electrical implant parameters through the clinical control module **154**. In some cases, the clinical control module may be utilized to initialize an individual's electrical implant after implantation through an USB interface to the patient controller module **156**.

(108) In some cases, the patient controller module **156** may comprise a direct interface to control aspects of their electrical stimulator as described herein. In some instances, the patient controller module **156** may comprise: medical information and communication band (MICS) communication platform, manual electrical stimulator control, enable or disable algorithm functionality, algorithm patient alerts, an inductive or wired charger for the implantable pulse generator rechargeable battery, or any combination thereof.

(109) In some cases, the patient controller module **156** may be configured to wirelessly and/or inductively charge the implantable pulse generator rechargeable battery with a recharger of the patient controller module **156**. In some cases, the patient controller module **156** may magnetically couple to the implantable pulse generator **160** from outside the subject's skin. In some cases, the magnetic coupling of the controller module **156** to the implantable pulse generator **160** may be made such that the coupling is ergonomic for the subject such that the subject may conduct him/herself as if the controller module **156** is not magnetically coupled to the implantable pulse generator **160**. In some cases, the patient controller module **156** may comprise a battery that may be recharged through wireless inductive charging via the recharger and/or wired charging. In some cases, the patient controller module **156** may comprise a rechargeable lithium-ion battery. In some instances, the recharger may comprise one or more inductive coils used when charging the patient controller module **156** and/or when using the patient controller module **156** to charge the implantable pulse generator **160**. In some cases, the patient controller module, shown as shown in FIG. 7B may be configured **722** to accept additional memory storage **723**. In some cases, the additional memory storage may be used to transport patient data and/or information between patient/subject and provider.

(110) In some cases, the patient controller module **156** may directly or automatically control the implantable pulse generator **160**. In some cases, the implantable pulse generator **160** may comprise a corresponding MICS-telemetry communication platform to that of the MICS communication platform of the patient controller module, enabling the communication between the two devices over an ad hoc Wi-Fi network **158**. In some instances, the implantable pulse generator **160** may further comprise a three-axis accelerometer biopotential amplifier that may be electrically coupled to one or more electrodes leads

214. In some cases the one or more electrode leads may comprise one or more stimulator and/or sensor electrodes. In some cases, a biopotential amplifier may be electrically coupled to a computation sub module comprising a classifier, control policy, real-time clock scheduler, microprocessor, or any combination thereof. In some instances, the biopotential amplifier, classifier, control policy, real-time clock scheduler, and microprocessor, or any combination thereof, may process and interpret detected myoelectric EMG signals in the patient to determine the necessary electrical stimulation pattern provided by the actuator to the one or more stimulator electrodes to prevent an episode of incontinence in an individual.

(111) In some instances, the patient controller module **156** may comprise a user interface, ports, indicators, or any combination thereof as seen in FIGS. **7A-7B**. The user interface and/or ports of the patient controller module may comprise an input charging socket **704**, keypad navigation button **706**, stimulation indicator, communication indicator, output charging adapter port, battery level indicator in both percentage and number of days **712**, manual excitation over-ride button **716**, or any combination thereof. In some cases, the input charging socket may be configured to accept a USB A, B, and/or C, Firewire, any micro versions thereof, or any combinations thereof. connections. In some instances, the patient controller module **156** may be connected to a power converter to through a corresponding cable adapted to the input charging socket to charge the patient controller module **156**.

(112) User Interface

(113) The patient controller module **156** may comprise a user interface **700** where the user interface may comprise one or more user interface objects (**701, 702, 705, 711, 712, 715, 716, 717, 721, 730, 723**), and/or views as seen in FIG. **7A-7B**. In some cases, the user interface **700** may comprise a touch screen display configured to receive touch or pressing input from a user, patient, and/or medical care personnel. In some cases, a user, patient, and/or medical care personnel may press and/or interact with button **716** based mixed graphic and text indicators, and/or switch mixed graphic and text indicators (**705, 711**). In some cases, the user, patient, and/or medical care personnel may double tap a user interface object to enable an emergency state. In some instances, the emergency state may enable the implanted stimulator to provide electrical stimulation immediately in response to the double tap command. In some cases, a parameter or setting value of the user interface objects may be modified and/or changed by tilting the patient controller module. In some cases, tilting the patient controller module in a first direction may increase the parameter and/or setting value of the user interface object whereas tilting the patient controller module in a second direction opposite the first direction may decrease the parameter and/or setting value.

(114) In some cases, the user interface between devices such as smart phones and tablets or other personal computing device may comprise a scaled version of the user interface. In some cases, the different user interface views e.g., the view shown in FIG. **7A** and FIG. **7B** may display varying user interface objects. In some cases, the user interface objects may comprise one or more buttons **716**, switches (**711,716**), and/or graphical or image based representation of data (**721,730, 723**).

(115) In some cases, users may customize the user interface object with a selection of one or more user interface objects (e.g., buttons, switch button to enable various device operation modes, graphical displays of device data, etc.). In some cases, the user views may be a predetermine set of views with set user interface object. In some instances, the user may customize and/or create one or more views accessible by the a menu icon **702**. In some cases, the menu icon **702** may be configured to display one or more submenu options. In some instances, the one or more submenu options may comprise personal identification, account information, device registration, customer support, or any combination thereof submenus. In some cases, one submenu may comprise information of how to connect the device platform to pre-existing health care providers. In some cases, the user interface may comprise a notification object **701**. The notification object may display a unique or highlighted state if a particular notification of device performance, detection of an incontinence event, or any combination thereof is to be provided to the user of the device. A user may interact with a press the notification object to view, in the form of a pop-up dialogue, the particular notification.

(116) In some cases, the one or more user interface objects may comprise text and/or mixed text and vector objects representations of the various API function calls and/or sub-user interface views, as seen in FIGS. **7A-7B**. In some cases, the user interface may comprise mixed text and vector objects that

permit the subject or user to activate **705** or de-activate **711** electrical stimulation of the device **716**, adjust device parameters **715**, indicate therapy state **717**, view device measured EMG signals **721**, view stimulator electrode electrical signal characteristics (e.g., frequency, amplitude, pulse width, etc.) delivered, view a medical portal to submit user data to a health care provider, log resulting incontinent events **719** overlaid on top of measured ENG/EMG signals, or any combination thereof. In some cases, the user interface may further comprise a battery **712** and wireless communication connectivity indicator for the users and/or subjects to visualize patient controller module **156** operating properties.

(117) In some cases, the keypad navigation button **706** may be configured to navigate between various user interface views e.g., the user interface views provided in FIG. 7A and FIG. 7B

(118) In some instances, the patient controller module **156** may comprise visual indicators (**715**, **717**, **712**), configured to indicate whether the implanted electrical stimulator is outputting electrical stimulation and/or the presence or lack thereof connectivity with a second or third device. In some cases, the patient controller module may comprise a device adjustment parameter, where the device adjustment parameter may comprise a stimulation indicator, or an activation of a stimulation mode **715**. In some cases, the stimulation indicator may be in electrical communication with a processor, described elsewhere herein, configured to display a visual indicator when the stimulator is providing an electrical stimulation to a subject. In some instances, the patient controller module may comprise a connectivity indicator. In some cases, the connectivity indicator may be in electrical communication with a processor, described elsewhere herein, and configured to provide a visual indicator when the patient controller module is connected to one or more discrete devices, data servers, local WIFI or ad-hoc WIFI networks, Bluetooth, medical implant communication system (MICS), or any combination thereof. In some cases, the connectivity indicator may indicate the wireless connection with the implanted electrical stimulator. In some cases, the connectivity indicator may comprise one or more states. In some instances, a first state may comprise a solid image indicator, where such a solid image indicator may notify a user, subject, individual, and/or health care personnel, a successfully established communication pairing between the patient controller module and a third device, server, etc. A second state may comprise a flashing image indicator, where such a flashing image indicator indicates a paired communication state between the patient controller module and a third device, server, etc. In some cases, the image indicator may comprise the universal symbol for Bluetooth that may be observed on smart devices and/or devices with Bluetooth connectivity. In some cases, the image indicator may comprise a graphic of the universal symbol indicator for Wi-Fi (e.g., a quarter circle of concentric rings) seen commonly on smart devices and/or devices with Wi-Fi connectivity.

(119) In some cases, device data (e.g., EMG/ENG, accelerometer, gyroscopic, magnetometer, 3-D spatial movement, global positioning system (GPS) data, or any combination thereof) may be transmitted **718** over Wi-Fi, Bluetooth, MICS, or other ad-hoc networks between one or more devices, as described elsewhere herein.

(120) FIG. 7B shows a different user interface view than that of the FIG. 7A. In some cases, the user interface of FIG. 7B may comprise one or more user interface objects (**721**, **730**, **723**), where each user interface object displays device data received **718** through wireless transmission **707** as described above. In some cases, one of the user interface objects may comprise a graphical therapy object **721**. The graphical therapy object may display detected EMG/ENG signals **726** and corresponding stimulation profiles **728**. In some cases, the graphical therapy object may display leak events **719** of where the user indicated an incontinent event but where the device did not provide stimulation. Another user interface object may comprise a GPS and motion activity object **730**. In some instances, the motion activity object **730** may display GPS and motion data of the subject over time. The GPS and motion data may be useful considerations when improving the classifier described elsewhere herein. Another user interface object may comprise a charging indicator object **723**. In some cases, the charging indicator object may display charge capacitance of the implanted stimulator over a period of time. In some cases, the charging indicator object may be used to monitor the health of the battery of the implanted stimulator. A user interacting with the display view shown in FIG. 7B may pinch, swipe, or otherwise interact with the data of each user interface object (**721**, **730**, **723**) to view other temporal regions of data or to zoom in on a particular scale of a measurement. In some cases, through the menu object **702** a user and/or subject may export their medical data to one or more provides.

(121) Method of Preventing an Incontinence Episode

(122) FIG. 6A illustrates a workflow of a method **216** for preventing an episode of incontinence in an individual sufferer. The method may comprise the steps of (a) implanting a sensor electrode and stimulator electrode within a body of an individual **218**; (b) sensing with the sensor electrode, a parameter that is associated with a response from the individual to prevent an episode of incontinence **220**; and (c) providing an electrical stimulation, with the stimulator electrode, that, together with the response, prevents the episode of incontinence **222**. Often, the methods described herein prevent the episode of urinary incontinence. In some cases, the urinary incontinence may comprise at least one of urge incontinence, stress incontinence, overflow incontinence, or mixed incontinence. In some instances, the methods described herein may prevent the episode of fecal incontinence. In some cases, the method may comprise a step of providing a constant electrical stimulation at a lower intensity level than the electrical stimulation provided in step (c). The constant electrical stimulation provided at a lower intensity may assist individual's suffering from urge incontinence. In some instances, the intensity or duration of the electrical stimulation provided in step (c) may vary according to the response that is sensed in step (b). The response may vary in step (c) due to stress events such as coughing, laughing, or exercising that may require an increase in electrical stimulation to prevent an episode of incontinence. In some cases, the time period between sensing a parameter **224** and providing an electrical stimulation **226** may be described as a response time **228**, as shown in FIG. 6B. In some cases, the stimulation may be provided for a duration of stimulation **227**.

(123) In some cases, the duration of stimulation **227** may comprise about 1 second to about 30 seconds. In some cases, the duration of stimulation **227** may comprise about 1 second to about 2 seconds, about 1 second to about 3 seconds, about 1 second to about 4 seconds, about 1 second to about 5 seconds, about 1 second to about 10 seconds, about 1 second to about 12 seconds, about 1 second to about 14 seconds, about 1 second to about 16 seconds, about 1 second to about 20 seconds, about 1 second to about 25 seconds, about 1 second to about 30 seconds, about 2 seconds to about 3 seconds, about 2 seconds to about 4 seconds, about 2 seconds to about 5 seconds, about 2 seconds to about 10 seconds, about 2 seconds to about 12 seconds, about 2 seconds to about 14 seconds, about 2 seconds to about 16 seconds, about 2 seconds to about 20 seconds, about 2 seconds to about 25 seconds, about 2 seconds to about 30 seconds, about 3 seconds to about 4 seconds, about 3 seconds to about 5 seconds, about 3 seconds to about 10 seconds, about 3 seconds to about 12 seconds, about 3 seconds to about 14 seconds, about 3 seconds to about 16 seconds, about 3 seconds to about 20 seconds, about 3 seconds to about 25 seconds, about 3 seconds to about 30 seconds, about 4 seconds to about 5 seconds, about 4 seconds to about 10 seconds, about 4 seconds to about 12 seconds, about 4 seconds to about 14 seconds, about 4 seconds to about 16 seconds, about 4 seconds to about 20 seconds, about 4 seconds to about 25 seconds, about 4 seconds to about 30 seconds, about 5 seconds to about 10 seconds, about 5 seconds to about 12 seconds, about 5 seconds to about 14 seconds, about 5 seconds to about 16 seconds, about 5 seconds to about 20 seconds, about 5 seconds to about 25 seconds, about 5 seconds to about 30 seconds, about 10 seconds to about 12 seconds, about 10 seconds to about 14 seconds, about 10 seconds to about 16 seconds, about 10 seconds to about 20 seconds, about 10 seconds to about 25 seconds, about 10 seconds to about 30 seconds, about 12 seconds to about 14 seconds, about 12 seconds to about 16 seconds, about 12 seconds to about 20 seconds, about 12 seconds to about 25 seconds, about 12 seconds to about 30 seconds, about 14 seconds to about 16 seconds, about 14 seconds to about 20 seconds, about 14 seconds to about 25 seconds, about 14 seconds to about 30 seconds, about 16 seconds to about 20 seconds, about 16 seconds to about 25 seconds, about 16 seconds to about 30 seconds, about 20 seconds to about 25 seconds, about 20 seconds to about 30 seconds, or about 25 seconds to about 30 seconds. In some cases, the duration of stimulation **227** may comprise about 1 second, about 2 seconds, about 3 seconds, about 4 seconds, about 5 seconds, about 10 seconds, about 12 seconds, about 14 seconds, about 16 seconds, about 20 seconds, about 25 seconds, or about 30 seconds. In some cases, the duration of stimulation **227** may comprise at least about 1 second, about 2 seconds, about 3 seconds, about 4 seconds, about 5 seconds, about 10 seconds, about 12 seconds, about 14 seconds, about 16 seconds, about 20 seconds, or about 25 seconds. In some cases, the duration of stimulation **227** may comprise at most about 2 seconds, about 3 seconds, about 4 seconds, about 5 seconds, about 10 seconds, about 12 seconds, about 14 seconds, about 16 seconds, about 20 seconds, about 25 seconds, or about 30 seconds.

(124) In some cases the response time **228** may comprise about 60 μ s to about 100 μ s. In some cases the response time **228** may comprise about 60 μ s to about 65 μ s, about 60 μ s to about 70 μ s, about 60 μ s to about 75 μ s, about 60 μ s to about 80 μ s, about 60 μ s to about 85 μ s, about 60 μ s to about 90 μ s, about 60 μ s to about 95 μ s, about 60 μ s to about 100 μ s, about 65 μ s to about 70 μ s, about 65 μ s to about 75 μ s, about 65 μ s to about 80 μ s, about 65 μ s to about 85 μ s, about 65 μ s to about 90 μ s, about 65 μ s to about 95 μ s, about 65 μ s to about 100 μ s, about 70 μ s to about 75 μ s, about 70 μ s to about 80 μ s, about 70 μ s to about 85 μ s, about 70 μ s to about 90 μ s, about 70 μ s to about 95 μ s, about 70 μ s to about 100 μ s, about 75 μ s to about 80 μ s, about 75 μ s to about 85 μ s, about 75 μ s to about 90 μ s, about 75 μ s to about 95 μ s, about 75 μ s to about 100 μ s, about 80 μ s to about 85 μ s, about 80 μ s to about 90 μ s, about 80 μ s to about 95 μ s, about 80 μ s to about 100 μ s, about 85 μ s to about 90 μ s, about 85 μ s to about 95 μ s, about 85 μ s to about 100 μ s, about 90 μ s to about 95 μ s, about 90 μ s to about 100 μ s, or about 95 μ s to about 100 μ s. In some cases the response time **228** may comprise about 60 μ s, about 65 μ s, about 70 μ s, about 75 μ s, about 80 μ s, about 85 μ s, about 90 μ s, about 95 μ s, or about 100 μ s. In some cases the response time **228** may comprise at least about 60 μ s, about 65 μ s, about 70 μ s, about 75 μ s, about 80 μ s, about 85 μ s, about 90 μ s, or about 95 μ s. In some cases the response time **228** may comprise at most about 65 μ s, about 70 μ s, about 75 μ s, about 80 μ s, about 85 μ s, about 90 μ s, about 95 μ s, or about 100 μ s. Through the development of iteratively trained machine learning classifiers the response time may be minimized, and improved incontinence prevention may be realized.

(125) FIGS. **12A-12B** illustrate a workflow of a method **1200** for preventing an episode of incontinence in an individual sufferer. The method may comprise the steps of (a) implanting a sensor electrode and stimulator electrode within a body of an individual **1202**; (b) sensing a parameter associated with a response from the individual, where the response is user induced stimulus **1204**; and (c) providing an electric stimulation with the stimulator electrode to prevent the incontinence event **1206**. In some cases, the user induced stimulus may be intended to prevent an episode of incontinence. As shown graphically in FIG. **12B**, a detected EMG signal **1208** may be analyzed by a classifier **1210**, described elsewhere herein, to determine when the EMG, ENG, accelerometer, gyroscope, magnetometer, pressure sensor signals or any combination thereof signals have crossed a pre-determined threshold, described elsewhere herein. Once the classifier **1210** has determined that the signal **1208** does represent an incontinent event, the processor, described elsewhere herein, may enable stimulation **1212** to prevent the incontinent event from occurring. In some cases, stimulation **1212** may comprise an extension of stimulation **1211** that may extend beyond the time the classifier **1210** determines there to be an incontinent event. In some cases the extension of stimulation **1211** may comprise a duration of time equal to the duration the subject continues to purposefully or with intent produce a muscle contraction.

(126) In some cases the extension of stimulation **1211** may comprise about 1 s to about 30 s. In some cases the extension of stimulation **1211** may comprise about 1 s to about 3 s, about 1 s to about 5 s, about 1 s to about 8 s, about 1 s to about 10 s, about 1 s to about 12 s, about 1 s to about 15 s, about 1 s to about 18 s, about 1 s to about 20 s, about 1 s to about 22 s, about 1 s to about 24 s, about 1 s to about 30 s, about 3 s to about 5 s, about 3 s to about 8 s, about 3 s to about 10 s, about 3 s to about 12 s, about 3 s to about 15 s, about 3 s to about 18 s, about 3 s to about 20 s, about 3 s to about 22 s, about 3 s to about 24 s, about 3 s to about 30 s, about 5 s to about 8 s, about 5 s to about 10 s, about 5 s to about 12 s, about 5 s to about 15 s, about 5 s to about 18 s, about 5 s to about 20 s, about 5 s to about 22 s, about 5 s to about 24 s, about 5 s to about 30 s, about 8 s to about 10 s, about 8 s to about 12 s, about 8 s to about 15 s, about 8 s to about 18 s, about 8 s to about 20 s, about 8 s to about 22 s, about 8 s to about 24 s, about 8 s to about 30 s, about 10 s to about 12 s, about 10 s to about 15 s, about 10 s to about 18 s, about 10 s to about 20 s, about 10 s to about 22 s, about 10 s to about 24 s, about 10 s to about 30 s, about 12 s to about 15 s, about 12 s to about 18 s, about 12 s to about 20 s, about 12 s to about 22 s, about 12 s to about 24 s, about 12 s to about 30 s, about 15 s to about 18 s, about 15 s to about 20 s, about 15 s to about 22 s, about 15 s to about 24 s, about 15 s to about 30 s, about 18 s to about 20 s, about 18 s to about 22 s, about 18 s to about 24 s, about 18 s to about 30 s, about 20 s to about 22 s, about 20 s to about 24 s, about 20 s to about 30 s, about 22 s to about 24 s, about 22 s to about 30 s, or about 24 s to about 30 s. In some cases the extension of stimulation **1211** may comprise about 1 s, about 3 s, about 5 s, about 8 s, about 10 s, about 12 s, about 15 s, about 18 s, about 20 s, about 22 s, about 24 s, or about 30 s. In some cases the extension of stimulation **1211** may comprise at least about 1 s, about 3 s, about 5 s,

about 8 s, about 10 s, about 12 s, about 15 s, about 18 s, about 20 s, about 22 s, or about 24 s. In some cases the extension of stimulation **1211** may comprise at most about 3 s, about 5 s, about 8 s, about 10 s, about 12 s, about 15 s, about 18 s, about 20 s, about 22 s, about 24 s, or about 30 s.

(127) In some instances, there may be a delay **1209** between the onset of the incontinent event in the raw EMG, ENG, accelerometer, gyroscope, magnetometer, pressure sensor, or any combination thereof signal data, and the onset of the electrical stimulation. Upon training the classifier **1210** on sufficiently large and varied datasets, such delay may be minimized further improving the device performance in prevent incontinent events. In some cases, sensing may comprise determining a global positioning system (GPS) location of the individual that in combination with the parameter associated with the response from the individual is used to prevent the episode of incontinence.

(128) Aspects of the disclosure provided herein may comprise a method of data processing. In some cases, the method of data processing may comprise: (i) receiving a measurement of a parameter previously measured by a sensor electrode, which parameter is predictive of an episode of incontinence in an individual; (ii) analyzing the parameter; and (iii) synthesizing an electrical stimulation signal for the individual, such that when the electrical stimulation signal is provided by a stimulator electrode to the individual, the electrical stimulation signal, together with an effort from the individual that is intended to prevent an episode of incontinence, prevents the episode of incontinence. In some cases, the parameter may be associated with a response from the individual intended to prevent an episode of incontinence. In some instances, the parameter may be associated with the individual's effort in trying to prevent an episode of incontinence, and where the electrical stimulation signal is synthesized so as to supplement the individual's effort with an electrical stimulation pattern that will, together with the effort from the individual, be sufficient to prevent an episode of incontinence. In some cases, the response from the individual may be insufficient on its own to prevent the episode of incontinence and the electrical stimulation signal is such that, when applied, it adds enough, together with the response, to prevent the episode of incontinence.

(129) In some instances, the episode of incontinence may comprise urinary incontinence. In some cases, the episode of incontinence may comprise fecal incontinence. In some cases, the episode of incontinence may comprise urinary stress incontinence. In some instances, the episode of incontinence is urinary incontinence and is urge incontinence type.

(130) In some instances, the parameter may comprise a signal from a sensor electrode that is configured to sense a contraction of a muscle of the individual related to a partial contraction of a sphincter that controls bladder or bowel voiding. In some cases, the electrical stimulation signal may comprise an electrical stimulation of the pudendal nerve. In some instances, the electrical stimulation signal may be synthesized to include a constant electrical stimulation component and a measurement parameter specific component. In some cases, an intensity or duration of the electrical stimulation that will be provided by the electrical stimulation signal may vary according to the value of the parameter that is received. In some instances, the parameter may comprise an EMG signal. In some cases, the EMG signal may determine that a contraction of at least one pelvic muscle has occurred. In some cases, a strength of the EMG signal may be proportional to the strength of the contraction of at least one pelvic muscle. In some instances, the electrical stimulation signal may comprise a first and second signals for the stimulation of a first pudendal nerve and a second pudendal nerve respectively. In some cases, the method may further comprise recording the signal previously measured by the sensor electrode. In some cases, the method may further comprise synthesizing an adjustment signal to adjust the sensor electrode in response to the recorded signal.

(131) Assessment of Treatment of Incontinence

(132) Ambulatory Assessments after Procedure

(133) In some embodiments, various ambulatory assessments may be taken to determine the effectiveness of the implantation procedure. In some embodiments, the implanted IPG permits telemetric downloading of data (inputs, outputs, and event classification). In some embodiments, the participant may be in an awake ambulatory setting and a series of resting and provoked electrophysiological data may be recorded. In some embodiments, at treatment initiation (24-48 hours post-implant) sensory and motor responses may be determined from the different sensor electrodes on the implanted leads. In some embodiments, based upon the responses, the electrodes with the most adequate response may be selected

to initiate treatment.

(134) In some embodiments, the patients may be subjected to different physiological events to program the IPG. In some embodiments, these events may comprise coughing, Valsalva maneuvers, picking up a 5 kg weight, or any combination thereof. In some embodiments, pelvic floor EMG may be measured with a transvaginal and/or anal probe. In some embodiments urethral pressures may be measured. In some embodiments, 1 hour continuous ‘resting’ recording of inputs and outputs (downloaded by telemetry) may be taken. In some embodiments, recording during controlled participant provoked events, such as coughing, Valsalva, lifting 5 Kg weight, may be obtained. In some embodiments, recording during pelvic floor surface EMG (from transvaginal probe: women only) to correlate inputs from lead vs. surface EMG may be obtained. In some embodiments, patient tolerances of basal stimulation ramping, and actuation parameters may be obtained. In some embodiments, standard urodynamic tests are performed at 48 hours. In some embodiments, UDCs (with or without reporting of urge) may be recorded during bladder filling to assess the acute effect of patient-actuation of device. In some embodiments, a standard 1 hour pad test may be performed.

(135) Clinical Outcomes

(136) In some embodiments, clinical outcomes may be assessed using a 5-day voiding diary recording number of voids, number of urgency episodes, number of leaks with severity of leaks to derive: stress and urge UI (summative) episodes per unit time; stress UI episodes per unit time; urge UI episodes per unit time; urgency to void episodes per unit time; total voiding frequency per unit time; responder rate: based on >50% decrease in UI episodes per unit time; functional cure rate defined as either >90% decrease in mean total UI episodes from baseline OR mean <1 UIE per week; ICIQ-SF-UI questionnaire; or any combination thereof.

(137) Definitions

(138) Unless defined otherwise, all terms of art, notations and other technical and scientific terms or terminology used herein are intended to have the same meaning as is commonly understood by one of ordinary skill in the art to which the claimed subject matter pertains. In some cases, terms with commonly understood meanings are defined herein for clarity and/or for ready reference, and the inclusion of such definitions herein should not necessarily be construed to represent a substantial difference over what is generally understood in the art.

(139) Throughout this application, various embodiments may be presented in a range format. It should be understood that the description in range format is merely for convenience and brevity and should not be construed as an inflexible limitation on the scope of the disclosure. Accordingly, the description of a range should be considered to have specifically disclosed all the possible subranges as well as individual numerical values within that range. For example, description of a range such as from 1 to 6 should be considered to have specifically disclosed subranges such as from 1 to 3, from 1 to 4, from 1 to 5, from 2 to 4, from 2 to 6, from 3 to 6 etc., as well as individual numbers within that range, for example, 1, 2, 3, 4, 5, and 6. This applies regardless of the breadth of the range.

(140) As used in the specification and claims, the singular forms “a”, “an” and “the” include plural references unless the context clearly dictates otherwise. For example, the term “a sample” includes a plurality of samples, including mixtures thereof.

(141) The terms “determining”, “measuring”, “evaluating”, “assessing,” “assaying,” and “analyzing” are often used interchangeably herein to refer to forms of measurement and include determining if an element is present or not (for example, detection). These terms can include quantitative, qualitative, or quantitative and qualitative determinations. Assessing is alternatively relative or absolute. “Detecting the presence of” includes determining the amount of something present, as well as determining whether it is present or absent.

(142) The terms “subject,” “individual,” or “patient” are often used interchangeably herein. A “subject” can be a biological entity containing expressed genetic materials. The biological entity can be an animal. The subject can be a mammal. The mammal can be a human. The subject may be diagnosed or suspected of being at high risk for a disease or a condition. The disease can be incontinence. In some cases, the disease can be urinary incontinence. In some cases, the disease is bowel incontinence. In some cases, the subject is not necessarily diagnosed or suspected of being at high risk for the disease.

(143) The term “in vivo” is used to describe an event that takes place in a subject's body.

(144) The term “ex vivo” is used to describe an event that takes place outside of a subject's body. An “ex vivo” assay is not performed on a subject. Rather, it is performed upon a sample separate from a subject. An example of an “ex vivo” assay performed on a sample is an “in vitro” assay.

(145) As used herein, the term ‘about’ a number refers to that number plus or minus 10% of that number. The term ‘about’ a range refers to that range minus 10% of its lowest value and plus 10% of its greatest value.

(146) As used herein, the terms “treatment” or “treating” are used in reference to an intervention regimen for obtaining beneficial or desired results in the recipient. Beneficial or desired results include but are not limited to a therapeutic benefit and/or a prophylactic benefit. A therapeutic benefit may refer to prevention or amelioration of symptoms or of an underlying disorder being treated. Also, a therapeutic benefit can be achieved with prevention or amelioration of one or more of the physiological symptoms associated with the underlying disorder such that an improvement is observed in the subject, notwithstanding that the subject may still be afflicted with the underlying disorder. A prophylactic effect includes delaying, preventing, or eliminating the appearance of a disease or a condition, delaying or eliminating the onset of symptoms of a disease or a condition, slowing, halting, or reversing the progression of a disease or a condition, or any combination thereof. For prophylactic benefit, a subject at risk of developing a particular disease or a condition, or to a subject reporting one or more of the physiological symptoms of a disease or a condition may undergo treatment.

(147) The section headings used herein are for organizational purposes only and are not to be construed as limiting the subject matter described.

(148) While preferred embodiments of the present invention have been shown and described herein, it will be obvious to those skilled in the art that such embodiments are provided by way of example only. Numerous variations, changes, and substitutions will now occur to those skilled in the art without departing from the invention. It should be understood that various alternatives to the embodiments of the invention described herein may be employed in practicing the invention. It is intended that the following claims define the scope of the invention and that methods and structures within the scope of these claims and their equivalents be covered thereby.

NUMERATED EMBODIMENTS

(149) Numbered embodiment 1 comprises a method for preventing an episode of incontinence in an individual in need thereof, the method comprising: (a) implanting a sensor and stimulator electrode within a body of the individual; (b) sensing, with the sensor electrode, a parameter that is associated with a response from the individual that is intended to prevent an episode of incontinence; and (c) providing an electrical stimulation, with the stimulator electrode, that, together with the response, prevents the episode of incontinence. Numbered embodiment 2 comprises the method of embodiment 1, wherein the episode of incontinence comprises urinary incontinence. Numbered embodiment 3 comprises the method of embodiment 1, wherein the episode of incontinence comprises fecal incontinence. Numbered embodiment 4 comprises the method of embodiment 1, wherein the episode of incontinence comprises urinary stress incontinence. Numbered embodiment 5 comprises the method of embodiment 1, wherein the sensor electrode is configured to sense a contraction of a muscle of the individual that results in a partial contraction of a sphincter that controls bladder or bowel voiding. Numbered embodiment 6 comprises the method of embodiment 5, wherein the sensor electrode is positioned within the pelvis of the individual. Numbered embodiment 7 comprises the method of embodiment 1, wherein the stimulator electrode provides an electrical stimulation to the pudendal nerve. Numbered embodiment 8 comprises the method of embodiment 7, wherein the sensor and the stimulator electrode are located on a single lead. Numbered embodiment 9 comprises the method of embodiment 1, comprising a step of providing a constant electrical stimulation at a lower intensity level than the electrical stimulation provided in step (c). Numbered embodiment 10 comprises the method of embodiment 9, wherein the episode of incontinence is urinary incontinence and is urge incontinence type. Numbered embodiment 11 comprises the method of embodiment 1, wherein the intensity or duration of the electrical stimulation provided in step (c) varies according to the response that is sensed in step (b). Numbered embodiment 12 comprises the method of embodiment 11, wherein the response that is sensed in step (b) is insufficient on its own to prevent the episode of incontinence and the electrical stimulation provided in step (b) adds just enough, together with the response, to prevent the episode of incontinence. Numbered embodiment 13 comprises

the method of embodiment 11, wherein the response in step (b) is insufficient on its own to prevent the episode of incontinence, and wherein the electrical stimulation provided in step (b) together with the response, prevents the episode of incontinence. Numbered embodiment 14 comprises the method of embodiment 1, wherein the sensor electrode is configured to sense an EMG signal. Numbered embodiment 15 comprises the method of embodiment 14, wherein the EMG signal determines that a contraction of at least one pelvic muscle has occurred. Numbered embodiment 16 comprises the method of embodiment 15, wherein a strength of the EMG signal is proportional to a strength of the contraction of at least one pelvic muscle. Numbered embodiment 17 comprises the method of embodiment 1, comprising a step of implanting a first stimulator electrode and second stimulator electrode, wherein the first stimulator electrode stimulates one region on a pudendal nerve and the second stimulator electrode stimulates a different region on the pudendal nerve. Numbered embodiment 18 comprises the method of embodiment 1, wherein the individual suffers from urinary incontinence of a mixed type. Numbered embodiment 19 comprises the method of embodiment 1, wherein the sensor and the stimulator electrode are operatively coupled to a processor and a non-transitory computer readable medium that includes software. Numbered embodiment 20 comprises the method of embodiment 19, wherein the sensor electrode is calibrated by the individual using an external input device that interfaces with the software. Numbered embodiment 21 comprises the method of embodiment 19, wherein the software is configured to record a signal from the sensor electrode. Numbered embodiment 22 comprises the method of embodiment 21, wherein the software is configured to adjust the sensor electrode's response to the signal. Numbered embodiment 23 comprises the method of embodiment 19, wherein the software comprises a machine learning model, and wherein the machine learning model is configured to classify signals detected by the sensor electrode and generate signals with the stimulator electrode. Numbered embodiment 24 comprises the method of embodiment 23, wherein the machine learning model is trained on prior data acquired from the individual or a set of individuals and the corresponding incontinence prevention or lack thereof information. Numbered embodiment 25 comprises the method of embodiment 24, wherein the prior data comprises the signals detected by the sensor electrode, the generated signals generated by the stimulator electrode, or any combination thereof. Numbered embodiment 26 comprises the method of embodiment 1, wherein sensing further comprises determining a global positioning system (GPS) location of the individual that in combination with the parameter associated with the response from the individual prevents the episode of incontinence.

(150) Numbered embodiment 27 comprises a system for preventing an episode of incontinence in an individual in need thereof, the apparatus comprising: (a) a sensor electrode configured to sense a parameter that is associated with a response from the individual that is intended to prevent the episode of incontinence; (b) a stimulator electrode configured to provide electrical stimulation; (c) a processor operably coupled to the sensor and stimulator electrode; and (d) a non-transitory computer readable storage medium including software configured to cause the processor to: (i) receive the parameter that is associated with the response from the individual that is intended to prevent the episode of incontinence; (ii) analyze the parameter that is associated with the response from the individual that is intended to prevent the episode of incontinence; and (iii) cause the stimulator electrode to provide the electrical stimulation to the individual such that the electrical stimulation together with the response from the individual that is intended to prevent the episode of incontinence prevents the episode of incontinence. Numbered embodiment 28 comprises the system of embodiment 27, wherein the episode of incontinence comprises urinary incontinence. Numbered embodiment 29 comprises the system of embodiment 27, wherein the episode of incontinence comprises fecal incontinence. Numbered embodiment 30 comprises the system of embodiment 27, wherein the episode of incontinence comprises urinary stress incontinence. Numbered embodiment 31 comprises the system of embodiment 27, wherein the sensor electrode is configured to sense a contraction of a muscle of the individual that results in a partial contraction of a sphincter that controls bladder or bowel voiding. Numbered embodiment 32 comprises the system of embodiment 27, wherein the sensor electrode is positioned within the pelvis of the individual. Numbered embodiment 33 comprises the system of embodiment 27, wherein the stimulator electrode provides the electrical stimulation to the pudendal nerve of the individual. Numbered embodiment 34 comprises the system of embodiment 27, wherein the sensor and stimulator electrodes are located on a single lead. Numbered embodiment 35 comprises the system of embodiment 27, wherein

the stimulator electrode is configured to provide an electrical stimulation at a lower intensity level than the electrical stimulation. Numbered embodiment 36 comprises the system of embodiment 27, wherein the episode of incontinence is urinary incontinence and is urge incontinence type. Numbered embodiment 37 comprises the system of embodiment 27, wherein an intensity or duration of the electrical stimulation varies according to the response that is sensed. Numbered embodiment 38 comprises the system of embodiment 27, wherein the response that is sensed by the sensor electrode is insufficient on its own to prevent the episode of incontinence and the electrical stimulation provided adds just enough, together with the response, to prevent the episode of incontinence. Numbered embodiment 39 comprises the system of embodiment 27, wherein the response that is sensed by the sensor electrode is insufficient on its own to prevent the episode of incontinence, and wherein the electrical stimulation provided in step (b), together with the response, prevents the episode of incontinence. Numbered embodiment 40, the system of embodiment 27, wherein the sensor electrode is configured to sense an EMG signal. Numbered embodiment 41 comprises the system of embodiment 40 wherein the EMG signal determines that a contraction of at least one pelvic muscle has occurred. Numbered embodiment 42 comprises the system of embodiment 41, wherein a strength of the EMG signal is proportional to a strength of the contraction of at least one pelvic muscle. Numbered embodiment 43 comprises the system of embodiment 27, wherein the stimulator electrode comprises a first stimulator electrode and a second stimulator electrode, wherein the first stimulator electrode stimulates a first pudendal nerve and the second stimulator electrode stimulates a second pudendal nerve. Numbered embodiment 44 comprises the system of embodiment 27, wherein the individual suffers from urinary incontinence of a mixed type. Numbered embodiment 45 comprises the system of embodiment 27, wherein the sensor electrode is calibrated by the individual using an external input device that interfaces with the software. Numbered embodiment 46 comprises the system of embodiment 27, wherein the software is configured to further cause the processor to record a signal from the sensor electrode. Numbered embodiment 47 comprises the system of embodiment 46, wherein the software is configured to adjust the sensor electrode's response to the signal. Numbered embodiment 48 comprises the system of embodiment 27, wherein the software comprises a machine learning model, and wherein the machine learning model is configured to classify signals detected by the sensor electrode and generate signals with the stimulator electrode. Numbered embodiment 49 comprises the system of embodiment 48, wherein the machine learning model is trained on prior data acquired from the individual or a set of individuals and the corresponding incontinence prevention or lack thereof information. Numbered embodiment 50 comprises the system of embodiment 49, wherein the prior data comprises the signals detected by the sensor electrode, the generated signals generated by the stimulator electrode, or any combination thereof. Numbered embodiment 51 comprises the system of embodiment 27, wherein the software comprises analyzing a global positioning system (GPS) location of the individual that in combination with the parameter associated with the response from the individual prevents the episode of incontinence.

(151) Numbered embodiment 52 comprises a non-transitory computer readable storage medium including software for preventing an episode of incontinence in an individual in need thereof, configured to cause a processor to: (i) receive a parameter by a sensor electrode that is associated with a response from the individual intended to prevent the episode of incontinence; (ii) analyze the parameter that is associated with the response from the individual that is intended to prevent the episode of incontinence; and (iii) cause a stimulator electrode to provide an electrical stimulation to the individual so that the electrical stimulation together with the response from the individual that is intended to prevent the episode of incontinence prevents the episode of incontinence. Numbered embodiment 53 comprises the non-transitory computer readable storage medium including software of embodiment 52, wherein the episode of incontinence comprises urinary incontinence. Numbered embodiment 54 comprises the non-transitory computer readable storage medium including software of embodiment 52, wherein the episode of incontinence comprises fecal incontinence. Numbered embodiment 55 comprises the non-transitory computer readable storage medium including software of embodiment 52, wherein the episode of incontinence comprises urinary stress incontinence. Numbered embodiment 56 comprises the non-transitory computer readable storage medium including software of embodiment 52, wherein the sensor electrode is configured to sense a contraction of a muscle of the individual that results in a partial contraction of a sphincter that controls bladder or bowel voiding. Numbered embodiment 57 comprises

the non-transitory computer readable storage medium including software of embodiment 52, wherein the sensor electrode is positioned within the pelvis of the individual. Numbered embodiment 58 comprises the non-transitory computer readable storage medium including software of embodiment 52, wherein the stimulator electrode provides the electrical stimulation to the individual's pudendal nerve. Numbered embodiment 59 comprises the non-transitory computer readable storage medium including software of embodiment 52, wherein the sensor and stimulator electrodes are located on a single lead. Numbered embodiment 60 comprises the non-transitory computer readable storage medium including software of embodiment 52, wherein the stimulator electrode is configured to provide a constant electrical stimulation at a lower intensity level than the electrical stimulation. Numbered embodiment 61 comprises the non-transitory computer readable storage medium including software of embodiment 52, wherein the episode of incontinence is urinary incontinence and is urge incontinence type. Numbered embodiment 62 comprises the non-transitory computer readable storage medium including software of embodiment 52, wherein an intensity or duration of the electrical stimulation varies according to the response that is sensed. Numbered embodiment 63 comprises the non-transitory computer readable storage medium including software of embodiment 62, wherein the response that is sensed by the sensor electrode is insufficient on its own to prevent the episode of incontinence and the electrical stimulation provided adds just enough, together with the response, to prevent the episode of incontinence. Numbered embodiment 64 comprises the non-transitory computer readable storage medium including software of embodiment 62, wherein the response that is sensed by the sensor is insufficient on its own to prevent the episode of incontinence and the electrical stimulation provided together with the response, prevents the episode of incontinence. Numbered embodiment 65 comprises the non-transitory computer readable storage medium including software of embodiment 52, wherein the sensor electrode is configured to sense an EMG signal. Numbered embodiment 66 comprises the non-transitory computer readable storage medium including software of embodiment 65, wherein the EMG signal determines that a contraction of at least one pelvic muscle has occurred. Numbered embodiment 67 comprises the non-transitory computer readable storage medium including software of embodiment 66, wherein a strength of the EMG signal is proportional to a strength of the contraction of at least one pelvic muscle. Numbered embodiment 68 comprises the non-transitory computer readable storage medium including software of embodiment 52, wherein the stimulator electrode comprises a first stimulator electrode and a second stimulator electrode, wherein the first stimulator electrode stimulates a first pudendal nerve and the second stimulator electrode stimulates a second pudendal nerve. Numbered embodiment 69 comprises the non-transitory computer readable storage medium including software of embodiment 52, wherein the individual suffers from urinary incontinence of a mixed type. Numbered embodiment 70 comprises the non-transitory computer readable storage medium including software of embodiment 52, wherein the sensor electrode is calibrated by the individual using an external input device that interfaces with the software. Numbered embodiment 71 comprises the non-transitory computer readable storage medium including software of embodiment 52, wherein the software is configured to further cause the processor to record a signal from the sensor electrode. Numbered embodiment 72 comprises the non-transitory computer readable storage medium including software of embodiment 71, wherein the software is configured to adjust the sensor electrode's response to the signal. Numbered embodiment 73 comprises the non-transitory computer readable storage medium of embodiment 52, wherein the software comprises a machine learning model, and wherein the machine learning model is configured to classify signals detected by the sensor electrode and generate signals with the stimulator electrode. Numbered embodiment 74 comprises the non-transitory computer readable storage medium of embodiment 73, wherein the machine learning model is trained on prior data acquired from the individual or a set of individuals and the corresponding incontinence prevention or lack thereof information. Numbered embodiment 75 comprises the non-transitory computer readable storage medium of embodiment 74, wherein the prior data comprise the signals detected by the sensor electrode, the generated signals, or any combination thereof. Numbered embodiment 76 comprises the non-transitory computer readable medium of embodiment 52, wherein the software comprises analyzing a global positioning system (GPS) location of the individual that in combination with the parameter associated with the response from the individual prevents the episode of incontinence.

(152) Numbered embodiment 77 comprises a method of data processing, said method comprising: (i)

receiving a measurement of a parameter previously measured by a sensor electrode, which parameter is predictive of an episode of incontinence in an individual; (ii) analyzing the parameter; and (iii) synthesizing an electrical stimulation signal for the individual, so that when the electrical stimulation signal is provided by a stimulator electrode to the individual, the electrical stimulation signal, together with an effort from the individual that is intended to prevent an episode of incontinence, prevents the episode of incontinence. Numbered embodiment 78 comprises the method of embodiment 77, wherein the parameter is associated with a response from the individual intended to prevent an episode of incontinence. Numbered embodiment 79 comprises the method of embodiment 78, wherein the parameter is associated with the individual's effort in trying to prevent an episode of incontinence, and wherein the electrical stimulation signal is synthesized so as to supplement the individual's effort with an electrical stimulation pattern that will, together with the effort from the individual, be sufficient to prevent an episode of incontinence. Numbered embodiment 80 comprises the method of embodiment 79, wherein the response from the individual is insufficient on its own to prevent the episode of incontinence and the electrical stimulation signal is such that, when applied, it adds enough, together with the response, to prevent the episode of incontinence. Numbered embodiment 81 comprises the method of any one of embodiments 77 to 80, wherein the episode of incontinence comprises urinary incontinence. Numbered embodiment 82 comprises the method of embodiments 77 or 80, wherein the episode of incontinence comprises fecal incontinence. Numbered embodiment 83 comprises the method of any one of embodiments 77 to 80, wherein the episode of incontinence comprises urinary stress incontinence. Numbered embodiment 84 comprises the method of any one of embodiments 77 to 83, wherein the parameter is a signal from a sensor electrode that is configured to sense a contraction of a muscle of the individual related to a partial contraction of a sphincter that controls bladder or bowel voiding. Numbered embodiment 85 comprises the method of any one of embodiments 77 to 84, wherein the electrical stimulation signal is for the electrical stimulation of the pudendal nerve. Numbered embodiment 86 comprises the method of any one of embodiments 77 to 85, wherein the electrical stimulation signal is synthesized to include a constant electrical stimulation component and a measurement parameter specific component. Numbered embodiment 87 comprises the method of any one of embodiments 77 to 86, wherein the episode of incontinence is urinary incontinence and is urge incontinence type. Numbered embodiment 88 comprises the method of any one of embodiments 77 to 87, wherein an intensity or duration of the electrical stimulation that will be provided by the electrical stimulation signal varies according to the value of the parameter that is received. Numbered embodiment 89 comprises the method of any one of embodiments 77 to 88, wherein the parameter is an EMG signal. Numbered embodiment 90 comprises the method of embodiment 89, wherein the EMG signal determines that a contraction of at least one pelvic muscle has occurred. Numbered embodiment 91 comprises the method of embodiment 90, wherein a strength of the EMG signal is proportional to the strength of the contraction of at least one pelvic muscle. Numbered embodiment 92 comprises the method of any one of embodiments 77 to 91, wherein the electrical stimulation signal comprises first and second signals for the stimulation of a first pudendal nerve and a second pudendal nerve respectively. Numbered embodiment 93 comprises the method of any one of embodiments 77 to 92, further comprising recording the signal previously measured by the sensor electrode. Numbered embodiment 94 comprises the method of embodiment 93, further comprising synthesizing an adjustment signal to adjust the sensor electrode in response to the recorded signal.

EXAMPLES

Example 1: EMG Measurement Reproducibility

(153) Using the systems, methods, and devices, described herein, patient EMG signals were measured and processed as patients performed muscle contractions, coughing, and Valsalva maneuvers, as can be seen in FIGS. **11A-11E**. For each patient, raw EMG data **1102** was recorded and amplified, as described elsewhere herein. Subsequently, the raw EMG data was filtered **1106**, rectified, and smoothed **1104**.

(154) From the data shown in FIGS. **11A-11C**, where a single patient iteratively performed the same sequence of muscle contractions three times, a distinct group of three signals may be observed in the temporal EMG data. Such a finding supports the reproducibility of the sensor electrodes of the devices and systems described herein and the potential ability of a trained classifier to distinguish EMG signals

temporally. Similarly data for a different patient shown in FIGS. 11D-11E shows similar results as the patient in FIGS. 11A-11C.

Claims

1. A system for preventing an episode of incontinence of a patient, the system comprising: a first lead and a second lead configured to be implanted bilaterally proximate a pudendal nerve; a first sensor electrode and a second sensor electrode configured to detect a parameter indicative of the patient's effort to prevent an episode of incontinence; a first stimulator electrode and a second stimulator electrode configured to provide an electrical stimulation at or near the pudendal nerve; wherein the first lead comprises the first sensor electrode and the first stimulator electrode; wherein the second lead comprises the second sensor electrode and the second stimulator electrode; and an implantable pulse generator operably coupled to the first and second stimulator electrodes and the first and second sensor electrodes; wherein the implantable pulse generator is configured to provide a base stimulation level; wherein in response to detecting the parameter indicative of the patient's effort to prevent the episode of incontinence, the implantable pulse generator is configured to provide an activated electrical stimulation level configured to prevent the episode of incontinence; wherein the activated electrical stimulation level is higher than the base stimulation level; and wherein the implantable pulse generator is further configured to provide the activated electrical stimulation level to the same lead that detected the parameter indicative of the patient's effort to prevent the episode of incontinence.
2. The system of claim 1, wherein each of the first sensor electrode and the second sensor electrode is configured to detect an EMG signal.
3. The system of claim 2, wherein the EMG signal is indicative of a contraction of a sphincter or a pelvic muscle.
4. The system of claim 2, wherein the EMG signal is indicative of a purposeful contraction of a pelvic muscle.
5. The system of claim 1, wherein the activated electrical stimulation level is higher than the base stimulation level in one or more of frequency, amplitude, or current.
6. The system of claim 1, further comprising an external input device configured to adjust parameters of the activated electrical stimulation level.
7. The system of claim 1, wherein the activated electrical stimulation level is based at least in part on a real-time clock scheduler.
8. A system for ameliorating an episode of incontinence of a patient, the system comprising: a first sensor electrode configured to detect an EMG signal indicative of the patient's effort to ameliorate an episode of incontinence; a second sensor configured to sense a second parameter indicative of the patient's effort to prevent the episode of incontinence; at least one stimulator electrode configured to provide an electrical stimulation at or near a pudendal nerve; and an implantable pulse generator operably coupled to the at least one stimulator electrode and the first sensor electrode; wherein the implantable pulse generator is configured to provide a base stimulation level; wherein in response to detecting the EMG signal indicative of the patient's effort to prevent the episode of incontinence and the second parameter, the implantable pulse generator is configured to provide an activated electrical stimulation level configured to ameliorate the episode of incontinence; wherein the activated electrical stimulation level is higher than the base stimulation level.
9. The system of claim 8, wherein the EMG signal is indicative of a contraction of a sphincter or a pelvic muscle.
10. The system of claim 8, wherein the EMG signal is indicative of a purposeful contraction of a pelvic muscle.
11. The system of claim 8, wherein an intensity of the activated electrical stimulation level is higher than an intensity of the base stimulation level.
12. The system of claim 8, wherein the activated electrical stimulation level is based at least in part on a real-time clock scheduler.
13. A system for ameliorating an episode of incontinence of a patient, the system comprising: a first

sensor electrode configured to detect an EMG signal indicative of the patient's effort to ameliorate an episode of incontinence; at least one stimulator electrode configured to provide an electrical stimulation at or near a pudendal nerve; and an implantable pulse generator operably coupled to the at least one stimulator electrode and the first sensor electrode; wherein the implantable pulse generator is configured to provide a base stimulation level, wherein the base stimulation level has a constant frequency; wherein in response to detecting the EMG signal indicative of the patient's effort to prevent the episode of incontinence, the implantable pulse generator is configured to provide an activated electrical stimulation level configured to ameliorate the episode of incontinence; wherein the activated electrical stimulation level is higher than the base stimulation level.

14. The system of claim 13, wherein the base stimulation level further has a constant amplitude or a constant current.

15. The system of claim 13, wherein the EMG signal is indicative of a contraction of a sphincter or a pelvic muscle.

16. The system of claim 13, wherein the EMG signal is indicative of a purposeful contraction of a pelvic muscle.

17. The system of claim 13, wherein an intensity of the activated electrical stimulation level is higher than an intensity of the base stimulation level.

18. The system of claim 13, wherein in response to detecting the EMG signal indicative of the patient's effort to prevent the episode of incontinence, the activated electrical stimulation level changes: a pulse width of electrical stimulation, a frequency of electrical stimulation, or an amplitude of electrical stimulation.

19. The system of claim 13, further comprising a second sensor configured to sense a second parameter indicative of the patient's effort to prevent the episode of incontinence, wherein in response to detecting the EMG signal and the second parameter, the implantable pulse generator is configured to provide the activated electrical stimulation level configured to ameliorate the episode of incontinence.

20. The system of claim 19, the second sensor is configured to sense physical movement of the patient.

21. The system of claim 19, wherein the second sensor comprises an accelerometer or a pressure sensor.

22. The system of claim 13, wherein the activated electrical stimulation level is based at least in part on a real-time clock scheduler.

23. A system for ameliorating an episode of incontinence of a patient, the system comprising: a first sensor electrode configured to detect an EMG signal indicative of the patient's effort to ameliorate an episode of incontinence; at least one stimulator electrode configured to provide an electrical stimulation at or near a pudendal nerve; and an implantable pulse generator operably coupled to the at least one stimulator electrode and the first sensor electrode; wherein the implantable pulse generator is configured to provide a base stimulation level, wherein the base stimulation level has a constant amplitude; wherein in response to detecting the EMG signal indicative of the patient's effort to prevent the episode of incontinence, the implantable pulse generator is configured to provide an activated electrical stimulation level configured to ameliorate the episode of incontinence; wherein the activated electrical stimulation level is higher than the base stimulation level.

24. The system of claim 23, wherein the base stimulation level further has a constant current.

25. The system of claim 23, wherein the EMG signal is indicative of a contraction of a sphincter or a pelvic muscle.

26. The system of claim 23, wherein the EMG signal is indicative of a purposeful contraction of a pelvic muscle.

27. The system of claim 23, wherein an intensity of the activated electrical stimulation level is higher than an intensity of the base stimulation level.

28. The system of claim 23, wherein in response to detecting the EMG signal indicative of the patient's effort to prevent the episode of incontinence, the activated electrical stimulation level changes: a pulse width of electrical stimulation, a frequency of electrical stimulation, or a frequency of electrical stimulation.

29. The system of claim 23, further comprising a second sensor configured to sense a second parameter indicative of the patient's effort to prevent the episode of incontinence, wherein in response to detecting the EMG signal and the second parameter, the implantable pulse generator is configured to provide the

activated electrical stimulation level configured to ameliorate the episode of incontinence.

30. The system of claim 29, the second sensor is configured to sense physical movement of the patient.

31. The system of claim 29, wherein the second sensor comprises an accelerometer or a pressure sensor.

32. The system of claim 23, wherein the activated electrical stimulation level is based at least in part on a real-time clock scheduler.

33. A system for ameliorating an episode of incontinence of a patient, the system comprising: a first sensor electrode configured to detect an EMG signal indicative of the patient's effort to ameliorate an episode of incontinence; at least one stimulator electrode configured to provide an electrical stimulation at or near a pudendal nerve; and an implantable pulse generator operably coupled to the at least one stimulator electrode and the first sensor electrode; wherein the implantable pulse generator is configured to provide a base stimulation level, wherein the base stimulation level has a constant current; wherein in response to detecting the EMG signal indicative of the patient's effort to prevent the episode of incontinence, the implantable pulse generator is configured to provide an activated electrical stimulation level configured to ameliorate the episode of incontinence; wherein the activated electrical stimulation level is higher than the base stimulation level.

34. The system of claim 33, wherein the EMG signal is indicative of a contraction of a sphincter or a pelvic muscle.

35. The system of claim 33, wherein the EMG signal is indicative of a purposeful contraction of a pelvic muscle.

36. The system of claim 33, wherein an intensity of the activated electrical stimulation level is higher than an intensity of the base stimulation level.

37. The system of claim 33, wherein in response to detecting the EMG signal indicative of the patient's effort to prevent the episode of incontinence, the activated electrical stimulation level changes: a pulse width of electrical stimulation, a frequency of electrical stimulation, or an amplitude of electrical stimulation.

38. The system of claim 33, further comprising a second sensor configured to sense a second parameter indicative of the patient's effort to prevent the episode of incontinence, wherein in response to detecting the EMG signal and the second parameter, the implantable pulse generator is configured to provide the activated electrical stimulation level configured to ameliorate the episode of incontinence.

39. The system of claim 38, the second sensor is configured to sense physical movement of the patient.

40. The system of claim 38, wherein the second sensor comprises an accelerometer or a pressure sensor.

41. The system of claim 33, wherein the activated electrical stimulation level is based at least in part on a real-time clock scheduler.

42. A system for ameliorating an episode of incontinence of a patient, the system comprising: a first sensor electrode configured to detect an EMG signal indicative of the patient's effort to ameliorate an episode of incontinence; at least one stimulator electrode configured to provide an electrical stimulation at or near a pudendal nerve; and an implantable pulse generator operably coupled to the at least one stimulator electrode and the first sensor electrode; wherein the implantable pulse generator is configured to provide a base stimulation level; wherein in response to detecting the EMG signal indicative of the patient's effort to prevent the episode of incontinence, the implantable pulse generator is configured to provide an activated electrical stimulation level configured to ameliorate the episode of incontinence and the activated electrical stimulation level changes a pulse width of electrical stimulation; wherein the activated electrical stimulation level is higher than the base stimulation level.

43. The system of claim 42, wherein the EMG signal is indicative of a contraction of a sphincter or a pelvic muscle.

44. The system of claim 42, wherein the EMG signal is indicative of a purposeful contraction of a pelvic muscle.

45. The system of claim 42, wherein an intensity of the activated electrical stimulation level is higher than an intensity of the base stimulation level.

46. The system of claim 42, wherein in response to detecting the EMG signal indicative of the patient's effort to prevent the episode of incontinence, the activated electrical stimulation level further changes a frequency of electrical stimulation or an amplitude of electrical stimulation.

47. The system of claim 42, further comprising a second sensor configured to sense a second parameter

indicative of the patient's effort to prevent the episode of incontinence, wherein in response to detecting the EMG signal and the second parameter, the implantable pulse generator is configured to provide the activated electrical stimulation level configured to ameliorate the episode of incontinence.

48. The system of claim 47, the second sensor is configured to sense physical movement of the patient.

49. The system of claim 47, wherein the second sensor comprises an accelerometer or a pressure sensor.

50. The system of claim 42, wherein the activated electrical stimulation level is based at least in part on a real-time clock scheduler.

51. A system for ameliorating an episode of incontinence of a patient, the system comprising: a first sensor electrode configured to detect an EMG signal indicative of the patient's effort to ameliorate an episode of incontinence; at least one stimulator electrode configured to provide an electrical stimulation at or near a pudendal nerve; and an implantable pulse generator operably coupled to the at least one stimulator electrode and the first sensor electrode; wherein the implantable pulse generator is configured to provide a base stimulation level; wherein in response to detecting the EMG signal indicative of the patient's effort to prevent the episode of incontinence, the implantable pulse generator is configured to provide an activated electrical stimulation level configured to ameliorate the episode of incontinence and the activated electrical stimulation level changes a frequency of electrical stimulation; wherein the activated electrical stimulation level is higher than the base stimulation level.

52. The system of claim 51, wherein the EMG signal is indicative of a contraction of a sphincter or a pelvic muscle.

53. The system of claim 51, wherein the EMG signal is indicative of a purposeful contraction of a pelvic muscle.

54. The system of claim 51, wherein an intensity of the activated electrical stimulation level is higher than an intensity of the base stimulation level.

55. The system of claim 51, wherein in response to detecting the EMG signal indicative of the patient's effort to prevent the episode of incontinence, the activated electrical stimulation level further changes an amplitude of electrical stimulation.

56. The system of claim 51, further comprising a second sensor configured to sense a second parameter indicative of the patient's effort to prevent the episode of incontinence, wherein in response to detecting the EMG signal and the second parameter, the implantable pulse generator is configured to provide the activated electrical stimulation level configured to ameliorate the episode of incontinence.

57. The system of claim 56, the second sensor is configured to sense physical movement of the patient.

58. The system of claim 56, wherein the second sensor comprises an accelerometer or a pressure sensor.

59. The system of claim 51, wherein the activated electrical stimulation level is based at least in part on a real-time clock scheduler.

60. A system for ameliorating an episode of incontinence of a patient, the system comprising: a first sensor electrode configured to detect an EMG signal indicative of the patient's effort to ameliorate an episode of incontinence; at least one stimulator electrode configured to provide an electrical stimulation at or near a pudendal nerve; and an implantable pulse generator operably coupled to the at least one stimulator electrode and the first sensor electrode; wherein the implantable pulse generator is configured to provide a base stimulation level; wherein in response to detecting the EMG signal indicative of the patient's effort to prevent the episode of incontinence, the implantable pulse generator is configured to provide an activated electrical stimulation level configured to ameliorate the episode of incontinence and the activated electrical stimulation level changes an amplitude of electrical stimulation; wherein the activated electrical stimulation level is higher than the base stimulation level.

61. The system of claim 60, wherein the EMG signal is indicative of a contraction of a sphincter or a pelvic muscle.

62. The system of claim 60, wherein the EMG signal is indicative of a purposeful contraction of a pelvic muscle.

63. The system of claim 60, wherein an intensity of the activated electrical stimulation level is higher than an intensity of the base stimulation level.

64. The system of claim 60, further comprising a second sensor configured to sense a second parameter indicative of the patient's effort to prevent the episode of incontinence, wherein in response to detecting the EMG signal and the second parameter, the implantable pulse generator is configured to provide the

activated electrical stimulation level configured to ameliorate the episode of incontinence.

65. The system of claim 64, the second sensor is configured to sense physical movement of the patient.

66. The system of claim 64, wherein the second sensor comprises an accelerometer or a pressure sensor.

67. The system of claim 64, wherein the activated electrical stimulation level is based at least in part on a real-time clock scheduler.

68. A system for ameliorating an episode of incontinence of a patient, the system comprising: a first sensor electrode configured to detect an EMG signal indicative of the patient's effort to ameliorate an episode of incontinence; a second sensor configured to sense a second parameter indicative of the patient's effort to prevent the episode of incontinence, the second sensor configured to sense physical movement of the patient; at least one stimulator electrode configured to provide an electrical stimulation at or near a pudendal nerve; and an implantable pulse generator operably coupled to the at least one stimulator electrode and the first sensor electrode; wherein the implantable pulse generator is configured to provide a base stimulation level; wherein in response to detecting the EMG signal indicative of the patient's effort to prevent the episode of incontinence and the second parameter, the implantable pulse generator is configured to provide an activated electrical stimulation level configured to ameliorate the episode of incontinence; wherein the activated electrical stimulation level is higher than the base stimulation level.

69. The system of claim 68, wherein the base stimulation level has: a constant frequency, a constant amplitude, or a constant current.

70. The system of claim 68, wherein the EMG signal is indicative of a contraction of a sphincter or a pelvic muscle.

71. The system of claim 68, wherein the EMG signal is indicative of a purposeful contraction of a pelvic muscle.

72. The system of claim 68, wherein an intensity of the activated electrical stimulation level is higher than an intensity of the base stimulation level.

73. The system of claim 68, wherein in response to detecting the EMG signal indicative of the patient's effort to prevent the episode of incontinence, the activated electrical stimulation level changes: a pulse width of electrical stimulation, a frequency of electrical stimulation, or an amplitude of electrical stimulation.

74. The system of claim 68, wherein the second sensor comprises an accelerometer or a pressure sensor.

75. The system of claim 68, wherein the activated electrical stimulation level is based at least in part on a real-time clock scheduler.

76. A system for ameliorating an episode of incontinence of a patient, the system comprising: a first sensor electrode configured to detect an EMG signal indicative of the patient's effort to ameliorate an episode of incontinence; a second sensor configured to sense a second parameter indicative of the patient's effort to prevent the episode of incontinence, wherein the second sensor comprises an accelerometer; at least one stimulator electrode configured to provide an electrical stimulation at or near a pudendal nerve; and an implantable pulse generator operably coupled to the at least one stimulator electrode and the first sensor electrode; wherein the implantable pulse generator is configured to provide a base stimulation level; wherein in response to detecting the EMG signal indicative of the patient's effort to prevent the episode of incontinence and the second parameter, the implantable pulse generator is configured to provide an activated electrical stimulation level configured to ameliorate the episode of incontinence; wherein the activated electrical stimulation level is higher than the base stimulation level.

77. The system of claim 76, wherein the base stimulation level has: a constant frequency, a constant amplitude, or a constant current.

78. The system of claim 76, wherein the EMG signal is indicative of a contraction of a sphincter or a pelvic muscle.

79. The system of claim 76, wherein the EMG signal is indicative of a purposeful contraction of a pelvic muscle.

80. The system of claim 76, wherein an intensity of the activated electrical stimulation level is higher than an intensity of the base stimulation level.

81. The system of claim 76, wherein in response to detecting the EMG signal indicative of the patient's effort to prevent the episode of incontinence, the activated electrical stimulation level changes: a pulse

width of electrical stimulation, a frequency of electrical stimulation, or an amplitude of electrical stimulation.

82. The system of claim 76, wherein the second sensor further comprises a pressure sensor.

83. The system of claim 76, wherein the activated electrical stimulation level is based at least in part on a real-time clock scheduler.

84. A system for ameliorating an episode of incontinence of a patient, the system comprising: a first sensor electrode configured to detect an EMG signal indicative of the patient's effort to ameliorate an episode of incontinence; a second sensor configured to sense a second parameter indicative of the patient's effort to prevent the episode of incontinence, wherein the second sensor comprises a pressure sensor; at least one stimulator electrode configured to provide an electrical stimulation at or near a pudendal nerve; and an implantable pulse generator operably coupled to the at least one stimulator electrode and the first sensor electrode; wherein the implantable pulse generator is configured to provide a base stimulation level; wherein in response to detecting the EMG signal indicative of the patient's effort to prevent the episode of incontinence and the second parameter, the implantable pulse generator is configured to provide an activated electrical stimulation level configured to ameliorate the episode of incontinence; wherein the activated electrical stimulation level is higher than the base stimulation level.

85. The system of claim 84, wherein the base stimulation level has: a constant frequency, a constant amplitude, or a constant current.

86. The system of claim 84, wherein the EMG signal is indicative of a contraction of a sphincter or a pelvic muscle.

87. The system of claim 84, wherein the EMG signal is indicative of a purposeful contraction of a pelvic muscle.

88. The system of claim 84, wherein an intensity of the activated electrical stimulation level is higher than an intensity of the base stimulation level.

89. The system of claim 84, wherein in response to detecting the EMG signal indicative of the patient's effort to prevent the episode of incontinence, the activated electrical stimulation level changes: a pulse width of electrical stimulation, a frequency of electrical stimulation, or an amplitude of electrical stimulation.

90. The system of claim 84, wherein the activated electrical stimulation level is based at least in part on a real-time clock scheduler.
