

LetsGetChecked myGene DNA Saliva Collection Activation and Return Guide

SPECS AND INSTRUCTIONS

Colour Reference:

CMYK

Size:

108.5mm wide by 152.5mm high folded

Paper:

135gsm silk

Fold type:

3 page accordion

Font:

Proxima Nova

(Regular, Semibold and Bold)

Part Number:

QR-IFU-115-R0 May-2024



*Visual reference for how to fold down the sheet so the front page is the first one to see. Artwork and page count may vary.

FOLD GUIDE

----- Fold

Accordion fold





Inside



Outside



Before you start

- Do not eat, drink, smoke, or chew gum for 30 minutes before collecting your sample.
- Collect your sample on a weekday and return on the same day.
- Do not remove the plastic film from the funnel lid.
- Plan your sample return in advance. Check the return delivery card in your kit for details.
- Read all instructions and warnings on the ORAgene®•Dx saliva collection kit.



Kit contents

Make sure you have all these items in your kit:*







Kit box

Return envelope

*The look or color of items may vary from what you see in these images,

Activate your kit



Activate your kit at

LetsGetChecked.com/activate.

Enter the unique alpha and
numerical codes on your biohazard
bag. Only activate when you are
ready to collect your sample.



Fill in the label on the biohazard bag.



Wash and dry your hands.

Please continue overleaf.



Activate your kit at LetsGetChecked.com/activate before collecting your sample

Prepare for sample collection



Remove all the contents from your kit box and open the packaging.



Follow the instructions inside the ORAgene®•Dx packaging to collect your sample.



Discard the ORAgene®•Dx packaging.

Package and return



Place the sample into the biohazard bag. Then, place the biohazard bag into the kit box. Double check you have filled in the biohazard bag label.



Place the kit box into the return envelope provided.



Send back on the same day. Check the return delivery card in your kit box for your return method.

LGC_ACC_MYGENE_SCREEN_DNA_SALIVA_QR_IFU_115_R0_ACTIVATION_US_3_PG_V10.indd 6-8