

Nick Lalic to: Mike Leffel

Mike,



This command gives me an error about an invalid separator (presumably the parenthesis). I have tried different permutations of 'MSH(P1)' (with the parenthesis, without, with an extra colon separating 'MSH:1', etc etc), but I cannot query it.

Note: There is another SCPI command menu that you can use without specifying the label of the kit (I don't know why, probably for historical reasons). I get the same result with these commands.

They are:

:CORR:CKIT:STAN:CAT? 'Port Specific'

Result:

'MSH(P1),MOP'

:CORR:CKIT:MOP? '3.5 mm','Port Specific'

Result:

'3.5 mm','Port Specific','Open Label',0,1E+012,0,0,50,0,0,0,0,0,0,0

:CORR:CKIT:MSH(P1)? '3.5 mm','Port Specific'

****ERROR****

It seems like whoever implemented this SCPI command did not take into account port specific standards. This needs to be fixed for me to finish Merge Cal.

Do you think this is a feasible project for Munich? We already have one request outstanding...

Have a happy 4th of July!

Best Regards,
Nick Lalic



ROHDE & SCHWARZ

Nick Lalic

VNA Software Developer

Rohde & Schwarz USA, Inc.

Mobile: 424-200-2846

E-Mail: Nick.Lalic@rsa.rohde-schwarz.com

Website: <http://www.rohde-schwarz.us/>