Detailed BrainSightAI MRI protocols - ***Surgical***

rs - fMRI	
Slice thickness	3 mm or less
Voxel size	3 mm x 3 mm or less
Number of slices	50 (Whole Brain Coverage)
Acquisition matrix	64 x 64 (Minimum)
Echo time	30- 50 ms
Repetition time	2000 to 2900
Flip Angle	90°
Slice Spacing	0
Number of Dynamic Scans	160-200
Acceleration Phase	Default 2
Ramp Scan	1
Freq Res	64
Phase Res	64
Phase FOV	100%
Patient position	HFS
Freq Direction	AP

dMRI	
Slice thickness	2 mm
Voxel size	2 mm x 2 mm x 2 mm or less
Number of slices	75 (Whole Brain Coverage)
Acquisition matrix	122 x 122 or above
Echo time	94 ms
Repetition time	5000 minimum or above
Flip Angle	90°
Slice Spacing/Gap	0
Number of Directions	64 -128 (if multiband)
B value	0, 1000
B0	2-3
Number of Avgs. (B100)	1
FOV	240mm
Patient position	HFS
Direction	AP

Detailed BrainSightAI MRI protocols - ***Surgical***

SEQUENCES	T1W	T2W	FLAIR	T1 with Contrast (T1C)
Slice Thickness	1mm (maximum)or (less)	1mm (maximum)or (less)	1mm (maximum)or (less)	1mm (maximum)or (less)
Voxel Size	1 × 1 × 1 mm or (less)	1 × 1 × 1 mm or (less)	1 × 1 × 1 mm or (less)	1 × 1 × 1 mm or (less)
Acquisition Type	3D	3D	3D	3D
Number of Slices	160 (minimum)	160 (minimum)	160 (minimum)	160 (minimum)
Acquisition Matrix	240 × 240 (minimum) or above	240 × 240 (minimum) or above	240 × 240 (minimum) or above	240 x240 (minimum) or above