



# **PDP90A**

## **2D Lateral Effect Position Sensor**

### **User Guide**



















# Table of Contents

<b>Chapter 1</b>	<b>Warning Symbol Definitions .....</b>	<b>1</b>
<b>Chapter 2</b>	<b>Description.....</b>	<b>2</b>
	<b>2.1. Setup Instructions .....</b>	<b>2</b>
	<b>2.2. Details of Operation .....</b>	<b>2</b>
	2.2.1. Overview.....	3
	2.2.2. Calculating the Position.....	3
	2.2.3. Position Resolution .....	4
<b>Chapter 3</b>	<b>Specifications .....</b>	<b>6</b>
<b>Chapter 4</b>	<b>Drawing .....</b>	<b>7</b>
<b>Chapter 5</b>	<b>Regulatory .....</b>	<b>8</b>
<b>Chapter 6</b>	<b>Thorlabs Worldwide Contacts.....</b>	<b>9</b>

# Chapter 1 Warning Symbol Definitions

Below is a list of warning symbols you may encounter in this manual or on your device.

Symbol	Description
	Direct Current
	Alternating Current
	Both Direct and Alternating Current
	Earth Ground Terminal
	Protective Conductor Terminal
	Frame or Chassis Terminal
	Equipotentiality
	On (Supply)
	Off (Supply)
	In Position of a Bi-Stable Push Control
	Out Position of a Bi-Stable Push Control
	Caution: Risk of Electric Shock
	Caution: Hot Surface
	Caution: Risk of Danger
	Warning: Laser Radiation
	Caution: Spinning Blades May Cause Harm

## Chapter 2 Description

The PDP90A is a two-dimensional, tetra-lateral position-sensing detector (PSD) that provides X- and Y-axis positional information. It is designed to work with Thorlabs' KPA101 K-Cube Auto-Aligner. This sensor can be used for general alignment purposes as well as precise position measurements. This sensor gives the position of any spot contained within the active region and supports spot sizes up to 9 mm in diameter. We recommend a spot between  $\varnothing 0.2$  mm and  $\varnothing 7$  mm.

Thorlabs also manufactures the PDQ80A quadrant position detector for precision alignment needs.

### 2.1. Setup Instructions

1. Unpack and mount the PDP90A position-sensing detector. An AS4M8E M4 to 8-32 thread adapter is included for customers using metric optomechanical components.
2. Carefully remove the protective film covering the lens.
3. Connect the sensor to a KPA101 K-Cube Auto-Aligner.
4. Follow the directions in the controller's operating manual to ensure that the position-sensing detector is properly communicating with the controller.
5. Send a beam onto the active area for measurement. The spot size input beam should be between  $\varnothing 0.2$  and  $\varnothing 7$  mm. For best results, the spot should be located within the center 80% of the surface of the detector (a 7.2 mm x 7.2 mm square). Adjust the power level so that the SUM output voltage is  $\leq 4$  V. This ensures the best signal-to-noise ratio and avoids signal saturation.

### 2.2. Details of Operation

This section of the manual discusses the operating principles of 2D lateral effect position-sensing detectors and includes key equations for understanding their output signals.

### 2.2.1. Overview

2D lateral effect sensors accurately measure displacement (movements, distances, or angles), providing feedback from systems that require precise alignment, such as steering mirrors, microscopy systems, and fiber launch systems. These sensors are light-controlled variable resistors that work by proportionally distributing photocurrent to determine position in the X and Y from the four electrodes when incident light is applied to its surface.

In addition, the PDP90A is a continuous position-sensing detector, which derives position by dividing photo-generated electrons within their substrate, rather than by profiling the intensity distribution on the surface as with segmented detectors. There are two different types of lateral effect sensors available: duo-lateral and tetra-lateral.

Duo-lateral sensors provide a resistive layer on both the anode and cathode photodiode connections. This isolates the X and Y positional information of the sensor and allows the sensor to be highly linear and very accurate. However, the added resistive layers significantly increase manufacturing cost.

Tetra-lateral sensors use a single resistive layer with a common cathode and an anode on each side of the detection area. This makes them very inexpensive to manufacture, but their linearity decreases as the spot moves away from the center. This is caused by the physical location of the anodes along the sides of the sensor, specifically in the corners where the anodes approach each other.

Thorlabs' PDP90A detector uses a variation of the tetra-lateral-type sensor commonly called a "pin cushion" position-sensing device. This sensor moves the anodes to the four corners of the sensors and reshapes the sensor area, to produce linearity comparable to a duo lateral sensor at a significantly lower cost.

The tetra-lateral-type, "pincushion" has a characteristic form of which we will use some important values to determine further calculations. The active area is 9x9 mm at the waist of the pincushion; however, at the outer most regions the area goes to 10mm, therefore the calculations are performed from the widest part of the sensor.

### 2.2.2. Calculating the Position

The PDP90A includes circuitry to decode the  $D_x$ ,  $D_y$ , and SUM signals as follows (see *Figure 1*):

$$\Delta x = (A + D) - (B + C)$$


$$\Delta y = (A + B) - (C + D)$$

$$SUM = (A + B + C + D)$$

**Note:** A, B, C and D are four electrodes on the upper surface, formed along each of the four edges

The following formulas will help you to determine the actual light beam position along the X- and Y-axis.

$$x = \frac{L_x(\Delta x)}{2 \text{ SUM}} ; y = \frac{L_y(\Delta y)}{2 \text{ SUM}}$$



**Figure 1** Layout of a Lateral Effect Position Sensor

### 2.2.3. Position Resolution

Position resolution is the minimum detectable displacement of a light spot incident on the PSD. Position resolution ( $\Delta R$ ) is a factor of both the resistance length ( $L_x$  or  $L_y$ ), or the X and Y length, and the signal-to-noise ratio (S/N). The signal to noise ratio for this system can be defined as the SUM output signal level ( $V_0$ ) divided by the output voltage noise ( $e_n$ ). The PDP90A output noise is  $<2 \text{ mV}_{pp}$ , or  $<300 \text{ } \mu\text{V}_{rms}$ :

$$\Delta R = L_x \left( \frac{e_n}{V_0} \right)$$

where  $\Delta R$  is the resolution,  $L_x$  (or  $L_y$ ) is the detector resistance length (For the PDP90A,  $L_x = L_y = 10 \text{ mm}$ ),  $e_n$  is the output noise voltage ( $300 \text{ } \mu\text{V}_{rms}$ ), and  $V_0$  is the SUM output voltage level (4 V).

$$\Delta R = 10 \text{ mm} \left( \frac{300 \text{ } \mu\text{V}_{rms}}{4 \text{ V}} \right) = 0.75 \text{ } \mu\text{m}$$

For best results,  $V_0$  should be maximized at 4 V, resulting in a position resolution of  $0.75 \text{ } \mu\text{m}$ . To do this, monitor the SUM output voltage of the sensor and adjust the optical intensity until the output is approximately 4 V. Outputs above 4 V will saturate the system, causing unreliable readings. The software will indicate the voltage level on a slide bar, and is available for download online at [thorlabs.com](http://thorlabs.com). To reach the download page go to Services, Software Downloads, Detectors, and then Position Sensing Detectors and click on the link. If the SUM is saturated, the slide bar color will turn red. Reduce the intensity until it switches to green. This will be equivalent to a 4 V SUM output.

Low power levels should be used for the position sensor. For the PDP90A, the max photocurrent before saturation is  $40 \mu\text{A}$ <sup>1</sup>. Using the photosensitivity curve below (Figure 2), the system saturation power can be calculated as:

$$P_{\max} = \frac{40 \mu\text{A}}{\text{Photosensitivity (in } \frac{\text{A}}{\text{W}})}$$


The KPA101 K-Cube PSD Auto Aligner has a sample resolution of 16-bits with an input voltage range of  $\pm 15 \text{ V}$  (30 V range), or a voltage resolution of:

$$V_{\text{step}} = \frac{30 \text{ V}}{2^{16}} = 0.458 \text{ mV}$$

$$I_{\text{step}} = \frac{V_{\text{step}}}{\text{Transimpedance Gain}} = \frac{V_{\text{step}}}{100 \frac{\text{kV}}{\text{A}}}$$

From this, the minimum power required can be determined based on the required accuracy. For example, if the user requires an accuracy of 1%, the minimum photocurrent is  $I_{\text{step}} \times 100 = 0.458 \mu\text{A}$ . Use the formula below to find the minimum optical power for a given wavelength.

$$P_{\min} = \frac{0.458 \mu\text{A}}{\text{Photosensitivity (in } \frac{\text{A}}{\text{W}})}$$



**Figure 2** Photosensitivity of PDP90A

<sup>1</sup> The amplifier would saturate before the saturation of the photodiode itself.

# Chapter 3 Specifications


Electrical Specifications	
Wavelength Range	320 - 1100 nm
Peak Responsivity	0.6 A/W at 960 nm
Transimpedance Gain	100 kV/A
Photocurrent (Max)	40 $\mu$ A
Output Voltage Range	-4 to +4 V (Pins 1 & 2) 0 to 4 V (Pin 3)
Signal Output Offset	0.3 mV (Typical) 7 mV (Max)
Bandwidth	15 kHz
Spot Size <sup>2</sup>	$\varnothing$ 0.2 to $\varnothing$ 7 mm (Recommended) $\varnothing$ 9 mm (Max)
Power Requirements	$\pm$ 5 VDC $\pm$ 5%, 35 mA

Physical Specifications	
Sensor Size	9 mm x 9 mm (0.35" x 0.35")
Mechanical Aperture	$\varnothing$ 0.50" ( $\varnothing$ 12.7 mm)
Position Displacement Error	$\pm$ 250 $\mu$ m
Aperture Thread	Internal SM05 (0.535"- 40)
Dimensions	2.00" x 1.20" x 0.65" (50.8 mm x 30.5 mm x 16.5 mm)
Mounting Thread	8-32, 0.25" Thread Depth (Min)
Included Metric Adapter	External 8-32 to Internal M4 (Thorlabs Item # AS4M8E)
Cable Length	5' (1.5 m) (Typ.)
Connector Plug	Hirose HR10A-7P-6P
Mating Receptacle	Hirose HR10A-7R-6S
Weight	0.25 lbs. (114 g)
Sensor Depth	3.05 mm (0.12")
Operating Temperature	10 to 40 $^{\circ}$ C
Storage Temperature	-20 to 80 $^{\circ}$ C

<sup>2</sup>  $\varnothing$ 9 mm (Max) at the waist of the pincushion.



# Chapter 4 Drawing




Pin	Description
1	X-Axis $[A + D] - [B + C]$
2	Y-Axis $[A + B] - [C + D]$
3	SUM $[A + B + C + D]$
4	+Vsupply (+5 V)
5	Common
6	-Vsupply (-5 V)

## Chapter 5 Regulatory

As required by the WEEE (Waste Electrical and Electronic Equipment Directive) of the European Community and the corresponding national laws, Thorlabs offers all end users in the EC the possibility to return "end of life" units without incurring disposal charges.

- This offer is valid for Thorlabs electrical and electronic equipment:
- Sold after August 13, 2005
- Marked correspondingly with the crossed out "wheelie bin" logo (see right)
- Sold to a company or institute within the EC
- Currently owned by a company or institute within the EC
- Still complete, not disassembled and not contaminated



**Wheelie Bin Logo**

As the WEEE directive applies to self-contained operational electrical and electronic products, this end of life take back service does not refer to other Thorlabs products, such as:

- Pure OEM products, that means assemblies to be built into a unit by the user (e.g. OEM laser driver cards)
- Components
- Mechanics and optics
- Left over parts of units disassembled by the user (PCB's, housings etc.).

If you wish to return a Thorlabs unit for waste recovery, please contact Thorlabs or your nearest dealer for further information.

### ***Waste Treatment is Your Own Responsibility***

If you do not return an "end of life" unit to Thorlabs, you must hand it to a company specialized in waste recovery. Do not dispose of the unit in a litter bin or at a public waste disposal site.

### ***Ecological Background***

It is well known that WEEE pollutes the environment by releasing toxic products during decomposition. The aim of the European RoHS directive is to reduce the content of toxic substances in electronic products in the future.

The intent of the WEEE directive is to enforce the recycling of WEEE. A controlled recycling of end of life products will thereby avoid negative impacts on the environment.

## Chapter 6 Thorlabs Worldwide Contacts

For technical support or sales inquiries, please visit us at [www.thorlabs.com/contact](http://www.thorlabs.com/contact) for our most up-to-date contact information.



### USA, Canada, and South America

Thorlabs, Inc.  
sales@thorlabs.com  
techsupport@thorlabs.com

### Europe

Thorlabs GmbH  
europe@thorlabs.com

### France

Thorlabs SAS  
sales.fr@thorlabs.com

### Japan

Thorlabs Japan, Inc.  
sales@thorlabs.jp

### UK and Ireland

Thorlabs Ltd.  
sales.uk@thorlabs.com  
techsupport.uk@thorlabs.com

### Scandinavia

Thorlabs Sweden AB  
scandinavia@thorlabs.com

### Brazil

Thorlabs Vendas de Fotônicos Ltda.  
brasil@thorlabs.com

### China

Thorlabs China  
chinasales@thorlabs.com



**THORLABS**

[www.thorlabs.com](http://www.thorlabs.com)

---