



γ ray absorption in matter by a γ -ray source

PH3105

2024-09-12

Debayan Sarkar
22MS002

Diptanuj Sarkar
22MS038

Sabarno Saha
22MS037



Contents

I. Introduction	1
II. Theory	1
II.1. Geiger–Müller Counter	1
II.1.A. Working Principle	1
II.1.B. Readouts	2
II.1.C. Limitations	3
II.2. Mass Attenuation	3
II.3. Inverse Square Law	3
III. Background Reading	4
III.1. Table 1 (Background reading)	4
IV. Determination of mass attenuation coefficient	4
IV.1. Measurement of thickness of plates	4
IV.2. Graph	4
V. Calculation of half value thickness	5
VI. Verification of inverse square law	6
VI.1. Graph	6
VII. Possible errors	6
VIII. Conclusion	7
IX. Supplementary	8
IX.1. Determination of Mass Attenuation Coefficient	8
IX.1.A. Table 1 (Thickness of Plates)	8
IX.1.B. Table 2 (Count rate vs. Thickness)	8
IX.2. Verification of Inverse Square Law Data	10

I. Introduction

In this experiment we have studied γ ray absorption by lead using a Na-22 γ ray source and a Geiger-Müller Counter. We also verified the inverse square law for γ radiation.

II. Theory

II.1. Geiger-Müller Counter

A Geiger-Müller Counter (GM counter) is an electronic device that lets us detect the presence of ionizing radiation. It detects radiations such as α , β , and γ rays using the ionisation effect produced by these rays in a GM tube.




Figure 1 :: A GM Counter. (Source: Wikipedia)

II.1.A. Working Principle

The GM tube is filled with an inert gas, such as Helium, Neon, Argon etc. at very low pressure. Then, a very high voltage is applied across the tube. This high voltage keeps the electrons from recombining with the formed ions during an ionisation event. When an ionising radiation falls on the tube, the ionisation process is considerably amplified due to an **ionisation avalanche**, which is also known as Townsend Discharge. Essentially, after the first ionisation process, the ejected electron, while accelerating through the applied potential collide with other gas molecules, and eject even more electrons.




Figure 2 :: Avalanche effect in gas subject to ionising radiation between two plate electrodes.
(Source: Wikipedia)

This leads to a significantly large amplification of the number of free electrons produced and hence, produces a large pulse, which is easily measured. We have seen what happens in a GM tube due to one radiation event. However, the tube must quickly return to its previous state, to detect subsequent radiations. However, when the ionised gaseous molecules reach the cathode to get gain electrons, they reach an excited state. And their de-excitation produces photons which cause ionisation again, causing false electric pulses. This would lead to a prolonged “dead time” for the tube during which, no new events can be detected. To counter this, we use a **quenching gas**. Generally, halogens like Bromine, and Chlorine are chosen as the quenching gas for GM tubes. When the ionised inert gas in the tube collides with the neutral Halogen molecules, the former accepts an electron from the latter. And when the ionised halogens reach the cathode, they also reach an excited state, but the de-excitation doesn't cause more ionisation for the inert gas, and hence, doesn't cause false pulses.




Figure 3 :: Detection of gamma in a G-M tube with a thick-walled stainless steel cathode.
(Source: Wikipedia)

II.1.B. Readouts

The GM tube counts these “pulses” and gives them as a readout, displayed either as total counts over a fixed amount of time, or count rates, like **counts per second** or **counts per minute**.

II.1.C. Limitations

As we can clearly see, each pulse detected using the GM tube will be of the same magnitude, regardless of the type of radiation. A GM counter is only good for detecting the presence of radiation, and the rate of radiation. It cannot give us any information about the energy of the radiation. Also, as discussed above, since a GM tube will have a “dead time”, at very high radiation rates, which will lead to a decreased reading being put out by the counter.

II.2. Mass Attenuation

The experiment is used to find the mass attenuation coefficient of Pb for γ rays. We define attenuation to be the gradual loss of flux intensity through a medium. Here we quantify the attenuation of γ rays when they move through Pb plates.

The linear attenuation coefficient characterizes how easily a volume of material can be penetrated by a beam of light, sound, particles, or other energy or matter, or in this case γ rays. The mass attenuation coefficient is then defined as the linear attenuation coefficient normalized by the density of the medium. We define the linear attenuation coefficient as

$$\mu = -\frac{1}{\Phi} \frac{d\Phi}{dz}$$

where

- μ = The Linear attenuation coefficient.
- ρ = The density of the medium.
- Φ = The flux at z .
- z = The path length travelled by the beam.


On solving we get our working formula as (assuming μ doesn't depend on z) to be:

$$I = I_0 e^{-\frac{\mu}{\rho} \rho z}$$

where we have the $\frac{\mu}{\rho}$ to be the mass attenuation coefficient.

II.3. Inverse Square Law

We can treat the used γ ray source as a point source, emitting photons in all directions. In a spherically symmetric manner. Then, as we change the distance between the source and the detector, the number of particles falling into the detector clearly reduces. This can be better illustrated using the diagram below.



Then, the number of particles falling on a detector of area A at a distance r is proportional to $\frac{A}{r^2}$. Hence, the counts should follow an inverse square law with respect to the distance from the detector.

III. Background Reading

Background radiation is a measure of the level of ionizing radiation present in the environment at a particular location which is not due to deliberate introduction of radiation sources. In the case of our experiment, the environment may have some amount of γ radiation present. It is important to measure the amount of background radiation so that we can correct our count rate reading from the GM counter by subtracting the count rate due to the background reading.

III.1. Table 1 (Background reading)

S. No.	Time (s)	Counts
1	600	161
2	600	158
3	600	187

Average count: 168.7 (in 600 s)

Average count rate: 0.281 s^{-1}

IV. Determination of mass attenuation coefficient

IV.1. Measurement of thickness of plates

Zero error on the screw gauge: 0.00 mm

Least count of the screw gauge: 0.01 mm

IV.2. Graph




Figure 5 :: Linear fit to determine the mass attenuation coefficient

The data plotted was fitted with the equation -

$$y(x) = ax + b$$

with the fit giving us the values of the coefficients -

$$a = -0.094 \pm 0.008$$

$$b = 0.06 \pm 0.07$$

The density of lead is known to be $\rho = 1.1343 \times 10^4 \text{ kg} \cdot \text{m}^{-3} = 1.1343 \times 10^{-2} \text{ g} \cdot \text{mm}^{-3}$

From analysing the expression for the mass attenuation coefficient, we find that -

$$\frac{\mu}{\rho} = (0.094) \left(\frac{1}{\rho} \right) = (0.094) \left(\frac{1}{1.1343} \right) (100) \text{ g}^{-1} \text{ mm}^2 = 8.287 \text{ g}^{-1} \text{ mm}^2$$

Thus, the value of the mass attenuation coefficient is found to be $(8.2 \pm 0.7) \text{ g}^{-1} \text{ mm}^2$.

V. Calculation of half value thickness

Half value thickness is the thickness of the absorber at which the value of the incident intensity is attenuated to half of its initial value.

We note that,

$$\frac{I_0}{2} = I_0 \exp \left(- \left(\frac{\mu}{\rho} \right) \rho z_h \right) \Rightarrow z_h = \frac{\ln \left(\frac{1}{2} \right)}{- \left(\frac{\mu}{\rho} \right) \rho} \quad [1]$$

Putting in the values, we get,

$$z_h = \frac{0.693}{0.094} \text{mm} = 7.372 \text{mm} \quad [2]$$

Thus, the value of the half-value thickness is $7.37 \pm 0.06 \text{mm}$.

VI. Verification of inverse square law

VI.1. Graph




Figure 6 :: Verification of Inverse Square Law

We represent the data in a $\log(\text{Counts})$ vs $\log(\text{distance})$ graph, given as

$$\begin{aligned} \log(I) &= a \log(x) + b \\ \Rightarrow y &= ax + b \end{aligned}$$

We fit the data with a linear fit to obtain a to be -1.72 ± 0.03 and b to be 8.52 ± 0.06 .

VII. Possible errors

The possible errors in this experiment are -

1. Introduction of secondary radiation sources in the environment of the experimental setup after taking the background radiation reading and while the experiment is being conducted - This can change the count rate due to sources other than the intended radiation source, thereby affecting the results.

2. Voltage fluctuation in the electronic instruments used - This may affect the count rate as expressed by the GM counter.
3. While measuring the mass attenuation coefficient, there were small air gaps between the lead plates. This leads to an error in the counts due to some surface scattering effects. Also, if all the lead plates are not parallel to each other, that would introduce an error in the measurement of the thickness. However, this error should be very minimal, since GM counter is very small.

VIII. Conclusion

We measured the mass attenuation coefficient of gamma rays when passing through Pb using a Geiger muller Counter . We also verified the Inverse square law within a marginal degree of error.

IX. Supplementary

IX.1. Determination of Mass Attenuation Coefficient

IX.1.A. Table 1 (Thickness of Plates)

Thickness of the plates					
S no.	MSR (mm)	CSR (mm)	LC (mm)	Total (mm)	Avg. Total
1	2	10	0.01	1.1	1.11
	2	13	0.01	1.13	
	2	10	0.01	1.1	
2	2	15	0.01	1.15	1.116666667
	2	11	0.01	1.11	
	2	9	0.01	1.09	
3	2	6	0.01	1.06	1.106666667
	2	15	0.01	1.15	
	2	11	0.01	1.11	
4	2	3	0.01	1.03	1.033333333
	2	3	0.01	1.03	
	2	4	0.01	1.04	
5	2	3	0.01	1.03	1.05
	2	8	0.01	1.08	
	2	4	0.01	1.04	
6	10	30	0.01	5.3	5.326666667
	10	37	0.01	5.37	
	10	31	0.01	5.31	
7	20	46	0.01	10.46	10.45666667
	20	45	0.01	10.45	
	20	46	0.01	10.46	
8	20	47	0.01	10.47	10.43
	20	40	0.01	10.4	
	20	42	0.01	10.42	

Figure 7 :: Linear fit to determine the mass attenuation coefficient

IX.1.B. Table 2 (Count rate vs. Thickness)

Time (s)	Thickness (mm)	Counts	Avg Count	Count rate (s^{-1})	Corrected rate (s^{-1})
45	0	396	398.5	8.856	8.574
		401			
120	1.11	139	145	1.208	0.927
		151			
120	2.23	137	138	1.150	0.869
		139			
120	3.33	119	124.5	1.038	0.756
		130			

Time (s)	Thickness (mm)	Counts	Avg Count	Count rate (s^{-1})	Corrected rate (s^{-1})
120	4.37	138	137.5	1.146	0.865
		137			
120	5.42	118	110	0.917	0.636
		102			
120	5.33	110	101	0.842	0.560
		92			
120	6.44	87	102	0.850	0.569
		117			
120	7.55	97	95	0.792	0.510
		93			
120	8.66	113	99.5	0.829	0.548
		86			
120	9.69	81	84	0.700	0.419
		87			
120	10.74	73	72.5	0.604	0.323
		72			
120	10.46	77	81	0.675	0.394
		85			
120	11.57	86	86.5	0.721	0.439
		87			
120	12.68	67	70.5	0.588	0.306
		74			

IX.2. Verification of Inverse Square Law Data

Time	Distance	Counts	Avg Count	Avg Count Rate
60	12	71	72.33333333	1.205555556
		76		
		70		
60	11.5	70	76.33333333	1.272222222
		71		
		88		
60	11	68	72.33333333	1.205555556
		68		
		81		
60	10.5	86	85.33333333	1.422222222
		85		
		85		
60	10	102	97.66666667	1.627777778
		107		
		84		
60	9.5	116	109.3333333	1.822222222
		116		
		96		
60	9	119	120	2
		117		
		124		
60	8.5	133	127.3333333	2.122222222
		116		
		133		
60	8	126	131.6666667	2.194444444
		134		
		135		
60	7.5	145	162.6666667	2.711111111
		171		
		172		
60	7	179	176.6666667	2.944444444
		172		
		179		
60	6.5	194	190.6666667	3.177777778
		195		
		183		
60	6	229	224	3.733333333
		225		
		218		
60	5.5	268	279	4.65
		281		
		288		
60	5	355	333.6666667	5.561111111
		313		
		333		
60	4.5	357	375.3333333	6.255555556
		376		
		393		
60	4	446	462.6666667	7.711111111
		473		
		469		
60	3.5	532	561.6666667	9.361111111
		573		
		580		