

20.022.NUSE.PH

Gait Monitoring and Analysis of Parkinson's Disease Patients

Ye Chen Rui, Prannaya Gupta and Nallapuraju Ananya
NUS High School of Mathematics and Science

RESEARCH PLAN

A) RATIONALE

This project is useful and relevant as Parkinson's Disease is becoming more and more prevalent amongst the elderly population in Singapore, and this means that more elderly are at a risk of falling due to the symptoms of Parkinson's Gait. Moreover, many elderly citizens in Singapore live in HDBs where due to cramped spaces, they may have a higher risk of severe injuries upon falling. Research has also shown that walking in cramped spaces can trigger the worsening of Freezing of Gait. Thus, we feel it is of utmost importance to try to understand more about the parameters that could possibly measure Freezing of Gait and how we could analyse these parameters to create a simple gait monitoring device. In the future, we hope that this device could help medical researchers identify Freezing of Gait easily for a more effective diagnosis of Parkinson's Disease and also aid elderly caregivers in their day-to-day lives.

B) RESEARCH QUESTION

Is it possible to make use of signal processing algorithms in extracting gait parameters from motion sensors and identifying most suitable parameters for classification of Freezing of Gait in Parkinson's Disease Patients?

HYPOTHESIS

Whether gait parameters can be derived from Inertia Motion Units, to classify freezing of gait (e.g. Accelerometers, Gyroscope)

ENGINEERING GOALS

- Monitor gait patterns for Parkinson's disease patients
- Analyse gait during freezing and non freezing events
- Investigating gait parameters to understand what is most suitable for classification for freezing of gait in PD patients.
- Create a prototype of the gait monitoring device

C) PROCEDURE

- Review and understand how to extract gait parameters from motion sensors.
- Investigate which algorithm is most suitable from different algorithms present in literature
- Make use of a PD patient public dataset and create a prototype to test out the algorithms

RISK ASSESSMENT

1. Usage of Soldering Iron to solder Arduino board
 - a. *Risks Involved:* Risk of Burn, Risk of Inhaling Toxic Gases
 - b. *Safety Precautions:* Wear gloves and goggles at all times when handling soldering equipment and use a fume fan to capture toxic gases released.

LITERATURE REVIEW

Utilisation of IMUs in analysis

This section reports previous studies which have explored the application of motion sensors on PD patients to accurately predict FOG. Ferster et al. [1] placed 9-axis IMUs (comprising 3D accelerometers, 3D gyroscopes and 3D magnetometers) on both ankles of the subjects to extract gait features such as stride length and stride duration. Moreover, as FOG exhibits unique frequency ranges, they introduce and discuss frequency features such as dominant frequency, dominant frequency amplitude and the inverse of the dominant frequency slope of the acceleration data to quantify changes in gait quality. Ferster was able to show specific changes in the stride duration, stride length, dominant frequency and the inverse of the dominant frequency slope with up to four seconds prior to FoG on all subjects.

Baechlin et al. [2] proposed placing accelerometers at three different parts of the body: the shank, thigh and lower back, where the wearable computer is attached to.




Figure 1: The Baechlin et al. 3-axial Accelerometer Wearable System

Alam et al. [3] analyzed the vertical ground reaction force using force insoles in patients' shoes to display gait cycles. Pinto et al. [4] again utilised accelerometers and gyroscopes to determine stride time, this time placing the accelerometer at the shank. In summary, a consolidated list of the main IMUs that were considered in this study is provided in Table 1.

Table 1: Table of possible IMUs

IMU	Purpose	Measured Parameter
Accelerometer	Measuring acceleration	Stride Length, Stride Duration
Gyroscope	Measuring angular velocity	Step Festination, Gait Asymmetry
Flexible Goniometer	Measuring body joint angles	Flat Foot Strike
Force- sensitive Insole	Measuring the tension and compression forces that act on the sensor	Gait Cycle (not accurate for PD patients who suffer from flat footedness)

Many works have utilised motion capture systems to annotate FOG events, synchronising sensor data and computer analysis to make way for machine learning algorithms. Kuhner et al. [5] performed this experiment, setting up 12 cameras as well as utilising an inertial measurement suit to create a 'live' system that reduces the latency of data processing.

The above literature confirms the success of utilising accelerometers, gyroscopes and force insoles to effectively differentiate FOG from normal gait and has greatly helped in designing the proposed approach for this study.

Gait parameters to be analysed

The frequency-based features discussed by Baechlin et al [2] are reliable, but they have one major drawback. The large amount of calculations required to perform the Fast Fourier Transform (FFT) algorithm needed to analyse frequency-based data means that a digital signal processing chip is required, making it impossible to fit in a small, lightweight unobtrusive system. This would make it impractical for real life detection. For the patients with gait problems, a bulky and obtrusive system may worsen the gait disturbances. Additionally, most experiments utilised a relatively long window length (4 seconds [2], 6 seconds [7], and 10 seconds [8]) for detection, leading to a long delay in FOG detection. Eom et al. [6] suggested a simple and fast time-domain method for FOG detection that was comparable to the traditional frequency-domain method with a calculation load of 1,154 times less. This has practical clinical applications.

Gait Monitoring and Analysis of Parkinson's Disease Patients

Previous researchers also have reported the importance of five parameters for the detection of PD. Most notably, Hollman et al.[9] have proposed five major domains of gait based on factor analysis:

1. Rhythm - step and stride time
2. Phase - gait cycle
3. Variability - step-to-step variability
4. Pace - gait speed, stride and step length
5. Base of support - step width

Others have confirmed the importance of determining the spatiotemporal parameters. Alcock et al.[10], Coste et al.[11] and Schlachetzki et al.[12] have all discussed the importance of stride length, stride time and gait velocity in distinguishing PD patients from healthy older adults. Overall, a summary of gait parameters found in this research is shown below. In this study, the focus will be on gait velocity, stride time and gait cycle.

Table 2: Summary of Gait parameters found in this research

Gait abnormalities	Definition
Step Festination	Shortening of steps
Decrease in Stride Duration	Decrease in the time difference between two consecutive detected acceleration peaks of the same leg
Postural Instability	Loss of balance
Increase in step-to-step time variability	Difference in time taken between each stride
Stride Length Reduction	Reduction in distance between each stride/step
Decreased Cadence	Decrease in number of steps taken per unit time
Gait Asymmetry	Difference in gait between the two legs of patient

Bibliography for Literature Review

- [1] Ferster, M. L., Mazilu, S., & Tröster, G. (2015). Gait Parameters Change Prior to Freezing in Parkinson's Disease: A Data-Driven Study with Wearable Inertial Units. Proceedings of the 10th EAI International Conference on Body Area Networks. <https://doi.org/10.4108/eai.28-9-2015.2261411>
- [1] Nutt, J. G., Bloem, B. R., Giladi, N., Hallett, M., Horak, F. B., & Nieuwboer, A. (2011). Freezing of gait: moving forward on a mysterious clinical phenomenon. The Lancet Neurology, 10(8), 734–744. [https://doi.org/10.1016/s1474-4422\(11\)70143-0](https://doi.org/10.1016/s1474-4422(11)70143-0)
- [2] Bachlin, M., Plotnik, M., Roggen, D., Maidan, I., Hausdorff, J. M., Giladi, N., & Troster, G. (2010). Wearable Assistant for Parkinson's Disease Patients With the Freezing of Gait Symptom. IEEE Transactions on Information Technology in Biomedicine, 14(2), 436–446. <https://doi.org/10.1109/titb.2009.2036165>
- [3] Alam, M. N., Garg, A., Munia, T. T. K., Fazel-Rezai, R., & Tavakolian, K. (2017). Vertical ground reaction force marker for Parkinson's disease. PLOS ONE, 12(5), e0175951. <https://doi.org/10.1371/journal.pone.0175951>
- [4] Pinto, C., Schuch, C. P., Balbinot, G., Salazar, A. P., Hennig, E. M., Kleiner, A. F. R., & Pagnussat, A. S. (2019). Movement smoothness during a functional mobility task in subjects with Parkinson's disease and freezing of gait – an analysis using inertial measurement units. Journal of NeuroEngineering and Rehabilitation, 16(1). <https://doi.org/10.1186/s12984-019-0579-8>
- [5] Kuhner, A., Schubert, T., Maurer, C., & Burgard, W. (2017). An online system for tracking the performance of Parkinson's patients. 2017 IEEE/RSJ International Conference on Intelligent Robots and Systems (IROS). <https://doi.org/10.1109/iros.2017.8205977>
- [6] Eom, G.-M., Kwon, Y., Park, S. H., Kim, J.-W., Ho, Y., Jeon, H.-M., Bang, M.-J., Jung, G.-I., Lee, S.-M., Koh, S.-B., Lee, J.-W., & Jeon, H. S. (2014). A practical method for the detection of freezing of gait in patients with Parkinson's disease. Clinical Interventions in Aging, 1709. <https://doi.org/10.2147/cia.s69773>

- [7] Moore, S. T., MacDougall, H. G., & Ondo, W. G. (2008). Ambulatory monitoring of freezing of gait in Parkinson's disease. *Journal of Neuroscience Methods*, 167(2), 340–348.
<https://doi.org/10.1016/j.jneumeth.2007.08.023>
- [8] Morris, T. R., Cho, C., Dilda, V., Shine, J. M., Naismith, S. L., Lewis, S. J. G., & Moore, S. T. (2012). A comparison of clinical and objective measures of freezing of gait in Parkinson's disease. *Parkinsonism & Related Disorders*, 18(5), 572–577. <https://doi.org/10.1016/j.parkreldis.2012.03.001>
- [9] Hollman, J. H., McDade, E. M., & Petersen, R. C. (2011). Normative spatiotemporal gait parameters in older adults. *Gait & Posture*, 34(1), 111–118. <https://doi.org/10.1016/j.gaitpost.2011.03.024>
- [10] Alcock, L., Galna, B., Perkins, R., Lord, S., & Rochester, L. (2018). Step length determines minimum toe clearance in older adults and people with Parkinson's disease. *Journal of Biomechanics*, 71, 30–36. <https://doi.org/10.1016/j.jbiomech.2017.12.002>
- [11] Azevedo Coste, C., Sijobert, B., Pissard-Gibollet, R., Pasquier, M., Espiau, B., & Geny, C. (2014). Detection of Freezing of Gait in Parkinson Disease: Preliminary Results. *Sensors*, 14(4), 6819–6827.
<https://doi.org/10.3390/s140406819>
- [12] Morris, T. R., Cho, C., Dilda, V., Shine, J. M., Naismith, S. L., Lewis, S. J. G., & Moore, S. T. (2012). A comparison of clinical and objective measures of freezing of gait in Parkinson's disease. *Parkinsonism & Related Disorders*, 18(5), 572–577. <https://doi.org/10.1016/j.parkreldis.2012.03.001>

PROJECT ABSTRACT

Parkinson's disease (PD) is a neurodegenerative disorder that affects the dopamine producing neurons in the substantia nigra, an area of the brain, leading to shaking, stiffness and difficulty walking. Parkinson's patients frequently exhibit the debilitating condition freezing of gait (FOG), which is when patients cannot move their feet forward despite the intention to walk. While the feet remain in place, the torso still has forward momentum, making falls very common. Several studies have reported methodologies to identify the PD patients at an early stage. The purpose of this project was to identify the parameter which is most suitable for classification of FOG in PD patients. IMUs, machine learning algorithms and gait parameters have been investigated. An algorithm has been developed to analyze the Public datasets of PD patients in order to extract the motion pattern of PD patients. Based on the analysis results, a combination of Freeze Index values is postulated and used to predict FOG based on these parameters. Ultimately, a prototype that fulfils all these requirements was created based on acceleration data from accelerometers placed on the thigh.

INTRODUCTION

Parkinson's Disease (PD) is a progressive neurodegenerative disorder that affects the substantia nigra, an area of the brain, resulting in the progressive loss of dopamine neurotransmitters and poorer connection between the central nervous system and muscles [13]. This leads to many gait abnormalities, mainly tremors, bradykinesia, rigidity, and a loss of postural reflexes. Secondary motor symptoms that are evoked include stride length reduction, shuffling gait, step festination and freezing [1,14]. Freezing of Gait (FOG) is one of the most debilitating effects of PD and is defined as a brief, episodic absence or marked reduction of forward progression of the feet despite the intention to walk [1]. It has serious social and clinical consequences for patients as it leads to falls [15] due to the torso still having forward momentum despite the fact that the feet remain in place, interferes with daily activities and significantly impacts quality of life [16]. While FOG can happen anytime, it happens most often during turns, before gait initiation, in tight quarters such as doorways and in stressful situations [17]. At the start, FOG can be triggered by stress, tight spaces or a sudden change in direction. It is generally triggered by visual stimulation.

Currently, FOG characterization is done using two main methods. In the first method, FOG is quantified by gait tests such as the timed up and go (TUG) test and the Hoehn and Yahr (H&Y) scale. The TUG test measures in seconds how long a patient takes to stand up from their chair, walk 3m past a line, turn around, walk back and sit down [18]. On the other hand, the H&Y scale has 5 stages to indicate levels of disability [19]:

1. Symptoms on one side of the body only.
2. Symptoms on both sides of the body; no impairment of balance.
3. Balance impairment; mild to moderate disease; physically independent.
4. Severe disability, but still able to walk or stand unassisted.
5. Wheelchair-bound or bedridden unless assisted.

In the second method, individual questionnaires are used. The Freezing Of Gait Questionnaire (FOG-Q) is a notable one, comprising 6 questions and utilising a 5-point scale to rank symptom severity [20].

However, both methods are highly inefficient in measuring FOG as it is highly sensitive to environmental triggers, medication and the patient's mental state. Therefore, there has been research into using wearable inertial measurement units (IMUs) to display exactly the gait types of PD patients.

This study aims to compare all the ways of classifying FOG in PD patients and determine the best parameter to utilise while creating an algorithm for data analysis. It also aims to compare multiple

machine learning models based on acceleration data from accelerometers placed on the thigh. Public datasets of PD patients will be analysed to extract the motion pattern of PD patients. A Freeze Index value is postulated and used to predict FOG based on these parameters. The comfort of the patient and the ease and accuracy in which the parameter can predict FOG will be taken into account in this study. Ultimately, a prototype that fulfils all these requirements will be made.

MATERIALS AND METHODS

Accelerometers have been chosen to be the main focus of the data analysis as they are very versatile. Research has shown that accelerometers can be used to determine stride time, gait cycle and gait velocity. Hence, they would be a good starting point.

DAPHNet Dataset

In the main investigation, the Daphnet Freezing in users with Parkinson's disease (hereafter known as the DAPHNet Dataset) is employed, of Gait Dataset which is the result of a study done by Baechlin et al.[2], carried out by the Laboratory for Gait and Neurodynamics, Department of Neurology, Tel Aviv Sourasky Medical Center (TASMC). In this experiment, 17 samples were derived from 10 PD patients with varying H&Y scales who were made to do various walking tasks, including walking back and forth in a straight line, doing several 180 degrees turns, random walking including a series of initiated stops and 360 degree turns and walking simulating daily activities. Daily activities refer to entering and leaving rooms, getting something to drink and returning to the starting room with the cup of water.

Data was recorded using three 3D acceleration sensors attached to the shank, thigh and the lower back of each subject. The sensors recorded at 64Hz and transmitted the acceleration data via a Bluetooth link to a wearable computing system that was located at the lower back of the subjects.

However, there are limitations to the dataset. Patient 08 and Patient 01 both suffered from walking difficulties due to disease severity and foot drop respectively. Hence, the system was unable to distinguish between walking periods and very short freezing events.




Figure 2: Acceleration Data from S03R02.txt. The blue line depicts the acceleration data provided.

Signal Processing

This study uses signal processing algorithms to compute a postulated freeze index (FI). Moore et al.[22] defined the FI as the power ratio of freeze band (0.5–3.0 Hz) to locomotor band (3–8 Hz) derived from the frequency spectrum and identified FOG episodes at the time periods when FI exceeds a certain threshold. The code uses an updated form of Moore's algorithm, where it calculates the FI from data from a specific axis (horizontal forward, horizontal lateral and vertical) from the sensors on the thigh by performing windowing of data with a length of 256 data points and steps up 32 data points. After performing a mean normalisation, the code finds the Fast Fourier Transform of the normalised window and uses that to compute the FI and supposed Power Spectral Density (PSD). After this, it accounts for a standing case where the $PSD_{threshold} = 2^{11.5}$, and in cases where it is below this threshold, the FI becomes zero. This postulated FI was considered.

In this experiment, a final function freeze(N) is defined, where freeze(N) = FI postulated from N-axis acceleration. For the investigation, the axes were defined as follows:

Axis	Axis with respect to the user
X	Horizontal Forward
Y	Vertical
Z	Horizontal Vertical

SVM Analysis

Unlike Baechlin et al.[2], where the FI with a threshold of 1.5 is used such that a FI greater than the threshold is considered FOG, SVM models have been tested against the derived FI values. This was in-line with a comprehensive review of four different types of machine learning algorithms (Support Vector Machine (SVM); k-Nearest Neighbour (kNN); Decision Tree (DT) and Naïve Bayes (NB)) in classifying patients with FOG or no FOG done by Aich et al [21]. They found that the SVM classifier with radial basis function provides the highest accuracy of 91.42% as well as the highest sensitivity and specificity of 90.89% and 91.21% respectively.

Due to the large amount of data amassed from the signal processing algorithms, the Google Colaboratory Tool [22] is used. It allows Graphics Processing Units (GPUs) of more powerful computers at Google to be utilised. This allows the SVM to run much faster such that results can be obtained faster for the investigation. The ThunderSVM utility [23] developed by the Xtra Computing Group at the NUS School of Computing (SoC) is also used to fully utilise the GPUs.

For the investigation, Linear and Gaussian Kernels have been employed and these SVMs are tested with either 2-axial freeze indices (i.e. taking freeze indices from 2 of the 3 axes defined) or 3-axial freeze indices (which took into account freeze(X), freeze(Y) and freeze(Z)).

Arduino Program

The program has been developed using Arduino's programming language based off the C++ programming language. To get the best SVM model, a program called sklearn-porter [24] developed by Darius Morawies, which converts the trained SVM model to C code, saved as a model.h file, has been utilised. This model is accessed by the Arduino Program, which also computes the freeze values.




Figure 3: The Arduino Nano 33 BLE Pinout Diagram. The built-in LED lights up upon detection of a freeze event.

This study utilises an Arduino Nano 33 BLE board that is attached to an elastic strap. It contains a 9-axial IMU, comprising a 3D accelerometer, 3D gyroscope and 3D magnetometer. In the program, upon a certain freeze event predicted by the SVM, the built-in light-emitting diode (LED) will turn on.

RESULTS

Upon performing the signal processing algorithms, the postulated function arrays for freeze(X), freeze(Y) and freeze(Z) are obtained. Below are the freeze indices based on the acceleration data shown in Figure 2.




Figure 4: FI Tabulations from S03R02.txt

All 17 samples are merged and taken as the main argument X in the supervised machine learning structure. From here, a provided freeze value for each point given by the dataset itself has been taken to function as the y argument. Below, all freeze indices are plotted against one another.




Figure 5: Plot of Freeze Indices

Performing an SVM analysis using ThunderSVM, the following results have been obtained:

Kernel	Parameters	Specificity	Sensitivity / Recall	Precision	Accuracy	F ₁ Score
Linear Kernel	freeze(X) and freeze(Y)	1.000	0.000	0.000	0.903	0.000
	freeze(X) and freeze(Z)	0.975	0.243	0.509	0.904	0.329
	freeze(Y) and freeze(Z)	1.000	0.006	0.889	0.904	0.011
	All 3 parameters	1.000	0.002	0.867	0.904	0.005
Radial Basis Kernel	freeze(X) and freeze(Y)	0.990	0.174	0.650	0.911	0.275
	freeze(X) and freeze(Z)	0.990	0.183	0.664	0.912	0.286
	freeze(Y) and freeze(Z)	0.991	0.163	0.665	0.911	0.262
	All 3 parameters	0.989	0.260	0.723	0.918	0.382

Table 3: Table of Algorithm Evaluation Functions with regards to kernels and parameters

Based on the results, it is believed that the Linear Kernel with parameters freeze(Y) and freeze(Z) is the best since it is highly accurate, specific and precise. After conducting an analysis, it is found that the weighted average value for precision and sensitivity/ recall as 0.90 and weighted average F₁ score as 0.86. The decision boundary function is shown below.




Figure 6: Decision Boundary of freeze(Y) and freeze(Z) SVM under Linear Kernel

Prototype

In the end, a prototype (shown in Figure 7) has been developed to verify the effectiveness of parameters based on public dataset analysis. It consists of an Arduino Nano 33 BLE board with a built-in 9-axial IMU consisting of a 3D accelerometer and 3D gyroscope attached to an elastic band. It is small, light, very comfortable and is meant to be wrapped around the thigh.



Figure 7: The Prototype

DISCUSSION & CONCLUSION

An algorithm has been developed to identify the most suitable parameter for the classification of FOG in PD patients. Multiple machine learning models have then been compared based on acceleration data from accelerometers placed on the thigh. After analyzing, the most suitable parameters for classification are freeze(Y) and freeze(Z) based on the acceleration data in the public datasets and the best model is the Linear Kernel model in terms of sensitivity. Furthermore, a prototype has been implemented using an Arduino Nano 33 BLE board. It can be utilized to test the performance of the identified most suitable parameters.

There are a few improvements that could have been done to the project. Due to time limitations, only accelerometers were studied in this project, which could have limited the sensitivity of the device. Multiple sensors can be combined to achieve maximum accuracy including gyroscopes to find out rotation motion parameters as well. Additionally, only postulated freeze index values have been analysed based on the public datasets. Parameters such as Stride Length and Stride Duration have not been studied. In future, such parameters can be taken into account when computing the general gait freeze moment. The sensitivity of the algorithm can also be improved. Since the prototype has already been built, the next phase is to test it in a laboratory with Parkinson's Disease patients in order to determine its accuracy.

Furthermore, by connecting this system to earbuds and implementing biofeedback via audio, the system will also be able to mediate FOG. Another method involves connecting the system to an App such that notifications can be sent to the caregiver's phone to alert them to a fall. This can be developed via Android App Development with Java/ Kotlin, iOS App Development with Swift/ Objective-C or via cross-platform frameworks like Kivy, React Native/ Ionic, Xamarin or Flutter for Python, Javascript, .NET/ C# and Dart Programming.

ACKNOWLEDGEMENTS

We would like to thank our mentors for their support and encouragement.

REFERENCES

- [13] Braak, H., Ghebremedhin, E., Rüb, U., Bratzke, H., & Del Tredici, K. (2004). Stages in the development of Parkinson's disease-related pathology. *Cell and Tissue Research*, 318(1), 121–134.
<https://doi.org/10.1007/s00441-004-0956-9>
- [14] Jankovic, J. (2008). Parkinson's disease: clinical features and diagnosis. *Journal of Neurology, Neurosurgery & Psychiatry*, 79(4), 368–376. <https://doi.org/10.1136/jnnp.2007.131045>
- [15] Bloem, B. R., Hausdorff, J. M., Visser, J. E., & Giladi, N. (2004). Falls and freezing of gait in Parkinson's disease: A review of two interconnected, episodic phenomena. *Movement Disorders*, 19(8), 871–884. <https://doi.org/10.1002/mds.20115>
- [16] Moore, O., Peretz, C., & Giladi, N. (2007). Freezing of gait affects quality of life of peoples with Parkinson's disease beyond its relationships with mobility and gait. *Movement Disorders*, 22(15), 2192–2195. <https://doi.org/10.1002/mds.21659>
- [17] Giladi, N., & Hausdorff, J. M. (2006). The role of mental function in the pathogenesis of freezing of gait in Parkinson's disease. *Journal of the Neurological Sciences*, 248(1–2), 173–176.
<https://doi.org/10.1016/j.jns.2006.05.015>
- [18] Podsiadlo, D., & Richardson, S. (1991). The Timed “Up & Go”: A Test of Basic Functional Mobility for Frail Elderly Persons. *Journal of the American Geriatrics Society*, 39(2), 142–148.
<https://doi.org/10.1111/j.1532-5415.1991.tb01616.x>
- [19] Hoehn, M. M., & Yahr, M. D. (1967). Parkinsonism: onset, progression, and mortality. *Neurology*, 17(5), 427. <https://doi.org/10.1212/wnl.17.5.427>
- [20] Rozenfeld, S., Miskovic, A., Nitsch, K. P., & Ehrlich-Jones, L. (2017). Measurement Characteristics and Clinical Utility of the Freezing of Gait Questionnaire in Individuals With Parkinson Disease. *Archives of Physical Medicine and Rehabilitation*, 98(10), 2106–2107.
<https://doi.org/10.1016/j.apmr.2017.04.027>

- [21] Aich, S., Pradhan, P., Park, J., Sethi, N., Vathsa, V., & Kim, H.-C. (2018). A Validation Study of Freezing of Gait (FoG) Detection and Machine-Learning-Based FoG Prediction Using Estimated Gait Characteristics with a Wearable Accelerometer. *Sensors*, 18(10), 3287. <https://doi.org/10.3390/s18103287>
- [22] Bisong E. (2019) Google Colaboratory. In: Building Machine Learning and Deep Learning Models on Google Cloud Platform. Apress, Berkeley, CA. https://doi.org/10.1007/978-1-4842-4470-8_7
- [23] Wen, Z., Shi, J., Li, Q., He, B., & Chen, J. (2018). ThunderSVM: A Fast SVM Library on GPUs and CPUs, *Journal of Machine Learning Research*, 19, 797–801.
- [24] Morawiec, D. (n.d.). sklearn-porter - Transpile trained scikit-learn estimators to C, Java, JavaScript and others. <https://github.com/nok/sklearn-porter>