

Command – Turning OFF the automatic search of the Michelangelo hand

Command	> S H ; 0 <
Decimal view	62 83 72 59 0 60
Hex view	0x3E 0x53 0x48 0x3B 0x00 0x3C
Command length	6
Answer	The TR-FB device will respond either with the ASCII message ">OK<" or ">ERR<".

On the other side, in order to turn on the automatic search of the hand, the following command should be sent:

Command – Turning ON the automatic search of the Michelangelo hand

Command	> S H ; 1 <
Decimal view	62 83 72 59 1 60
Hex view	0x3E 0x53 0x48 0x3B 0x01 0x3C
Command length	6
Answer	The TR-FB device will respond either with the ASCII message ">OK<" or ">ERR<".

Command – Trigger all vibromotors with their own amplitude

Command	> V B A ; a b c d e f g h <
Decimal view	62 86 66 65 59 a b c d e f g h 60
Hex view	0x3E 0x56 0x42 0x41 0x3B a b c d e f g h 0x3C
Command length	14
Description	14 bytes total, where: a - one byte of the intensity value for the vibromotor 1 (0 - 100% => 0x00 - 0x64) b - one byte of the intensity value for the vibromotor 2 (0 - 100% => 0x00 - 0x64) c - one byte of the intensity value for the vibromotor 3 (0 - 100% => 0x00 - 0x64) d - one byte of the intensity value for the vibromotor 4 (0 - 100% => 0x00 - 0x64) e - one byte of the intensity value for the vibromotor 5 (0 - 100% => 0x00 - 0x64) f - one byte of the intensity value for the vibromotor 6 (0 - 100% => 0x00 - 0x64) g - one byte of the intensity value for the vibromotor 7 (0 - 100% => 0x00 - 0x64) h - one byte of the intensity value for the vibromotor 8 (0 - 100% => 0x00 - 0x64)
Answer	The TR-FB device will respond either with the ASCII message ">OK<" or ">ERR<".