


Pirani_V_to_Pa.vi
C:\Users\IITH\Desktop\RK\Lab_view_automation\Ludweig_tube_automation_new\Pirani_V_to_Pa.vi
Last modified on 25-Sep-26 at 11:20 AM
Printed on 10-Oct-26 at 2:04 AM

Pirani_V_to_Pa.vi





Pirani_V_to_Pa.vi
C:\Users\IITH\Desktop\RK\Lab_view_automation\Ludweig_tube_automation_new\Pirani_V_to_Pa.vi
Last modified on 25-Sep-26 at 11:20 AM
Printed on 10-Oct-26 at 2:04 AM



V_table

0
1.320000
1.330000
1.360000
1.400000
1.430000
1.470000
1.510000
1.540000
1.570000
1.960000
2.230000
2.360000
2.510000
2.660000
2.820000
2.940000
3.080000
3.160000
3.260000
3.940000
4.450000
4.910000
5.310000
5.630000
5.910000
6.160000
6.350000
6.500000
7.230000
7.810000
8.150000
8.340000
8.490000
8.590000
8.690000
8.760000
8.880000
9.150000
9.220000
9.280000
9.310000
9.330000
9.350000
9.360000
9.380000
9.390000
9.410000
9.450000
9.510000
9.570000
9.650000
9.750000
9.850000
9.940000
10.010000

logP_table

0
-3.000000
-2.698970
-2.522879
-2.397940
-2.301030
-2.221849
-2.154902
-2.096910
-2.045757
-2.000000
-1.698970
-1.522879
-1.397940
-1.301030
-1.221849
-1.154902
-1.096910
-1.045757
-1.000000
-0.698970
-0.522879
-0.397940
-0.301030
-0.221849
-0.154902
-0.096910
-0.045757
0.000000
0.301030
0.477121
0.602060
0.698970
0.778151
0.845098
0.903090
0.954243
1.000000
1.301030
1.477121
1.602060
1.698970
1.778151
1.845098
1.903090
1.954243
2.000000
2.301030
2.477121
2.602060
2.698970
2.778151
2.845098
2.903090
2.954243
3.000000