A Compact Closed-Loop Optogenetics System Based on Artifact-Free Transparent Graphene Electrodes

Paper: Liu et al. (2018). Electrode and feedback-hardware validation.

DOI: 10.3389/fnins.2018.00132

Page references below use PDF page numbers, including front matter.

To which general question does this paper relate?

Practical measurement during optogenetic stimulation and implementation latency. Relevant to the sensor/actuator interface, rather than the origin of neurological symptoms.

How does it answer the general questions?

Measured and controlled. Transparent graphene electrodes reduced light-induced recording artifacts compared with gold electrodes under the tested conditions. A portable acquisition/microcontroller/LED system triggered light using a threshold detector (PDF pp. 2–10).

Validation boundary. Closed-loop tests used artificial electrical signals applied in saline, not demonstrated feedback regulation of an animal's neural activity. The reported approximately 40 μs processing response is a hardware-test result, not the total physiological sensor-to-neuron delay (PDF pp. 10–11).

Model relevance. Supports separating electrical sensing, signal processing and optical actuation. Residual artifacts and total loop delay should be characterized in the actual preparation. The equivalent circuit explains electrode artifacts; it is not a fitted neural-network model. Seizure-control applications discussed in the paper are proposed uses or prior work, so this paper does not establish beta suppression or another disease-specific control objective.

What papers does it cite that could provide more information?

- Paz et al. (2013), *Closed-loop optogenetic control of thalamus as a tool for interrupting seizures after cortical injury*: biological feedback validation (references, PDF p. 13).
- Siegle and Wilson (2014), *Enhancement of encoding and retrieval functions through theta phase-specific manipulation of hippocampus*: phase-dependent intervention (PDF p. 13).