



3 Phase Motor Analysis.vi

C:\Users\Public\Desktop\LabVIEW_Applications_2018\3 Phase Motor Analysis.exe\3 Phase Motor Analysis.vi




Last modified on 2022-11-16 at 10:17 AM

Printed on 2022-11-16 at 10:17 AM

3 Phase Motor Analysis

Main View

Phasor View



Electrical Measurements

Ch1 Trigger Source

Amps Volts

Fundamental Frequency

59.8 Hz

V1 RMS	A1 RMS	θ1 (deg)	Ch1 Watts
19.33	5.93	-74.7	28.1
V2 RMS	A2 RMS	θ2 (deg)	Ch2 Watts
19.28	5.85	-74.8	29.1
V3 RMS	A3 RMS	θ3 (deg)	Ch3 Watts
19.33	5.87	-74.8	28.9
Avg V	Avg A	Avg θ	Tot Watt
19.31	5.88	-74.8	86.1

Measurement Window (mS)



Filter

ON

Cutoff Frequency (Hz)



Torque

0.24

Zero Torque

RPM (Analog)

1793

Zero RPM

Zero current/
torque/speed
while motor is
not running


RPM (Digital)

1789

Pause

Save Data

Zoom



Position

