HPV Direct Flow Chip Kit

LOTS

PCR:	MOCKP001	4/1/2030
Chips:	MOCKE001.1	9/1/2030
Reagent:	MOCKH001	1/1/2030

SAMPLE DETAILS

ID SAMPLE: 100-00-0001-0001_1-019-0_20240715 **SAMPLE TYPE:**

ID PATIENT: **PATIENT:**

SEX: - **BIRTHDATE:** **AGE:**

REPORT

HPV NEGATIVE

The sample is negative for all the genotypes included in the HPV direct flow chip test.

PROTOCOL

Detection and genotyping of HPV viral DNA by PCR and reverse dot blot hybridization:
- High risk genotypes: 16, 18, 26, 31, 33, 35, 39, 45, 51, 52, 53, 56, 58, 59, 66, 68, 73, 82.
- Low risk genotypes: 6, 11, 40, 42, 43, 44/55, 54, 61, 62/81, 67, 69, 70, 71, 72, 84.

Automatic analysis of results

NOTES

FACULTATIVE: Default Doctor, doctor **Validated:** 2/10/2025

Performed by: Default Tech, tech **Processed:** 2/10/2025

Instr. : HS12 **Serial Nº:** 100000 **hybriSoft:** MOCK 1.0.0 (HS12a)

HPV Direct Flow Chip Kit

LOTS

PCR:	MOCKP001	4/1/2030
Chips:	MOCKE001.1	9/1/2030
Reagent:	MOCKH001	1/1/2030

SAMPLE DETAILS

ID SAMPLE: 100-00-0001-0001_1-019-0_20240715 **SAMPLE TYPE:**

ID PATIENT: **PATIENT:**

SEX: - **BIRTHDATE:** **AGE:**

REPORT

ANALYSIS INFORMATION

- Spot B: Hybridization control (Biotin signals to orientate the CHIP)
- Spot C: Internal DNA control (Genomic human DNA probe)
- Spot U: HPV Universal probe
- Spot #: Genotype specific probes

All the spots are printed in duplicate.

Threshold: 6

NOTES

FACULTATIVE: Default Doctor, doctor **Validated:** 2/10/2025

Performed by: Default Tech, tech **Processed:** 2/10/2025

Instr. : HS12 **Serial Nº:** 100000 **hybriSoft:** MOCK 1.0.0 (HS12a)