SYNTHETIC TEST REPORT - NOT REAL DATA

HPV Direct Flow Chip Kit

LOTS

PCR:	MOCKP001	4/1/2030
Chips:	MOCKE001.1	9/1/2030
Reagent:	MOCKH001	1/1/2030

SAMPLE DETAILS

ID SAMPLE:	100-00-0002-0002_1-010-0_20240415			**SAMPLE TYPE:**	
ID PATIENT:		**PATIENT:**			
SEX:	-	**BIRTHDATE:**		**AGE:**	

REPORT

HPV POSITIVE

The sample is positive for the following genotypes: 16, 52.

PROTOCOL

Detection and genotyping of HPV viral DNA by PCR and reverse dot blot hybridization:
- High risk genotypes: 16, 18, 26, 31, 33, 35, 39, 45, 51, 52, 53, 56, 58, 59, 66, 68, 73, 82.
- Low risk genotypes: 6, 11, 40, 42, 43, 44/55, 54, 61, 62/81, 67, 69, 70, 71, 72, 84.

Automatic analysis of results

NOTES

FACULTATIVE:	Default Doctor, doctor	**Validated:**	2/10/2025
Performed by:	Default Tech, tech	**Processed:**	2/10/2025

Instr. : HS12 **Serial Nº:** 100000 **hybriSoft:** MOCK 1.0.0 (HS12a)

SYNTHETIC TEST REPORT - NOT REAL DATA

HPV Direct Flow Chip Kit

LOTS

PCR:	MOCKP001	4/1/2030
Chips:	MOCKE001.1	9/1/2030
Reagent:	MOCKH001	1/1/2030

SAMPLE DETAILS

ID SAMPLE:	100-00-0002-0002_1-010-0_20240415		**SAMPLE TYPE:**
ID PATIENT:		**PATIENT:**	
SEX:	-	**BIRTHDATE:**	**AGE:**

REPORT

ANALYSIS INFORMATION

Threshold: 6

- Spot B: Hybridization control (5 signals to orientate the CHIP)
- Spot C: Internal DNA control (Genomic human DNA probe)
- Spot U: HPV Universal probe
- Spot #: Genotype specific probes

All the spots are printed in duplicate.

NOTES

FACULTATIVE:	Default Doctor, doctor	**Validated:**	2/10/2025
Performed by:	Default Tech, tech	**Processed:**	2/10/2025

Instr. : HS12 **Serial Nº:** 100000 **hybriSoft:** MOCK 1.0.0 (HS12a)