HPV Direct Flow Chip Kit

LOTS

PCR:	MOCKP001	4/1/2030
Chips:	MOCKE001.1	9/1/2030
Reagent:	MOCKH001	1/1/2030

SAMPLE DETAILS

ID SAMPLE: 100-00-0004-0004_1-010-0_20240415-1 **SAMPLE TYPE:**

ID PATIENT: **PATIENT:**

SEX: - **BIRTHDATE:** **AGE:**

REPORT

HPV POSITIVE

The sample is positive for the following genotypes: 18.

PROTOCOL

Detection and genotyping of HPV viral DNA by PCR and reverse dot blot hybridization:
- High risk genotypes: 16, 18, 26, 31, 33, 35, 39, 45, 51, 52, 53, 56, 58, 59, 66, 68, 73, 82.
- Low risk genotypes: 6, 11, 40, 42, 43, 44/55, 54, 61, 62/81, 67, 69, 70, 71, 72, 84.

Automatic analysis of results

NOTES

FACULTATIVE: Default Doctor, doctor **Validated:** 2/11/2025

Performed by: Default Tech, tech **Processed:** 2/11/2025

Instr. : HS12 **Serial Nº:** 100000 **hybriSoft:** MOCK 1.0.0 (HS12a)

HPV Direct Flow Chip Kit

LOTS		
PCR:	MOCKP001	4/1/2030
Chips:	MOCKE001.1	9/1/2030
Reagent:	MOCKH001	1/1/2030

SAMPLE DETAILS

ID SAMPLE: 100-00-0004-0004_1-010-0_20240415-1 **SAMPLE TYPE:**

ID PATIENT: **PATIENT:**

SEX: - **BIRTHDATE:** **AGE:**

REPORT

ANALYSIS INFORMATION

- Spot B: Hybridization control (Biotin signals to orientate the CHIP)
- Spot C: Internal DNA control (Genomic human DNA probe)
- Spot U: HPV Universal probe
- Spot #: Genotype specific probes

All the spots are printed in duplicate.

Threshold: 6

NOTES

FACULTATIVE: Default Doctor, doctor **Validated:** 2/11/2025

Performed by: Default Tech, tech **Processed:** 2/11/2025

Instr. : HS12 **Serial Nº:** 100000 **hybriSoft:** MOCK 1.0.0 (HS12a)