

Advanced mDSC software package: helping people unfamiliar with programming to unravel complex mDSC data

Tom Konings^{1*}, Julia Bandera^{2*}, and Guy Van den Mooter¹

¹ Drug Delivery and Disposition, KU Leuven, Department of Pharmaceutical and Pharmacological Sciences, Campus Gasthuisberg ON2, Herestraat 49 b921, 3000 Leuven, Belgium. ² VIB-KU Leuven Center for Brain & Disease Research, Herestraat 49 Box 602, Leuven, 3000, Belgium; Department of Neurosciences, Leuven Brain Institute, KU Leuven, Herestraat 49 Box 602, Leuven, 3000, Belgium. ¶ Corresponding author * These authors contributed equally.

DOI: [10.xxxxxx/draft](https://doi.org/10.xxxxxx/draft)

Software

- [Review](#)
- [Repository](#)
- [Archive](#)

Editor: [Open Journals](#)

Reviewers:

- [@openjournals](#)

Submitted: 01 January 1970

Published: unpublished

License

Authors of papers retain copyright and release the work under a Creative Commons Attribution 4.0 International License ([CC BY 4.0](#)).

Summary

The Advanced mDSC software package is a collection of four user-friendly RShiny apps that helps users without programming knowledge to decode their modulated differential scanning calorimetry (mDSC) data. Differential scanning calorimetry (DSC) is a thermal analysis technique that is commonly used in pharmaceutical and material science to detect the different transitions and thermal events undergone by a material when it is heated. mDSC is an improved version of DSC in the sense that it can be used to deconvolute different events, but it also presents additional complexity and technical difficulty. Moreover, the data generated using this method must be analyzed by computational methods, since it generally involves datasets with hundreds of thousands of datapoints.

Background

Differential scanning calorimetry (DSC) is one of the most common methods to study the thermal properties of materials. (Knopp et al., 2016; Leyva-Porras et al., 2019) It is of crucial importance in polymer chemistry and physics, material science, pharmaceutical science, and so forth. It allows the user to characterize material properties such as glass transitions, crystallization and melting events, solvent evaporation, degradation, or any other detectable event that involves a change in enthalpy or heat capacity. A significant upgrade with respect to unmodulated DSC is modulated DSC, which allows for deconvoluting different signals. Where unmodulated DSC uses a simple constant heating rate, mDSC superimposes a sinusoidal signal. Certain thermal events (such as glass transitions) can react to this faster heating rate, but others (such as most crystallisation events) can not, allowing the user to deconvolute the signal. (Craig & Royall, 1998; Rabel et al., 1999; Reading et al., 1993; Royall et al., 1998) The Advanced mDSC software package groups several sub-apps that help the user gain a better understanding not only of their mDSC data, but also of the technique in general.

Despite the potential of mDSC, the technique has some limitations. Since the technique is more complex, blind interpretation of the results can be risky. A Fourier transform is generally used to deconvolute the data, but this can, in some cases, lead to artifacts in the deconvoluted signals. (Thomas, 2005) Furthermore, an important assumption in mDSC is the steady-state assumption, namely that the material is at complete equilibrium at any point during the heating procedure. However, this assumption is generally not met; since the material is being heated continuously and thermal events occur continuously, it is almost a given that the material is not in equilibrium. It is for this reason that the user might choose to perform a quasi-isothermal

mDSC analysis, where only the sinusoidal temperature modulation is applied for extended periods of time at a fixed average temperature. (Wunderlich, 2005) This results in complex analyses containing hundreds of thousands of data points.

Statement of need

The software presented here tackles the two problems presented above. First, the “Regular modulated DSC deconvolution” application offers different tools and methods for deconvoluting mDSC data, allowing the user to avoid performing a Fourier transform. Furthermore, the “Quasi-Isothermal modulated DSC deconvolution” app helps users to analyse complex quasi-isothermal mDSC data by performing automatic segmentation and heat capacity calculations.

Additionally, it may be useful to predict the result of the deconvolution procedure when different sets of parameters are used. The “Modulated DSC deconvolution simulation” app helps with this. It is merely a mathematical tool (as opposed to a physical model) in the sense that it requires complete knowledge of the thermal events in a material, but it can predict what the deconvoluted signals will look like. This can be used to solve complex problems but can also be used as an educational “look under the hood” to see what really happens in an mDSC deconvolution procedure.

Finally, TRIOS® by TA Instruments is a commonly used software package for the analysis of mDSC results. In order to speed up the analysis of data generated using TRIOS®, the “DSC descriptive statistics” app was included into the overall package as well. It simply combines several tables in an interactive manner, allowing the user to very quickly calculate averages, standard deviations, and relative standard deviations.

All aforementioned applications are currently being used to fully understand very elusive thermal behaviour of a polymer (more precisely, a polyoxazoline). All mDSC runs of this material have consistently produced an impossible peak in the resulting thermograms, but the development of the different applications (especially the first three mentioned above) has allowed us to start to understand its otherwise unexplainable and unique behaviour. This is why the software package was developed; to allow users to quickly and efficiently analyze their mDSC data without any coding knowledge, all while allowing for exporting results in .xlsx format and different image formats. This software package is, to the best of our knowledge, unique in the sense that it is completely open source, user-friendly, and combines powerful calculations that absolutely require code to execute.

State of the field

We are confident that these applications will constitute an improvement when compared to the currently available software. There is commercial software available (TRIOS® by TA instruments, STARe® by Mettler-Toledo, Pyris® by PerkinElmer, and so forth), but these packages are not open source and a license is required to use them. Moreover, they do not necessarily offer the same capabilities as the Advanced mDSC software package. To the best of our knowledge, none of them offers an option comparable to the “Modulated DSC deconvolution simulation” app. Furthermore, quasi-isothermal mDSC data analysis might not be as user-friendly as the “Quasi-Isothermal modulated DSC deconvolution” app, requiring manual data segmentation rather than doing this automatically. The “DSC descriptive statistics” app was specifically built to cover a shortcoming of the TRIOS® software. Finally, it is not clear what the options are for these software packages when it comes to performing mDSC deconvolution without Fourier transform, but it is likely that it is not possible in some of them (it is impossible in TRIOS® for instance). Indeed, it is difficult to find out exactly what these packages can and can not do, since all require a license in order to install them. An open source alternative called pyDSC exists, but this focuses on integrating and correcting mDSC data, and thus the scope is not the same as the Advanced mDSC software package.

Software design

The core principle guiding most of the software's design was its user friendliness, especially towards people without coding experience. This is reflected in the use of a ShinyR UI that helps users to navigate the software, upload files, run analyses, and download results. Special consideration was given to make this as intuitive as in commercial software both for uploads and downloads by using a file picker rather than asking the user to use absolute paths. Another important aspect of this has been to separate the UI for the four apps constituting the software by using the shiny.router package, keeping everything clearly separated. Clear instructions as well as mathematical explanations are given in the documentation, which is included as a pdf in the github but also within the software, once more separated into relevant parts for each app. The software was written in R since it is open-source and has a wide community with many different packages and clear documentation, thus inviting contribution. Furthermore, the code is separated into four folders for the four apps, and then separated into different files for the UI, error handling, general functions, and small specific functions for each app. This, together with common naming conventions across the software, was once more done to further invite contributions.

Research impact statement

The software was initially developed in order to elucidate a physically unexplainable signal on the reversing heat flow when analysing a polymer from the polyoxazoline class, which required the use of all the apps present within the software. This research will be submitted to the Journal of Thermal Analysis and Calorimetry. This being said, the software's value lies in the fact that the four different apps can be used in other contexts as well. For instance, quasi-isothermal mDSC is a more widely used technique that requires software for data analysis (Wunderlich, 2005). The mDSC deconvolution simulator can be used to simulate certain thermal events and investigate them, as was done in the previously mentioned research. The app that deconvolutes mDSC data without performing a Fourier transform can be used to check the validity of a wide variety of mDSC analyses, especially when it comes to the question as to whether enough modulations were present over a thermal event. Furthermore, the software is not tied to any specific file format and requires Excel files to run analyses, which can always be exported using any commercial software. This makes the software compatible with mDSC systems used across research groups to solve any research question that requires advanced mDSC analysis.

Mathematics

The detailed mathematics used within the software package are described in the documentation, as it would be too extensive to give a full overview here. However, three equations are crucial for all applications and are thus mentioned here. When deconvoluting an mDSC signal (also called the modulated heat flow, MHF), this normally results in a total heat flow (THF), a reversing heat flow (RHF), and a non-reversing heat flow (NRHF). The equations used to calculate these three quantities are, respectively,

$$THF = \langle MHF \rangle = \left\langle \frac{dQ}{dt} \right\rangle,$$

$$RHF = -\beta \frac{A_{MHF}}{\frac{2\pi}{T} A_{Temp}},$$

$$NRHF = THF - RHF,$$

where $\frac{dQ}{dt}$ is the modulated heat flow, β is the underlying (constant) heating rate, T is the period of the temperature modulation, and A_{Temp} is the amplitude of the temperature modulation.

AI usage disclosure

No generative AI was used in the design, the mathematics (which are based on literature), the manuscript writing, the testing, and most of the code writing for the app. Some generative AI was used to speed up the writing of small blocks of code or to help find relevant documentation.

Acknowledgments

The authors would like to acknowledge the help from Els Verdonck (TA Instruments) and Guy Van Assche (Vrije Universiteit Brussel) for providing help with the theoretical background of the software. Additionally, this research was funded through an FWO grant (1SH0S24N).

References

- Craig, D. Q. M., & Royall, P. G. (1998). *Pharmaceutical Research*, 15(8), 1152–1153. <https://doi.org/10.1023/a:1011967202972>
- Knopp, M., Löbmann, K., Elder, D., Rades, T., & Holm, R. (2016). Recent advances and potential applications of modulated differential scanning calorimetry (mDSC) in drug development. *European Journal of Pharmaceutical Sciences*, 87, 164–173. <https://doi.org/j.ejps.2015.12.024>
- Leyva-Porras, C., Cruz-Alcantar, P., Espinosa-Solís, V., Martínez-Guerra, E., Piñón-Balderrama, C. I., Compean Martínez, I., & Saavedra-Leos, M. Z. (2019). Application of differential scanning calorimetry (DSC) and modulated differential scanning calorimetry (MDSC) in food and drug industries. *Polymers*, 12(1), 5. <https://doi.org/10.3390/polym12010005>
- Rabel, S. R., Jona, J. A., & Maurin, M. B. (1999). Applications of modulated differential scanning calorimetry in preformulation studies. *Journal of Pharmaceutical and Biomedical Analysis*, 21(2), 339–345. [https://doi.org/10.1016/s0731-7085\(99\)00142-9](https://doi.org/10.1016/s0731-7085(99)00142-9)
- Reading, M., Elliott, D., & Hill, V. L. (1993). A new approach to the calorimetric investigation of physical and chemical transitions. *Journal of Thermal Analysis*, 40(3), 949–955. <https://doi.org/10.1007/bf02546854>
- Royall, P. G., Craig, D. Q. M., & Doherty, C. (1998). *Pharmaceutical Research*, 15(7), 1117–1121. <https://doi.org/10.1023/a:1011902816175>
- Thomas, L. (2005). TA Instruments.
- Wunderlich, B. (2005). *Thermal analysis of polymeric materials*. Springer-Verlag. <https://doi.org/10.1007/b137476>