

Toni Guan
College of Information Technology
Shenyang Institute of Technology
Shenyang, Liaoning, China
guantong490@gmail.com

September 14, 2026

Editor-in-Chief
Biomedical Signal Processing and Control
Editorial Office

Dear Editor,

We are pleased to submit our original research article entitled “**When Does Temperature Scaling Pay Off in Cross-Corpus ECG Transfer? A Multi-Seed Calibration Boundary Study**” for your consideration as a **Full Length Article** in *Biomedical Signal Processing and Control*. This manuscript has not been published elsewhere and is not under consideration by any other journal.

Summary of contributions. Temperature scaling (TS) is the default post-hoc recalibration for deep classifiers, yet whether its in-distribution (ID) benefit transfers to out-of-distribution (OOD) target corpora remains the central open question for cross-institution ECG deployment. We pre-registered and executed a multi-seed calibration boundary study comprising **60 experiments** (6 directed transfer pairs $\times$ 2 architectures $\times$ 5 seeds, fully balanced) on PTB-XL, Chapman-Shaoxing, and CPSC2018+2019, with a patient-level cluster paired bootstrap (the operating CI is BCa at $B = 10,000$, fully re-run on the entire 60-experiment grid) and eight recalibration methods. TS yields a positive OOD calibration benefit $\Delta ECE_{\text{OOD}} = ECE_{\text{raw}}^{\text{OOD}} - ECE_{\text{TS}}^{\text{OOD}}$ in **51/60 = 85.0%** of seed experiments (InceptionTime 27/30, ResNet1D 24/30), while the remaining **9 counter-examples are transparently disclosed** and mechanistically explained case by case (six of nine co-occur with low OOD accuracy; seven of nine with a large ID–OOD accuracy gap). The paper delivers **five independent contributions**: **C1** (Primary, OOD benefit quantification: 51/60 seed support), **C2** (Boundary condition, pre-registered branch triggered: only 23/60 = 38.3% of cells have BCa 95% CIs containing zero, below the $\geq 80\%$ criterion, so the “ID nonzero boundary” branch applies and C1 is reported with this qualifier), **C3** (Methodological, Shapley decomposition validation: 27/27 cells pass at $n = 20,000$), **C4** (Practical, method-selection boundary: TS robust vs. EM prior and Matrix scaling weakest, governed by a prior-likelihood mismatch mechanism), and **C5** (Clinical, discrimination-aware calibration gate: 7/14 strata pass both layers, decoupling calibration effectiveness from clinical deployability).

Why this journal. Cross-corpus ECG transfer calibration is a critical bottleneck for clinical deployment of deep ECG models across institutions with different acquisition protocols, lead sets, and label vocabularies. Our methodological contribution—an explicit boundary-condition characterization of when TS pays off, combined with transparent reporting of nine counter-examples and a pre-registered failure branch—aligns directly with the journal’s scope on biomedical signal processing, clinical machine learning reliability, and reproducible computational methods in medicine.

Transparency statement. All hypotheses in this manuscript are tagged *exploratory + robustness validation*, not confirmatory, consistent with the open-science best practices of Nosek et al. (2018). The nine counter-examples (15.0% of the experiment grid) are disclosed in full in the main text and supplements; no non-supporting result is omitted or post-hoc excluded. Two pre-registration amendments are recorded with reason and date: A1 re-locates the primary endpoint (protocol v2.0 → v2.1-A1) and A2 reduces the confirmatory family from 156 to 12 hypotheses, re-designating the L2 shift levels as exploratory dose-response. The protocol, the revision documents, and the file-level hash manifest are included with the submission as supplementary files; the manifest records the snapshot taken on 2026-09-05. The protocol, both amendments, and a per-file archival-status report are publicly archived at <https://doi.org/10.17605/OSF.IO/53H64>. A five-item honesty declaration (adversarial audit product) is included in the manuscript.

Data and code availability. The full pipeline code, the patient-level split indices, the protocol documents, and the per-experiment result artifacts are already publicly available at <https://github.com/gt17641001169-design/ecg-calibration-transportability>, with trained model checkpoints for the 60-experiment grid provided as release assets of the same repository. The repository will be archived on Zenodo with a DOI upon acceptance. All three ECG corpora are public (PhysioNet/Kaggle).

This is a single-author submission; the author is responsible for all aspects of the work (see the CRediT statement in the manuscript). Highlights, the CRediT contribution statement, a data statement, and the generative-AI declaration are provided with the manuscript. We look forward to your feedback.

Sincerely,

Toni Guan
College of Information Technology
Shenyang Institute of Technology
Shenyang, Liaoning, China
guantong490@gmail.com