

Critical Component Audit – Sensirion AG (Schweiz)

Critical Component Audit Report

Component: Feuchte- und Temperatursensoren (SHT-Serie)
Supplier: Sensirion AG (Schweiz)
Audit Date: 14. September 2023
Auditor: Dipl.-Ing. Stefan Hoffmann

Findings

Finding	Severity	Corrective Action	Due Date
Calibration records incomplete	Minor	Update SOP, re-train staff	1. November 2023
Supplier change notification delay	Observation	Revise supplier agreement	1. Dezember 2023

Conclusion

The supplier Sensirion AG (Schweiz) demonstrates adequate quality management for critical medical device components. Two minor findings require corrective action. Overall qualification status: APPROVED WITH CONDITIONS.