


CALIBRATION CERTIFICATE

Date	06.04.2016
Project No.	31455-1
Mass Flowmeter Tubes	RHM 015 GNT Parallel
Transmitter Power Supply Program	RHE 11 230 VAC M221299 (V1.24)
	Ser. No. RHM-21436 TAG No. -
	Ser. No. Test TAG No. -

calibration results




flow rate [g/min]	error [%]	pulses [pulse/g]
600,00	0,05	10,005
400,00	-0,09	9,991
200,00	0,03	10,003
50,00	-0,11	9,989
8,00	-0,09	9,991

checked by :

controller :

CALIBRATION CERTIFICATE



Date	06.04.2016		
Project No.	31455-1		
Mass Flowmeter Tubes	RHM 015 Parallel	GNT	Ser. No. RHM-21436 TAG No. -
Transmitter	RHE 11		Ser. No. Test
Power Supply	230 VAC		TAG No. -

Program settings at calibration

<u>Setup Menu</u>		<u>Adjust Menu</u>
Sensor	RHM 01	AdjTemp 0 °C
Tubes	Parallel	RefTemp 23,0 °C
MaxTemp	120 °C	AdjDens 0 g/l
FiltArray	64	
Tflow	0,80 s	
FiltBand	off	
CutLimit	1,0 %	
DensCut	off	

Flow Calibration

Int.Gates	20
Scale-mul	2
Scale-div	2212
ZeroPnt	-101
TempCor	3,60 %/100°C

Reference Conditions

Calibration	Medium	Water	Test Stand No. 03	scale No. 04
-------------	--------	-------	-------------------	--------------

Standards: electronic weightscale according ISO 4185 - 1980 (E)
used scales are custody transfer calibrated instruments

Calibration performed with one of the following scales:

scale No.	manufacturer	type	range	serial No.
01	Mettler	KD 1500 / 9	1500 kg	1939101
02	Sartorius	F 150 S	150 kg	35070003
03	Sartorius	Quintix	5100 g	31810591
04	Sartorius	E 1200 S	1200 g	40050016
05	Sartorius	Quintix	5100 g	31810590