

Burkhart 6/26/2014
11879 NE GLENN 4:12:58 PM
WIDING DRIVE Type: CALL
PORTLAND, OR 97220 503-252-9777
Work Order Number 393902 RP
40*496918 0040
WD TUALATIN
17130 SW UP BOONESFERRY R

DURHAM, OR 97224

503-639-6620

Customer Purchase Order: 0626TUAL

Requested Date: 06-26

Location:

Contact: RENEE

Symptoms:

MONTHLY

MFG: MISC

Prod Desc: MONTHLY

Model #: S/N#:

Service Performed:

1. EYE WASH STATION WILL NEED TO ORDER
MISSING CAPS. 2. REPLACED AIR FILTER ON
MICROETCHER 3. WILL SHIP IN ULTRASONIC
CLEANER FOR REPAIR 4. MOUNTED SHARPS
BRACKET IN OPS 6-12 5. OP8 REPLACED HVE
AND SE AS NEEDED 6. ADJUSTED PANO AND
TESTED 7. OP4 CLEAN AND REPLACED BOTH
3WAY BUTTONS

Call Status: Complete Charge

Service Charges

Technician	Vend/Cat
Type	Hrs.Min
4091 DAVID CRONAN	22384450
WD CONTRACT SVC	1.00

Part/Qty	Loc	Price	Total
07301492*40	TRIP CHARGE		
1		19.97	19.97
01202150*40	VALVE ASSY BUTTON, AUTO		
2 91*40		30.54	61.08


Service Tech

Customer

Date 6-26-14

Burkhart Repair Log

Clinic Name: Tualatin

MONTH: June 2014

Date	Equipment	Room Location	Person Noting Problem	Priority: High=H Med=M Low=L	Description	Tech Initials & Date Completed
	Eyewash	Darkroom	Hilda	M	Replacement for missing cap	6-24 Parts needed
	Micro etch filter	Lab	Hilda	M	Please replace filter in microetcher. I have part	6-24
	Ultrasonic	Darkroom	Hilda	M	Please repair broken ultrasonic	6-24
	Sharps Bracket	6-12	Hilda	H	Please mount sharps bracket	6-24
	Highspeed suction	8	Hilda	M	Lever slips in and out. Please replace	6-24
	Pano Machine	Xray Room	Vicki	M-H	X-ray platform does not work. Intermittent. Bite tab does not raise up.	6-24
	Air water tips (both)	4	Hilda	m	Both leak water	6-24

Please look at list on clipboard. Additional repairs may have been added. Also initial and date when the repairs are completed. Thank you

Location: Washinton County, Oregon, United States

[illegible]