
Plate Results Report

16S_qPCR_PL1_20230920_115804.eds



Summary

Property	Details
Bar Code	-
File Name	16S_qPCR_PL1_20230920_115804.eds
Run Start Date/Time	Sep 20, 2023, 11:58:47 AM
Run End Date/Time	Sep 20, 2023, 2:15:22 PM
Run Duration	2 hours, 16 minutes, and 35 seconds
Operator	Admin
Instrument Name	QS7Pro-2778719090068
Instrument Type	QuantStudio™ 7 Pro System
Instrument Serial Number	2778719090068
Block Type	96-Well 0.2-mL
Block Serial Number	2778219050095
Heated Cover Serial Number	2779219090141
PCR Stage/Step Number	Stage 2, Step 3
Quantification Cycle Method	Baseline Threshold
Comment	-
Software Name and Version	Design & Analysis Software v2.7.0
Plugin Name and Version	Primary Analysis v1.8.0, Standard Curve v1.7.0
Analysis Date/Time	Oct 25, 2023, 2:38:24 PM

Well Table

Well	Sample	Target	Task	Cq	Cq Confidence	Amp Score	Amp Status	Cq Threshold	Baseline Start/End
A1	STD_1	Bacteria	Standard	12.965	0.996	3.276	AMP	241,184.978	3-7
A2	STD_1	Bacteria	Standard	12.93	0.995	3.243	AMP	241,184.978	3-7
A3	STD_1	Bacteria	Standard	12.916	0.996	3.25	AMP	241,184.978	3-7
A4	KC1	Bacteria	Unknown	35.259	0.989	3.112	AMP	241,184.978	3-24
A5	KC1	Bacteria	Unknown	36.041	0.985	3.013	AMP	241,184.978	3-25
A6	KC1	Bacteria	Unknown	35.426	0.991	3.109	AMP	241,184.978	3-22
A7	KC9	Bacteria	Unknown	37.54	0.971	2.825	Inconclusive	241,184.978	3-24
A8	KC9	Bacteria	Unknown	37.873	0.972	2.788	Inconclusive	241,184.978	3-24
A9	KC9	Bacteria	Unknown	36.936	0.982	2.904	AMP	241,184.978	3-22
A10	FC1	Bacteria	Unknown	23.107	0.995	3.246	AMP	241,184.978	3-13
A11	FC1	Bacteria	Unknown	23.17	0.996	3.258	AMP	241,184.978	3-14
A12	FC1	Bacteria	Unknown	24.153	0.996	3.242	AMP	241,184.978	3-13
B1	STD_2	Bacteria	Standard	16.669	0.996	3.28	AMP	241,184.978	3-9
B2	STD_2	Bacteria	Standard	17.077	0.995	3.221	AMP	241,184.978	3-9
B3	STD_2	Bacteria	Standard	16.728	0.997	3.26	AMP	241,184.978	3-9
B4	KC2	Bacteria	Unknown	33.898	0.994	3.196	AMP	241,184.978	3-22
B5	KC2	Bacteria	Unknown	34.429	0.993	3.154	AMP	241,184.978	3-24
B6	KC2	Bacteria	Unknown	33.263	0.994	3.22	AMP	241,184.978	3-23
B7	KC10	Bacteria	Unknown	36.497	0.984	2.951	AMP	241,184.978	3-21
B8	KC10	Bacteria	Unknown	37.17	0.975	2.869	Inconclusive	241,184.978	3-24
B9	KC10	Bacteria	Unknown	36.658	0.983	2.941	AMP	241,184.978	3-20
B10	FC2	Bacteria	Unknown	22.938	0.997	3.245	AMP	241,184.978	3-13
B11	FC2	Bacteria	Unknown	23.182	0.995	3.225	AMP	241,184.978	3-14
B12	FC2	Bacteria	Unknown	23.349	0.991	3.25	AMP	241,184.978	3-15
C1	STD_3	Bacteria	Standard	20.183	0.995	3.281	AMP	241,184.978	3-11
C2	STD_3	Bacteria	Standard	20.154	0.995	3.27	AMP	241,184.978	3-10

Well	Sample	Target	Task	Cq	Cq Confidence	Amp Score	Amp Status	Cq Threshold	Baseline Start/End
C3	STD_3	Bacteria	Standard	20.457	0.996	3.251	AMP	241,184.978	3-10
C4	KC3	Bacteria	Unknown	33.363	0.994	3.214	AMP	241,184.978	3-24
C5	KC3	Bacteria	Unknown	33.11	0.994	3.225	AMP	241,184.978	3-19
C6	KC3	Bacteria	Unknown	34.122	0.991	3.179	AMP	241,184.978	3-23
C7	KC11	Bacteria	Unknown	37.491	0.975	2.831	Inconclusive	241,184.978	3-27
C8	KC11	Bacteria	Unknown	36.755	0.986	2.928	AMP	241,184.978	3-23
C9	KC11	Bacteria	Unknown	37.825	0.974	2.79	Inconclusive	241,184.978	3-23
C10	FC3	Bacteria	Unknown	23.137	0.996	3.24	AMP	241,184.978	3-13
C11	FC3	Bacteria	Unknown	23.221	0.996	3.236	AMP	241,184.978	3-14
C12	FC3	Bacteria	Unknown	23.584	0.997	3.229	AMP	241,184.978	3-12
D1	STD_4	Bacteria	Standard	24.062	0.996	3.286	AMP	241,184.978	3-12
D2	STD_4	Bacteria	Standard	24.188	0.995	3.244	AMP	241,184.978	3-12
D3	STD_4	Bacteria	Standard	24.389	0.996	3.244	AMP	241,184.978	3-13
D4	KC4	Bacteria	Unknown	35.1	0.989	3.117	AMP	241,184.978	3-20
D5	KC4	Bacteria	Unknown	36.321	0.985	2.98	AMP	241,184.978	3-23
D6	KC4	Bacteria	Unknown	36.217	0.987	3	AMP	241,184.978	3-22
D7	KC12	Bacteria	Unknown	37.798	0.971	2.794	Inconclusive	241,184.978	3-26
D8	KC12	Bacteria	Unknown	36.128	0.982	2.984	AMP	241,184.978	3-21
D9	KC12	Bacteria	Unknown	35.986	0.987	3.008	AMP	241,184.978	3-22
D10	FC4	Bacteria	Unknown	23.183	0.996	3.233	AMP	241,184.978	3-12
D11	FC4	Bacteria	Unknown	23.232	0.996	3.222	AMP	241,184.978	3-13
D12	FC4	Bacteria	Unknown	23.447	0.996	3.237	AMP	241,184.978	3-14
E1	STD_5	Bacteria	Standard	27.412	0.996	3.279	AMP	241,184.978	3-16
E2	STD_5	Bacteria	Standard	27.571	0.997	3.245	AMP	241,184.978	3-17
E3	STD_5	Bacteria	Standard	27.627	0.996	3.241	AMP	241,184.978	3-17
E4	KC5	Bacteria	Unknown	33.975	0.995	3.181	AMP	241,184.978	3-22
E5	KC5	Bacteria	Unknown	33.53	0.994	3.197	AMP	241,184.978	3-18
E6	KC5	Bacteria	Unknown	33.64	0.994	3.206	AMP	241,184.978	3-22

Well	Sample	Target	Task	Cq	Cq Confidence	Amp Score	Amp Status	Cq Threshold	Baseline Start/End
E7	KC13	Bacteria	Unknown	36.458	0.986	2.959	AMP	241,184.978	3-23
E8	KC13	Bacteria	Unknown	35.833	0.99	3.029	AMP	241,184.978	3-23
E9	KC13	Bacteria	Unknown	35.154	0.989	3.124	AMP	241,184.978	3-21
E10	FC5	Bacteria	Unknown	23.196	0.995	3.226	AMP	241,184.978	3-13
E11	FC5	Bacteria	Unknown	23.298	0.995	3.233	AMP	241,184.978	3-12
E12	FC5	Bacteria	Unknown	23.202	0.996	3.253	AMP	241,184.978	3-14
F1	STD_6	Bacteria	Standard	31.78	0.996	3.255	AMP	241,184.978	3-21
F2	STD_6	Bacteria	Standard	31.389	0.996	3.242	AMP	241,184.978	3-20
F3	STD_6	Bacteria	Standard	31.553	0.996	3.238	AMP	241,184.978	3-17
F4	KC6	Bacteria	Unknown	35.88	0.991	3.022	AMP	241,184.978	3-23
F5	KC6	Bacteria	Unknown	34.432	0.993	3.153	AMP	241,184.978	3-24
F6	KC6	Bacteria	Unknown	35.377	0.991	3.114	AMP	241,184.978	3-24
F7	KC14	Bacteria	Unknown	34.501	0.993	3.131	AMP	241,184.978	3-21
F8	KC14	Bacteria	Unknown	34.87	0.992	3.123	AMP	241,184.978	3-23
F9	KC14	Bacteria	Unknown	35.078	0.987	3.122	AMP	241,184.978	3-23
F10	FC6	Bacteria	Unknown	23.206	0.995	3.227	AMP	241,184.978	3-13
F11	FC6	Bacteria	Unknown	23.391	0.996	3.213	AMP	241,184.978	3-13
F12	FC6	Bacteria	Unknown	23.351	0.996	3.242	AMP	241,184.978	3-14
G1	STD_7	Bacteria	Standard	33.74	0.974	3.215	AMP	241,184.978	3-24
G2	STD_7	Bacteria	Standard	34.26	0.993	3.179	AMP	241,184.978	3-21
G3	STD_7	Bacteria	Standard	34.5	0.993	3.163	AMP	241,184.978	3-22
G4	KC7	Bacteria	Unknown	34.15	0.992	3.145	AMP	241,184.978	3-24
G5	KC7	Bacteria	Unknown	34.58	0.994	3.137	AMP	241,184.978	3-21
G6	KC7	Bacteria	Unknown	34.719	0.994	3.149	AMP	241,184.978	3-21
G7	KC15	Bacteria	Unknown	37.019	0.971	2.876	AMP	241,184.978	3-26
G8	KC15	Bacteria	Unknown	37.842	0.964	2.784	AMP	241,184.978	3-24
G9	KC15	Bacteria	Unknown	38	0.947	2.772	AMP	241,184.978	3-21
G10	FC7	Bacteria	Unknown	23.217	0.995	3.25	AMP	241,184.978	3-13

Well	Sample	Target	Task	Cq	Cq Confidence	Amp Score	Amp Status	Cq Threshold	Baseline Start/End
G11	FC7	Bacteria	Unknown	23.203	0.996	3.251	AMP	241,184.978	3-13
G12	FC7	Bacteria	Unknown	24.65	0.997	3.225	AMP	241,184.978	3-14
H1	NTC	Bacteria	Negative Control	37.163	0.978	2.878	Inconclusive	241,184.978	3-25
H2	NTC	Bacteria	Negative Control	35.183	0.991	3.143	AMP	241,184.978	3-26
H3	NTC	Bacteria	Negative Control	35.371	0.991	3.138	AMP	241,184.978	3-26
H4	KC8	Bacteria	Unknown	34.273	0.991	3.165	AMP	241,184.978	3-21
H5	KC8	Bacteria	Unknown	35.386	0.989	3.08	AMP	241,184.978	3-24
H6	KC8	Bacteria	Unknown	34.402	0.992	3.155	AMP	241,184.978	3-24
H7	KC16	Bacteria	Unknown	37.479	0.971	2.83	AMP	241,184.978	3-26
H8	KC16	Bacteria	Unknown	37.189	0.972	2.86	AMP	241,184.978	3-24
H9	KC16	Bacteria	Unknown	38.663	0.375	2.692	Inconclusive	241,184.978	3-23
H10	FC8	Bacteria	Unknown	23.247	0.995	3.223	AMP	241,184.978	3-13
H11	FC8	Bacteria	Unknown	23.287	0.996	3.242	AMP	241,184.978	3-14
H12	FC8	Bacteria	Unknown	23.86	0.997	3.24	AMP	241,184.978	3-14

Replicate Group Table


Sample	Target	No. of Replicates	Cq Mean	Cq SD
FC1	Bacteria	3	23.476	0.587
FC2	Bacteria	3	23.156	0.207
FC3	Bacteria	3	23.314	0.237
FC4	Bacteria	3	23.287	0.14
FC5	Bacteria	3	23.232	0.057
FC6	Bacteria	3	23.316	0.097
FC7	Bacteria	3	23.69	0.831
FC8	Bacteria	3	23.465	0.343
KC1	Bacteria	3	35.575	0.412
KC2	Bacteria	3	33.864	0.584
KC3	Bacteria	3	33.531	0.527
KC4	Bacteria	3	35.879	0.677
KC5	Bacteria	3	33.715	0.232
KC6	Bacteria	3	35.229	0.735
KC7	Bacteria	3	34.483	0.297
KC8	Bacteria	3	34.687	0.609
KC9	Bacteria	3	37.45	0.475
KC10	Bacteria	3	36.775	0.351
KC11	Bacteria	3	37.357	0.548
KC12	Bacteria	3	36.637	1.008
KC13	Bacteria	3	35.815	0.652
KC14	Bacteria	3	34.816	0.293
KC15	Bacteria	3	37.62	0.527
KC16	Bacteria	3	37.777	0.781
NTC	Bacteria	3	35.906	1.093
STD_1	Bacteria	3	12.937	0.026
STD_2	Bacteria	3	16.825	0.221
STD_3	Bacteria	3	20.265	0.167
STD_4	Bacteria	3	24.213	0.165
STD_5	Bacteria	3	27.537	0.112

Sample	Target	No. of Replicates	Cq Mean	Cq SD
STD_6	Bacteria	3	31.574	0.196
STD_7	Bacteria	3	34.166	0.389


Plate Layout

	1	2	3	4	5	6	7	8	9	10	11	12
A	STD_1 Bacteria (12.965)	STD_1 Bacteria (12.93)	STD_1 Bacteria (12.916)	KC1 Bacteria (35.259)	KC1 Bacteria (36.041)	KC1 Bacteria (35.426)	KC9 Bacteria (37.54)	KC9 Bacteria (37.873)	KC9 Bacteria (36.936)	FC1 Bacteria (23.107)	FC1 Bacteria (23.17)	FC1 Bacteria (24.153)
B	STD_2 Bacteria (16.669)	STD_2 Bacteria (17.077)	STD_2 Bacteria (16.728)	KC2 Bacteria (33.898)	KC2 Bacteria (34.429)	KC2 Bacteria (33.263)	KC10 Bacteria (36.497)	KC10 Bacteria (37.17)	KC10 Bacteria (36.658)	FC2 Bacteria (22.938)	FC2 Bacteria (23.182)	FC2 Bacteria (23.349)
C	STD_3 Bacteria (20.183)	STD_3 Bacteria (20.154)	STD_3 Bacteria (20.457)	KC3 Bacteria (33.363)	KC3 Bacteria (33.11)	KC3 Bacteria (34.122)	KC11 Bacteria (37.491)	KC11 Bacteria (36.755)	KC11 Bacteria (37.825)	FC3 Bacteria (23.137)	FC3 Bacteria (23.221)	FC3 Bacteria (23.584)
D	STD_4 Bacteria (24.062)	STD_4 Bacteria (24.188)	STD_4 Bacteria (24.389)	KC4 Bacteria (35.1)	KC4 Bacteria (36.321)	KC4 Bacteria (36.217)	KC12 Bacteria (37.798)	KC12 Bacteria (36.128)	KC12 Bacteria (35.986)	FC4 Bacteria (23.183)	FC4 Bacteria (23.232)	FC4 Bacteria (23.447)
E	STD_5 Bacteria (27.412)	STD_5 Bacteria (27.571)	STD_5 Bacteria (27.627)	KC5 Bacteria (33.975)	KC5 Bacteria (33.53)	KC5 Bacteria (33.64)	KC13 Bacteria (36.458)	KC13 Bacteria (35.833)	KC13 Bacteria (35.154)	FC5 Bacteria (23.196)	FC5 Bacteria (23.298)	FC5 Bacteria (23.202)
F	STD_6 Bacteria (31.78)	STD_6 Bacteria (31.389)	STD_6 Bacteria (31.553)	KC6 Bacteria (35.88)	KC6 Bacteria (34.432)	KC6 Bacteria (35.377)	KC14 Bacteria (34.501)	KC14 Bacteria (34.87)	KC14 Bacteria (35.078)	FC6 Bacteria (23.206)	FC6 Bacteria (23.391)	FC6 Bacteria (23.351)
G	STD_7 Bacteria (33.74)	STD_7 Bacteria (34.26)	STD_7 Bacteria (34.5)	KC7 Bacteria (34.15)	KC7 Bacteria (34.58)	KC7 Bacteria (34.719)	KC15 Bacteria (37.019)	KC15 Bacteria (37.842)	KC15 Bacteria (38)	FC7 Bacteria (23.217)	FC7 Bacteria (23.203)	FC7 Bacteria (24.65)
H	NTC Bacteria (37.163)	NTC Bacteria (35.183)	NTC Bacteria (35.371)	KC8 Bacteria (34.273)	KC8 Bacteria (35.386)	KC8 Bacteria (34.402)	KC16 Bacteria (37.479)	KC16 Bacteria (37.189)	KC16 Bacteria (38.663)	FC8 Bacteria (23.247)	FC8 Bacteria (23.287)	FC8 Bacteria (23.86)

Amplification Plot (dRn)



Amplification Plot (Rn)



Run Method

Block Type	96-Well 0.2-mL Block
Sample Volume	20.0
Cover Temperature	105.0
Run mode	STANDARD

Stage	Collection Flag	Ramp Rate	Temperature	Hold Time	Starting Cycle	Auto Delta Temperature	Auto Delta Hold Time
Hold Stage	false	1.6°C/sec	95.0°C	600	-	-	-
PCR Stage (40 cycles)	false	1.6°C/sec	95.0°C	30	-	-	-
	true	1.6°C/sec	50.0°C	30	-	-	-
	true	1.6°C/sec	72.0°C	60	-	-	-
Hold Stage	false	1.6°C/sec	72.0°C	420	-	-	-

Primary Analysis Settings

General

PCR Stage/Step	Stage 2, Step 3
Quantification Cycle Method	Baseline Threshold

Target	Auto Threshold	Threshold	Auto Baseline	Baseline Start	Baseline End
DEFAULT	Yes	AUTO	Yes	AUTO	AUTO

QC Alerts

Curve Quality Alert Enabled	No
Results Quality Alert Enabled	Yes

Advanced

Set the Delta-Rn below which curves will be considered Non-Amplified	No
Primary Analysis Variant	N/A
Reduce dye signal crosstalk by algorithm	No

Standard Curve Settings

Standard Curve	On Plate

Standard Curve Results

Well	Sample	Sample Type	Target	Dyes	Task	Cq	Quantity
A1	STD_1	Standard	Bacteria	SYBR	Standard	12.965	2.0000e+7
A2	STD_1	Standard	Bacteria	SYBR	Standard	12.93	2.0000e+7
A3	STD_1	Standard	Bacteria	SYBR	Standard	12.916	2.0000e+7
A4	KC1	Unknown	Bacteria	SYBR	Unknown	35.259	1.3924e+1
A5	KC1	Unknown	Bacteria	SYBR	Unknown	36.041	8.4302e+0
A6	KC1	Unknown	Bacteria	SYBR	Unknown	35.426	1.2503e+1
A7	KC9	Unknown	Bacteria	SYBR	Unknown	37.54	3.2204e+0
A8	KC9	Unknown	Bacteria	SYBR	Unknown	37.873	2.6010e+0
A9	KC9	Unknown	Bacteria	SYBR	Unknown	36.936	4.7444e+0
A10	FC1	Unknown	Bacteria	SYBR	Unknown	23.107	3.3943e+4
A11	FC1	Unknown	Bacteria	SYBR	Unknown	23.17	3.2595e+4
A12	FC1	Unknown	Bacteria	SYBR	Unknown	24.153	1.7345e+4
B1	STD_2	Standard	Bacteria	SYBR	Standard	16.669	2.0000e+6
B2	STD_2	Standard	Bacteria	SYBR	Standard	17.077	2.0000e+6
B3	STD_2	Standard	Bacteria	SYBR	Standard	16.728	2.0000e+6
B4	KC2	Unknown	Bacteria	SYBR	Unknown	33.898	3.3339e+1
B5	KC2	Unknown	Bacteria	SYBR	Unknown	34.429	2.3713e+1
B6	KC2	Unknown	Bacteria	SYBR	Unknown	33.263	5.0106e+1
B7	KC10	Unknown	Bacteria	SYBR	Unknown	36.497	6.2899e+0
B8	KC10	Unknown	Bacteria	SYBR	Unknown	37.17	4.0838e+0
B9	KC10	Unknown	Bacteria	SYBR	Unknown	36.658	5.6734e+0
B10	FC2	Unknown	Bacteria	SYBR	Unknown	22.938	3.7830e+4
B11	FC2	Unknown	Bacteria	SYBR	Unknown	23.182	3.2333e+4
B12	FC2	Unknown	Bacteria	SYBR	Unknown	23.349	2.9055e+4
C1	STD_3	Standard	Bacteria	SYBR	Standard	20.183	2.0000e+5
C2	STD_3	Standard	Bacteria	SYBR	Standard	20.154	2.0000e+5
C3	STD_3	Standard	Bacteria	SYBR	Standard	20.457	2.0000e+5
C4	KC3	Unknown	Bacteria	SYBR	Unknown	33.363	4.7015e+1
C5	KC3	Unknown	Bacteria	SYBR	Unknown	33.11	5.5303e+1
C6	KC3	Unknown	Bacteria	SYBR	Unknown	34.122	2.8877e+1

Well	Sample	Sample Type	Target	Dyes	Task	Cq	Quantity
C7	KC11	Unknown	Bacteria	SYBR	Unknown	37.491	3.3237e+0
C8	KC11	Unknown	Bacteria	SYBR	Unknown	36.755	5.3316e+0
C9	KC11	Unknown	Bacteria	SYBR	Unknown	37.825	2.6823e+0
C10	FC3	Unknown	Bacteria	SYBR	Unknown	23.137	3.3277e+4
C11	FC3	Unknown	Bacteria	SYBR	Unknown	23.221	3.1543e+4
C12	FC3	Unknown	Bacteria	SYBR	Unknown	23.584	2.4992e+4
D1	STD_4	Standard	Bacteria	SYBR	Standard	24.062	2.0000e+4
D2	STD_4	Standard	Bacteria	SYBR	Standard	24.188	2.0000e+4
D3	STD_4	Standard	Bacteria	SYBR	Standard	24.389	2.0000e+4
D4	KC4	Unknown	Bacteria	SYBR	Unknown	35.1	1.5419e+1
D5	KC4	Unknown	Bacteria	SYBR	Unknown	36.321	7.0435e+0
D6	KC4	Unknown	Bacteria	SYBR	Unknown	36.217	7.5262e+0
D7	KC12	Unknown	Bacteria	SYBR	Unknown	37.798	2.7297e+0
D8	KC12	Unknown	Bacteria	SYBR	Unknown	36.128	7.9714e+0
D9	KC12	Unknown	Bacteria	SYBR	Unknown	35.986	8.7305e+0
D10	FC4	Unknown	Bacteria	SYBR	Unknown	23.183	3.2310e+4
D11	FC4	Unknown	Bacteria	SYBR	Unknown	23.232	3.1317e+4
D12	FC4	Unknown	Bacteria	SYBR	Unknown	23.447	2.7280e+4
E1	STD_5	Standard	Bacteria	SYBR	Standard	27.412	2.0000e+3
E2	STD_5	Standard	Bacteria	SYBR	Standard	27.571	2.0000e+3
E3	STD_5	Standard	Bacteria	SYBR	Standard	27.627	2.0000e+3
E4	KC5	Unknown	Bacteria	SYBR	Unknown	33.975	3.1741e+1
E5	KC5	Unknown	Bacteria	SYBR	Unknown	33.53	4.2229e+1
E6	KC5	Unknown	Bacteria	SYBR	Unknown	33.64	3.9359e+1
E7	KC13	Unknown	Bacteria	SYBR	Unknown	36.458	6.4510e+0
E8	KC13	Unknown	Bacteria	SYBR	Unknown	35.833	9.6287e+0
E9	KC13	Unknown	Bacteria	SYBR	Unknown	35.154	1.4892e+1
E10	FC5	Unknown	Bacteria	SYBR	Unknown	23.196	3.2054e+4
E11	FC5	Unknown	Bacteria	SYBR	Unknown	23.298	3.0026e+4
E12	FC5	Unknown	Bacteria	SYBR	Unknown	23.202	3.1933e+4
F1	STD_6	Standard	Bacteria	SYBR	Standard	31.78	2.0000e+2


Well	Sample	Sample Type	Target	Dyes	Task	Cq	Quantity
F2	STD_6	Standard	Bacteria	SYBR	Standard	31.389	2.0000e+2
F3	STD_6	Standard	Bacteria	SYBR	Standard	31.553	2.0000e+2
F4	KC6	Unknown	Bacteria	SYBR	Unknown	35.88	9.3441e+0
F5	KC6	Unknown	Bacteria	SYBR	Unknown	34.432	2.3673e+1
F6	KC6	Unknown	Bacteria	SYBR	Unknown	35.377	1.2910e+1
F7	KC14	Unknown	Bacteria	SYBR	Unknown	34.501	2.2649e+1
F8	KC14	Unknown	Bacteria	SYBR	Unknown	34.87	1.7871e+1
F9	KC14	Unknown	Bacteria	SYBR	Unknown	35.078	1.5633e+1
F10	FC6	Unknown	Bacteria	SYBR	Unknown	23.206	3.1839e+4
F11	FC6	Unknown	Bacteria	SYBR	Unknown	23.391	2.8284e+4
F12	FC6	Unknown	Bacteria	SYBR	Unknown	23.351	2.9005e+4
G1	STD_7	Standard	Bacteria	SYBR	Standard	33.74	2.0000e+1
G2	STD_7	Standard	Bacteria	SYBR	Standard	34.26	2.0000e+1
G3	STD_7	Standard	Bacteria	SYBR	Standard	34.5	2.0000e+1
G4	KC7	Unknown	Bacteria	SYBR	Unknown	34.15	2.8367e+1
G5	KC7	Unknown	Bacteria	SYBR	Unknown	34.58	2.1530e+1
G6	KC7	Unknown	Bacteria	SYBR	Unknown	34.719	1.9686e+1
G7	KC15	Unknown	Bacteria	SYBR	Unknown	37.019	4.5006e+0
G8	KC15	Unknown	Bacteria	SYBR	Unknown	37.842	2.6529e+0
G9	KC15	Unknown	Bacteria	SYBR	Unknown	38	2.3973e+0
G10	FC7	Unknown	Bacteria	SYBR	Unknown	23.217	3.1621e+4
G11	FC7	Unknown	Bacteria	SYBR	Unknown	23.203	3.1898e+4
G12	FC7	Unknown	Bacteria	SYBR	Unknown	24.65	1.2609e+4
H1	NTC	Negative Control	Bacteria	SYBR	Negative Control	37.163	-
H2	NTC	Negative Control	Bacteria	SYBR	Negative Control	35.183	-
H3	NTC	Negative Control	Bacteria	SYBR	Negative Control	35.371	-
H4	KC8	Unknown	Bacteria	SYBR	Unknown	34.273	2.6212e+1
H5	KC8	Unknown	Bacteria	SYBR	Unknown	35.386	1.2831e+1
H6	KC8	Unknown	Bacteria	SYBR	Unknown	34.402	2.4124e+1
H7	KC16	Unknown	Bacteria	SYBR	Unknown	37.479	3.3484e+0
H8	KC16	Unknown	Bacteria	SYBR	Unknown	37.189	4.0332e+0

Well	Sample	Sample Type	Target	Dyes	Task	Cq	Quantity
H9	KC16	Unknown	Bacteria	SYBR	Unknown	38.663	1.5667e+0
H10	FC8	Unknown	Bacteria	SYBR	Unknown	23.247	3.1016e+4
H11	FC8	Unknown	Bacteria	SYBR	Unknown	23.287	3.0240e+4
H12	FC8	Unknown	Bacteria	SYBR	Unknown	23.86	2.0927e+4

Standard Curve Results - Replicate

Sample	Target	Replication Number	Quantity Mean	Quantity SD
FC1	Bacteria	3	2.7961e+4	9.2183e+3
FC2	Bacteria	3	3.3073e+4	4.4337e+3
FC3	Bacteria	3	2.9937e+4	4.3695e+3
FC4	Bacteria	3	3.0302e+4	2.6639e+3
FC5	Bacteria	3	3.1337e+4	1.1376e+3
FC6	Bacteria	3	2.9710e+4	1.8788e+3
FC7	Bacteria	3	2.5376e+4	1.1058e+4
FC8	Bacteria	3	2.7394e+4	5.6138e+3
KC1	Bacteria	3	1.1619e+1	2.8514e+0
KC2	Bacteria	3	3.5719e+1	1.3357e+1
KC3	Bacteria	3	4.3732e+1	1.3516e+1
KC4	Bacteria	3	9.9961e+0	4.7022e+0
KC5	Bacteria	3	3.7776e+1	5.4202e+0
KC6	Bacteria	3	1.5309e+1	7.4594e+0
KC7	Bacteria	3	2.3194e+1	4.5736e+0
KC8	Bacteria	3	2.1056e+1	7.1990e+0
KC9	Bacteria	3	3.5219e+0	1.1031e+0
KC10	Bacteria	3	5.3491e+0	1.1383e+0
KC11	Bacteria	3	3.7792e+0	1.3821e+0
KC12	Bacteria	3	6.4772e+0	3.2675e+0
KC13	Bacteria	3	1.0324e+1	4.2634e+0
KC14	Bacteria	3	1.8718e+1	3.5837e+0
KC15	Bacteria	3	3.1836e+0	1.1477e+0
KC16	Bacteria	3	2.9828e+0	1.2733e+0
NTC	Bacteria	3	-	-

Standard Curve Plot



Target:Bacteria Slope:-3.588 R^2:0.998 Y-Inter:39.362 Eff:89.985 Error:0.040

- End of Report -