
Plate Results Report

16S_qPCR_PL3_20230921_145833.eds



Summary

Property	Details
Bar Code	-
File Name	16S_qPCR_PL3_20230921_145833.eds
Run Start Date/Time	Sep 21, 2023, 2:59:23 PM
Run End Date/Time	Sep 21, 2023, 5:15:59 PM
Run Duration	2 hours, 16 minutes, and 36 seconds
Operator	Admin
Instrument Name	QS7Pro-2778719090068
Instrument Type	QuantStudio™ 7 Pro System
Instrument Serial Number	2778719090068
Block Type	96-Well 0.2-mL
Block Serial Number	2778219050095
Heated Cover Serial Number	2779219090141
PCR Stage/Step Number	Stage 2, Step 3
Quantification Cycle Method	Baseline Threshold
Comment	-
Software Name and Version	Design & Analysis Software v2.7.0
Plugin Name and Version	Primary Analysis v1.8.0, Standard Curve v1.7.0
Analysis Date/Time	Oct 25, 2023, 2:32:31 PM

Well Table

Well	Sample	Target	Task	Cq	Cq Confidence	Amp Score	Amp Status	Cq Threshold	Baseline Start/End
A1	STD_1	Bacteria	Standard	13.378	0.993	3.262	AMP	376,260.839	3-7
A2	STD_1	Bacteria	Standard	13.81	0.993	3.201	AMP	376,260.839	3-7
A3	STD_1	Bacteria	Standard	13.693	0.994	3.211	AMP	376,260.839	3-7
A4	NL147-6	Bacteria	Unknown	26.727	0.994	3.177	AMP	376,260.839	3-15
A5	NL147-6	Bacteria	Unknown	27.031	0.991	3.193	AMP	376,260.839	3-15
A6	NL147-6	Bacteria	Unknown	26.714	0.994	3.201	AMP	376,260.839	3-14
A7	NL100-LUM	Bacteria	Unknown	24.438	0.988	3.173	AMP	376,260.839	3-14
A8	NL100-LUM	Bacteria	Unknown	24.03	0.99	3.172	AMP	376,260.839	3-14
A9	NL100-LUM	Bacteria	Unknown	23.979	0.994	3.176	AMP	376,260.839	3-12
A10	NL149-LUM	Bacteria	Unknown	25.508	0.993	3.173	AMP	376,260.839	3-14
A11	NL149-LUM	Bacteria	Unknown	25.088	0.994	3.242	AMP	376,260.839	3-15
A12	NL149-LUM	Bacteria	Unknown	25.43	0.994	3.234	AMP	376,260.839	3-14
B1	STD_2	Bacteria	Standard	16.77	0.995	3.289	AMP	376,260.839	3-8
B2	STD_2	Bacteria	Standard	17.251	0.992	3.206	AMP	376,260.839	3-8
B3	STD_2	Bacteria	Standard	16.972	0.995	3.248	AMP	376,260.839	3-8
B4	NL149-6	Bacteria	Unknown	34.138	0.986	3.152	AMP	376,260.839	3-21
B5	NL149-6	Bacteria	Unknown	34.937	0.988	3.148	AMP	376,260.839	3-21
B6	NL149-6	Bacteria	Unknown	34.123	0.984	3.17	AMP	376,260.839	3-21
B7	NL103-LUM	Bacteria	Unknown	25.205	0.991	3.177	AMP	376,260.839	3-14
B8	NL103-LUM	Bacteria	Unknown	25.449	0.992	3.182	AMP	376,260.839	3-16
B9	NL103-LUM	Bacteria	Unknown	25.251	0.993	3.213	AMP	376,260.839	3-13
B10	NL150-LUM	Bacteria	Unknown	28.996	0.994	3.195	AMP	376,260.839	3-16
B11	NL150-LUM	Bacteria	Unknown	28.979	0.995	3.171	AMP	376,260.839	3-16
B12	NL150-LUM	Bacteria	Unknown	29.007	0.982	3.202	AMP	376,260.839	3-20
C1	STD_3	Bacteria	Standard	20.448	0.995	3.277	AMP	376,260.839	3-10
C2	STD_3	Bacteria	Standard	20.652	0.995	3.251	AMP	376,260.839	3-10

Well	Sample	Target	Task	Cq	Cq Confidence	Amp Score	Amp Status	Cq Threshold	Baseline Start/End
C3	STD_3	Bacteria	Standard	20.784	0.995	3.241	AMP	376,260.839	3-10
C4	NL150-6	Bacteria	Unknown	31.143	0.99	3.181	AMP	376,260.839	3-15
C5	NL150-6	Bacteria	Unknown	31.615	0.99	3.168	AMP	376,260.839	3-20
C6	NL150-6	Bacteria	Unknown	30.68	0.993	3.202	AMP	376,260.839	3-16
C7	NL113-LUM	Bacteria	Unknown	37.81	0.964	2.881	AMP	376,260.839	3-19
C8	NL113-LUM	Bacteria	Unknown	37.562	0.96	2.888	AMP	376,260.839	3-22
C9	NL113-LUM	Bacteria	Unknown	37.9	0.962	2.863	AMP	376,260.839	3-21
C10	NL153-LUM	Bacteria	Unknown	28.446	0.993	3.194	AMP	376,260.839	3-16
C11	NL153-LUM	Bacteria	Unknown	28.543	0.993	3.188	AMP	376,260.839	3-17
C12	NL153-LUM	Bacteria	Unknown	29.08	0.989	3.171	AMP	376,260.839	3-20
D1	STD_4	Bacteria	Standard	24.322	0.994	3.275	AMP	376,260.839	3-13
D2	STD_4	Bacteria	Standard	24.79	0.994	3.205	AMP	376,260.839	3-15
D3	STD_4	Bacteria	Standard	24.7	0.994	3.217	AMP	376,260.839	3-13
D4	NL153-6	Bacteria	Unknown	22.716	0.993	3.156	AMP	376,260.839	3-11
D5	NL153-6	Bacteria	Unknown	22.557	0.993	3.162	AMP	376,260.839	3-11
D6	NL153-6	Bacteria	Unknown	22.558	0.993	3.179	AMP	376,260.839	3-11
D7	NL126-LUM	Bacteria	Unknown	24.923	0.994	3.184	AMP	376,260.839	3-12
D8	NL126-LUM	Bacteria	Unknown	24.994	0.995	3.2	AMP	376,260.839	3-13
D9	NL126-LUM	Bacteria	Unknown	24.986	0.994	3.187	AMP	376,260.839	3-15
D10	NL155-LUM	Bacteria	Unknown	25.755	0.993	3.175	AMP	376,260.839	3-14
D11	NL155-LUM	Bacteria	Unknown	25.939	0.995	3.174	AMP	376,260.839	3-15
D12	NL155-LUM	Bacteria	Unknown	26.206	0.989	3.121	AMP	376,260.839	3-16
E1	STD_5	Bacteria	Standard	28.062	0.992	3.225	AMP	376,260.839	3-17
E2	STD_5	Bacteria	Standard	28.262	0.992	3.213	AMP	376,260.839	3-16
E3	STD_5	Bacteria	Standard	28.259	0.992	3.202	AMP	376,260.839	3-16
E4	NL155-6	Bacteria	Unknown	14.029	0.989	3.159	AMP	376,260.839	3-7
E5	NL155-6	Bacteria	Unknown	13.948	0.993	3.178	AMP	376,260.839	3-7
E6	NL155-6	Bacteria	Unknown	13.614	0.992	3.206	AMP	376,260.839	3-7

Well	Sample	Target	Task	Cq	Cq Confidence	Amp Score	Amp Status	Cq Threshold	Baseline Start/End
E7	NL130-LUM	Bacteria	Unknown	27.605	0.992	3.169	AMP	376,260.839	3-14
E8	NL130-LUM	Bacteria	Unknown	27.639	0.993	3.189	AMP	376,260.839	3-15
E9	NL130-LUM	Bacteria	Unknown	27.81	0.994	3.191	AMP	376,260.839	3-15
E10	NL157-LUM	Bacteria	Unknown	26.884	0.993	3.17	AMP	376,260.839	3-16
E11	NL157-LUM	Bacteria	Unknown	27.023	0.99	3.159	AMP	376,260.839	3-16
E12	NL157-LUM	Bacteria	Unknown	27.261	0.994	3.206	AMP	376,260.839	3-16
F1	STD_6	Bacteria	Standard	31.795	0.994	3.215	AMP	376,260.839	3-19
F2	STD_6	Bacteria	Standard	31.977	0.994	3.21	AMP	376,260.839	3-18
F3	STD_6	Bacteria	Standard	32.515	0.992	3.201	AMP	376,260.839	3-18
F4	NL157-6	Bacteria	Unknown	31.573	0.991	3.157	AMP	376,260.839	3-15
F5	NL157-6	Bacteria	Unknown	31.227	0.99	3.186	AMP	376,260.839	3-18
F6	NL157-6	Bacteria	Unknown	31.34	0.989	3.16	AMP	376,260.839	3-18
F7	NL136-LUM	Bacteria	Unknown	24.906	0.994	3.157	AMP	376,260.839	3-12
F8	NL136-LUM	Bacteria	Unknown	24.842	0.994	3.169	AMP	376,260.839	3-14
F9	NL136-LUM	Bacteria	Unknown	24.901	0.995	3.181	AMP	376,260.839	3-14
F10	NL161-LUM	Bacteria	Unknown	28.131	0.989	3.151	AMP	376,260.839	3-18
F11	NL161-LUM	Bacteria	Unknown	27.936	0.993	3.142	AMP	376,260.839	3-15
F12	NL161-LUM	Bacteria	Unknown	28.493	0.991	3.171	AMP	376,260.839	3-16
G1	STD_7	Bacteria	Standard	34.674	0.976	3.188	AMP	376,260.839	3-24
G2	STD_7	Bacteria	Standard	36.623	0.982	3.017	AMP	376,260.839	3-20
G3	STD_7	Bacteria	Standard	36.271	0.983	3.1	AMP	376,260.839	3-25
G4	NL161-6	Bacteria	Unknown	31.142	0.991	3.196	AMP	376,260.839	3-19
G5	NL161-6	Bacteria	Unknown	31.119	0.991	3.188	AMP	376,260.839	3-20
G6	NL161-6	Bacteria	Unknown	31.113	0.99	3.198	AMP	376,260.839	3-16
G7	NL144-LUM	Bacteria	Unknown	26.151	0.991	3.189	AMP	376,260.839	3-14
G8	NL144-LUM	Bacteria	Unknown	26.09	0.991	3.18	AMP	376,260.839	3-15
G9	NL144-LUM	Bacteria	Unknown	26.127	0.992	3.201	AMP	376,260.839	3-18
G10	NL164-LUM	Bacteria	Unknown	25.219	0.991	3.213	AMP	376,260.839	3-14

Well	Sample	Target	Task	Cq	Cq Confidence	Amp Score	Amp Status	Cq Threshold	Baseline Start/End
G11	NL164-LUM	Bacteria	Unknown	25.232	0.992	3.21	AMP	376,260.839	3-15
G12	NL164-LUM	Bacteria	Unknown	25.739	0.994	3.195	AMP	376,260.839	3-15
H1	NTC	Bacteria	Negative Control	38.133	0.958	2.853	Inconclusive	376,260.839	3-26
H2	NTC	Bacteria	Negative Control	36.845	0.985	2.987	AMP	376,260.839	3-24
H3	NTC	Bacteria	Negative Control	38.141	0.95	2.843	AMP	376,260.839	3-21
H4	NL164-5	Bacteria	Unknown	24.02	0.985	3.161	AMP	376,260.839	3-12
H5	NL164-5	Bacteria	Unknown	23.92	0.992	3.133	AMP	376,260.839	3-11
H6	NL164-5	Bacteria	Unknown	23.394	0.992	3.197	AMP	376,260.839	3-12
H7	NL147-LUM	Bacteria	Unknown	31.528	0.99	3.15	AMP	376,260.839	3-21
H8	NL147-LUM	Bacteria	Unknown	31.544	0.991	3.148	AMP	376,260.839	3-20
H9	NL147-LUM	Bacteria	Unknown	33.251	0.985	3.106	AMP	376,260.839	3-21
H10	NL100-MUC	Bacteria	Unknown	29.954	0.992	3.139	AMP	376,260.839	3-16
H11	NL100-MUC	Bacteria	Unknown	28.828	0.994	3.204	AMP	376,260.839	3-15
H12	NL100-MUC	Bacteria	Unknown	29.723	0.994	3.21	AMP	376,260.839	3-19

Replicate Group Table


Sample	Target	No. of Replicates	Cq Mean	Cq SD
NL100-LUM	Bacteria	3	24.149	0.251
NL100-MUC	Bacteria	3	29.501	0.595
NL103-LUM	Bacteria	3	25.302	0.13
NL113-LUM	Bacteria	3	37.757	0.175
NL126-LUM	Bacteria	3	24.968	0.039
NL130-LUM	Bacteria	3	27.685	0.11
NL136-LUM	Bacteria	3	24.883	0.036
NL144-LUM	Bacteria	3	26.123	0.031
NL147-6	Bacteria	3	26.824	0.179
NL147-LUM	Bacteria	3	32.108	0.99
NL149-6	Bacteria	3	34.399	0.466
NL149-LUM	Bacteria	3	25.342	0.224
NL150-6	Bacteria	3	31.146	0.467
NL150-LUM	Bacteria	3	28.994	0.014
NL153-6	Bacteria	3	22.61	0.092
NL153-LUM	Bacteria	3	28.69	0.341
NL155-6	Bacteria	3	13.864	0.22
NL155-LUM	Bacteria	3	25.967	0.227
NL157-6	Bacteria	3	31.38	0.176
NL157-LUM	Bacteria	3	27.056	0.191
NL161-6	Bacteria	3	31.125	0.015
NL161-LUM	Bacteria	3	28.186	0.283
NL164-5	Bacteria	3	23.778	0.336
NL164-LUM	Bacteria	3	25.397	0.297
NTC	Bacteria	3	37.706	0.746
STD_1	Bacteria	3	13.627	0.223
STD_2	Bacteria	3	16.998	0.241
STD_3	Bacteria	3	20.628	0.169
STD_4	Bacteria	3	24.604	0.248
STD_5	Bacteria	3	28.194	0.114

Sample	Target	No. of Replicates	Cq Mean	Cq SD
STD_6	Bacteria	3	32.096	0.374
STD_7	Bacteria	3	35.856	1.038


Plate Layout

	1	2	3	4	5	6	7	8	9	10	11	12
A	STD_1 Bacteria (13.378)	STD_1 Bacteria (13.81)	STD_1 Bacteria (13.693)	NL147-6 Bacteria (26.727)	NL147-6 Bacteria (27.031)	NL147-6 Bacteria (26.714)	NL100-LUM Bacteria (24.438)	NL100-LUM Bacteria (24.03)	NL100-LUM Bacteria (23.979)	NL149-LUM Bacteria (25.508)	NL149-LUM Bacteria (25.088)	NL149-LUM Bacteria (25.43)
B	STD_2 Bacteria (16.77)	STD_2 Bacteria (17.251)	STD_2 Bacteria (16.972)	NL149-6 Bacteria (34.138)	NL149-6 Bacteria (34.937)	NL149-6 Bacteria (34.123)	NL103-LUM Bacteria (25.205)	NL103-LUM Bacteria (25.449)	NL103-LUM Bacteria (25.251)	NL150-LUM Bacteria (28.996)	NL150-LUM Bacteria (28.979)	NL150-LUM Bacteria (29.007)
C	STD_3 Bacteria (20.448)	STD_3 Bacteria (20.652)	STD_3 Bacteria (20.784)	NL150-6 Bacteria (31.143)	NL150-6 Bacteria (31.615)	NL150-6 Bacteria (30.68)	NL113-LUM Bacteria (37.81)	NL113-LUM Bacteria (37.562)	NL113-LUM Bacteria (37.9)	NL153-LUM Bacteria (28.446)	NL153-LUM Bacteria (28.543)	NL153-LUM Bacteria (29.08)
D	STD_4 Bacteria (24.322)	STD_4 Bacteria (24.79)	STD_4 Bacteria (24.7)	NL153-6 Bacteria (22.716)	NL153-6 Bacteria (22.557)	NL153-6 Bacteria (22.558)	NL126-LUM Bacteria (24.923)	NL126-LUM Bacteria (24.994)	NL126-LUM Bacteria (24.986)	NL155-LUM Bacteria (25.755)	NL155-LUM Bacteria (25.939)	NL155-LUM Bacteria (26.206)
E	STD_5 Bacteria (28.062)	STD_5 Bacteria (28.262)	STD_5 Bacteria (28.259)	NL155-6 Bacteria (14.029)	NL155-6 Bacteria (13.948)	NL155-6 Bacteria (13.614)	NL130-LUM Bacteria (27.605)	NL130-LUM Bacteria (27.639)	NL130-LUM Bacteria (27.81)	NL157-LUM Bacteria (26.884)	NL157-LUM Bacteria (27.023)	NL157-LUM Bacteria (27.261)
F	STD_6 Bacteria (31.795)	STD_6 Bacteria (31.977)	STD_6 Bacteria (32.515)	NL157-6 Bacteria (31.573)	NL157-6 Bacteria (31.227)	NL157-6 Bacteria (31.34)	NL136-LUM Bacteria (24.906)	NL136-LUM Bacteria (24.842)	NL136-LUM Bacteria (24.901)	NL161-LUM Bacteria (28.131)	NL161-LUM Bacteria (27.936)	NL161-LUM Bacteria (28.493)
G	STD_7 Bacteria (34.674)	STD_7 Bacteria (36.623)	STD_7 Bacteria (36.271)	NL161-6 Bacteria (31.142)	NL161-6 Bacteria (31.119)	NL161-6 Bacteria (31.113)	NL144-LUM Bacteria (26.151)	NL144-LUM Bacteria (26.09)	NL144-LUM Bacteria (26.127)	NL164-LUM Bacteria (25.219)	NL164-LUM Bacteria (25.232)	NL164-LUM Bacteria (25.739)
H	NTC Bacteria (38.133)	NTC Bacteria (36.845)	NTC Bacteria (38.141)	NL164-5 Bacteria (24.02)	NL164-5 Bacteria (23.92)	NL164-5 Bacteria (23.394)	NL147-LUM Bacteria (31.528)	NL147-LUM Bacteria (31.544)	NL147-LUM Bacteria (33.251)	NL100-MUC Bacteria (29.954)	NL100-MUC Bacteria (28.828)	NL100-MUC Bacteria (29.723)

Amplification Plot (dRn)



Amplification Plot (Rn)



Run Method

Block Type	96-Well 0.2-mL Block
Sample Volume	20.0
Cover Temperature	105.0
Run mode	STANDARD

Stage	Collection Flag	Ramp Rate	Temperature	Hold Time	Starting Cycle	Auto Delta Temperature	Auto Delta Hold Time
Hold Stage	false	1.6°C/sec	95.0°C	600	-	-	-
PCR Stage (40 cycles)	false	1.6°C/sec	95.0°C	30	-	-	-
	true	1.6°C/sec	50.0°C	30	-	-	-
	true	1.6°C/sec	72.0°C	60	-	-	-
Hold Stage	false	1.6°C/sec	72.0°C	420	-	-	-

Primary Analysis Settings

General

PCR Stage/Step	Stage 2, Step 3
Quantification Cycle Method	Baseline Threshold

Target	Auto Threshold	Threshold	Auto Baseline	Baseline Start	Baseline End
DEFAULT	Yes	AUTO	Yes	AUTO	AUTO

QC Alerts

Curve Quality Alert Enabled	No
Results Quality Alert Enabled	Yes

Advanced

Set the Delta-Rn below which curves will be considered Non-Amplified	No
Primary Analysis Variant	N/A
Reduce dye signal crosstalk by algorithm	No

Standard Curve Settings

Standard Curve	On Plate

Standard Curve Results

Well	Sample	Sample Type	Target	Dyes	Task	Cq	Quantity
A1	STD_1	Standard	Bacteria	SYBR	Standard	13.378	2.0000e+7
A2	STD_1	Standard	Bacteria	SYBR	Standard	13.81	2.0000e+7
A3	STD_1	Standard	Bacteria	SYBR	Standard	13.693	2.0000e+7
A4	NL147-6	Unknown	Bacteria	SYBR	Unknown	26.727	5.2882e+3
A5	NL147-6	Unknown	Bacteria	SYBR	Unknown	27.031	4.3842e+3
A6	NL147-6	Unknown	Bacteria	SYBR	Unknown	26.714	5.3314e+3
A7	NL100-LUM	Unknown	Bacteria	SYBR	Unknown	24.438	2.1729e+4
A8	NL100-LUM	Unknown	Bacteria	SYBR	Unknown	24.03	2.7946e+4
A9	NL100-LUM	Unknown	Bacteria	SYBR	Unknown	23.979	2.8845e+4
A10	NL149-LUM	Unknown	Bacteria	SYBR	Unknown	25.508	1.1221e+4
A11	NL149-LUM	Unknown	Bacteria	SYBR	Unknown	25.088	1.4548e+4
A12	NL149-LUM	Unknown	Bacteria	SYBR	Unknown	25.43	1.1778e+4
B1	STD_2	Standard	Bacteria	SYBR	Standard	16.77	2.0000e+6
B2	STD_2	Standard	Bacteria	SYBR	Standard	17.251	2.0000e+6
B3	STD_2	Standard	Bacteria	SYBR	Standard	16.972	2.0000e+6
B4	NL149-6	Unknown	Bacteria	SYBR	Unknown	34.138	5.4527e+1
B5	NL149-6	Unknown	Bacteria	SYBR	Unknown	34.937	3.3292e+1
B6	NL149-6	Unknown	Bacteria	SYBR	Unknown	34.123	5.5041e+1
B7	NL103-LUM	Unknown	Bacteria	SYBR	Unknown	25.205	1.3528e+4
B8	NL103-LUM	Unknown	Bacteria	SYBR	Unknown	25.449	1.1636e+4
B9	NL103-LUM	Unknown	Bacteria	SYBR	Unknown	25.251	1.3155e+4
B10	NL150-LUM	Unknown	Bacteria	SYBR	Unknown	28.996	1.3037e+3
B11	NL150-LUM	Unknown	Bacteria	SYBR	Unknown	28.979	1.3168e+3
B12	NL150-LUM	Unknown	Bacteria	SYBR	Unknown	29.007	1.2941e+3
C1	STD_3	Standard	Bacteria	SYBR	Standard	20.448	2.0000e+5
C2	STD_3	Standard	Bacteria	SYBR	Standard	20.652	2.0000e+5
C3	STD_3	Standard	Bacteria	SYBR	Standard	20.784	2.0000e+5
C4	NL150-6	Unknown	Bacteria	SYBR	Unknown	31.143	3.4642e+2
C5	NL150-6	Unknown	Bacteria	SYBR	Unknown	31.615	2.5883e+2
C6	NL150-6	Unknown	Bacteria	SYBR	Unknown	30.68	4.6084e+2

Well	Sample	Sample Type	Target	Dyes	Task	Cq	Quantity
C7	NL113-LUM	Unknown	Bacteria	SYBR	Unknown	37.81	5.6510e+0
C8	NL113-LUM	Unknown	Bacteria	SYBR	Unknown	37.562	6.5882e+0
C9	NL113-LUM	Unknown	Bacteria	SYBR	Unknown	37.9	5.3482e+0
C10	NL153-LUM	Unknown	Bacteria	SYBR	Unknown	28.446	1.8304e+3
C11	NL153-LUM	Unknown	Bacteria	SYBR	Unknown	28.543	1.7236e+3
C12	NL153-LUM	Unknown	Bacteria	SYBR	Unknown	29.08	1.2377e+3
D1	STD_4	Standard	Bacteria	SYBR	Standard	24.322	2.0000e+4
D2	STD_4	Standard	Bacteria	SYBR	Standard	24.79	2.0000e+4
D3	STD_4	Standard	Bacteria	SYBR	Standard	24.7	2.0000e+4
D4	NL153-6	Unknown	Bacteria	SYBR	Unknown	22.716	6.2884e+4
D5	NL153-6	Unknown	Bacteria	SYBR	Unknown	22.557	6.9386e+4
D6	NL153-6	Unknown	Bacteria	SYBR	Unknown	22.558	6.9311e+4
D7	NL126-LUM	Unknown	Bacteria	SYBR	Unknown	24.923	1.6100e+4
D8	NL126-LUM	Unknown	Bacteria	SYBR	Unknown	24.994	1.5414e+4
D9	NL126-LUM	Unknown	Bacteria	SYBR	Unknown	24.986	1.5487e+4
D10	NL155-LUM	Unknown	Bacteria	SYBR	Unknown	25.755	9.6338e+3
D11	NL155-LUM	Unknown	Bacteria	SYBR	Unknown	25.939	8.5991e+3
D12	NL155-LUM	Unknown	Bacteria	SYBR	Unknown	26.206	7.2925e+3
E1	STD_5	Standard	Bacteria	SYBR	Standard	28.062	2.0000e+3
E2	STD_5	Standard	Bacteria	SYBR	Standard	28.262	2.0000e+3
E3	STD_5	Standard	Bacteria	SYBR	Standard	28.259	2.0000e+3
E4	NL155-6	Unknown	Bacteria	SYBR	Unknown	14.029	1.3410e+7
E5	NL155-6	Unknown	Bacteria	SYBR	Unknown	13.948	1.4095e+7
E6	NL155-6	Unknown	Bacteria	SYBR	Unknown	13.614	1.7321e+7
E7	NL130-LUM	Unknown	Bacteria	SYBR	Unknown	27.605	3.0748e+3
E8	NL130-LUM	Unknown	Bacteria	SYBR	Unknown	27.639	3.0125e+3
E9	NL130-LUM	Unknown	Bacteria	SYBR	Unknown	27.81	2.7098e+3
E10	NL157-LUM	Unknown	Bacteria	SYBR	Unknown	26.884	4.7990e+3
E11	NL157-LUM	Unknown	Bacteria	SYBR	Unknown	27.023	4.4045e+3
E12	NL157-LUM	Unknown	Bacteria	SYBR	Unknown	27.261	3.8028e+3
F1	STD_6	Standard	Bacteria	SYBR	Standard	31.795	2.0000e+2


Well	Sample	Sample Type	Target	Dyes	Task	Cq	Quantity
F2	STD_6	Standard	Bacteria	SYBR	Standard	31.977	2.0000e+2
F3	STD_6	Standard	Bacteria	SYBR	Standard	32.515	2.0000e+2
F4	NL157-6	Unknown	Bacteria	SYBR	Unknown	31.573	2.6568e+2
F5	NL157-6	Unknown	Bacteria	SYBR	Unknown	31.227	3.2886e+2
F6	NL157-6	Unknown	Bacteria	SYBR	Unknown	31.34	3.0677e+2
F7	NL136-LUM	Unknown	Bacteria	SYBR	Unknown	24.906	1.6274e+4
F8	NL136-LUM	Unknown	Bacteria	SYBR	Unknown	24.842	1.6931e+4
F9	NL136-LUM	Unknown	Bacteria	SYBR	Unknown	24.901	1.6325e+4
F10	NL161-LUM	Unknown	Bacteria	SYBR	Unknown	28.131	2.2236e+3
F11	NL161-LUM	Unknown	Bacteria	SYBR	Unknown	27.936	2.5081e+3
F12	NL161-LUM	Unknown	Bacteria	SYBR	Unknown	28.493	1.7783e+3
G1	STD_7	Standard	Bacteria	SYBR	Standard	34.674	2.0000e+1
G2	STD_7	Standard	Bacteria	SYBR	Standard	36.623	2.0000e+1
G3	STD_7	Standard	Bacteria	SYBR	Standard	36.271	2.0000e+1
G4	NL161-6	Unknown	Bacteria	SYBR	Unknown	31.142	3.4651e+2
G5	NL161-6	Unknown	Bacteria	SYBR	Unknown	31.119	3.5153e+2
G6	NL161-6	Unknown	Bacteria	SYBR	Unknown	31.113	3.5278e+2
G7	NL144-LUM	Unknown	Bacteria	SYBR	Unknown	26.151	7.5453e+3
G8	NL144-LUM	Unknown	Bacteria	SYBR	Unknown	26.09	7.8339e+3
G9	NL144-LUM	Unknown	Bacteria	SYBR	Unknown	26.127	7.6604e+3
G10	NL164-LUM	Unknown	Bacteria	SYBR	Unknown	25.219	1.3417e+4
G11	NL164-LUM	Unknown	Bacteria	SYBR	Unknown	25.232	1.3305e+4
G12	NL164-LUM	Unknown	Bacteria	SYBR	Unknown	25.739	9.7318e+3
H1	NTC	Negative Control	Bacteria	SYBR	Negative Control	38.133	-
H2	NTC	Negative Control	Bacteria	SYBR	Negative Control	36.845	-
H3	NTC	Negative Control	Bacteria	SYBR	Negative Control	38.141	-
H4	NL164-5	Unknown	Bacteria	SYBR	Unknown	24.02	2.8121e+4
H5	NL164-5	Unknown	Bacteria	SYBR	Unknown	23.92	2.9903e+4
H6	NL164-5	Unknown	Bacteria	SYBR	Unknown	23.394	4.1375e+4
H7	NL147-LUM	Unknown	Bacteria	SYBR	Unknown	31.528	2.7305e+2
H8	NL147-LUM	Unknown	Bacteria	SYBR	Unknown	31.544	2.7040e+2

Well	Sample	Sample Type	Target	Dyes	Task	Cq	Quantity
H9	NL147-LUM	Unknown	Bacteria	SYBR	Unknown	33.251	9.4275e+1
H10	NL100-MUC	Unknown	Bacteria	SYBR	Unknown	29.954	7.2160e+2
H11	NL100-MUC	Unknown	Bacteria	SYBR	Unknown	28.828	1.4460e+3
H12	NL100-MUC	Unknown	Bacteria	SYBR	Unknown	29.723	8.3223e+2

Standard Curve Results - Replicate

Sample	Target	Replication Number	Quantity Mean	Quantity SD
NL100-LUM	Bacteria	3	2.6173e+4	3.8749e+3
NL100-MUC	Bacteria	3	9.9994e+2	3.9022e+2
NL103-LUM	Bacteria	3	1.2773e+4	1.0021e+3
NL113-LUM	Bacteria	3	5.8625e+0	6.4649e-1
NL126-LUM	Bacteria	3	1.5667e+4	3.7664e+2
NL130-LUM	Bacteria	3	2.9324e+3	1.9525e+2
NL136-LUM	Bacteria	3	1.6510e+4	3.6526e+2
NL144-LUM	Bacteria	3	7.6799e+3	1.4530e+2
NL147-6	Bacteria	3	5.0013e+3	5.3485e+2
NL147-LUM	Bacteria	3	2.1258e+2	1.0246e+2
NL149-6	Bacteria	3	4.7620e+1	1.2411e+1
NL149-LUM	Bacteria	3	1.2516e+4	1.7818e+3
NL150-6	Bacteria	3	3.5536e+2	1.0130e+2
NL150-LUM	Bacteria	3	1.3049e+3	1.1363e+1
NL153-6	Bacteria	3	6.7194e+4	3.7324e+3
NL153-LUM	Bacteria	3	1.5972e+3	3.1588e+2
NL155-6	Bacteria	3	1.4942e+7	2.0883e+6
NL155-LUM	Bacteria	3	8.5084e+3	1.1733e+3
NL157-6	Bacteria	3	3.0043e+2	3.2065e+1
NL157-LUM	Bacteria	3	4.3354e+3	5.0166e+2
NL161-6	Bacteria	3	3.5027e+2	3.3214e+0
NL161-LUM	Bacteria	3	2.1700e+3	3.6783e+2
NL164-5	Bacteria	3	3.3133e+4	7.1933e+3
NL164-LUM	Bacteria	3	1.2151e+4	2.0962e+3
NTC	Bacteria	3	-	-

Standard Curve Plot



Target:Bacteria Slope:-3.730 R^2:0.997 Y-Inter:40.616 Eff:85.384 Error:0.045

- End of Report -