
Plate Results Report

16S_qPCR_PL4_20230922_095926.eds



Summary

Property	Details
Bar Code	-
File Name	16S_qPCR_PL4_20230922_095926.eds
Run Start Date/Time	Sep 22, 2023, 9:59:39 AM
Run End Date/Time	Sep 22, 2023, 12:16:13 PM
Run Duration	2 hours, 16 minutes, and 33 seconds
Operator	Admin
Instrument Name	QS7Pro-2778719090068
Instrument Type	QuantStudio™ 7 Pro System
Instrument Serial Number	2778719090068
Block Type	96-Well 0.2-mL
Block Serial Number	2778219050095
Heated Cover Serial Number	2779219090141
PCR Stage/Step Number	Stage 2, Step 3
Quantification Cycle Method	Baseline Threshold
Comment	-
Software Name and Version	Design & Analysis Software v2.7.0
Plugin Name and Version	Primary Analysis v1.8.0, Standard Curve v1.7.0
Analysis Date/Time	Oct 25, 2023, 2:41:08 PM

Well Table

Well	Sample	Target	Task	Cq	Cq Confidence	Amp Score	Amp Status	Cq Threshold	Baseline Start/End
A1	STD_1	Bacteria	Standard	12.394	0.996	3.271	AMP	148,597.876	3-7
A2	STD_1	Bacteria	Standard	12.264	0.996	3.239	AMP	148,597.876	3-7
A3	STD_1	Bacteria	Standard	12.285	0.997	3.264	AMP	148,597.876	3-7
A4	NL103-MUC	Bacteria	Unknown	30.465	0.997	3.212	AMP	148,597.876	3-20
A5	NL103-MUC	Bacteria	Unknown	30.193	0.997	3.232	AMP	148,597.876	3-22
A6	NL103-MUC	Bacteria	Unknown	30.52	0.997	3.214	AMP	148,597.876	3-19
A7	NL150-MUC	Bacteria	Unknown	38.109	0.947	2.653	NO_AMP	148,597.876	3-27
A8	NL150-MUC	Bacteria	Unknown	37.313	0.978	2.807	NO_AMP	148,597.876	3-25
A9	NL150-MUC	Bacteria	Unknown	37.636	0.972	2.709	Inconclusive	148,597.876	3-26
A10	NL107-6	Bacteria	Unknown	31.999	0.997	3.208	AMP	148,597.876	3-20
A11	NL107-6	Bacteria	Unknown	31.98	0.998	3.205	AMP	148,597.876	3-23
A12	NL107-6	Bacteria	Unknown	32.731	0.997	3.229	AMP	148,597.876	3-22
B1	STD_2	Bacteria	Standard	15.576	0.998	3.281	AMP	148,597.876	3-8
B2	STD_2	Bacteria	Standard	16.112	0.997	3.233	AMP	148,597.876	3-9
B3	STD_2	Bacteria	Standard	15.865	0.998	3.253	AMP	148,597.876	3-9
B4	NL113-MUC	Bacteria	Unknown	34.383	0.994	3.131	AMP	148,597.876	3-22
B5	NL113-MUC	Bacteria	Unknown	34.858	0.995	3.052	AMP	148,597.876	3-22
B6	NL113-MUC	Bacteria	Unknown	34.328	0.994	3.137	AMP	148,597.876	3-25
B7	NL153-MUC	Bacteria	Unknown	38.936	0.072	2.562	NO_AMP	148,597.876	3-29
B8	NL153-MUC	Bacteria	Unknown	38.796	0.234	2.576	NO_AMP	148,597.876	3-28
B9	NL153-MUC	Bacteria	Unknown	38.107	0.951	2.653	NO_AMP	148,597.876	3-24
B10	NL118-6	Bacteria	Unknown	36.145	0.986	2.904	AMP	148,597.876	3-24
B11	NL118-6	Bacteria	Unknown	36.726	0.987	2.832	AMP	148,597.876	3-23
B12	NL118-6	Bacteria	Unknown	36.777	0.962	2.837	AMP	148,597.876	3-28
C1	STD_3	Bacteria	Standard	18.927	0.998	3.276	AMP	148,597.876	3-10
C2	STD_3	Bacteria	Standard	19.187	0.998	3.27	AMP	148,597.876	3-10

Well	Sample	Target	Task	Cq	Cq Confidence	Amp Score	Amp Status	Cq Threshold	Baseline Start/End
C3	STD_3	Bacteria	Standard	19.306	0.998	3.25	AMP	148,597.876	3-11
C4	NL126-MUC	Bacteria	Unknown	Undetermined	-	2.267	NO_AMP	148,597.876	3-24
C5	NL126-MUC	Bacteria	Unknown	Undetermined	-	2.387	NO_AMP	148,597.876	3-24
C6	NL126-MUC	Bacteria	Unknown	Undetermined	-	2.411	NO_AMP	148,597.876	3-28
C7	NL155-MUC	Bacteria	Unknown	39.251	-	2.564	NO_AMP	148,597.876	3-27
C8	NL155-MUC	Bacteria	Unknown	Undetermined	-	2.42	NO_AMP	148,597.876	3-27
C9	NL155-MUC	Bacteria	Unknown	Undetermined	-	2.346	NO_AMP	148,597.876	3-27
C10	NL123-6	Bacteria	Unknown	31.558	0.997	3.217	AMP	148,597.876	3-20
C11	NL123-6	Bacteria	Unknown	31.722	0.997	3.21	AMP	148,597.876	3-18
C12	NL123-6	Bacteria	Unknown	31.976	0.997	3.2	AMP	148,597.876	3-23
D1	STD_4	Bacteria	Standard	22.973	0.998	3.289	AMP	148,597.876	3-14
D2	STD_4	Bacteria	Standard	23.457	0.998	3.26	AMP	148,597.876	3-12
D3	STD_4	Bacteria	Standard	23.263	0.997	3.254	AMP	148,597.876	3-12
D4	NL130-MUC	Bacteria	Unknown	31.984	0.997	3.188	AMP	148,597.876	3-18
D5	NL130-MUC	Bacteria	Unknown	31.967	0.998	3.209	AMP	148,597.876	3-20
D6	NL130-MUC	Bacteria	Unknown	31.409	0.997	3.199	AMP	148,597.876	3-18
D7	NL157-MUC	Bacteria	Unknown	35.009	0.991	3.065	AMP	148,597.876	3-25
D8	NL157-MUC	Bacteria	Unknown	35.671	0.989	2.943	AMP	148,597.876	3-24
D9	NL157-MUC	Bacteria	Unknown	36.611	0.982	2.837	AMP	148,597.876	3-25
D10	NL124-6	Bacteria	Unknown	25.259	0.997	3.214	AMP	148,597.876	3-16
D11	NL124-6	Bacteria	Unknown	25.472	0.997	3.19	AMP	148,597.876	3-15
D12	NL124-6	Bacteria	Unknown	26.058	0.996	3.195	AMP	148,597.876	3-15
E1	STD_5	Bacteria	Standard	26.65	0.998	3.272	AMP	148,597.876	3-18
E2	STD_5	Bacteria	Standard	26.871	0.998	3.258	AMP	148,597.876	3-17
E3	STD_5	Bacteria	Standard	26.765	0.998	3.261	AMP	148,597.876	3-16
E4	NL136-MUC	Bacteria	Unknown	32.468	0.996	3.193	AMP	148,597.876	3-19
E5	NL136-MUC	Bacteria	Unknown	31.828	0.997	3.234	AMP	148,597.876	3-22
E6	NL136-MUC	Bacteria	Unknown	32.088	0.994	3.218	AMP	148,597.876	3-24

Well	Sample	Target	Task	Cq	Cq Confidence	Amp Score	Amp Status	Cq Threshold	Baseline Start/End
E7	NL161-MUC	Bacteria	Unknown	33.382	0.995	3.169	AMP	148,597.876	3-21
E8	NL161-MUC	Bacteria	Unknown	38.28	0.952	2.665	NO_AMP	148,597.876	3-22
E9	NL161-MUC	Bacteria	Unknown	38	0.949	2.661	NO_AMP	148,597.876	3-26
E10	NL141-6	Bacteria	Unknown	30.714	0.996	3.2	AMP	148,597.876	3-21
E11	NL141-6	Bacteria	Unknown	30.682	0.996	3.189	AMP	148,597.876	3-21
E12	NL141-6	Bacteria	Unknown	30.669	0.997	3.224	AMP	148,597.876	3-19
F1	STD_6	Bacteria	Standard	30.734	0.997	3.26	AMP	148,597.876	3-20
F2	STD_6	Bacteria	Standard	30.658	0.995	3.256	AMP	148,597.876	3-17
F3	STD_6	Bacteria	Standard	30.815	0.998	3.252	AMP	148,597.876	3-21
F4	NL144-MUC	Bacteria	Unknown	32.574	0.997	3.212	AMP	148,597.876	3-22
F5	NL144-MUC	Bacteria	Unknown	31.998	0.998	3.227	AMP	148,597.876	3-20
F6	NL144-MUC	Bacteria	Unknown	32.2	0.996	3.212	AMP	148,597.876	3-22
F7	NL164-MUC	Bacteria	Unknown	37.079	0.973	2.775	AMP	148,597.876	3-26
F8	NL164-MUC	Bacteria	Unknown	37.056	0.975	2.779	Inconclusive	148,597.876	3-24
F9	NL164-MUC	Bacteria	Unknown	36.974	0.984	2.798	Inconclusive	148,597.876	3-21
F10	NL143-6	Bacteria	Unknown	27.466	0.996	3.205	AMP	148,597.876	3-15
F11	NL143-6	Bacteria	Unknown	27.627	0.996	3.173	AMP	148,597.876	3-17
F12	NL143-6	Bacteria	Unknown	27.828	0.997	3.207	AMP	148,597.876	3-18
G1	STD_7	Bacteria	Standard	33.537	0.967	3.218	AMP	148,597.876	3-26
G2	STD_7	Bacteria	Standard	33.913	0.997	3.165	AMP	148,597.876	3-22
G3	STD_7	Bacteria	Standard	33.533	0.996	3.18	AMP	148,597.876	3-22
G4	NL147-MUC	Bacteria	Unknown	35.915	0.99	2.927	AMP	148,597.876	3-18
G5	NL147-MUC	Bacteria	Unknown	35.645	0.991	2.962	AMP	148,597.876	3-22
G6	NL147-MUC	Bacteria	Unknown	36.064	0.985	2.911	AMP	148,597.876	3-21
G7	NL093-6	Bacteria	Unknown	30.539	0.997	3.21	AMP	148,597.876	3-20
G8	NL093-6	Bacteria	Unknown	30.435	0.997	3.198	AMP	148,597.876	3-21
G9	NL093-6	Bacteria	Unknown	30.285	0.997	3.207	AMP	148,597.876	3-20
G10	NL146-6	Bacteria	Unknown	19.363	0.996	3.218	AMP	148,597.876	3-11

Well	Sample	Target	Task	Cq	Cq Confidence	Amp Score	Amp Status	Cq Threshold	Baseline Start/End
G11	NL146-6	Bacteria	Unknown	19.608	0.997	3.207	AMP	148,597.876	3-11
G12	NL146-6	Bacteria	Unknown	19.885	0.998	3.206	AMP	148,597.876	3-12
H1	NTC	Bacteria	Negative Control	38.467	0.786	2.668	NO_AMP	148,597.876	3-29
H2	NTC	Bacteria	Negative Control	35.965	0.994	2.977	NO_AMP	148,597.876	3-27
H3	NTC	Bacteria	Negative Control	37.669	0.932	2.72	NO_AMP	148,597.876	3-27
H4	NL149-MUC	Bacteria	Unknown	33.378	0.995	3.18	AMP	148,597.876	3-19
H5	NL149-MUC	Bacteria	Unknown	33.48	0.994	3.148	AMP	148,597.876	3-18
H6	NL149-MUC	Bacteria	Unknown	32.553	0.996	3.204	AMP	148,597.876	3-19
H7	NL094-6	Bacteria	Unknown	32.225	0.995	3.203	AMP	148,597.876	3-22
H8	NL094-6	Bacteria	Unknown	32.474	0.996	3.199	AMP	148,597.876	3-20
H9	NL094-6	Bacteria	Unknown	32.637	0.995	3.186	AMP	148,597.876	3-23
H10	NL151-6	Bacteria	Unknown	29.061	0.996	3.201	AMP	148,597.876	3-18
H11	NL151-6	Bacteria	Unknown	28.853	0.997	3.198	AMP	148,597.876	3-20
H12	NL151-6	Bacteria	Unknown	29.526	0.997	3.197	AMP	148,597.876	3-20

Replicate Group Table


Sample	Target	No. of Replicates	Cq Mean	Cq SD
NL093-6	Bacteria	3	30.42	0.127
NL094-6	Bacteria	3	32.445	0.207
NL103-MUC	Bacteria	3	30.393	0.175
NL107-6	Bacteria	3	32.237	0.428
NL113-MUC	Bacteria	3	34.523	0.292
NL118-6	Bacteria	3	36.55	0.351
NL123-6	Bacteria	3	31.752	0.211
NL124-6	Bacteria	3	25.596	0.414
NL126-MUC	Bacteria	3	-	-
NL130-MUC	Bacteria	3	31.787	0.328
NL136-MUC	Bacteria	3	32.128	0.322
NL141-6	Bacteria	3	30.688	0.023
NL143-6	Bacteria	3	27.64	0.182
NL144-MUC	Bacteria	3	32.258	0.292
NL146-6	Bacteria	3	19.619	0.261
NL147-MUC	Bacteria	3	35.875	0.212
NL149-MUC	Bacteria	3	33.137	0.509
NL150-MUC	Bacteria	3	37.636	-
NL151-6	Bacteria	3	29.147	0.345
NL153-MUC	Bacteria	3	-	-
NL155-MUC	Bacteria	3	-	-
NL157-MUC	Bacteria	3	35.764	0.805
NL161-MUC	Bacteria	3	33.382	-
NL164-MUC	Bacteria	3	37.036	0.055
NTC	Bacteria	3	-	-
STD_1	Bacteria	3	12.314	0.07
STD_2	Bacteria	3	15.851	0.269
STD_3	Bacteria	3	19.14	0.194
STD_4	Bacteria	3	23.231	0.244
STD_5	Bacteria	3	26.762	0.111

Sample	Target	No. of Replicates	Cq Mean	Cq SD
STD_6	Bacteria	3	30.736	0.079
STD_7	Bacteria	3	33.661	0.218


Plate Layout

	1	2	3	4	5	6	7	8	9	10	11	12
A	STD_1 Bacteria (12.394)	STD_1 Bacteria (12.264)	STD_1 Bacteria (12.285)	NL103-MUC Bacteria (30.465)	NL103-MUC Bacteria (30.193)	NL103-MUC Bacteria (30.52)	NL150-MUC Bacteria (38.109)	NL150-MUC Bacteria (37.313)	NL150-MUC Bacteria (37.636)	NL107-6 Bacteria (31.999)	NL107-6 Bacteria (31.98)	NL107-6 Bacteria (32.731)
B	STD_2 Bacteria (15.576)	STD_2 Bacteria (16.112)	STD_2 Bacteria (15.865)	NL113-MUC Bacteria (34.383)	NL113-MUC Bacteria (34.858)	NL113-MUC Bacteria (34.328)	NL153-MUC Bacteria (38.936)	NL153-MUC Bacteria (38.796)	NL153-MUC Bacteria (38.107)	NL118-6 Bacteria (36.145)	NL118-6 Bacteria (36.726)	NL118-6 Bacteria (36.777)
C	STD_3 Bacteria (18.927)	STD_3 Bacteria (19.187)	STD_3 Bacteria (19.306)	NL126-MUC Bacteria (n/a)	NL126-MUC Bacteria (n/a)	NL126-MUC Bacteria (n/a)	NL155-MUC Bacteria (39.251)	NL155-MUC Bacteria (n/a)	NL155-MUC Bacteria (n/a)	NL123-6 Bacteria (31.558)	NL123-6 Bacteria (31.722)	NL123-6 Bacteria (31.976)
D	STD_4 Bacteria (22.973)	STD_4 Bacteria (23.457)	STD_4 Bacteria (23.263)	NL130-MUC Bacteria (31.984)	NL130-MUC Bacteria (31.967)	NL130-MUC Bacteria (31.409)	NL157-MUC Bacteria (35.009)	NL157-MUC Bacteria (35.671)	NL157-MUC Bacteria (36.611)	NL124-6 Bacteria (25.259)	NL124-6 Bacteria (25.472)	NL124-6 Bacteria (26.058)
E	STD_5 Bacteria (26.65)	STD_5 Bacteria (26.871)	STD_5 Bacteria (26.765)	NL136-MUC Bacteria (32.468)	NL136-MUC Bacteria (31.828)	NL136-MUC Bacteria (32.088)	NL161-MUC Bacteria (33.382)	NL161-MUC Bacteria (38.28)	NL161-MUC Bacteria (38)	NL141-6 Bacteria (30.714)	NL141-6 Bacteria (30.682)	NL141-6 Bacteria (30.669)
F	STD_6 Bacteria (30.734)	STD_6 Bacteria (30.658)	STD_6 Bacteria (30.815)	NL144-MUC Bacteria (32.574)	NL144-MUC Bacteria (31.998)	NL144-MUC Bacteria (32.2)	NL164-MUC Bacteria (37.079)	NL164-MUC Bacteria (37.056)	NL164-MUC Bacteria (36.974)	NL143-6 Bacteria (27.466)	NL143-6 Bacteria (27.627)	NL143-6 Bacteria (27.828)
G	STD_7 Bacteria (33.537)	STD_7 Bacteria (33.913)	STD_7 Bacteria (33.533)	NL147-MUC Bacteria (35.915)	NL147-MUC Bacteria (35.645)	NL147-MUC Bacteria (36.064)	NL093-6 Bacteria (30.539)	NL093-6 Bacteria (30.435)	NL093-6 Bacteria (30.285)	NL146-6 Bacteria (19.363)	NL146-6 Bacteria (19.608)	NL146-6 Bacteria (19.885)
H	NTC Bacteria (38.467)	NTC Bacteria (35.965)	NTC Bacteria (37.669)	NL149-MUC Bacteria (33.378)	NL149-MUC Bacteria (33.48)	NL149-MUC Bacteria (32.553)	NL094-6 Bacteria (32.225)	NL094-6 Bacteria (32.474)	NL094-6 Bacteria (32.637)	NL151-6 Bacteria (29.061)	NL151-6 Bacteria (28.853)	NL151-6 Bacteria (29.526)

Amplification Plot (dRn)



Amplification Plot (Rn)



Run Method

Block Type	96-Well 0.2-mL Block
Sample Volume	20.0
Cover Temperature	105.0
Run mode	STANDARD

Stage	Collection Flag	Ramp Rate	Temperature	Hold Time	Starting Cycle	Auto Delta Temperature	Auto Delta Hold Time
Hold Stage	false	1.6°C/sec	95.0°C	600	-	-	-
PCR Stage (40 cycles)	false	1.6°C/sec	95.0°C	30	-	-	-
	true	1.6°C/sec	50.0°C	30	-	-	-
	true	1.6°C/sec	72.0°C	60	-	-	-
Hold Stage	false	1.6°C/sec	72.0°C	420	-	-	-

Primary Analysis Settings

General

PCR Stage/Step	Stage 2, Step 3
Quantification Cycle Method	Baseline Threshold

Target	Auto Threshold	Threshold	Auto Baseline	Baseline Start	Baseline End
DEFAULT	Yes	AUTO	Yes	AUTO	AUTO

QC Alerts

Curve Quality Alert Enabled	No
Results Quality Alert Enabled	Yes

Advanced

Set the Delta-Rn below which curves will be considered Non-Amplified	No
Primary Analysis Variant	N/A
Reduce dye signal crosstalk by algorithm	No

Standard Curve Settings

Standard Curve	On Plate

Standard Curve Results

Well	Sample	Sample Type	Target	Dyes	Task	Cq	Quantity
A1	STD_1	Standard	Bacteria	SYBR	Standard	12.394	2.0000e+7
A2	STD_1	Standard	Bacteria	SYBR	Standard	12.264	2.0000e+7
A3	STD_1	Standard	Bacteria	SYBR	Standard	12.285	2.0000e+7
A4	NL103-MUC	Unknown	Bacteria	SYBR	Unknown	30.465	1.8527e+2
A5	NL103-MUC	Unknown	Bacteria	SYBR	Unknown	30.193	2.2017e+2
A6	NL103-MUC	Unknown	Bacteria	SYBR	Unknown	30.52	1.7890e+2
A7	NL150-MUC	Unknown	Bacteria	SYBR	Unknown	38.109	1.4376e+0
A8	NL150-MUC	Unknown	Bacteria	SYBR	Unknown	37.313	2.3840e+0
A9	NL150-MUC	Unknown	Bacteria	SYBR	Unknown	37.636	1.9421e+0
A10	NL107-6	Unknown	Bacteria	SYBR	Unknown	31.999	6.9871e+1
A11	NL107-6	Unknown	Bacteria	SYBR	Unknown	31.98	7.0725e+1
A12	NL107-6	Unknown	Bacteria	SYBR	Unknown	32.731	4.3874e+1
B1	STD_2	Standard	Bacteria	SYBR	Standard	15.576	2.0000e+6
B2	STD_2	Standard	Bacteria	SYBR	Standard	16.112	2.0000e+6
B3	STD_2	Standard	Bacteria	SYBR	Standard	15.865	2.0000e+6
B4	NL113-MUC	Unknown	Bacteria	SYBR	Unknown	34.383	1.5356e+1
B5	NL113-MUC	Unknown	Bacteria	SYBR	Unknown	34.858	1.1350e+1
B6	NL113-MUC	Unknown	Bacteria	SYBR	Unknown	34.328	1.5901e+1
B7	NL153-MUC	Unknown	Bacteria	SYBR	Unknown	38.936	8.4967e-1
B8	NL153-MUC	Unknown	Bacteria	SYBR	Unknown	38.796	9.2876e-1
B9	NL153-MUC	Unknown	Bacteria	SYBR	Unknown	38.107	1.4393e+0
B10	NL118-6	Unknown	Bacteria	SYBR	Unknown	36.145	5.0087e+0
B11	NL118-6	Unknown	Bacteria	SYBR	Unknown	36.726	3.4619e+0
B12	NL118-6	Unknown	Bacteria	SYBR	Unknown	36.777	3.3516e+0
C1	STD_3	Standard	Bacteria	SYBR	Standard	18.927	2.0000e+5
C2	STD_3	Standard	Bacteria	SYBR	Standard	19.187	2.0000e+5
C3	STD_3	Standard	Bacteria	SYBR	Standard	19.306	2.0000e+5
C4	NL126-MUC	Unknown	Bacteria	SYBR	Unknown	-	-
C5	NL126-MUC	Unknown	Bacteria	SYBR	Unknown	-	-
C6	NL126-MUC	Unknown	Bacteria	SYBR	Unknown	-	-

Well	Sample	Sample Type	Target	Dyes	Task	Cq	Quantity
C7	NL155-MUC	Unknown	Bacteria	SYBR	Unknown	39.251	6.9550e-1
C8	NL155-MUC	Unknown	Bacteria	SYBR	Unknown	-	-
C9	NL155-MUC	Unknown	Bacteria	SYBR	Unknown	-	-
C10	NL123-6	Unknown	Bacteria	SYBR	Unknown	31.558	9.2473e+1
C11	NL123-6	Unknown	Bacteria	SYBR	Unknown	31.722	8.3334e+1
C12	NL123-6	Unknown	Bacteria	SYBR	Unknown	31.976	7.0880e+1
D1	STD_4	Standard	Bacteria	SYBR	Standard	22.973	2.0000e+4
D2	STD_4	Standard	Bacteria	SYBR	Standard	23.457	2.0000e+4
D3	STD_4	Standard	Bacteria	SYBR	Standard	23.263	2.0000e+4
D4	NL130-MUC	Unknown	Bacteria	SYBR	Unknown	31.984	7.0522e+1
D5	NL130-MUC	Unknown	Bacteria	SYBR	Unknown	31.967	7.1283e+1
D6	NL130-MUC	Unknown	Bacteria	SYBR	Unknown	31.409	1.0168e+2
D7	NL157-MUC	Unknown	Bacteria	SYBR	Unknown	35.009	1.0312e+1
D8	NL157-MUC	Unknown	Bacteria	SYBR	Unknown	35.671	6.7686e+0
D9	NL157-MUC	Unknown	Bacteria	SYBR	Unknown	36.611	3.7253e+0
D10	NL124-6	Unknown	Bacteria	SYBR	Unknown	25.259	5.0689e+3
D11	NL124-6	Unknown	Bacteria	SYBR	Unknown	25.472	4.4260e+3
D12	NL124-6	Unknown	Bacteria	SYBR	Unknown	26.058	3.0506e+3
E1	STD_5	Standard	Bacteria	SYBR	Standard	26.65	2.0000e+3
E2	STD_5	Standard	Bacteria	SYBR	Standard	26.871	2.0000e+3
E3	STD_5	Standard	Bacteria	SYBR	Standard	26.765	2.0000e+3
E4	NL136-MUC	Unknown	Bacteria	SYBR	Unknown	32.468	5.1843e+1
E5	NL136-MUC	Unknown	Bacteria	SYBR	Unknown	31.828	7.7906e+1
E6	NL136-MUC	Unknown	Bacteria	SYBR	Unknown	32.088	6.6037e+1
E7	NL161-MUC	Unknown	Bacteria	SYBR	Unknown	33.382	2.9010e+1
E8	NL161-MUC	Unknown	Bacteria	SYBR	Unknown	38.28	1.2893e+0
E9	NL161-MUC	Unknown	Bacteria	SYBR	Unknown	38	1.5406e+0
E10	NL141-6	Unknown	Bacteria	SYBR	Unknown	30.714	1.5816e+2
E11	NL141-6	Unknown	Bacteria	SYBR	Unknown	30.682	1.6140e+2
E12	NL141-6	Unknown	Bacteria	SYBR	Unknown	30.669	1.6269e+2
F1	STD_6	Standard	Bacteria	SYBR	Standard	30.734	2.0000e+2


Well	Sample	Sample Type	Target	Dyes	Task	Cq	Quantity
F2	STD_6	Standard	Bacteria	SYBR	Standard	30.658	2.0000e+2
F3	STD_6	Standard	Bacteria	SYBR	Standard	30.815	2.0000e+2
F4	NL144-MUC	Unknown	Bacteria	SYBR	Unknown	32.574	4.8464e+1
F5	NL144-MUC	Unknown	Bacteria	SYBR	Unknown	31.998	6.9916e+1
F6	NL144-MUC	Unknown	Bacteria	SYBR	Unknown	32.2	6.1469e+1
F7	NL164-MUC	Unknown	Bacteria	SYBR	Unknown	37.079	2.7671e+0
F8	NL164-MUC	Unknown	Bacteria	SYBR	Unknown	37.056	2.8064e+0
F9	NL164-MUC	Unknown	Bacteria	SYBR	Unknown	36.974	2.9580e+0
F10	NL143-6	Unknown	Bacteria	SYBR	Unknown	27.466	1.2466e+3
F11	NL143-6	Unknown	Bacteria	SYBR	Unknown	27.627	1.1252e+3
F12	NL143-6	Unknown	Bacteria	SYBR	Unknown	27.828	9.8997e+2
G1	STD_7	Standard	Bacteria	SYBR	Standard	33.537	2.0000e+1
G2	STD_7	Standard	Bacteria	SYBR	Standard	33.913	2.0000e+1
G3	STD_7	Standard	Bacteria	SYBR	Standard	33.533	2.0000e+1
G4	NL147-MUC	Unknown	Bacteria	SYBR	Unknown	35.915	5.7991e+0
G5	NL147-MUC	Unknown	Bacteria	SYBR	Unknown	35.645	6.8812e+0
G6	NL147-MUC	Unknown	Bacteria	SYBR	Unknown	36.064	5.2752e+0
G7	NL093-6	Unknown	Bacteria	SYBR	Unknown	30.539	1.7678e+2
G8	NL093-6	Unknown	Bacteria	SYBR	Unknown	30.435	1.8879e+2
G9	NL093-6	Unknown	Bacteria	SYBR	Unknown	30.285	2.0764e+2
G10	NL146-6	Unknown	Bacteria	SYBR	Unknown	19.363	2.1504e+5
G11	NL146-6	Unknown	Bacteria	SYBR	Unknown	19.608	1.8402e+5
G12	NL146-6	Unknown	Bacteria	SYBR	Unknown	19.885	1.5431e+5
H1	NTC	Negative Control	Bacteria	SYBR	Negative Control	38.467	-
H2	NTC	Negative Control	Bacteria	SYBR	Negative Control	35.965	-
H3	NTC	Negative Control	Bacteria	SYBR	Negative Control	37.669	-
H4	NL149-MUC	Unknown	Bacteria	SYBR	Unknown	33.378	2.9088e+1
H5	NL149-MUC	Unknown	Bacteria	SYBR	Unknown	33.48	2.7247e+1
H6	NL149-MUC	Unknown	Bacteria	SYBR	Unknown	32.553	4.9143e+1
H7	NL094-6	Unknown	Bacteria	SYBR	Unknown	32.225	6.0512e+1
H8	NL094-6	Unknown	Bacteria	SYBR	Unknown	32.474	5.1657e+1

Well	Sample	Sample Type	Target	Dyes	Task	Cq	Quantity
H9	NL094-6	Unknown	Bacteria	SYBR	Unknown	32.637	4.6588e+1
H10	NL151-6	Unknown	Bacteria	SYBR	Unknown	29.061	4.5230e+2
H11	NL151-6	Unknown	Bacteria	SYBR	Unknown	28.853	5.1608e+2
H12	NL151-6	Unknown	Bacteria	SYBR	Unknown	29.526	3.3651e+2

Standard Curve Results - Replicate

Sample	Target	Replication Number	Quantity Mean	Quantity SD
NL093-6	Bacteria	3	1.9107e+2	1.5559e+1
NL094-6	Bacteria	3	5.2919e+1	7.0469e+0
NL103-MUC	Bacteria	3	1.9478e+2	2.2215e+1
NL107-6	Bacteria	3	6.1490e+1	1.5262e+1
NL113-MUC	Bacteria	3	1.4202e+1	2.4853e+0
NL118-6	Bacteria	3	3.9408e+0	9.2654e-1
NL123-6	Bacteria	3	8.2229e+1	1.0839e+1
NL124-6	Bacteria	3	4.1818e+3	1.0311e+3
NL126-MUC	Bacteria	3	-	-
NL130-MUC	Bacteria	3	8.1161e+1	1.7772e+1
NL136-MUC	Bacteria	3	6.5262e+1	1.3049e+1
NL141-6	Bacteria	3	1.6075e+2	2.3325e+0
NL143-6	Bacteria	3	1.1206e+3	1.2837e+2
NL144-MUC	Bacteria	3	5.9950e+1	1.0807e+1
NL146-6	Bacteria	3	1.8446e+5	3.0365e+4
NL147-MUC	Bacteria	3	5.9852e+0	8.1902e-1
NL149-MUC	Bacteria	3	3.5159e+1	1.2145e+1
NL150-MUC	Bacteria	3	1.9213e+0	4.7353e-1
NL151-6	Bacteria	3	4.3497e+2	9.1031e+1
NL153-MUC	Bacteria	3	1.0726e+0	3.2005e-1
NL155-MUC	Bacteria	3	-	-
NL157-MUC	Bacteria	3	6.9353e+0	3.2966e+0
NL161-MUC	Bacteria	3	1.0613e+1	1.5933e+1
NL164-MUC	Bacteria	3	2.8439e+0	1.0082e-1
NTC	Bacteria	3	-	-

Standard Curve Plot



Target:Bacteria Slope:-3.623 R^2:0.999 Y-Inter:38.680 Eff:88.821 Error:0.032

- End of Report -