
Plate Results Report

16S_qPCR_PL5_20230922_122043.eds



Summary

Property	Details
Bar Code	-
File Name	16S_qPCR_PL5_20230922_122043.eds
Run Start Date/Time	Sep 22, 2023, 12:21:07 PM
Run End Date/Time	Sep 22, 2023, 2:37:45 PM
Run Duration	2 hours, 16 minutes, and 38 seconds
Operator	Admin
Instrument Name	QS7Pro-2778719090068
Instrument Type	QuantStudio™ 7 Pro System
Instrument Serial Number	2778719090068
Block Type	96-Well 0.2-mL
Block Serial Number	2778219050095
Heated Cover Serial Number	2779219090141
PCR Stage/Step Number	Stage 2, Step 3
Quantification Cycle Method	Baseline Threshold
Comment	-
Software Name and Version	Design & Analysis Software v2.7.0
Plugin Name and Version	Primary Analysis v1.8.0, Standard Curve v1.7.0
Analysis Date/Time	Oct 25, 2023, 2:45:27 PM

Well Table

Well	Sample	Target	Task	Cq	Cq Confidence	Amp Score	Amp Status	Cq Threshold	Baseline Start/End
A1	STD_1	Bacteria	Standard	12.2	0.995	3.261	AMP	174,010.243	3-7
A2	STD_1	Bacteria	Standard	12.471	0.99	3.232	AMP	174,010.243	3-7
A3	STD_1	Bacteria	Standard	12.371	0.996	3.246	AMP	174,010.243	3-7
A4	NL152-6	Bacteria	Unknown	25.414	0.995	3.175	AMP	174,010.243	3-14
A5	NL152-6	Bacteria	Unknown	24.904	0.998	3.229	AMP	174,010.243	3-14
A6	NL152-6	Bacteria	Unknown	25.189	0.995	3.211	AMP	174,010.243	3-16
A7	NL124-LUM	Bacteria	Unknown	23.849	0.997	3.227	AMP	174,010.243	3-15
A8	NL124-LUM	Bacteria	Unknown	23.466	0.996	3.206	AMP	174,010.243	3-15
A9	NL124-LUM	Bacteria	Unknown	23.294	0.997	3.227	AMP	174,010.243	3-13
A10	NL093-MUC	Bacteria	Unknown	34.313	0.993	3.134	AMP	174,010.243	3-22
A11	NL093-MUC	Bacteria	Unknown	34.656	0.994	3.135	AMP	174,010.243	3-24
A12	NL093-MUC	Bacteria	Unknown	35.414	0.988	3.017	AMP	174,010.243	3-25
B1	STD_2	Bacteria	Standard	15.671	0.998	3.271	AMP	174,010.243	3-8
B2	STD_2	Bacteria	Standard	16.123	0.996	3.219	AMP	174,010.243	3-8
B3	STD_2	Bacteria	Standard	15.839	0.998	3.246	AMP	174,010.243	3-8
B4	NL158-6	Bacteria	Unknown	27.768	0.997	3.207	AMP	174,010.243	3-16
B5	NL158-6	Bacteria	Unknown	27.467	0.997	3.227	AMP	174,010.243	3-17
B6	NL158-6	Bacteria	Unknown	27.466	0.996	3.226	AMP	174,010.243	3-16
B7	NL141-LUM	Bacteria	Unknown	25.885	0.997	3.22	AMP	174,010.243	3-15
B8	NL141-LUM	Bacteria	Unknown	24.969	0.997	3.226	AMP	174,010.243	3-15
B9	NL141-LUM	Bacteria	Unknown	24.903	0.997	3.25	AMP	174,010.243	3-15
B10	NL094-MUC	Bacteria	Unknown	31.594	0.995	3.21	AMP	174,010.243	3-21
B11	NL094-MUC	Bacteria	Unknown	31.709	0.996	3.194	AMP	174,010.243	3-21
B12	NL094-MUC	Bacteria	Unknown	31.709	0.99	3.217	AMP	174,010.243	3-23
C1	STD_3	Bacteria	Standard	19.355	0.997	3.28	AMP	174,010.243	3-10
C2	STD_3	Bacteria	Standard	19.518	0.997	3.25	AMP	174,010.243	3-10

Well	Sample	Target	Task	Cq	Cq Confidence	Amp Score	Amp Status	Cq Threshold	Baseline Start/End
C3	STD_3	Bacteria	Standard	19.57	0.997	3.244	AMP	174,010.243	3-10
C4	NL165-6	Bacteria	Unknown	26.844	0.997	3.207	AMP	174,010.243	3-16
C5	NL165-6	Bacteria	Unknown	26.773	0.996	3.216	AMP	174,010.243	3-17
C6	NL165-6	Bacteria	Unknown	26.653	0.996	3.221	AMP	174,010.243	3-17
C7	NL143-LUM	Bacteria	Unknown	22.652	0.997	3.24	AMP	174,010.243	3-13
C8	NL143-LUM	Bacteria	Unknown	22.399	0.996	3.221	AMP	174,010.243	3-13
C9	NL143-LUM	Bacteria	Unknown	22.327	0.997	3.229	AMP	174,010.243	3-13
C10	NL107-MUC	Bacteria	Unknown	33.169	0.994	3.21	AMP	174,010.243	3-20
C11	NL107-MUC	Bacteria	Unknown	32.661	0.996	3.202	AMP	174,010.243	3-22
C12	NL107-MUC	Bacteria	Unknown	33.212	0.994	3.192	AMP	174,010.243	3-22
D1	STD_4	Bacteria	Standard	22.941	0.998	3.274	AMP	174,010.243	3-14
D2	STD_4	Bacteria	Standard	23.227	0.997	3.245	AMP	174,010.243	3-12
D3	STD_4	Bacteria	Standard	23.268	0.997	3.239	AMP	174,010.243	3-13
D4	NL093-LUM	Bacteria	Unknown	28.901	0.997	3.192	AMP	174,010.243	3-16
D5	NL093-LUM	Bacteria	Unknown	28.492	0.996	3.198	AMP	174,010.243	3-17
D6	NL093-LUM	Bacteria	Unknown	28.54	0.997	3.208	AMP	174,010.243	3-17
D7	NL146-LUM	Bacteria	Unknown	22.839	0.996	3.21	AMP	174,010.243	3-13
D8	NL146-LUM	Bacteria	Unknown	21.926	0.996	3.199	AMP	174,010.243	3-11
D9	NL146-LUM	Bacteria	Unknown	22.298	0.996	3.204	AMP	174,010.243	3-11
D10	NL118-MUC	Bacteria	Unknown	37.366	0.955	2.739	AMP	174,010.243	3-22
D11	NL118-MUC	Bacteria	Unknown	37.629	0.961	2.716	AMP	174,010.243	3-24
D12	NL118-MUC	Bacteria	Unknown	38	0.947	2.739	Inconclusive	174,010.243	3-27
E1	STD_5	Bacteria	Standard	26.675	0.996	3.272	AMP	174,010.243	3-16
E2	STD_5	Bacteria	Standard	26.826	0.998	3.246	AMP	174,010.243	3-14
E3	STD_5	Bacteria	Standard	26.849	0.998	3.249	AMP	174,010.243	3-18
E4	NL094-LUM	Bacteria	Unknown	22.937	0.997	3.213	AMP	174,010.243	3-14
E5	NL094-LUM	Bacteria	Unknown	22.665	0.997	3.227	AMP	174,010.243	3-12
E6	NL094-LUM	Bacteria	Unknown	22.705	0.997	3.238	AMP	174,010.243	3-12

Well	Sample	Target	Task	Cq	Cq Confidence	Amp Score	Amp Status	Cq Threshold	Baseline Start/End
E7	NL151-LUM	Bacteria	Unknown	28.523	0.996	3.179	AMP	174,010.243	3-17
E8	NL151-LUM	Bacteria	Unknown	28.103	0.995	3.206	AMP	174,010.243	3-18
E9	NL151-LUM	Bacteria	Unknown	27.916	0.997	3.195	AMP	174,010.243	3-16
E10	NL123-MUC	Bacteria	Unknown	33.065	0.991	3.163	AMP	174,010.243	3-17
E11	NL123-MUC	Bacteria	Unknown	32.846	0.994	3.171	AMP	174,010.243	3-19
E12	NL123-MUC	Bacteria	Unknown	33.161	0.994	3.191	AMP	174,010.243	3-21
F1	STD_6	Bacteria	Standard	30.48	0.997	3.252	AMP	174,010.243	3-21
F2	STD_6	Bacteria	Standard	30.651	0.997	3.247	AMP	174,010.243	3-19
F3	STD_6	Bacteria	Standard	30.283	0.996	3.244	AMP	174,010.243	3-17
F4	NL107-LUM	Bacteria	Unknown	23.116	0.996	3.213	AMP	174,010.243	3-15
F5	NL107-LUM	Bacteria	Unknown	22.903	0.998	3.227	AMP	174,010.243	3-12
F6	NL107-LUM	Bacteria	Unknown	22.93	0.998	3.219	AMP	174,010.243	3-14
F7	NL152-LUM	Bacteria	Unknown	24.534	0.996	3.192	AMP	174,010.243	3-14
F8	NL152-LUM	Bacteria	Unknown	24.438	0.996	3.204	AMP	174,010.243	3-16
F9	NL152-LUM	Bacteria	Unknown	24.523	0.997	3.221	AMP	174,010.243	3-14
F10	NL124-MUC	Bacteria	Unknown	31.973	0.995	3.166	AMP	174,010.243	3-17
F11	NL124-MUC	Bacteria	Unknown	31.815	0.995	3.148	AMP	174,010.243	3-17
F12	NL124-MUC	Bacteria	Unknown	32.263	0.994	3.157	AMP	174,010.243	3-20
G1	STD_7	Bacteria	Standard	33.326	0.995	3.209	AMP	174,010.243	3-23
G2	STD_7	Bacteria	Standard	33.529	0.995	3.176	AMP	174,010.243	3-21
G3	STD_7	Bacteria	Standard	34.106	0.993	3.164	AMP	174,010.243	3-22
G4	NL118-LUM	Bacteria	Unknown	25.566	0.997	3.197	AMP	174,010.243	3-13
G5	NL118-LUM	Bacteria	Unknown	25.402	0.997	3.197	AMP	174,010.243	3-15
G6	NL118-LUM	Bacteria	Unknown	25.124	0.996	3.21	AMP	174,010.243	3-14
G7	NL158-LUM	Bacteria	Unknown	22.645	0.997	3.194	AMP	174,010.243	3-13
G8	NL158-LUM	Bacteria	Unknown	22.153	0.995	3.181	AMP	174,010.243	3-12
G9	NL158-LUM	Bacteria	Unknown	22.131	0.995	3.204	AMP	174,010.243	3-13
G10	NL141-MUC	Bacteria	Unknown	34.532	0.987	3.107	AMP	174,010.243	3-24

Well	Sample	Target	Task	Cq	Cq Confidence	Amp Score	Amp Status	Cq Threshold	Baseline Start/End
G11	NL141-MUC	Bacteria	Unknown	34.05	0.992	3.125	AMP	174,010.243	3-24
G12	NL141-MUC	Bacteria	Unknown	34.536	0.994	3.105	AMP	174,010.243	3-24
H1	NTC	Bacteria	Negative Control	35.489	0.991	3.043	AMP	174,010.243	3-25
H2	NTC	Bacteria	Negative Control	36.782	0.985	2.867	Inconclusive	174,010.243	3-23
H3	NTC	Bacteria	Negative Control	27.961	0.092	1.813	Inconclusive	174,010.243	3-20
H4	NL123-LUM	Bacteria	Unknown	26.888	0.995	3.142	AMP	174,010.243	3-16
H5	NL123-LUM	Bacteria	Unknown	26.349	0.995	3.139	AMP	174,010.243	3-15
H6	NL123-LUM	Bacteria	Unknown	25.975	0.997	3.194	AMP	174,010.243	3-16
H7	NL165-LUM	Bacteria	Unknown	24	0.997	3.211	AMP	174,010.243	3-12
H8	NL165-LUM	Bacteria	Unknown	23.888	0.997	3.198	AMP	174,010.243	3-13
H9	NL165-LUM	Bacteria	Unknown	24.008	0.996	3.207	AMP	174,010.243	3-12
H10	NL143-MUC	Bacteria	Unknown	32.465	0.995	3.185	AMP	174,010.243	3-22
H11	NL143-MUC	Bacteria	Unknown	32.748	0.996	3.194	AMP	174,010.243	3-23
H12	NL143-MUC	Bacteria	Unknown	32.514	0.996	3.187	AMP	174,010.243	3-19

Replicate Group Table


Sample	Target	No. of Replicates	Cq Mean	Cq SD
NL093-LUM	Bacteria	3	28.645	0.224
NL093-MUC	Bacteria	3	34.794	0.563
NL094-LUM	Bacteria	3	22.769	0.147
NL094-MUC	Bacteria	3	31.671	0.066
NL107-LUM	Bacteria	3	22.983	0.116
NL107-MUC	Bacteria	3	33.014	0.306
NL118-LUM	Bacteria	3	25.364	0.223
NL118-MUC	Bacteria	3	37.665	0.319
NL123-LUM	Bacteria	3	26.404	0.459
NL123-MUC	Bacteria	3	33.024	0.161
NL124-LUM	Bacteria	3	23.536	0.284
NL124-MUC	Bacteria	3	32.017	0.227
NL141-LUM	Bacteria	3	25.253	0.549
NL141-MUC	Bacteria	3	34.373	0.279
NL143-LUM	Bacteria	3	22.459	0.17
NL143-MUC	Bacteria	3	32.575	0.151
NL146-LUM	Bacteria	3	22.354	0.459
NL151-LUM	Bacteria	3	28.181	0.31
NL152-6	Bacteria	3	25.169	0.256
NL152-LUM	Bacteria	3	24.498	0.053
NL158-6	Bacteria	3	27.567	0.174
NL158-LUM	Bacteria	3	22.31	0.291
NL165-6	Bacteria	3	26.757	0.096
NL165-LUM	Bacteria	3	23.965	0.067
NTC	Bacteria	3	33.411	4.763
STD_1	Bacteria	3	12.348	0.137
STD_2	Bacteria	3	15.878	0.228
STD_3	Bacteria	3	19.481	0.112
STD_4	Bacteria	3	23.145	0.178
STD_5	Bacteria	3	26.784	0.094

Sample	Target	No. of Replicates	Cq Mean	Cq SD
STD_6	Bacteria	3	30.471	0.184
STD_7	Bacteria	3	33.653	0.405


Plate Layout

	1	2	3	4	5	6	7	8	9	10	11	12
A	STD_1 Bacteria (12.2)	STD_1 Bacteria (12.471)	STD_1 Bacteria (12.371)	NL152-6 Bacteria (25.414)	NL152-6 Bacteria (24.904)	NL152-6 Bacteria (25.189)	NL124-LUM Bacteria (23.849)	NL124-LUM Bacteria (23.466)	NL124-LUM Bacteria (23.294)	NL093-MUC Bacteria (34.313)	NL093-MUC Bacteria (34.656)	NL093-MUC Bacteria (35.414)
B	STD_2 Bacteria (15.671)	STD_2 Bacteria (16.123)	STD_2 Bacteria (15.839)	NL158-6 Bacteria (27.768)	NL158-6 Bacteria (27.467)	NL158-6 Bacteria (27.466)	NL141-LUM Bacteria (25.885)	NL141-LUM Bacteria (24.969)	NL141-LUM Bacteria (24.903)	NL094-MUC Bacteria (31.594)	NL094-MUC Bacteria (31.709)	NL094-MUC Bacteria (31.709)
C	STD_3 Bacteria (19.355)	STD_3 Bacteria (19.518)	STD_3 Bacteria (19.57)	NL165-6 Bacteria (26.844)	NL165-6 Bacteria (26.773)	NL165-6 Bacteria (26.653)	NL143-LUM Bacteria (22.652)	NL143-LUM Bacteria (22.399)	NL143-LUM Bacteria (22.327)	NL107-MUC Bacteria (33.169)	NL107-MUC Bacteria (32.661)	NL107-MUC Bacteria (33.212)
D	STD_4 Bacteria (22.941)	STD_4 Bacteria (23.227)	STD_4 Bacteria (23.268)	NL093-LUM Bacteria (28.901)	NL093-LUM Bacteria (28.492)	NL093-LUM Bacteria (28.54)	NL146-LUM Bacteria (22.839)	NL146-LUM Bacteria (21.926)	NL146-LUM Bacteria (22.298)	NL118-MUC Bacteria (37.366)	NL118-MUC Bacteria (37.629)	NL118-MUC Bacteria (38)
E	STD_5 Bacteria (26.675)	STD_5 Bacteria (26.826)	STD_5 Bacteria (26.849)	NL094-LUM Bacteria (22.937)	NL094-LUM Bacteria (22.665)	NL094-LUM Bacteria (22.705)	NL151-LUM Bacteria (28.523)	NL151-LUM Bacteria (28.103)	NL151-LUM Bacteria (27.916)	NL123-MUC Bacteria (33.065)	NL123-MUC Bacteria (32.846)	NL123-MUC Bacteria (33.161)
F	STD_6 Bacteria (30.48)	STD_6 Bacteria (30.651)	STD_6 Bacteria (30.283)	NL107-LUM Bacteria (23.116)	NL107-LUM Bacteria (22.903)	NL107-LUM Bacteria (22.93)	NL152-LUM Bacteria (24.534)	NL152-LUM Bacteria (24.438)	NL152-LUM Bacteria (24.523)	NL124-MUC Bacteria (31.973)	NL124-MUC Bacteria (31.815)	NL124-MUC Bacteria (32.263)
G	STD_7 Bacteria (33.326)	STD_7 Bacteria (33.529)	STD_7 Bacteria (34.106)	NL118-LUM Bacteria (25.566)	NL118-LUM Bacteria (25.402)	NL118-LUM Bacteria (25.124)	NL158-LUM Bacteria (22.645)	NL158-LUM Bacteria (22.153)	NL158-LUM Bacteria (22.131)	NL141-MUC Bacteria (34.532)	NL141-MUC Bacteria (34.05)	NL141-MUC Bacteria (34.536)
H	NTC Bacteria (35.489)	NTC Bacteria (36.782)	NTC Bacteria (27.961)	NL123-LUM Bacteria (26.888)	NL123-LUM Bacteria (26.349)	NL123-LUM Bacteria (25.975)	NL165-LUM Bacteria (24)	NL165-LUM Bacteria (23.888)	NL165-LUM Bacteria (24.008)	NL143-MUC Bacteria (32.465)	NL143-MUC Bacteria (32.748)	NL143-MUC Bacteria (32.514)

Amplification Plot (dRn)



Amplification Plot (Rn)



Run Method

Block Type	96-Well 0.2-mL Block
Sample Volume	20.0
Cover Temperature	105.0
Run mode	STANDARD

Stage	Collection Flag	Ramp Rate	Temperature	Hold Time	Starting Cycle	Auto Delta Temperature	Auto Delta Hold Time
Hold Stage	false	1.6°C/sec	95.0°C	600	-	-	-
PCR Stage (40 cycles)	false	1.6°C/sec	95.0°C	30	-	-	-
	true	1.6°C/sec	50.0°C	30	-	-	-
	true	1.6°C/sec	72.0°C	60	-	-	-
Hold Stage	false	1.6°C/sec	72.0°C	420	-	-	-

Primary Analysis Settings

General

PCR Stage/Step	Stage 2, Step 3
Quantification Cycle Method	Baseline Threshold

Target	Auto Threshold	Threshold	Auto Baseline	Baseline Start	Baseline End
DEFAULT	Yes	AUTO	Yes	AUTO	AUTO

QC Alerts

Curve Quality Alert Enabled	No
Results Quality Alert Enabled	Yes

Advanced

Set the Delta-Rn below which curves will be considered Non-Amplified	No
Primary Analysis Variant	N/A
Reduce dye signal crosstalk by algorithm	No

Standard Curve Settings

Standard Curve	On Plate

Standard Curve Results

Well	Sample	Sample Type	Target	Dyes	Task	Cq	Quantity
A1	STD_1	Standard	Bacteria	SYBR	Standard	12.2	2.0000e+7
A2	STD_1	Standard	Bacteria	SYBR	Standard	12.471	2.0000e+7
A3	STD_1	Standard	Bacteria	SYBR	Standard	12.371	2.0000e+7
A4	NL152-6	Unknown	Bacteria	SYBR	Unknown	25.414	4.5500e+3
A5	NL152-6	Unknown	Bacteria	SYBR	Unknown	24.904	6.3131e+3
A6	NL152-6	Unknown	Bacteria	SYBR	Unknown	25.189	5.2581e+3
A7	NL124-LUM	Unknown	Bacteria	SYBR	Unknown	23.849	1.2433e+4
A8	NL124-LUM	Unknown	Bacteria	SYBR	Unknown	23.466	1.5898e+4
A9	NL124-LUM	Unknown	Bacteria	SYBR	Unknown	23.294	1.7759e+4
A10	NL093-MUC	Unknown	Bacteria	SYBR	Unknown	34.313	1.5009e+1
A11	NL093-MUC	Unknown	Bacteria	SYBR	Unknown	34.656	1.2046e+1
A12	NL093-MUC	Unknown	Bacteria	SYBR	Unknown	35.414	7.4056e+0
B1	STD_2	Standard	Bacteria	SYBR	Standard	15.671	2.0000e+6
B2	STD_2	Standard	Bacteria	SYBR	Standard	16.123	2.0000e+6
B3	STD_2	Standard	Bacteria	SYBR	Standard	15.839	2.0000e+6
B4	NL158-6	Unknown	Bacteria	SYBR	Unknown	27.768	1.0038e+3
B5	NL158-6	Unknown	Bacteria	SYBR	Unknown	27.467	1.2179e+3
B6	NL158-6	Unknown	Bacteria	SYBR	Unknown	27.466	1.2188e+3
B7	NL141-LUM	Unknown	Bacteria	SYBR	Unknown	25.885	3.3632e+3
B8	NL141-LUM	Unknown	Bacteria	SYBR	Unknown	24.969	6.0543e+3
B9	NL141-LUM	Unknown	Bacteria	SYBR	Unknown	24.903	6.3173e+3
B10	NL094-MUC	Unknown	Bacteria	SYBR	Unknown	31.594	8.6020e+1
B11	NL094-MUC	Unknown	Bacteria	SYBR	Unknown	31.709	7.9915e+1
B12	NL094-MUC	Unknown	Bacteria	SYBR	Unknown	31.709	7.9913e+1
C1	STD_3	Standard	Bacteria	SYBR	Standard	19.355	2.0000e+5
C2	STD_3	Standard	Bacteria	SYBR	Standard	19.518	2.0000e+5
C3	STD_3	Standard	Bacteria	SYBR	Standard	19.57	2.0000e+5
C4	NL165-6	Unknown	Bacteria	SYBR	Unknown	26.844	1.8173e+3
C5	NL165-6	Unknown	Bacteria	SYBR	Unknown	26.773	1.9019e+3
C6	NL165-6	Unknown	Bacteria	SYBR	Unknown	26.653	2.0534e+3

Well	Sample	Sample Type	Target	Dyes	Task	Cq	Quantity
C7	NL143-LUM	Unknown	Bacteria	SYBR	Unknown	22.652	2.6812e+4
C8	NL143-LUM	Unknown	Bacteria	SYBR	Unknown	22.399	3.1538e+4
C9	NL143-LUM	Unknown	Bacteria	SYBR	Unknown	22.327	3.3022e+4
C10	NL107-MUC	Unknown	Bacteria	SYBR	Unknown	33.169	3.1301e+1
C11	NL107-MUC	Unknown	Bacteria	SYBR	Unknown	32.661	4.3369e+1
C12	NL107-MUC	Unknown	Bacteria	SYBR	Unknown	33.212	3.0452e+1
D1	STD_4	Standard	Bacteria	SYBR	Standard	22.941	2.0000e+4
D2	STD_4	Standard	Bacteria	SYBR	Standard	23.227	2.0000e+4
D3	STD_4	Standard	Bacteria	SYBR	Standard	23.268	2.0000e+4
D4	NL093-LUM	Unknown	Bacteria	SYBR	Unknown	28.901	4.8487e+2
D5	NL093-LUM	Unknown	Bacteria	SYBR	Unknown	28.492	6.3051e+2
D6	NL093-LUM	Unknown	Bacteria	SYBR	Unknown	28.54	6.1129e+2
D7	NL146-LUM	Unknown	Bacteria	SYBR	Unknown	22.839	2.3782e+4
D8	NL146-LUM	Unknown	Bacteria	SYBR	Unknown	21.926	4.2746e+4
D9	NL146-LUM	Unknown	Bacteria	SYBR	Unknown	22.298	3.3652e+4
D10	NL118-MUC	Unknown	Bacteria	SYBR	Unknown	37.366	2.1148e+0
D11	NL118-MUC	Unknown	Bacteria	SYBR	Unknown	37.629	1.7856e+0
D12	NL118-MUC	Unknown	Bacteria	SYBR	Unknown	38	1.4071e+0
E1	STD_5	Standard	Bacteria	SYBR	Standard	26.675	2.0000e+3
E2	STD_5	Standard	Bacteria	SYBR	Standard	26.826	2.0000e+3
E3	STD_5	Standard	Bacteria	SYBR	Standard	26.849	2.0000e+3
E4	NL094-LUM	Unknown	Bacteria	SYBR	Unknown	22.937	2.2331e+4
E5	NL094-LUM	Unknown	Bacteria	SYBR	Unknown	22.665	2.6584e+4
E6	NL094-LUM	Unknown	Bacteria	SYBR	Unknown	22.705	2.5907e+4
E7	NL151-LUM	Unknown	Bacteria	SYBR	Unknown	28.523	6.1834e+2
E8	NL151-LUM	Unknown	Bacteria	SYBR	Unknown	28.103	8.0948e+2
E9	NL151-LUM	Unknown	Bacteria	SYBR	Unknown	27.916	9.1261e+2
E10	NL123-MUC	Unknown	Bacteria	SYBR	Unknown	33.065	3.3461e+1
E11	NL123-MUC	Unknown	Bacteria	SYBR	Unknown	32.846	3.8505e+1
E12	NL123-MUC	Unknown	Bacteria	SYBR	Unknown	33.161	3.1450e+1
F1	STD_6	Standard	Bacteria	SYBR	Standard	30.48	2.0000e+2


Well	Sample	Sample Type	Target	Dyes	Task	Cq	Quantity
F2	STD_6	Standard	Bacteria	SYBR	Standard	30.651	2.0000e+2
F3	STD_6	Standard	Bacteria	SYBR	Standard	30.283	2.0000e+2
F4	NL107-LUM	Unknown	Bacteria	SYBR	Unknown	23.116	1.9903e+4
F5	NL107-LUM	Unknown	Bacteria	SYBR	Unknown	22.903	2.2815e+4
F6	NL107-LUM	Unknown	Bacteria	SYBR	Unknown	22.93	2.2423e+4
F7	NL152-LUM	Unknown	Bacteria	SYBR	Unknown	24.534	8.0051e+3
F8	NL152-LUM	Unknown	Bacteria	SYBR	Unknown	24.438	8.5159e+3
F9	NL152-LUM	Unknown	Bacteria	SYBR	Unknown	24.523	8.0659e+3
F10	NL124-MUC	Unknown	Bacteria	SYBR	Unknown	31.973	6.7472e+1
F11	NL124-MUC	Unknown	Bacteria	SYBR	Unknown	31.815	7.4640e+1
F12	NL124-MUC	Unknown	Bacteria	SYBR	Unknown	32.263	5.5987e+1
G1	STD_7	Standard	Bacteria	SYBR	Standard	33.326	2.0000e+1
G2	STD_7	Standard	Bacteria	SYBR	Standard	33.529	2.0000e+1
G3	STD_7	Standard	Bacteria	SYBR	Standard	34.106	2.0000e+1
G4	NL118-LUM	Unknown	Bacteria	SYBR	Unknown	25.566	4.1288e+3
G5	NL118-LUM	Unknown	Bacteria	SYBR	Unknown	25.402	4.5873e+3
G6	NL118-LUM	Unknown	Bacteria	SYBR	Unknown	25.124	5.4826e+3
G7	NL158-LUM	Unknown	Bacteria	SYBR	Unknown	22.645	2.6928e+4
G8	NL158-LUM	Unknown	Bacteria	SYBR	Unknown	22.153	3.6929e+4
G9	NL158-LUM	Unknown	Bacteria	SYBR	Unknown	22.131	3.7464e+4
G10	NL141-MUC	Unknown	Bacteria	SYBR	Unknown	34.532	1.3046e+1
G11	NL141-MUC	Unknown	Bacteria	SYBR	Unknown	34.05	1.7773e+1
G12	NL141-MUC	Unknown	Bacteria	SYBR	Unknown	34.536	1.3010e+1
H1	NTC	Negative Control	Bacteria	SYBR	Negative Control	35.489	-
H2	NTC	Negative Control	Bacteria	SYBR	Negative Control	36.782	-
H3	NTC	Negative Control	Bacteria	SYBR	Negative Control	27.961	-
H4	NL123-LUM	Unknown	Bacteria	SYBR	Unknown	26.888	1.7667e+3
H5	NL123-LUM	Unknown	Bacteria	SYBR	Unknown	26.349	2.4968e+3
H6	NL123-LUM	Unknown	Bacteria	SYBR	Unknown	25.975	3.1752e+3
H7	NL165-LUM	Unknown	Bacteria	SYBR	Unknown	24	1.1286e+4
H8	NL165-LUM	Unknown	Bacteria	SYBR	Unknown	23.888	1.2122e+4

Well	Sample	Sample Type	Target	Dyes	Task	Cq	Quantity
H9	NL165-LUM	Unknown	Bacteria	SYBR	Unknown	24.008	1.1226e+4
H10	NL143-MUC	Unknown	Bacteria	SYBR	Unknown	32.465	4.9183e+1
H11	NL143-MUC	Unknown	Bacteria	SYBR	Unknown	32.748	4.1024e+1
H12	NL143-MUC	Unknown	Bacteria	SYBR	Unknown	32.514	4.7665e+1

Standard Curve Results - Replicate

Sample	Target	Replication Number	Quantity Mean	Quantity SD
NL093-LUM	Bacteria	3	5.7556e+2	7.9124e+1
NL093-MUC	Bacteria	3	1.1487e+1	3.8325e+0
NL094-LUM	Bacteria	3	2.4941e+4	2.2857e+3
NL094-MUC	Bacteria	3	8.1949e+1	3.5256e+0
NL107-LUM	Bacteria	3	2.1714e+4	1.5805e+3
NL107-MUC	Bacteria	3	3.5041e+1	7.2250e+0
NL118-LUM	Bacteria	3	4.7329e+3	6.8856e+2
NL118-MUC	Bacteria	3	1.7691e+0	3.5412e-1
NL123-LUM	Bacteria	3	2.4796e+3	7.0444e+2
NL123-MUC	Bacteria	3	3.4472e+1	3.6346e+0
NL124-LUM	Bacteria	3	1.5363e+4	2.7025e+3
NL124-MUC	Bacteria	3	6.6033e+1	9.4095e+0
NL141-LUM	Bacteria	3	5.2449e+3	1.6349e+3
NL141-MUC	Bacteria	3	1.4610e+1	2.7393e+0
NL143-LUM	Bacteria	3	3.0457e+4	3.2433e+3
NL143-MUC	Bacteria	3	4.5958e+1	4.3392e+0
NL146-LUM	Bacteria	3	3.3393e+4	9.4845e+3
NL151-LUM	Bacteria	3	7.8014e+2	1.4931e+2
NL152-6	Bacteria	3	5.3737e+3	8.8723e+2
NL152-LUM	Bacteria	3	8.1957e+3	2.7903e+2
NL158-6	Bacteria	3	1.1468e+3	1.2387e+2
NL158-LUM	Bacteria	3	3.3774e+4	5.9343e+3
NL165-6	Bacteria	3	1.9242e+3	1.1965e+2
NL165-LUM	Bacteria	3	1.1544e+4	5.0070e+2
NTC	Bacteria	3	-	-

Standard Curve Plot



Target:Bacteria Slope:-3.586 R^2:0.999 Y-Inter:38.532 Eff:90.048 Error:0.024

- End of Report -