
Plate Results Report

16S_qPCR_PL6_20230922_143932.eds



Summary

Property	Details
Bar Code	-
File Name	16S_qPCR_PL6_20230922_143932.eds
Run Start Date/Time	Sep 22, 2023, 2:39:46 PM
Run End Date/Time	Sep 22, 2023, 4:56:22 PM
Run Duration	2 hours, 16 minutes, and 36 seconds
Operator	Admin
Instrument Name	QS7Pro-2778719090068
Instrument Type	QuantStudio™ 7 Pro System
Instrument Serial Number	2778719090068
Block Type	96-Well 0.2-mL
Block Serial Number	2778219050095
Heated Cover Serial Number	2779219090141
PCR Stage/Step Number	Stage 2, Step 3
Quantification Cycle Method	Baseline Threshold
Comment	-
Software Name and Version	Design & Analysis Software v2.7.0
Plugin Name and Version	Primary Analysis v1.8.0, Standard Curve v1.7.0
Analysis Date/Time	Oct 25, 2023, 2:47:50 PM

Well Table

Well	Sample	Target	Task	Cq	Cq Confidence	Amp Score	Amp Status	Cq Threshold	Baseline Start/End
A1	STD_1	Bacteria	Standard	13.913	0.995	3.271	AMP	479,922.141	3-7
A2	STD_1	Bacteria	Standard	13.897	0.995	3.258	AMP	479,922.141	3-7
A3	STD_1	Bacteria	Standard	13.935	0.994	3.248	AMP	479,922.141	3-7
A4	NL146-MUC	Bacteria	Unknown	25.291	0.991	3.231	AMP	479,922.141	3-15
A5	NL146-MUC	Bacteria	Unknown	24.963	0.995	3.246	AMP	479,922.141	3-12
A6	NL146-MUC	Bacteria	Unknown	25.188	0.99	3.219	AMP	479,922.141	3-15
A7	Y150333-LUM	Bacteria	Unknown	18.911	0.994	3.25	AMP	479,922.141	3-10
A8	Y150333-LUM	Bacteria	Unknown	18.698	0.992	3.212	AMP	479,922.141	3-9
A9	Y150333-LUM	Bacteria	Unknown	18.534	0.993	3.246	AMP	479,922.141	3-9
A10	Y150331-MUC	Bacteria	Unknown	26.184	0.991	3.241	AMP	479,922.141	3-16
A11	Y150331-MUC	Bacteria	Unknown	26.077	0.992	3.249	AMP	479,922.141	3-16
A12	Y150331-MUC	Bacteria	Unknown	26.59	0.993	3.26	AMP	479,922.141	3-14
B1	STD_2	Bacteria	Standard	17.323	0.993	3.281	AMP	479,922.141	3-8
B2	STD_2	Bacteria	Standard	17.753	0.993	3.226	AMP	479,922.141	3-8
B3	STD_2	Bacteria	Standard	17.419	0.993	3.254	AMP	479,922.141	3-8
B4	NL151-MUC	Bacteria	Unknown	32.839	0.99	3.231	AMP	479,922.141	3-16
B5	NL151-MUC	Bacteria	Unknown	32.514	0.99	3.248	AMP	479,922.141	3-17
B6	NL151-MUC	Bacteria	Unknown	33.278	0.986	3.219	AMP	479,922.141	3-17
B7	Y150346-LUM	Bacteria	Unknown	31.353	0.989	3.204	AMP	479,922.141	3-19
B8	Y150346-LUM	Bacteria	Unknown	31.798	0.992	3.198	AMP	479,922.141	3-20
B9	Y150346-LUM	Bacteria	Unknown	31.041	0.988	3.216	AMP	479,922.141	3-19
B10	Y150332-MUC	Bacteria	Unknown	27.574	0.993	3.228	AMP	479,922.141	3-16
B11	Y150332-MUC	Bacteria	Unknown	27.776	0.993	3.205	AMP	479,922.141	3-17
B12	Y150332-MUC	Bacteria	Unknown	27.8	0.988	3.233	AMP	479,922.141	3-17
C1	STD_3	Bacteria	Standard	21.379	0.992	3.292	AMP	479,922.141	3-11
C2	STD_3	Bacteria	Standard	21.266	0.992	3.264	AMP	479,922.141	3-10

Well	Sample	Target	Task	Cq	Cq Confidence	Amp Score	Amp Status	Cq Threshold	Baseline Start/End
C3	STD_3	Bacteria	Standard	21.344	0.993	3.259	AMP	479,922.141	3-11
C4	NL152-MUC	Bacteria	Unknown	34.588	0.987	3.199	AMP	479,922.141	3-23
C5	NL152-MUC	Bacteria	Unknown	34.306	0.985	3.216	AMP	479,922.141	3-18
C6	NL152-MUC	Bacteria	Unknown	34.238	0.984	3.218	AMP	479,922.141	3-20
C7	Y150347-LUM	Bacteria	Unknown	23.169	0.989	3.245	AMP	479,922.141	3-12
C8	Y150347-LUM	Bacteria	Unknown	23.404	0.99	3.216	AMP	479,922.141	3-13
C9	Y150347-LUM	Bacteria	Unknown	23.178	0.989	3.215	AMP	479,922.141	3-12
C10	Y150333-MUC	Bacteria	Unknown	25.116	0.991	3.237	AMP	479,922.141	3-14
C11	Y150333-MUC	Bacteria	Unknown	25.129	0.991	3.238	AMP	479,922.141	3-13
C12	Y150333-MUC	Bacteria	Unknown	25.348	0.992	3.225	AMP	479,922.141	3-12
D1	STD_4	Bacteria	Standard	24.771	0.995	3.282	AMP	479,922.141	3-13
D2	STD_4	Bacteria	Standard	24.791	0.993	3.244	AMP	479,922.141	3-13
D3	STD_4	Bacteria	Standard	24.828	0.989	3.249	AMP	479,922.141	3-14
D4	NL158-MUC	Bacteria	Unknown	Undetermined	-	2.276	NO_AMP	479,922.141	3-22
D5	NL158-MUC	Bacteria	Unknown	Undetermined	-	2.638	NO_AMP	479,922.141	3-19
D6	NL158-MUC	Bacteria	Unknown	Undetermined	-	2.585	NO_AMP	479,922.141	3-19
D7	Y150348-LUM	Bacteria	Unknown	23.895	0.992	3.192	AMP	479,922.141	3-11
D8	Y150348-LUM	Bacteria	Unknown	23.8	0.992	3.187	AMP	479,922.141	3-11
D9	Y150348-LUM	Bacteria	Unknown	23.916	0.992	3.173	AMP	479,922.141	3-12
D10	Y150346-MUC	Bacteria	Unknown	20.712	0.992	3.219	AMP	479,922.141	3-11
D11	Y150346-MUC	Bacteria	Unknown	20.684	0.991	3.199	AMP	479,922.141	3-10
D12	Y150346-MUC	Bacteria	Unknown	20.816	0.987	3.227	AMP	479,922.141	3-11
E1	STD_5	Bacteria	Standard	28.421	0.994	3.277	AMP	479,922.141	3-16
E2	STD_5	Bacteria	Standard	28.312	0.992	3.253	AMP	479,922.141	3-17
E3	STD_5	Bacteria	Standard	28.283	0.992	3.256	AMP	479,922.141	3-18
E4	NL165-MUC	Bacteria	Unknown	Undetermined	-	2.419	NO_AMP	479,922.141	3-20
E5	NL165-MUC	Bacteria	Unknown	Undetermined	-	2.458	NO_AMP	479,922.141	3-19
E6	NL165-MUC	Bacteria	Unknown	Undetermined	-	2.043	NO_AMP	479,922.141	3-17

Well	Sample	Target	Task	Cq	Cq Confidence	Amp Score	Amp Status	Cq Threshold	Baseline Start/End
E7	Y150354-LUM	Bacteria	Unknown	23.634	0.992	3.206	AMP	479,922.141	3-13
E8	Y150354-LUM	Bacteria	Unknown	23.689	0.991	3.189	AMP	479,922.141	3-13
E9	Y150354-LUM	Bacteria	Unknown	23.592	0.991	3.186	AMP	479,922.141	3-11
E10	Y150347-MUC	Bacteria	Unknown	28.558	0.99	3.204	AMP	479,922.141	3-16
E11	Y150347-MUC	Bacteria	Unknown	28.588	0.991	3.21	AMP	479,922.141	3-18
E12	Y150347-MUC	Bacteria	Unknown	28.574	0.991	3.215	AMP	479,922.141	3-17
F1	STD_6	Bacteria	Standard	32.399	0.992	3.257	AMP	479,922.141	3-20
F2	STD_6	Bacteria	Standard	32.47	0.991	3.25	AMP	479,922.141	3-22
F3	STD_6	Bacteria	Standard	32.397	0.991	3.25	AMP	479,922.141	3-18
F4	Y150330-LUM	Bacteria	Unknown	22.43	0.99	3.224	AMP	479,922.141	3-12
F5	Y150330-LUM	Bacteria	Unknown	22.261	0.991	3.234	AMP	479,922.141	3-11
F6	Y150330-LUM	Bacteria	Unknown	22.422	0.991	3.224	AMP	479,922.141	3-11
F7	Y150355-LUM	Bacteria	Unknown	26.722	0.992	3.205	AMP	479,922.141	3-15
F8	Y150355-LUM	Bacteria	Unknown	26.557	0.991	3.201	AMP	479,922.141	3-14
F9	Y150355-LUM	Bacteria	Unknown	26.575	0.991	3.205	AMP	479,922.141	3-15
F10	Y150348-MUC	Bacteria	Unknown	28.725	0.992	3.217	AMP	479,922.141	3-15
F11	Y150348-MUC	Bacteria	Unknown	28.755	0.991	3.198	AMP	479,922.141	3-14
F12	Y150348-MUC	Bacteria	Unknown	28.823	0.992	3.209	AMP	479,922.141	3-14
G1	STD_7	Bacteria	Standard	36.73	0.984	3.074	AMP	479,922.141	3-23
G2	STD_7	Bacteria	Standard	35.838	0.989	3.163	AMP	479,922.141	3-22
G3	STD_7	Bacteria	Standard	35.714	0.988	3.17	AMP	479,922.141	3-22
G4	Y150331-LUM	Bacteria	Unknown	29.498	0.989	3.158	AMP	479,922.141	3-14
G5	Y150331-LUM	Bacteria	Unknown	28.732	0.991	3.167	AMP	479,922.141	3-16
G6	Y150331-LUM	Bacteria	Unknown	28.504	0.989	3.172	AMP	479,922.141	3-17
G7	Y150356-LUM	Bacteria	Unknown	21.439	0.992	3.202	AMP	479,922.141	3-10
G8	Y150356-LUM	Bacteria	Unknown	21.386	0.991	3.191	AMP	479,922.141	3-11
G9	Y150356-LUM	Bacteria	Unknown	21.3	0.99	3.204	AMP	479,922.141	3-11
G10	Y150354-MUC	Bacteria	Unknown	30.045	0.989	3.222	AMP	479,922.141	3-16

Well	Sample	Target	Task	Cq	Cq Confidence	Amp Score	Amp Status	Cq Threshold	Baseline Start/End
G11	Y150354-MUC	Bacteria	Unknown	30.134	0.99	3.225	AMP	479,922.141	3-18
G12	Y150354-MUC	Bacteria	Unknown	30.538	0.979	3.217	AMP	479,922.141	3-17
H1	NTC	Bacteria	Negative Control	Undetermined	-	2.57	NO_AMP	479,922.141	3-29
H2	NTC	Bacteria	Negative Control	39.19	-	2.828	NO_AMP	479,922.141	3-22
H3	NTC	Bacteria	Negative Control	38.535	0.507	2.841	Inconclusive	479,922.141	3-31
H4	Y150332-LUM	Bacteria	Unknown	21.674	0.989	3.173	AMP	479,922.141	3-11
H5	Y150332-LUM	Bacteria	Unknown	21.565	0.989	3.169	AMP	479,922.141	3-10
H6	Y150332-LUM	Bacteria	Unknown	21.178	0.987	3.222	AMP	479,922.141	3-11
H7	Y150330-MUC	Bacteria	Unknown	25.676	0.994	3.239	AMP	479,922.141	3-13
H8	Y150330-MUC	Bacteria	Unknown	25.797	0.991	3.224	AMP	479,922.141	3-16
H9	Y150330-MUC	Bacteria	Unknown	25.774	0.994	3.223	AMP	479,922.141	3-13
H10	Y150355-MUC	Bacteria	Unknown	26.673	0.991	3.205	AMP	479,922.141	3-13
H11	Y150355-MUC	Bacteria	Unknown	26.472	0.992	3.219	AMP	479,922.141	3-14
H12	Y150355-MUC	Bacteria	Unknown	27.017	0.968	3.207	AMP	479,922.141	3-15

Replicate Group Table


Sample	Target	No. of Replicates	Cq Mean	Cq SD
NL146-MUC	Bacteria	3	25.148	0.168
NL151-MUC	Bacteria	3	32.877	0.383
NL152-MUC	Bacteria	3	34.377	0.186
NL158-MUC	Bacteria	3	-	-
NL165-MUC	Bacteria	3	-	-
NTC	Bacteria	3	38.535	-
STD_1	Bacteria	3	13.915	0.019
STD_2	Bacteria	3	17.498	0.225
STD_3	Bacteria	3	21.33	0.058
STD_4	Bacteria	3	24.796	0.029
STD_5	Bacteria	3	28.339	0.073
STD_6	Bacteria	3	32.422	0.042
STD_7	Bacteria	3	36.094	0.554
Y150330-LUM	Bacteria	3	22.371	0.095
Y150330-MUC	Bacteria	3	25.749	0.064
Y150331-LUM	Bacteria	3	28.911	0.521
Y150331-MUC	Bacteria	3	26.284	0.271
Y150332-LUM	Bacteria	3	21.472	0.26
Y150332-MUC	Bacteria	3	27.717	0.124
Y150333-LUM	Bacteria	3	18.714	0.189
Y150333-MUC	Bacteria	3	25.198	0.131
Y150346-LUM	Bacteria	3	31.397	0.38
Y150346-MUC	Bacteria	3	20.737	0.069
Y150347-LUM	Bacteria	3	23.251	0.133
Y150347-MUC	Bacteria	3	28.573	0.015
Y150348-LUM	Bacteria	3	23.87	0.062
Y150348-MUC	Bacteria	3	28.768	0.05
Y150354-LUM	Bacteria	3	23.638	0.049
Y150354-MUC	Bacteria	3	30.239	0.263
Y150355-LUM	Bacteria	3	26.618	0.091

Sample	Target	No. of Replicates	Cq Mean	Cq SD
Y150355-MUC	Bacteria	3	26.721	0.275
Y150356-LUM	Bacteria	3	21.375	0.07


Plate Layout

	1	2	3	4	5	6	7	8	9	10	11	12
A	STD_1 Bacteria (13.913)	STD_1 Bacteria (13.897)	STD_1 Bacteria (13.935)	NL146-MUC Bacteria (25.291)	NL146-MUC Bacteria (24.963)	NL146-MUC Bacteria (25.188)	Y150333-LUM Bacteria (18.911)	Y150333-LUM Bacteria (18.698)	Y150333-LUM Bacteria (18.534)	Y150331-MUC Bacteria (26.184)	Y150331-MUC Bacteria (26.077)	Y150331-MUC Bacteria (26.59)
B	STD_2 Bacteria (17.323)	STD_2 Bacteria (17.753)	STD_2 Bacteria (17.419)	NL151-MUC Bacteria (32.839)	NL151-MUC Bacteria (32.514)	NL151-MUC Bacteria (33.278)	Y150346-LUM Bacteria (31.353)	Y150346-LUM Bacteria (31.798)	Y150346-LUM Bacteria (31.041)	Y150332-MUC Bacteria (27.574)	Y150332-MUC Bacteria (27.776)	Y150332-MUC Bacteria (27.8)
C	STD_3 Bacteria (21.379)	STD_3 Bacteria (21.266)	STD_3 Bacteria (21.344)	NL152-MUC Bacteria (34.588)	NL152-MUC Bacteria (34.306)	NL152-MUC Bacteria (34.238)	Y150347-LUM Bacteria (23.169)	Y150347-LUM Bacteria (23.404)	Y150347-LUM Bacteria (23.178)	Y150333-MUC Bacteria (25.116)	Y150333-MUC Bacteria (25.129)	Y150333-MUC Bacteria (25.348)
D	STD_4 Bacteria (24.771)	STD_4 Bacteria (24.791)	STD_4 Bacteria (24.828)	NL158-MUC Bacteria (n/a)	NL158-MUC Bacteria (n/a)	NL158-MUC Bacteria (n/a)	Y150348-LUM Bacteria (23.895)	Y150348-LUM Bacteria (23.8)	Y150348-LUM Bacteria (23.916)	Y150346-MUC Bacteria (20.712)	Y150346-MUC Bacteria (20.684)	Y150346-MUC Bacteria (20.816)
E	STD_5 Bacteria (28.421)	STD_5 Bacteria (28.312)	STD_5 Bacteria (28.283)	NL165-MUC Bacteria (n/a)	NL165-MUC Bacteria (n/a)	NL165-MUC Bacteria (n/a)	Y150354-LUM Bacteria (23.634)	Y150354-LUM Bacteria (23.689)	Y150354-LUM Bacteria (23.592)	Y150347-MUC Bacteria (28.558)	Y150347-MUC Bacteria (28.588)	Y150347-MUC Bacteria (28.574)
F	STD_6 Bacteria (32.399)	STD_6 Bacteria (32.47)	STD_6 Bacteria (32.397)	Y150330-LUM Bacteria (22.43)	Y150330-LUM Bacteria (22.261)	Y150330-LUM Bacteria (22.422)	Y150355-LUM Bacteria (26.722)	Y150355-LUM Bacteria (26.557)	Y150355-LUM Bacteria (26.575)	Y150348-MUC Bacteria (28.725)	Y150348-MUC Bacteria (28.755)	Y150348-MUC Bacteria (28.823)
G	STD_7 Bacteria (36.73)	STD_7 Bacteria (35.838)	STD_7 Bacteria (35.714)	Y150331-LUM Bacteria (29.498)	Y150331-LUM Bacteria (28.732)	Y150331-LUM Bacteria (28.504)	Y150356-LUM Bacteria (21.439)	Y150356-LUM Bacteria (21.386)	Y150356-LUM Bacteria (21.3)	Y150354-MUC Bacteria (30.045)	Y150354-MUC Bacteria (30.134)	Y150354-MUC Bacteria (30.538)
H	NTC Bacteria (n/a)	NTC Bacteria (39.19)	NTC Bacteria (38.535)	Y150332-LUM Bacteria (21.674)	Y150332-LUM Bacteria (21.565)	Y150332-LUM Bacteria (21.178)	Y150330-MUC Bacteria (25.676)	Y150330-MUC Bacteria (25.797)	Y150330-MUC Bacteria (25.774)	Y150355-MUC Bacteria (26.673)	Y150355-MUC Bacteria (26.472)	Y150355-MUC Bacteria (27.017)

Amplification Plot (dRn)



Amplification Plot (Rn)



Run Method

Block Type	96-Well 0.2-mL Block
Sample Volume	20.0
Cover Temperature	105.0
Run mode	STANDARD

Stage	Collection Flag	Ramp Rate	Temperature	Hold Time	Starting Cycle	Auto Delta Temperature	Auto Delta Hold Time
Hold Stage	false	1.6°C/sec	95.0°C	600	-	-	-
PCR Stage (40 cycles)	false	1.6°C/sec	95.0°C	30	-	-	-
	true	1.6°C/sec	50.0°C	30	-	-	-
	true	1.6°C/sec	72.0°C	60	-	-	-
Hold Stage	false	1.6°C/sec	72.0°C	420	-	-	-

Primary Analysis Settings

General

PCR Stage/Step	Stage 2, Step 3
Quantification Cycle Method	Baseline Threshold

Target	Auto Threshold	Threshold	Auto Baseline	Baseline Start	Baseline End
DEFAULT	Yes	AUTO	Yes	AUTO	AUTO

QC Alerts

Curve Quality Alert Enabled	No
Results Quality Alert Enabled	Yes

Advanced

Set the Delta-Rn below which curves will be considered Non-Amplified	No
Primary Analysis Variant	N/A
Reduce dye signal crosstalk by algorithm	No

Standard Curve Settings

Standard Curve	On Plate

Standard Curve Results

Well	Sample	Sample Type	Target	Dyes	Task	Cq	Quantity
A1	STD_1	Standard	Bacteria	SYBR	Standard	13.913	2.0000e+7
A2	STD_1	Standard	Bacteria	SYBR	Standard	13.897	2.0000e+7
A3	STD_1	Standard	Bacteria	SYBR	Standard	13.935	2.0000e+7
A4	NL146-MUC	Unknown	Bacteria	SYBR	Unknown	25.291	1.5803e+4
A5	NL146-MUC	Unknown	Bacteria	SYBR	Unknown	24.963	1.9388e+4
A6	NL146-MUC	Unknown	Bacteria	SYBR	Unknown	25.188	1.6849e+4
A7	Y150333-LUM	Unknown	Bacteria	SYBR	Unknown	18.911	8.4471e+5
A8	Y150333-LUM	Unknown	Bacteria	SYBR	Unknown	18.698	9.6467e+5
A9	Y150333-LUM	Unknown	Bacteria	SYBR	Unknown	18.534	1.0685e+6
A10	Y150331-MUC	Unknown	Bacteria	SYBR	Unknown	26.184	9.0563e+3
A11	Y150331-MUC	Unknown	Bacteria	SYBR	Unknown	26.077	9.6826e+3
A12	Y150331-MUC	Unknown	Bacteria	SYBR	Unknown	26.59	7.0289e+3
B1	STD_2	Standard	Bacteria	SYBR	Standard	17.323	2.0000e+6
B2	STD_2	Standard	Bacteria	SYBR	Standard	17.753	2.0000e+6
B3	STD_2	Standard	Bacteria	SYBR	Standard	17.419	2.0000e+6
B4	NL151-MUC	Unknown	Bacteria	SYBR	Unknown	32.839	1.4282e+2
B5	NL151-MUC	Unknown	Bacteria	SYBR	Unknown	32.514	1.7486e+2
B6	NL151-MUC	Unknown	Bacteria	SYBR	Unknown	33.278	1.0863e+2
B7	Y150346-LUM	Unknown	Bacteria	SYBR	Unknown	31.353	3.6073e+2
B8	Y150346-LUM	Unknown	Bacteria	SYBR	Unknown	31.798	2.7333e+2
B9	Y150346-LUM	Unknown	Bacteria	SYBR	Unknown	31.041	4.3813e+2
B10	Y150332-MUC	Unknown	Bacteria	SYBR	Unknown	27.574	3.8068e+3
B11	Y150332-MUC	Unknown	Bacteria	SYBR	Unknown	27.776	3.3570e+3
B12	Y150332-MUC	Unknown	Bacteria	SYBR	Unknown	27.8	3.3060e+3
C1	STD_3	Standard	Bacteria	SYBR	Standard	21.379	2.0000e+5
C2	STD_3	Standard	Bacteria	SYBR	Standard	21.266	2.0000e+5
C3	STD_3	Standard	Bacteria	SYBR	Standard	21.344	2.0000e+5
C4	NL152-MUC	Unknown	Bacteria	SYBR	Unknown	34.588	4.7984e+1
C5	NL152-MUC	Unknown	Bacteria	SYBR	Unknown	34.306	5.7207e+1
C6	NL152-MUC	Unknown	Bacteria	SYBR	Unknown	34.238	5.9684e+1

Well	Sample	Sample Type	Target	Dyes	Task	Cq	Quantity
C7	Y150347-LUM	Unknown	Bacteria	SYBR	Unknown	23.169	5.9341e+4
C8	Y150347-LUM	Unknown	Bacteria	SYBR	Unknown	23.404	5.1257e+4
C9	Y150347-LUM	Unknown	Bacteria	SYBR	Unknown	23.178	5.9010e+4
C10	Y150333-MUC	Unknown	Bacteria	SYBR	Unknown	25.116	1.7631e+4
C11	Y150333-MUC	Unknown	Bacteria	SYBR	Unknown	25.129	1.7486e+4
C12	Y150333-MUC	Unknown	Bacteria	SYBR	Unknown	25.348	1.5251e+4
D1	STD_4	Standard	Bacteria	SYBR	Standard	24.771	2.0000e+4
D2	STD_4	Standard	Bacteria	SYBR	Standard	24.791	2.0000e+4
D3	STD_4	Standard	Bacteria	SYBR	Standard	24.828	2.0000e+4
D4	NL158-MUC	Unknown	Bacteria	SYBR	Unknown	-	-
D5	NL158-MUC	Unknown	Bacteria	SYBR	Unknown	-	-
D6	NL158-MUC	Unknown	Bacteria	SYBR	Unknown	-	-
D7	Y150348-LUM	Unknown	Bacteria	SYBR	Unknown	23.895	3.7745e+4
D8	Y150348-LUM	Unknown	Bacteria	SYBR	Unknown	23.8	4.0050e+4
D9	Y150348-LUM	Unknown	Bacteria	SYBR	Unknown	23.916	3.7244e+4
D10	Y150346-MUC	Unknown	Bacteria	SYBR	Unknown	20.712	2.7467e+5
D11	Y150346-MUC	Unknown	Bacteria	SYBR	Unknown	20.684	2.7947e+5
D12	Y150346-MUC	Unknown	Bacteria	SYBR	Unknown	20.816	2.5746e+5
E1	STD_5	Standard	Bacteria	SYBR	Standard	28.421	2.0000e+3
E2	STD_5	Standard	Bacteria	SYBR	Standard	28.312	2.0000e+3
E3	STD_5	Standard	Bacteria	SYBR	Standard	28.283	2.0000e+3
E4	NL165-MUC	Unknown	Bacteria	SYBR	Unknown	-	-
E5	NL165-MUC	Unknown	Bacteria	SYBR	Unknown	-	-
E6	NL165-MUC	Unknown	Bacteria	SYBR	Unknown	-	-
E7	Y150354-LUM	Unknown	Bacteria	SYBR	Unknown	23.634	4.4404e+4
E8	Y150354-LUM	Unknown	Bacteria	SYBR	Unknown	23.689	4.2927e+4
E9	Y150354-LUM	Unknown	Bacteria	SYBR	Unknown	23.592	4.5601e+4
E10	Y150347-MUC	Unknown	Bacteria	SYBR	Unknown	28.558	2.0614e+3
E11	Y150347-MUC	Unknown	Bacteria	SYBR	Unknown	28.588	2.0224e+3
E12	Y150347-MUC	Unknown	Bacteria	SYBR	Unknown	28.574	2.0410e+3
F1	STD_6	Standard	Bacteria	SYBR	Standard	32.399	2.0000e+2


Well	Sample	Sample Type	Target	Dyes	Task	Cq	Quantity
F2	STD_6	Standard	Bacteria	SYBR	Standard	32.47	2.0000e+2
F3	STD_6	Standard	Bacteria	SYBR	Standard	32.397	2.0000e+2
F4	Y150330-LUM	Unknown	Bacteria	SYBR	Unknown	22.43	9.4110e+4
F5	Y150330-LUM	Unknown	Bacteria	SYBR	Unknown	22.261	1.0456e+5
F6	Y150330-LUM	Unknown	Bacteria	SYBR	Unknown	22.422	9.4546e+4
F7	Y150355-LUM	Unknown	Bacteria	SYBR	Unknown	26.722	6.4754e+3
F8	Y150355-LUM	Unknown	Bacteria	SYBR	Unknown	26.557	7.1793e+3
F9	Y150355-LUM	Unknown	Bacteria	SYBR	Unknown	26.575	7.0955e+3
F10	Y150348-MUC	Unknown	Bacteria	SYBR	Unknown	28.725	1.8566e+3
F11	Y150348-MUC	Unknown	Bacteria	SYBR	Unknown	28.755	1.8226e+3
F12	Y150348-MUC	Unknown	Bacteria	SYBR	Unknown	28.823	1.7475e+3
G1	STD_7	Standard	Bacteria	SYBR	Standard	36.73	2.0000e+1
G2	STD_7	Standard	Bacteria	SYBR	Standard	35.838	2.0000e+1
G3	STD_7	Standard	Bacteria	SYBR	Standard	35.714	2.0000e+1
G4	Y150331-LUM	Unknown	Bacteria	SYBR	Unknown	29.498	1.1471e+3
G5	Y150331-LUM	Unknown	Bacteria	SYBR	Unknown	28.732	1.8494e+3
G6	Y150331-LUM	Unknown	Bacteria	SYBR	Unknown	28.504	2.1314e+3
G7	Y150356-LUM	Unknown	Bacteria	SYBR	Unknown	21.439	1.7461e+5
G8	Y150356-LUM	Unknown	Bacteria	SYBR	Unknown	21.386	1.8045e+5
G9	Y150356-LUM	Unknown	Bacteria	SYBR	Unknown	21.3	1.9042e+5
G10	Y150354-MUC	Unknown	Bacteria	SYBR	Unknown	30.045	8.1533e+2
G11	Y150354-MUC	Unknown	Bacteria	SYBR	Unknown	30.134	7.7130e+2
G12	Y150354-MUC	Unknown	Bacteria	SYBR	Unknown	30.538	5.9949e+2
H1	NTC	Negative Control	Bacteria	SYBR	Negative Control	-	-
H2	NTC	Negative Control	Bacteria	SYBR	Negative Control	39.19	-
H3	NTC	Negative Control	Bacteria	SYBR	Negative Control	38.535	-
H4	Y150332-LUM	Unknown	Bacteria	SYBR	Unknown	21.674	1.5082e+5
H5	Y150332-LUM	Unknown	Bacteria	SYBR	Unknown	21.565	1.6141e+5
H6	Y150332-LUM	Unknown	Bacteria	SYBR	Unknown	21.178	2.0542e+5
H7	Y150330-MUC	Unknown	Bacteria	SYBR	Unknown	25.676	1.2430e+4
H8	Y150330-MUC	Unknown	Bacteria	SYBR	Unknown	25.797	1.1526e+4

Well	Sample	Sample Type	Target	Dyes	Task	Cq	Quantity
H9	Y150330-MUC	Unknown	Bacteria	SYBR	Unknown	25.774	1.1696e+4
H10	Y150355-MUC	Unknown	Bacteria	SYBR	Unknown	26.673	6.6754e+3
H11	Y150355-MUC	Unknown	Bacteria	SYBR	Unknown	26.472	7.5680e+3
H12	Y150355-MUC	Unknown	Bacteria	SYBR	Unknown	27.017	5.3883e+3

Standard Curve Results - Replicate

Sample	Target	Replication Number	Quantity Mean	Quantity SD
NL146-MUC	Bacteria	3	1.7347e+4	1.8437e+3
NL151-MUC	Bacteria	3	1.4211e+2	3.3121e+1
NL152-MUC	Bacteria	3	5.4958e+1	6.1653e+0
NL158-MUC	Bacteria	3	-	-
NL165-MUC	Bacteria	3	-	-
NTC	Bacteria	3	-	-
Y150330-LUM	Bacteria	3	9.7739e+4	5.9129e+3
Y150330-MUC	Bacteria	3	1.1884e+4	4.8070e+2
Y150331-LUM	Bacteria	3	1.7093e+3	5.0692e+2
Y150331-MUC	Bacteria	3	8.5892e+3	1.3871e+3
Y150332-LUM	Bacteria	3	1.7255e+5	2.8955e+4
Y150332-MUC	Bacteria	3	3.4899e+3	2.7557e+2
Y150333-LUM	Bacteria	3	9.5931e+5	1.1201e+5
Y150333-MUC	Bacteria	3	1.6789e+4	1.3343e+3
Y150346-LUM	Bacteria	3	3.5740e+2	8.2451e+1
Y150346-MUC	Bacteria	3	2.7053e+5	1.1575e+4
Y150347-LUM	Bacteria	3	5.6536e+4	4.5751e+3
Y150347-MUC	Bacteria	3	2.0416e+3	1.9524e+1
Y150348-LUM	Bacteria	3	3.8346e+4	1.4967e+3
Y150348-MUC	Bacteria	3	1.8089e+3	5.5828e+1
Y150354-LUM	Bacteria	3	4.4311e+4	1.3394e+3
Y150354-MUC	Bacteria	3	7.2871e+2	1.1405e+2
Y150355-LUM	Bacteria	3	6.9167e+3	3.8451e+2
Y150355-MUC	Bacteria	3	6.5439e+3	1.0958e+3
Y150356-LUM	Bacteria	3	1.8183e+5	7.9943e+3

Standard Curve Plot



Target:Bacteria Slope:-3.693 R^2:0.999 Y-Inter:40.796 Eff:86.557 Error:0.027

- End of Report -