
Plate Results Report

16S_qPCR_PL7_20230921_100914.eds



Summary

Property	Details
Bar Code	-
File Name	16S_qPCR_PL7_20230921_100914.eds
Run Start Date/Time	Sep 21, 2023, 10:09:37 AM
Run End Date/Time	Sep 21, 2023, 12:26:11 PM
Run Duration	2 hours, 16 minutes, and 34 seconds
Operator	Admin
Instrument Name	QS7Pro-2778719090068
Instrument Type	QuantStudio™ 7 Pro System
Instrument Serial Number	2778719090068
Block Type	96-Well 0.2-mL
Block Serial Number	2778219050095
Heated Cover Serial Number	2779219090141
PCR Stage/Step Number	Stage 2, Step 3
Quantification Cycle Method	Baseline Threshold
Comment	-
Software Name and Version	Design & Analysis Software v2.7.0
Plugin Name and Version	Primary Analysis v1.8.0, Standard Curve v1.7.0
Analysis Date/Time	Oct 25, 2023, 2:50:16 PM

Well Table

Well	Sample	Target	Task	Cq	Cq Confidence	Amp Score	Amp Status	Cq Threshold	Baseline Start/End
A1	STD_1	Bacteria	Standard	13.72	0.995	3.278	AMP	428,974.791	3-7
A2	STD_1	Bacteria	Standard	13.749	0.995	3.254	AMP	428,974.791	3-7
A3	STD_1	Bacteria	Standard	13.752	0.994	3.252	AMP	428,974.791	3-7
A4	Y150356-MUC	Bacteria	Unknown	28.353	0.992	3.227	AMP	428,974.791	3-15
A5	Y150356-MUC	Bacteria	Unknown	28.177	0.992	3.245	AMP	428,974.791	3-17
A6	Y150356-MUC	Bacteria	Unknown	28.238	0.992	3.231	AMP	428,974.791	3-17
A7	2301-INOC	Bacteria	Unknown	25.335	0.992	3.216	AMP	428,974.791	3-15
A8	2301-INOC	Bacteria	Unknown	25.046	0.99	3.204	AMP	428,974.791	3-13
A9	2301-INOC	Bacteria	Unknown	25.03	0.99	3.216	AMP	428,974.791	3-15
B1	STD_2	Bacteria	Standard	17.421	0.994	3.282	AMP	428,974.791	3-9
B2	STD_2	Bacteria	Standard	17.769	0.993	3.229	AMP	428,974.791	3-9
B3	STD_2	Bacteria	Standard	17.57	0.994	3.254	AMP	428,974.791	3-9
B4	Y150331-7	Bacteria	Unknown	25.226	0.989	3.183	AMP	428,974.791	3-12
B5	Y150331-7	Bacteria	Unknown	25.247	0.99	3.215	AMP	428,974.791	3-14
B6	Y150331-7	Bacteria	Unknown	25.308	0.991	3.202	AMP	428,974.791	3-14
B7	2302-INOC	Bacteria	Unknown	20.895	0.993	3.208	AMP	428,974.791	3-10
B8	2302-INOC	Bacteria	Unknown	20.914	0.993	3.2	AMP	428,974.791	3-10
B9	2302-INOC	Bacteria	Unknown	20.609	0.994	3.238	AMP	428,974.791	3-10
C1	STD_3	Bacteria	Standard	21.24	0.994	3.282	AMP	428,974.791	3-11
C2	STD_3	Bacteria	Standard	21.314	0.994	3.269	AMP	428,974.791	3-11
C3	STD_3	Bacteria	Standard	21.498	0.994	3.244	AMP	428,974.791	3-11
C4	Y150332-7	Bacteria	Unknown	29.826	0.992	3.177	AMP	428,974.791	3-14
C5	Y150332-7	Bacteria	Unknown	29.656	0.991	3.188	AMP	428,974.791	3-18
C6	Y150332-7	Bacteria	Unknown	29.675	0.99	3.201	AMP	428,974.791	3-16
C7	2304-INOC	Bacteria	Unknown	28.384	0.992	3.216	AMP	428,974.791	3-16
C8	2304-INOC	Bacteria	Unknown	28.5	0.991	3.191	AMP	428,974.791	3-17

Well	Sample	Target	Task	Cq	Cq Confidence	Amp Score	Amp Status	Cq Threshold	Baseline Start/End
C9	2304-INOC	Bacteria	Unknown	28.304	0.991	3.208	AMP	428,974.791	3-16
D1	STD_4	Bacteria	Standard	25.35	0.994	3.284	AMP	428,974.791	3-14
D2	STD_4	Bacteria	Standard	25.496	0.994	3.247	AMP	428,974.791	3-14
D3	STD_4	Bacteria	Standard	25.595	0.994	3.239	AMP	428,974.791	3-15
D4	Y150333-7	Bacteria	Unknown	31.543	0.991	3.183	AMP	428,974.791	3-17
D5	Y150333-7	Bacteria	Unknown	31.14	0.989	3.188	AMP	428,974.791	3-15
D6	Y150333-7	Bacteria	Unknown	31.686	0.99	3.189	AMP	428,974.791	3-17
D7	2305-INOC	Bacteria	Unknown	23.12	0.989	3.159	AMP	428,974.791	3-12
D8	2305-INOC	Bacteria	Unknown	22.843	0.993	3.157	AMP	428,974.791	3-12
D9	2305-INOC	Bacteria	Unknown	22.949	0.993	3.164	AMP	428,974.791	3-11
E1	STD_5	Bacteria	Standard	29.011	0.99	3.276	AMP	428,974.791	3-18
E2	STD_5	Bacteria	Standard	29.084	0.992	3.252	AMP	428,974.791	3-18
E3	STD_5	Bacteria	Standard	29.043	0.992	3.247	AMP	428,974.791	3-14
E4	Y150347-7	Bacteria	Unknown	29.044	0.988	3.183	AMP	428,974.791	3-17
E5	Y150347-7	Bacteria	Unknown	28.511	0.991	3.206	AMP	428,974.791	3-15
E6	Y150347-7	Bacteria	Unknown	28.681	0.991	3.207	AMP	428,974.791	3-17
E7	2306-INOC	Bacteria	Unknown	18.005	0.99	3.203	AMP	428,974.791	3-9
E8	2306-INOC	Bacteria	Unknown	17.859	0.993	3.193	AMP	428,974.791	3-9
E9	2306-INOC	Bacteria	Unknown	17.801	0.994	3.204	AMP	428,974.791	3-9
F1	STD_6	Bacteria	Standard	32.4	0.991	3.262	AMP	428,974.791	3-21
F2	STD_6	Bacteria	Standard	32.13	0.992	3.247	AMP	428,974.791	3-21
F3	STD_6	Bacteria	Standard	32.161	0.991	3.249	AMP	428,974.791	3-19
F4	Y150348-7	Bacteria	Unknown	30.988	0.994	3.208	AMP	428,974.791	3-17
F5	Y150348-7	Bacteria	Unknown	30.634	0.993	3.221	AMP	428,974.791	3-18
F6	Y150348-7	Bacteria	Unknown	30.538	0.991	3.21	AMP	428,974.791	3-17
F7	2311-INOC	Bacteria	Unknown	27.498	0.991	3.196	AMP	428,974.791	3-14
F8	2311-INOC	Bacteria	Unknown	26.977	0.993	3.172	AMP	428,974.791	3-16
F9	2311-INOC	Bacteria	Unknown	27.094	0.99	3.214	AMP	428,974.791	3-15

Well	Sample	Target	Task	Cq	Cq Confidence	Amp Score	Amp Status	Cq Threshold	Baseline Start/End
G1	STD_7	Bacteria	Standard	35.761	0.988	3.185	AMP	428,974.791	3-23
G2	STD_7	Bacteria	Standard	36.471	0.985	3.114	AMP	428,974.791	3-23
G3	STD_7	Bacteria	Standard	35.811	0.99	3.158	AMP	428,974.791	3-21
G4	Y150355-7	Bacteria	Unknown	22.317	0.991	3.183	AMP	428,974.791	3-11
G5	Y150355-7	Bacteria	Unknown	22.112	0.989	3.181	AMP	428,974.791	3-13
G6	Y150355-7	Bacteria	Unknown	22.04	0.988	3.184	AMP	428,974.791	3-12
G7	2317-INOC	Bacteria	Unknown	24.233	0.989	3.148	AMP	428,974.791	3-13
G8	2317-INOC	Bacteria	Unknown	23.875	0.993	3.138	AMP	428,974.791	3-11
G9	2317-INOC	Bacteria	Unknown	23.628	0.993	3.17	AMP	428,974.791	3-11
H1	NTC	Bacteria	Negative Control	38.835	0.19	2.8	Inconclusive	428,974.791	3-25
H2	NTC	Bacteria	Negative Control	38.294	0.808	2.86	Inconclusive	428,974.791	3-25
H3	NTC	Bacteria	Negative Control	Undetermined	-	2.481	NO_AMP	428,974.791	3-24
H4	Y150356-7	Bacteria	Unknown	18.998	0.993	3.154	AMP	428,974.791	3-9
H5	Y150356-7	Bacteria	Unknown	18.873	0.992	3.156	AMP	428,974.791	3-9
H6	Y150356-7	Bacteria	Unknown	18.485	0.993	3.21	AMP	428,974.791	3-9
H7	2318-INOC	Bacteria	Unknown	22.122	0.989	3.17	AMP	428,974.791	3-12
H8	2318-INOC	Bacteria	Unknown	22.203	0.989	3.159	AMP	428,974.791	3-11
H9	2318-INOC	Bacteria	Unknown	22.253	0.988	3.177	AMP	428,974.791	3-12


Replicate Group Table

Sample	Target	No. of Replicates	Cq Mean	Cq SD
2301-INOC	Bacteria	3	25.137	0.172
2302-INOC	Bacteria	3	20.806	0.171
2304-INOC	Bacteria	3	28.396	0.099
2305-INOC	Bacteria	3	22.971	0.14
2306-INOC	Bacteria	3	17.888	0.105
2311-INOC	Bacteria	3	27.189	0.273
2317-INOC	Bacteria	3	23.912	0.304
2318-INOC	Bacteria	3	22.193	0.066
NTC	Bacteria	3	38.564	0.382
STD_1	Bacteria	3	13.74	0.018
STD_2	Bacteria	3	17.587	0.175
STD_3	Bacteria	3	21.351	0.133
STD_4	Bacteria	3	25.48	0.124
STD_5	Bacteria	3	29.046	0.036
STD_6	Bacteria	3	32.23	0.148
STD_7	Bacteria	3	36.014	0.396
Y150331-7	Bacteria	3	25.26	0.043
Y150332-7	Bacteria	3	29.719	0.093
Y150333-7	Bacteria	3	31.457	0.283
Y150347-7	Bacteria	3	28.745	0.272
Y150348-7	Bacteria	3	30.72	0.237
Y150355-7	Bacteria	3	22.157	0.144
Y150356-7	Bacteria	3	18.785	0.268
Y150356-MUC	Bacteria	3	28.256	0.089


Plate Layout

	1	2	3	4	5	6	7	8	9	10	11	12
A	STD_1 Bacteria (13.72)	STD_1 Bacteria (13.749)	STD_1 Bacteria (13.752)	Y150356-MUC Bacteria (28.353)	Y150356-MUC Bacteria (28.177)	Y150356-MUC Bacteria (28.238)	2301-INOC Bacteria (25.335)	2301-INOC Bacteria (25.046)	2301-INOC Bacteria (25.03)			
B	STD_2 Bacteria (17.421)	STD_2 Bacteria (17.769)	STD_2 Bacteria (17.57)	Y150331-7 Bacteria (25.226)	Y150331-7 Bacteria (25.247)	Y150331-7 Bacteria (25.308)	2302-INOC Bacteria (20.895)	2302-INOC Bacteria (20.914)	2302-INOC Bacteria (20.609)			
C	STD_3 Bacteria (21.24)	STD_3 Bacteria (21.314)	STD_3 Bacteria (21.498)	Y150332-7 Bacteria (29.826)	Y150332-7 Bacteria (29.656)	Y150332-7 Bacteria (29.675)	2304-INOC Bacteria (28.384)	2304-INOC Bacteria (28.5)	2304-INOC Bacteria (28.304)			
D	STD_4 Bacteria (25.35)	STD_4 Bacteria (25.496)	STD_4 Bacteria (25.595)	Y150333-7 Bacteria (31.543)	Y150333-7 Bacteria (31.14)	Y150333-7 Bacteria (31.686)	2305-INOC Bacteria (23.12)	2305-INOC Bacteria (22.843)	2305-INOC Bacteria (22.949)			
E	STD_5 Bacteria (29.011)	STD_5 Bacteria (29.084)	STD_5 Bacteria (29.043)	Y150347-7 Bacteria (29.044)	Y150347-7 Bacteria (28.511)	Y150347-7 Bacteria (28.681)	2306-INOC Bacteria (18.005)	2306-INOC Bacteria (17.859)	2306-INOC Bacteria (17.801)			
F	STD_6 Bacteria (32.4)	STD_6 Bacteria (32.13)	STD_6 Bacteria (32.161)	Y150348-7 Bacteria (30.988)	Y150348-7 Bacteria (30.634)	Y150348-7 Bacteria (30.538)	2311-INOC Bacteria (27.498)	2311-INOC Bacteria (26.977)	2311-INOC Bacteria (27.094)			
G	STD_7 Bacteria (35.761)	STD_7 Bacteria (36.471)	STD_7 Bacteria (35.811)	Y150355-7 Bacteria (22.317)	Y150355-7 Bacteria (22.112)	Y150355-7 Bacteria (22.04)	2317-INOC Bacteria (24.233)	2317-INOC Bacteria (23.875)	2317-INOC Bacteria (23.628)			
H	NTC Bacteria (38.835)	NTC Bacteria (38.294)	NTC Bacteria (n/a)	Y150356-7 Bacteria (18.998)	Y150356-7 Bacteria (18.873)	Y150356-7 Bacteria (18.485)	2318-INOC Bacteria (22.122)	2318-INOC Bacteria (22.203)	2318-INOC Bacteria (22.253)			

Amplification Plot (dRn)



Amplification Plot (Rn)



Run Method

Block Type	96-Well 0.2-mL Block
Sample Volume	20.0
Cover Temperature	105.0
Run mode	STANDARD

Stage	Collection Flag	Ramp Rate	Temperature	Hold Time	Starting Cycle	Auto Delta Temperature	Auto Delta Hold Time
Hold Stage	false	1.6°C/sec	95.0°C	600	-	-	-
PCR Stage (40 cycles)	false	1.6°C/sec	95.0°C	30	-	-	-
	true	1.6°C/sec	50.0°C	30	-	-	-
	true	1.6°C/sec	72.0°C	60	-	-	-
Hold Stage	false	1.6°C/sec	72.0°C	420	-	-	-

Primary Analysis Settings

General

PCR Stage/Step	Stage 2, Step 3
Quantification Cycle Method	Baseline Threshold

Target	Auto Threshold	Threshold	Auto Baseline	Baseline Start	Baseline End
DEFAULT	Yes	AUTO	Yes	AUTO	AUTO

QC Alerts

Curve Quality Alert Enabled	No
Results Quality Alert Enabled	Yes

Advanced

Set the Delta-Rn below which curves will be considered Non-Amplified	No
Primary Analysis Variant	N/A
Reduce dye signal crosstalk by algorithm	No

Standard Curve Settings

Standard Curve	On Plate

Standard Curve Results

Well	Sample	Sample Type	Target	Dyes	Task	Cq	Quantity
A1	STD_1	Standard	Bacteria	SYBR	Standard	13.72	2.0000e+7
A2	STD_1	Standard	Bacteria	SYBR	Standard	13.749	2.0000e+7
A3	STD_1	Standard	Bacteria	SYBR	Standard	13.752	2.0000e+7
A4	Y150356-MUC	Unknown	Bacteria	SYBR	Unknown	28.353	2.5932e+3
A5	Y150356-MUC	Unknown	Bacteria	SYBR	Unknown	28.177	2.8925e+3
A6	Y150356-MUC	Unknown	Bacteria	SYBR	Unknown	28.238	2.7853e+3
A7	2301-INOC	Unknown	Bacteria	SYBR	Unknown	25.335	1.6902e+4
A8	2301-INOC	Unknown	Bacteria	SYBR	Unknown	25.046	2.0225e+4
A9	2301-INOC	Unknown	Bacteria	SYBR	Unknown	25.03	2.0430e+4
B1	STD_2	Standard	Bacteria	SYBR	Standard	17.421	2.0000e+6
B2	STD_2	Standard	Bacteria	SYBR	Standard	17.769	2.0000e+6
B3	STD_2	Standard	Bacteria	SYBR	Standard	17.57	2.0000e+6
B4	Y150331-7	Unknown	Bacteria	SYBR	Unknown	25.226	1.8089e+4
B5	Y150331-7	Unknown	Bacteria	SYBR	Unknown	25.247	1.7853e+4
B6	Y150331-7	Unknown	Bacteria	SYBR	Unknown	25.308	1.7183e+4
B7	2302-INOC	Unknown	Bacteria	SYBR	Unknown	20.895	2.6633e+5
B8	2302-INOC	Unknown	Bacteria	SYBR	Unknown	20.914	2.6327e+5
B9	2302-INOC	Unknown	Bacteria	SYBR	Unknown	20.609	3.1818e+5
C1	STD_3	Standard	Bacteria	SYBR	Standard	21.24	2.0000e+5
C2	STD_3	Standard	Bacteria	SYBR	Standard	21.314	2.0000e+5
C3	STD_3	Standard	Bacteria	SYBR	Standard	21.498	2.0000e+5
C4	Y150332-7	Unknown	Bacteria	SYBR	Unknown	29.826	1.0391e+3
C5	Y150332-7	Unknown	Bacteria	SYBR	Unknown	29.656	1.1544e+3
C6	Y150332-7	Unknown	Bacteria	SYBR	Unknown	29.675	1.1408e+3
C7	2304-INOC	Unknown	Bacteria	SYBR	Unknown	28.384	2.5432e+3
C8	2304-INOC	Unknown	Bacteria	SYBR	Unknown	28.5	2.3664e+3
C9	2304-INOC	Unknown	Bacteria	SYBR	Unknown	28.304	2.6740e+3
D1	STD_4	Standard	Bacteria	SYBR	Standard	25.35	2.0000e+4
D2	STD_4	Standard	Bacteria	SYBR	Standard	25.496	2.0000e+4
D3	STD_4	Standard	Bacteria	SYBR	Standard	25.595	2.0000e+4


Well	Sample	Sample Type	Target	Dyes	Task	Cq	Quantity
D4	Y150333-7	Unknown	Bacteria	SYBR	Unknown	31.543	3.5747e+2
D5	Y150333-7	Unknown	Bacteria	SYBR	Unknown	31.14	4.5920e+2
D6	Y150333-7	Unknown	Bacteria	SYBR	Unknown	31.686	3.2718e+2
D7	2305-INOC	Unknown	Bacteria	SYBR	Unknown	23.12	6.6880e+4
D8	2305-INOC	Unknown	Bacteria	SYBR	Unknown	22.843	7.9438e+4
D9	2305-INOC	Unknown	Bacteria	SYBR	Unknown	22.949	7.4366e+4
E1	STD_5	Standard	Bacteria	SYBR	Standard	29.011	2.0000e+3
E2	STD_5	Standard	Bacteria	SYBR	Standard	29.084	2.0000e+3
E3	STD_5	Standard	Bacteria	SYBR	Standard	29.043	2.0000e+3
E4	Y150347-7	Unknown	Bacteria	SYBR	Unknown	29.044	1.6884e+3
E5	Y150347-7	Unknown	Bacteria	SYBR	Unknown	28.511	2.3514e+3
E6	Y150347-7	Unknown	Bacteria	SYBR	Unknown	28.681	2.1153e+3
E7	2306-INOC	Unknown	Bacteria	SYBR	Unknown	18.005	1.6031e+6
E8	2306-INOC	Unknown	Bacteria	SYBR	Unknown	17.859	1.7563e+6
E9	2306-INOC	Unknown	Bacteria	SYBR	Unknown	17.801	1.8201e+6
F1	STD_6	Standard	Bacteria	SYBR	Standard	32.4	2.0000e+2
F2	STD_6	Standard	Bacteria	SYBR	Standard	32.13	2.0000e+2
F3	STD_6	Standard	Bacteria	SYBR	Standard	32.161	2.0000e+2
F4	Y150348-7	Unknown	Bacteria	SYBR	Unknown	30.988	5.0487e+2
F5	Y150348-7	Unknown	Bacteria	SYBR	Unknown	30.634	6.2900e+2
F6	Y150348-7	Unknown	Bacteria	SYBR	Unknown	30.538	6.6733e+2
F7	2311-INOC	Unknown	Bacteria	SYBR	Unknown	27.498	4.4114e+3
F8	2311-INOC	Unknown	Bacteria	SYBR	Unknown	26.977	6.0960e+3
F9	2311-INOC	Unknown	Bacteria	SYBR	Unknown	27.094	5.6678e+3
G1	STD_7	Standard	Bacteria	SYBR	Standard	35.761	2.0000e+1
G2	STD_7	Standard	Bacteria	SYBR	Standard	36.471	2.0000e+1
G3	STD_7	Standard	Bacteria	SYBR	Standard	35.811	2.0000e+1
G4	Y150355-7	Unknown	Bacteria	SYBR	Unknown	22.317	1.1012e+5
G5	Y150355-7	Unknown	Bacteria	SYBR	Unknown	22.112	1.2511e+5
G6	Y150355-7	Unknown	Bacteria	SYBR	Unknown	22.04	1.3079e+5
G7	2317-INOC	Unknown	Bacteria	SYBR	Unknown	24.233	3.3507e+4

Well	Sample	Sample Type	Target	Dyes	Task	Cq	Quantity
G8	2317-INOC	Unknown	Bacteria	SYBR	Unknown	23.875	4.1847e+4
G9	2317-INOC	Unknown	Bacteria	SYBR	Unknown	23.628	4.8803e+4
H1	NTC	Negative Control	Bacteria	SYBR	Negative Control	38.835	-
H2	NTC	Negative Control	Bacteria	SYBR	Negative Control	38.294	-
H3	NTC	Negative Control	Bacteria	SYBR	Negative Control	-	-
H4	Y150356-7	Unknown	Bacteria	SYBR	Unknown	18.998	8.6545e+5
H5	Y150356-7	Unknown	Bacteria	SYBR	Unknown	18.873	9.3512e+5
H6	Y150356-7	Unknown	Bacteria	SYBR	Unknown	18.485	1.1903e+6
H7	2318-INOC	Unknown	Bacteria	SYBR	Unknown	22.122	1.2433e+5
H8	2318-INOC	Unknown	Bacteria	SYBR	Unknown	22.203	1.1820e+5
H9	2318-INOC	Unknown	Bacteria	SYBR	Unknown	22.253	1.1462e+5

Standard Curve Results - Replicate

Sample	Target	Replication Number	Quantity Mean	Quantity SD
2301-INOC	Bacteria	3	1.9186e+4	1.9802e+3
2302-INOC	Bacteria	3	2.8259e+5	3.0857e+4
2304-INOC	Bacteria	3	2.5279e+3	1.5436e+2
2305-INOC	Bacteria	3	7.3561e+4	6.3177e+3
2306-INOC	Bacteria	3	1.7265e+6	1.1153e+5
2311-INOC	Bacteria	3	5.3917e+3	8.7559e+2
2317-INOC	Bacteria	3	4.1385e+4	7.6584e+3
2318-INOC	Bacteria	3	1.1905e+5	4.9120e+3
NTC	Bacteria	3	-	-
Y150331-7	Bacteria	3	1.7708e+4	4.7043e+2
Y150332-7	Bacteria	3	1.1114e+3	6.3026e+1
Y150333-7	Bacteria	3	3.8128e+2	6.9155e+1
Y150347-7	Bacteria	3	2.0517e+3	3.3601e+2
Y150348-7	Bacteria	3	6.0040e+2	8.4923e+1
Y150355-7	Bacteria	3	1.2201e+5	1.0681e+4
Y150356-7	Bacteria	3	9.9696e+5	1.7104e+5
Y150356-MUC	Bacteria	3	2.7570e+3	1.5167e+2

Standard Curve Plot



Target:Bacteria Slope:-3.707 R^2:0.999 Y-Inter:41.009 Eff:86.097 Error:0.032

- End of Report -