CHAPTER 4 INFUSION THERAPY

Check the infusion rate every 15 minutes.

Report error E-21 to biomedical engineering on ext 2270.

4.2 Cleaning

Wipe the casing with 70% isopropyl alcohol.