
PROTOCOL

1. For each read, **top-label a 1.5 mL tube** with **the exact name you put on the submission form. Not PCR tubes.** Not an abbreviation, and not a name that only makes sense to you — the facility matches tube to form by that string.
2. Set up each reaction:
 - 3 G00101 sequencing oligo (2.66 μ M)
 - 10 miniprep DNA

13 μ L total
 - Our sequencing oligos are standardised to **2.66 μ M**, so **3 μ L is exactly one reaction's worth.**
 - **Medium copy** — pBR322, pAC, and **pP6: 10 μ L** miniprep.
 - **High copy** — pUC: **4 μ L** miniprep + 6 μ L water.
 - **Low copy** — pSC101, BACs: **PCR the region first** and sequence the product.
3. Specify the **dGTP protocol** on the submission form. Standard chemistry **dies inside a hairpin** and will lose the read.
4. Submit. Results come back in **1–2 days.**

NOTES

- Pick a primer **upstream** of what you want to see — the read starts **20–50 bp downstream** and gives **400–1000 bp.**
- **Sequence everything you picked**, including the boring ones. The middle of the range is the informative part.