

DeltaVision OMX™

Customer Instructions

Photokinetics



DeltaVision OMX™ Photokinetics

- ◆ Discusses important use and safety information
- ◆ Introduces Photokinetics
- ◆ Describes the Ring TIRF/PK Hardware Module
- ◆ Describes how to acquire and analyze PK (FRAP and FRET) data

Important user information



WARNING

Using controls, making adjustments, or performing procedures other than those specified in the DeltaVision OMX Imaging System's documentation can result in hazardous exposure to high voltage, laser radiation, or moving parts. Exposure to these hazards can cause severe personal injury. Do not operate the DeltaVision OMX Imaging System in any other way than described in the user documentation.



IMPORTANT!

Prior to reading this document, you must read and understand the safety and operating information described in the *DeltaVision OMX Operating Instructions* and the *DeltaVision OMX Getting Started Guide*.

Do not operate the DeltaVision OMX system in any way other than described in the user documentation. If you do, you may be exposed to hazards that can lead to personal injury and you may cause damage to the equipment.

Intended Use

Use of the DeltaVision OMX Imaging System assumes that you are familiar with the basics of fluorescence microscopy. In addition, an understanding of image processing basics will help you use the system to its full potential. To manage the computer systems, familiarity with Linux workstations and Windows-based personal computers is also helpful.

About Photokinetics

Photokinetics (PK) refers to the reactivity of fluorescent molecules while they are in an excited state. Photokinetics can be used to study the interactions of molecules within living cells. Photo-bleaching, FRET, and photo-activation are examples of PK reactions.

The following table shows photokinetic experiment methods and the biological applications that can be studied with each of these methods.

Table .1 Photokinetic Methods and Applications


Photokinetic Methods	Biological Applications
FRAP (Fluorescence Recovery After Photo-Bleaching) - Single-Point and Multi-Point	Affinity, Biomolecular Cycling, Biomolecular Environment, Structural Kinetics
Pattern Bleaching	Compartmental Analysis, Biomolecular Cycling, Transport
FLIP	Compartmental Analysis, Biomolecular Cycling, Transport, Structural Visualization
Background Reduction	Structural Visualization
Combinations: - FRAP/FRET - Repeat during cell cycle - Rapid Repeat	Compartmental Analysis, Biomolecular Cycling, Transport Affinity, Biomolecular Cycling, Biomolecular Environment
FRET: - Sensitized emission - Donor Photo-bleaching - Acceptor Depletion	Affinity, Biomolecular Environment
Photo Activation	Compartmental Analysis, Affinity, Biomolecular Cycling, Biomolecular Environment, Transport, Cell Fate, Structural Kinetics, Structural Visualization

The Ring TIRF/PK Module

Light is delivered to the sample for PK applications using a by-pass system incorporated into the DeltaVision OMX Ring TIRF Module. The same mirror system is used to control both the TIRF and PK beams, making it possible to bleach areas of different sizes and configurations, depending on how the PK beam is directed. The following figure illustrates the PK light path.

NOTE Based on the system hardware configuration, the light path may vary from that shown in the following figure.

Figure 1. Ring TIRF Module - PK Light Path



Optical Components

- F1: Fiber Launch
- M1: Mirror 1
- A1: Field Aperture
- L1: Collimating Lens
- G1: X-Galvo
- G2: Y-Galvo
- P1: Polarizing Cube
- L2: Focusing Lens
- M2: Mirror2
- D1: Polychroic Mirror
- D2: HWAf Dichroic Mirror
- OBJ: Objective Lens
- M3: PK Mirror 3
- M4: PK Mirror 4
- M5: PK Mirror 5
- M6: PK Mirror 6
- M7: PK Mirror 7
- M8: PK Mirror 8
- M9: PK Mirror 9
- L3: Focusing Lens

The Ring TIRF system utilized on the DeltaVision OMX system was developed by researchers at Yale University and is exclusively licensed to GE Healthcare Company.

Ring TIRF/PK Hardware

The Ring TIRF/PK module is mounted on the DeltaVision OMX system block, as shown in the following figures.

NOTE Based on the system configuration, the hardware may vary from that shown in the following figures.

Figure 2. Ring TIRF/PK Module Mounted on DeltaVision OMX System Block (Yellow-Bordered Area)




Figure 3. Ring TIRF/PK Module (Close Up)



Ring TIRF/PK Laser Safety Considerations

Due to the Ring TIRF/PK illumination optics, the light being emitted from the DeltaVision OMX objective is collimated and has high power density. The TIRF/PK system also has the ability to direct this light to sharp off-axis angles relative to the objective axis.



WARNING!

The DeltaVision OMX Imaging System is a Class I laser system. No access to laser radiation is permitted during operation or maintenance. During service, however, Class IV radiation is accessible.

Due to the potential for personal injury, particularly to the eyes, service on the DeltaVision OMX Imaging System should **ONLY** be performed by GE personnel or persons trained by GE specifically for this purpose. Unauthorized service by any other personnel may violate the warranty.



WARNING! (Service Only) When servicing the TIRF/PK system, use extreme caution that the emitted light is not directed into anyone's eyes. Appropriate laser safety goggles selected for the specific wavelength being tested are mandatory.

Running PK Experiments

This section describes how to run a basic photokinetic experiment, such as FRAP or FRET.

To run a basic PK experiment:

1. Mount the sample slide and focus on the sample using the appropriate camera and excitation settings for your experiment.



Tips

- #1 Selecting a "Conventional" **Light Path** setting may help you focus on the sample.
- #2 Raising the stage brings the coverslip more closely into focus so, to focus near the coverslip, raise the stage to the highest position at which the sample remains in focus.

2. In the **Image Mode** field, select either the "Simultaneous" or the "Sequential" imaging mode. For multi-channel PK experiments, select "Simultaneous" for faster imaging after the PK event.
3. In the **Light Path** field, select "TIRF" or "Conventional."

Figure 4. Select the TIRF or Conventional Light Path



- Set the **%T** and the **Exposure** fields to produce imaging conditions at the desired signal level.
- On the Experiment tab on the main program window, select the “Photo Kinetic” experiment **Type**. (The PK-specific fields will appear at the bottom of the tab, in a section titled Laser Event, as shown in the following figure.)

Figure 5. Photo Kinetic Experiment Tab

Experiment **Spiral Mosaic** DV Points

Experiment Type
Type **Photo Kinetic** Save Exp... Open Exp...

☐ Sectioning
Focus point when scan starts middle
Optical section spacing 0.200
Number of optical sections 1
Sample thickness 0 Get thickness

☒ Time-lapse
Hours Minutes Seconds Msecs
Time-lapse 0 0 1 0
Time points 10 Total time 0 0 10 0
☐ Subset Time-lapse
Every 0 time points
Image:

Laser Event
Laser channel 488 %T 100.0% Duration(secs) 0.050
Sequence Repeats 1 Event before timepoint 4
☐ Spot ☒ Rectangle Undo Delete All ☒ HW UltimateFocus ...

Data File Save As...
Title Run

- Set up the **Sectioning** parameters as desired for your photokinetic experiment. If sectioning **is not** enabled, the laser event will occur at the imaging plane. If sectioning **is** enabled, the laser event will occur at the top, middle, or bottom of the sample, as defined in the **Focus point when scan starts** field.
- Set up the **Time-lapse** parameters as desired for your photokinetic experiment.
- In the Laser Event section, select the **Laser channel** to use for the laser event.
- Adjust the **%T** and **Duration (secs)** fields as required for your experiment.

NOTES

#1 When setting up these fields, we recommend that you use alternative (sacrificial) cells for testing the bleaching/activation effects.

#2 If you are imaging in TIRF mode and using the same laser line for both illumination and the laser event, the **%T** setting will remain at the same level for both imaging and the PK event.

- Specify how often the photokinetic event sequence repeats in the **Sequence Repeats** field.


NOTE Typical FRAP experiments will use a Sequence Repeats value of “1”. If another value (**x**) is used, the entire timelapse sequence will be repeated **x** times. For example,

if **Sequence Repeats** was set to "3" in the experiment above, a total of 30 timepoints would be acquired with the laser events occurring three times.

11. Specify the number of time points before the laser event occurs in the **Event before timepoint** field. For the experiment shown in the previous figure, the laser event will occur between time points 3 and 4.
12. Laser events can be a single spot or rectangle, multiple spots or rectangles, or a combination of both.

NOTE The PK targeting accuracy is best near the middle of the field of view and will decrease as you move outwards. For best results, position the areas to bleach (activate) near the center of the image. For reference, expected accuracy for the image below (512x512 field of view) is about 0.5 μm in area 1, about 1 μm in area 2, and about 2 μm in area 3,

Figure 6. PK Targeting Accuracy




For spots, click the **Spot** radio button then click on the image where you want the bleach event to occur. The spot is roughly 0.5 microns in diameter and will change based on the objective used. The bleach spot is formed by moving the galvos to the defined position while the stage remains static throughout the laser event.

For rectangles, click the **Rectangle** radio button and then click and drag the mouse cursor on the image to draw a rectangle of the desired size. The bleach rectangle is formed by continuously moving the galvos during the laser event to form a Lissajous pattern. The density of the Lissajous pattern is defined by clicking the **Rectangle** button (which appears when you select the **Rectangle** radio button). In the PK Mode Setup dialog box that appears, make the appropriate **PK Mode** selection based on the following definitions:

- **Dense.** Provides a denser fill for larger rectangles.
- **Sparse.** Provides a sparser fill for smaller rectangles.
- **Manual.** Allows you to specify the X and Y frequency for the pattern. The ideal settings are when $Y = X - 1$ or $X + 1$.

As the Lissajous parameters are changed, an estimation of the laser pattern is drawn in the window. If necessary, click the **Undo** or **Delete All** buttons to remove the bleach patterns and start over.

Figure 7. Defining a Rectangular Event Pattern




13. If you would like to use **HW UltimateFocus**, select the check box and click the ... button to set up the parameters.
14. Enter a **Data File** name and **Title** (if desired) and then click **Run** to start the experiment.

Analyzing FRAP Data

The FRAP experiment method consists of photo-bleaching a point (or points) of interest and then observing the recovery of the fluorescence in the bleached area.

An example of a single-point FRAP experiment is shown in the following figure.

Figure 8. A Point of Interest Is Photo-Bleached and Monitored



There are two types of FRAP data:


- Single-Point FRAP data is collected in experiments that monitor a single location or that monitor several locations in a sequential fashion.
- Multi-Point FRAP data is collected in experiments that monitor several locations in the sample at the same time.

The following procedure describes how to analyze FRAP data using the softWoRx Imaging Workstation.

To analyze FRAP data:

1. Open an image in the Image window. From the softWoRx main menu, choose **Measure:PK Analysis** to display the Photokinetic Data Analysis window.

Figure 9. PK Data Analysis Window




NOTE You can use the Photokinetic Data Analysis window to specify a recovery model, type of ROI, beam profile shape, and number of sites. You can also use it to remove background intensity, select which Z sections and wavelengths to include in the analysis, and specify other options.

2. Drag the window number into the **Input Image** field.
3. Select the desired **Response** type, **Recovery Model**, and **Recovery Data From** options.
4. If you are analyzing a Multi-point FRAP data set, enter the number of laser sites into the **Number of Laser Sites** field.
5. If you are using background subtraction, enter a background value into the **Background Intensity** field.
6. To determine a number to enter into the **Spot Radius** field, use the Measure Distances tool to make an approximate measurement of the bleach spot.
7. Click **Do It** to run the analysis.

The software will generate the following files:


- A three channel output image file containing 1) the original time lapse image, 2) the ratio of the current time point to average pre-bleach time points, and 3) the ratio data at the location used for analysis.

Figure 10. Three Channel Image Generated by the Photokinetic Data Analysis Tool




- A JPEG file containing a recovery graph that plots the fluorescence intensity before and after the event.

Figure 11. Fluorescence Recovery Graph



- A log file containing the analysis results.

Figure 12. FRAP Analysis Log File



For local office contact information, visit:

www.gelifesciences.com/contact

GE Healthcare UK Limited
Amersham Place, Little Chalfont
Buckinghamshire, HP7 9NA
UK

www.gelifesciences.com



imagination at work

GE, imagination at work, and the GE monogram are trademarks of General Electric Company.

DeltaVision, DeltaVision OMX, softWoRx, UltimateFocus, and Blaze are trademarks of GE Healthcare companies.

Microsoft and Windows are either registered trademarks or trademarks of Microsoft Corporation in the United States and/or other countries.

All other third party names and trademarks referred to in this document are the property of their respective owners.

Any use of this software is subject to GE Healthcare Standard Software End-User License Agreement for Life Sciences Software Products. A copy of this Standard Software End-User License is available on request.

© 2014 General Electric Company – All rights reserved. First published June 2014.

All goods and services are sold subject to the terms and conditions of sale of the company within GE Healthcare which supplies them. A copy of these terms and conditions is available on request. Contact your local GE Healthcare representative for the most current information.

Notice to purchaser

DeltaVision OMX Imaging Systems are for research use only – not for use in diagnostic procedures.

GE Healthcare Bio-Science AB
Björkgatan 30
751 84 Uppsala
Sweden

GE Healthcare Europe, GmbH
Munzinger Strasse 5
D79111 Freiburg
Germany

GE Healthcare Bio-Sciences Corp.
800 Centennial Avenue, P.O. Box 1327
Piscataway, NJ 08855-1327

GE Healthcare Japan Corporation
Sanken Bldg., 3-25-1, Hyakunincho
Shinjuku-ku, Tokyo 169-0073
Japan