

Set up Bright Field for Color imaging in Elements

1. Change to the objective you will be imaging/sticking the tissue with
2. Go to White area on slide (no tissue)
3. Under the Color DS-Ri2 Setting tab:
 - a) Click the “AE button
 - b) Click Auto White
 - Check the “LUTS” window – the red green and blue histograms (little hills) should line up
4. Click the triangle next to “BF” in the OC panel to make the computer remember

Note: Only need to do this once at the beginning of imaging session!

