



ECMPS Client Tool

Version 1.* 2022 Q2

QA/Cert Test Evaluation Report

November 18, 2022 05:02 PM

Facility Name: Barry

Facility Details

Facility ID (ORISPL): 3
State: AL
County: Mobile County

Evaluation Results

Unit/Stack Identifier	Test Type	Test Number Test Date/Hour	Sys / Comp ID / Type	Severity	Check Code	Result
6B	CYCLE	1 05/01/2019 10:07	CO2/CO2	INFORM	CYCLE-5-C	This test has already been submitted and will not be resubmitted. If you wish to resubmit this test, please contact EPA for approval.
					CYCLE-11-D	The BeginMonitorValue for Gas Level HIGH was greater than or equal to the EndMonitorValue. These values are unexpected for a 'high' level gas injection.
		2 05/01/2019 10:07	NX1/NOX	CRIT1	CYCLE-12-C	The reported ComponentCycleTime for Gas Level ZERO does not equal the value of the component cycle time recalculated from the reported starting and ending times.
					CYCLE-13-A	You reported a test Begin Date, Hour, and Minute that is not the same as the Begin Date, Hour, and Minute of the first injection in the cycle time test.
					CYCLE-17-C	The recalculated total cycle time does not equal the reported TotalCycleTime for the test.
					CYCLE-18-D	The TestResultCode indicates a passing test, but the result recalculated from the cycle time injection records indicates a failing test.
				INFORM	CYCLE-5-C	This test has already been submitted and will not be resubmitted. If you wish to resubmit this test, please contact EPA for approval.