

National Human Exposure Assessment Survey (NHEXAS)

Region 5 Study

Quality Systems and Implementation Plan for Human Exposure Assessment

Research Triangle Institute
Research Triangle Park, NC 27079
Cooperative Agreement CR 821902

Laboratory Operations Protocol

RTI/ACS-AP-209-121

Title: Analysis of House Dust for Arsenic

Source: Research Triangle Institute

U.S. Environmental Protection Agency
Office of Research and Development
Human Exposure & Atmospheric Sciences Division
Human Exposure Research Branch

Notice: The U.S. Environmental Protection Agency (EPA), through its Office of Research and Development (ORD), partially funded and collaborated in the research described here. This protocol is part of the Quality Systems Implementation Plan (QSIP) that was reviewed by the EPA and approved for use in this demonstration/scoping study. Mention of trade names or commercial products does not constitute endorsement or recommendation by EPA for use.

TITLE: ANALYSIS OF HOUSE DUST FOR ARSENIC

SOURCE: Research Triangle Institute
Post Office Box 12194
Analytical and Chemical Sciences
Research Triangle Park, NC 27709-2194

AUTHOR(s):

Ruando Date: 05.31.95

_____ Date: _____
_____ Date: _____

APPROVED BY:

Principal Investigator: E. Pelligrini Date: 6/2/95

QA Officer: HJ Smith Date: 6/10/95

<u>STATUS:</u>	IN PROGRESS:	<input type="checkbox"/>
	DRAFT:	<input type="checkbox"/>
	FINAL VERSION:	<input checked="" type="checkbox"/>

REVISIONS:

No.	Date	No.	Date
0	†	6	
1		7	
2		8	
3		9	
4		10	
5		11	

† Effective date of this version is the date of the last approval signature;
revision 0 is the original version.

ANALYSIS OF HOUSE DUST FOR ARSENIC

TABLE OF CONTENTS

<u>Section</u>		<u>Page</u>
<u>1.0</u>	<u>Scope and Application</u>	4
<u>2.0</u>	<u>Summary of the Method</u>	4
<u>3.0</u>	<u>Interferences</u>	4
<u>4.0</u>	<u>Safety</u>	5
<u>5.0</u>	<u>Equipment</u>	5
	<u>5.1 Laboratory Equipment</u>	5
	<u>5.2 Hydride Generation Atomic Fluorescence Spectrometer (HGAF)</u>	6
<u>6.0</u>	<u>Reagents and Standards</u>	7
	<u>6.1 Reagents</u>	7
	<u>6.2 Stock Standard Solutions</u>	7
	<u>6.3 Working Standard Solutions</u>	7
	<u>6.4 Calibration Standards</u>	8
<u>7.0</u>	<u>Sample Storage</u>	8
<u>8.0</u>	<u>Quality Assurance and Quality Control Procedures</u>	8
	<u>8.1 Field Blanks</u>	8
	<u>8.2 Method Blanks</u>	8
	<u>8.3 Method Controls</u>	9
	<u>8.4 Laboratory Controls</u>	9
	<u>8.5 Performance Evaluation Samples</u>	9
	<u>8.6 Custody and Sample Tracking Procedures</u>	9
<u>9.0</u>	<u>Sample Preparation</u>	10
<u>10.0</u>	<u>Calibration and Standardization</u>	11
	<u>10.1 Initial Calibration</u>	11
	<u>10.2 Initial Calibration Check</u>	12
<u>11.0</u>	<u>Analysis Procedure</u>	13
	<u>11.1 Analysis Conditions</u>	13
	<u>11.2 Analysis Batches</u>	13
	<u>11.3 Data Acquisition</u>	14
	<u>11.4 Identification of the Analyte</u>	14
<u>12.0</u>	<u>Method Performance</u>	15
	<u>12.1 Method Detection Limit (MDL)</u>	15
	<u>12.2 Method Quantitation Limit (MQL)</u>	15
<u>13.0</u>	<u>Data Management</u>	15
	<u>13.1 Processing of Data Output</u>	15
	<u>13.2 Data Storage</u>	16
	<u>13.3 Data Transmission</u>	16
<u>14.0</u>	<u>Calculations</u>	16

LIST OF FIGURES

<u>Number</u>		<u>Page</u>
<u>1</u>	<u>Batch Sample Submission Form</u>	21
<u>2</u>	<u>Example of a Printed Custody Record</u>	22

LIST OF TABLES

<u>Number</u>		<u>Page</u>
<u>1</u>	<u>Suggested Calibration Solutions</u>	18
<u>2</u>	<u>Operating Parameters for the Instrument</u>	19
<u>3</u>	<u>Example of An ASCII Data File Generated by the HGAF Software</u>	20

1.0 SCOPE AND APPLICATION

The purpose of this protocol is to provide guidelines for the analysis of wipes, that were used in Liroy-Weisel surface samplers to collect house dust, for arsenic (As) by hydride generation atomic fluorescence spectrometry (HGAF). This protocol describes the methodology and all other analytical aspects involved in the analysis of carpet dust samples for arsenic by hydride generation atomic fluorescence spectrometry (HGAF).

2.0 SUMMARY OF THE METHOD

This method involves the extraction of the analyte from wipe samples using 50% ultra-pure nitric acid, and subsequent analysis by hydride generation atomic fluorescence spectrometry. The analyte is identified by the presence of fluorescence at its specific wavelength. Quantitation of the analyte is carried out using a standard calibration curve. Recovery of the analyte is monitored using samples that are spiked with standards prior to the extraction process.

3.0 INTERFERENCES

3.1 In most cases contamination occurs through reagents, sample handling and sample collection materials. These sources of contamination are monitored by the analysis of both field blanks and method blanks as described in Sections 8.1 and 8.2.

3.2 Memory effects are a common source of contamination in HGAF. Flushing of sample delivery system and a proper cleaning cycle must be employed to ensure the complete removal of any traces of analytes from the sample (atomizer) cell and sample transport system.

3.3 Once samples are brought to the analytical lab, all sample handling is performed in a clean laboratory environment under HEPA filtered air. All sample preparation steps must be carried out in the class 100 clean lab environment to eliminate contamination that may result from the environment. If samples need to be carried out from the class 100 clean lab at any

stage of the sample processing, they should be tightly capped before taken out from the clean lab and never be opened to outside environment. All sample analyses must be performed in the class 10,000 clean lab to eliminate any contamination that may occur during analysis. Proper procedures for the use of ACS clean lab facility are described in NHX/SOP-300-006, "Standard Operating Procedures for the ACS Inorganic Class 100/10,000 Clean Lab Facility".

3.4 Contaminants in reagents, solvents, labware (glassware and plasticware) and other components of sample processing apparatus can also cause interferences which may lead to increased analyte and/or background signals. Atomic fluorescence reagents and apparatus are checked routinely to make sure they are free from interferences under the conditions of the analysis by method blanks as described in Section 8.2.

4.0 SAFETY

Since the toxicity of the chemicals used in this method is not clearly defined, they should be treated as potential health hazards at all times. All laboratory personnel are advised to take full precautions (wearing gloves, eye protection etc.) when handling these chemicals and personal exposure to these chemicals should be minimized. Safety issues in the clean lab facility are discussed in the NHX/SOP-300-002, "Standard Operating Procedures for the Personnel Safety in the ACS Inorganic Clean Lab Facility" in detail.

5.0 EQUIPMENT

5.1 Laboratory Equipment

5.1.1 All labware (glassware and plasticware) must be thoroughly cleaned by washing with detergent and water followed by rinsing with deionized water. Once the washing is completed, all labware must be soaked in 50% HNO₃ for not less than 24 hours, or heated in 50% HNO₃ for not less than 2 hours, rinsed with deionized water and dried in HEPA filtered air. All cleaned labware must be stored in air tight plastic bags if they are not used

immediately. Cleaning of labware in the ACS inorganic clean lab facility is described in NHX/SOP-300-007, "Standard Operating Procedures for Cleaning Labware in the ACS Inorganic Clean 100/10,000 Clean Lab Facility".

5.1.2 Labware

Beakers, Teflon and glass

Volumetric flasks, polypropylene and glass

Pipettes, glass

Extraction tubes, polypropylene

Storage bottles, plastic

Filter paper, Whatman

5.1.3 Apparatus

Hot plates

Variable speed horizontal shaker

Automatic pipettors, fixed and variable volume

Refrigerator

5.1.4 Analytical balance with 0.01 mg to 205 g weighing range. Capable of 0.01 mg readability.

5.2 Hydride Generation Atomic Fluorescence Spectrometer (HGAF)

5.2.1 The spectrometer must have the capability of automation to reduce personnel errors to a minimum. Lamp current setting and PMT gain setting should be able to be controlled individually. This is necessary to optimize these parameters during method development and also to provide consistency of these parameters during the analysis from sample to sample.

5.2.2 Vapor generator must have the capability to generate hydride species in both continuous mode and flow injection mode. It should have an active pumping system to deliver the sample, blank and reagent to the mixing chamber with high reproducibility. The presence of an efficient separation technique to separate gaseous species from the solution is also important and finally, there should be an effective transport system to transport the gaseous species to the atomizer cell.

- 5.2.3 An autosampler is required to deliver reproducible volumes of sample and reagents into the reaction chamber or mixing vessel in a consistent manner (addition of reagents in proper order).
- 5.2.4 The data system must be capable of data acquisition, storage and processing including producing an output. Data acquisition software must allow the use of different integration periods and also be capable of acquiring data under both peak height and peak area modes. Data processing software must be capable of analyzing the raw data, calculating net fluorescence, calculating statistics (mean, standard deviation, relative standard deviation), constructing calibration curves and calculating analyte concentrations. It is also recommended that the data system be able to transfer its data from the instrument to peripheral computers for further processing.

6.0 REAGENTS AND STANDARDS

6.1 Reagents

Hydrochloric acid ("Trace Metal" grade or better)

Nitric acid ("Optima" grade or better)

Sodium borohydrate (ACS reagent grade)

Sodium hydroxide (ACS reagent grade)

Potassium iodide (ACS reagent grade)

Ascorbic acid (ACS reagent grade)

6.2 Stock Standard Solutions

Commercially available, spectrometric grade, certified arsenic reference solution (NIST traceable) at 1000 mg/L is used as the stock solution.

6.3 Working Standard Solutions

Working standard solutions are prepared by diluting the stock standard solution using 2% HCl to stabilize the analyte. Working standard solutions are stored in acid washed polypropylene containers and prepared freshly when necessary.

6.4 Calibration Standards

A series of calibration standards are prepared on a daily basis by dilution of working standard solutions with deionized water. A suitable matrix component may also be added to calibration standards before they are made up to their final volumes, to stabilize the analyte and also to match the matrix to that of prepared samples. At least five different concentration standards must be used to define the linear portion of the calibration curve, and a calibration blank should also be prepared and included in the calibration. Calibration blank should contain all the components in the calibration standards except the analyte. Calibration standards should cover the full range of analyte concentrations expected to be analyzed by the instrument. The concentrations of calibration standards are based on the concentrations found in typical dust samples, and are given in Table 1.

7.0 SAMPLE STORAGE

All samples are placed in appropriate airtight plastic containers and are stored in a refrigerator prior to sample preparation.

All prepared samples are stored in plastic containers with air tight screw caps, in a refrigerator.

8.0 QUALITY ASSURANCE AND QUALITY CONTROL PROCEDURES

8.1 Field Blanks

Field blanks are unspiked wipes taken to the field along with the other samples except that they are not used for sample collection. These field blanks help in monitoring contamination, if any, from handling and transportation.

8.2 Method Blanks

Method blanks are extracting solvent carried throughout the sample preparation procedure along with each batch of samples analyzed. Analysis of these will provide information on contamination that result from sample extraction and preparation steps.

8.3 Method Controls

Method controls are extracting solvents spiked with an NIST certified calibration standard at levels described in Section 6.4 and treated the same as method blanks. Analysis of these will provide information on recoveries of the target analyte from sample extraction and preparation steps.

8.4 Laboratory Controls

Laboratory controls are blank wipes spiked with an NIST certified calibration standard at levels described in Section 6.4 and treated the same as field samples.

Laboratory control samples will be processed and analyzed prior to processing any samples. These are used to demonstrate the performance of the method before analyzing field samples.

Laboratory controls are designed to represent field samples. These samples undergo all extraction and sample preparation steps that field samples undergo. The recovery of the analyte will be monitored as well.

8.5 Performance Evaluation Samples

Once acceptable performance of the method is demonstrated for the analyte by laboratory controls, a performance evaluation sample is analyzed to further evaluate the overall performance of the method. The performance evaluation sample must contain the target analyte with the level certified by an approved material certification agency, in a matrix almost identical to the sample matrix. An example of this type of material is NIST SRM 1648, a certified urban particulate. The arsenic concentration in this material is 115 ± 10 Fg/g.

Performance of the method may also be evaluated by the analysis of a simulated sample provided by NIST.

8.6 Custody and Sample Tracking Procedures

8.6.1 Sample Tracking Procedures

Each and every sample will be identified by a unique bar code. These bar codes are clearly placed on all samples and will be inspected at various stages, from the transportation to the field, to their final analysis.

Detailed information regarding sample transportation to the monitoring sites (sampling sites), and to sample analysis site (analytical lab), sample weighing, sample extraction and sample analysis will be recorded on RTI log books. When a batch of samples is submitted to the analytical lab for analysis, the analyst is responsible for recording sample information along with the analysis information in the sample submission form. An example for this type of record is given in Figure 1.

Method blanks and method control samples will be linked to sample batches by a date code. Sample preparation and analysis dates will be recorded in laboratory notebooks.

8.6.2 Custody

Custody procedures will be designed to track down the sample flow within the study. Each sample is accompanied by a custody document which provides the information on that sample. An example is given in Figure 2.

9.0 SAMPLE PREPARATION

Samples are placed in separate 35 mL polypropylene test tubes and 10 mL of 50% HNO_3 is added to each test tube. Sample containers are rinsed with 1 mL of deionized water to recover any dislodged particles and added to respective test tubes. Tubes are sealed with airtight plastic screw caps and are placed on a horizontal shaker for 30 minutes. The extract is decanted into a 50 mL Teflon beaker and the extraction is repeated twice with a fresh portion of acid each time (total extract vol. = 30 mL). All three extracts are combined in a 50 mL beaker and evaporated down to about 0.5 mL on a hot plate. The beaker is removed from the hot plate, allowed to cool down, and 12.5 mL of 50% HCl is added. The solution is warmed and filtered if necessary using a Whatman filter paper into a 25 mL polypropylene volumetric flask and diluted with deionized water. The final solution is transferred to a clean plastic storage bottle and is stored in a refrigerator until analysis.

10.0 CALIBRATION AND STANDARDIZATION

An acceptable performance of the instrument must be demonstrated and documented as described in Section 2.3 of NHX/SOP-300-001, "Standard Operating Procedure for the Operation of PS Analytical Hydride Generation Atomic Fluorescence Spectrometer (HGAF)", prior to any sample analysis and the demonstrated acceptable level must be maintained throughout the analysis. A new calibration curve is constructed at the beginning of each day of analysis. An additional calibration checks are performed periodically, when necessary.

10.1 Initial Calibration

- 10.1.1 An instrument performance test is performed prior to the analysis of any samples as described in Section 2.3 of the NHX/SOP-300-001, "Standard Operating Procedure for the Operation of the PS Analytical Hydride Generation Atomic Fluorescence Spectrometer (HGAF)". If the performance test meets the acceptance criteria defined in the SOP, proceed with the initial calibration (starting from Section 10.1.2). If the performance test does not meet the acceptance criteria, necessary steps (Section 10.2.5) will be taken to bring the performance of the instrument to an acceptable level.
- 10.1.2 Calibration blank and calibration standards are analyzed for arsenic starting from the lowest concentration to the highest. Each solution (blank and standards) is analyzed at least in duplicate in order to perform statistical evaluation of the analysis. For each calibration standard and blank, the mean, the standard deviation (SD) and, the percent relative standard deviation (%RSD) is calculated. If the RSD is greater than 10% for any standard, further aliquots of that solution will be analyzed or necessary steps will be taken to improve the performance of the instrument.
- 10.1.3 Fluorescence is measured for each calibration standard and a least squares linear regression calibration curve is constructed as follows for each analyte by the instrument software.

$$y = a + bx$$

where:

y = net fluorescence intensity

a = y intercept

b = slope

x = analyte concentration in ng/mL

10.2 Initial Calibration Check

At the beginning of each day of analysis, a new calibration curve is constructed. The performance of the initial calibration must be verified prior to any sample analysis as follows.

- 10.2.1 Linearity of the calibration curve must be verified both visually and mathematically. The correlation coefficient (r) must be greater than 0.99.
- 10.2.2 Concentration of the analyte in the calibration standards calculated using the measured signal and the calibration curve must be $\pm 20\%$ of the nominal concentration for lowest calibration standard and $\pm 10\%$ for all other calibration standards.
- 10.2.3 Analyze a predetermined QC check and compare the new results with the previous results obtained from Section 10.1.2.
- 10.2.4 Calculate the difference between the two values and verify that it has not changed by more than 10% from the initial calibration. If it has changed more than the above limits, adjustments must be made to restore system sensitivity (Section 10.2.7) and recalibration is required.
- 10.2.5 Analyze a NIST certified standard to ensure that the analytical accuracy and precision remains within acceptable limits. If the results differ more than 10% from the certified value or from one of the value limits, a second aliquot will be analyzed or the NIST standard will be re-prepared and re-analyzed. Questionable situations will be reviewed with the laboratory manager or facility supervisor and the QA Officer, and a final decision taken at that time.

10.2.6 Control Charts

Daily performance test results will be plotted against time (in days) for the HGAF instrument. The sensitivity (Pk. Ht/ng.mL^{-1}) will be used as the performance parameter. These graphs will be used as a control chart to continuously monitor the performance of the instrument over time. The operator of the instrument is responsible for maintaining the chart. The chart must be presented to the laboratory manager or to the task leader for his/her review at least once a month. The control chart and data must be stored along with the sample analysis data.

10.2.7 Remedial Actions

Possible remedial actions are described in the NHX/SOP-300-005, "Standard Operating Procedure for the Maintenance of PS Analytical Hydride Generation Atomic Fluorescence Spectrometer" and in the NHX/SOP-300-001, "Standard Operating Procedure for the Operation of the PS Analytical Hydride Generation Atomic Fluorescence Spectrometer".

11.0 ANALYSIS PROCEDURE

11.1 Analysis Conditions

The instrumental parameters for the analyte are given in Table 2. Each sample is analyzed similar to standards as described in Section 10.1.2.

11.2 Analysis Batches

Sample preparation and analysis are carried out in batches of samples. A batch is defined as a fixed number of samples that can be handled conveniently and efficiently in the laboratory. A batch includes 30 samples, 2 method blanks, and 1 method control. Field blanks are also included in the batch but, as they are unknown to the analyst, they will be treated as samples. The order of analysis is:

QC check standard

Calibration blank

NIST reference standard

Method blank

Method control

Samples

Periodic QC checks

Prior to any sample analysis, a set of calibration standards will be analyzed and a calibration curve constructed. A pre-determined calibration standard will be selected as the QC check standard and analyzed periodically along with the calibration blank immediately after the calibration, every 10 samples, and at the end of the analysis. All QC check standards must be within 10% of their nominal value or initial values for the data to be accepted.

Method blanks will be analyzed with each batch of samples and subtracted from method controls/samples when they are found to be significant. The analyst will make this determination.

A method control will be run with each batch of samples and the recovery of analyte(s) will be monitored. If the recoveries are outside the range 90 to 100% from the certified value or from one of the value limits, a second aliquot will be analyzed. Questionable results will be reviewed with the laboratory manager or facility supervisor and the QA Officer and a final decision is made at that time.

11.3 Data Acquisition

Data acquisition is carried out by the instrument software. The fluorescence signal given out by excited analyte atoms is measured during data acquisition. Once the data acquisition is completed, the concentration of the analyte in the sample is calculated based on the standard calibration curve.

11.4 Identification of the Analyte

Analyte is identified by the presence of fluorescence signal, at wavelength specific to the analyte. The analyte is present in the sample, if the concentration of the analyte is greater than its method detection limit (MDL).

12.0 METHOD PERFORMANCE

12.1 Method Detection Limit (MDL)

Method detection limit (MDL) is defined as 3 times the standard deviation of method blanks. Method detection limit is calculated for the analyte using the following equation;

$$MDL = 3 \times SD_{bl}$$

MDL = Method detection limit, ng/mL or ppb.

SD_{bl} = Standard deviation of method blanks, ng/mL or ppb.

12.2 Method Quantitation Limit (MQL)

Method quantitation limit (MQL) is defined as 10 times the standard deviation of method blanks. Method quantitation limit is calculated for the analyte using the following equation;

$$MQL = 10 \times SD_{bl}$$

MQL = Method quantitation limit, ng/mL or ppb.

SD_{bl} = Standard deviation of the method blank, ng/mL or ppb.

13.0 DATA MANAGEMENT

13.1 Processing of Data Output

Instrument software is capable of processing the raw data (fluorescence intensities) and generating printed output. The output contains information on sample I.D., analyte tested, fluorescence intensity, concentration of the analyte in the sample, and statistical data. This type of output helps in visual examination of data for any unusual behavior and/or any inadequate method performance.

13.2 Data Storage

All raw data acquired during the analysis are stored on the hard drive of the computer dedicated to the instrument, along with the processed data. At the end of each day of analysis all data are transferred to floppy disks. These floppy disks are labelled and stored in a central storage area along with the printed outputs.

13.3 Data Transmission

Once reviewed and accepted by the laboratory manager, raw and processed data generated by the instrument are converted to ASCII format and handed over to the database manager for further processing. Example of an ASCII data file generated by HGAF instrument software is given in Table 3.

14.0 CALCULATIONS

The concentration of the analyte in the sample is calculated during the analysis by the instrument software against the reference calibration curve (Section 10.1.3) using the measured fluorescence as follows:

$$C_x = \frac{(y-a)}{b}$$

where:

- C_x = Uncorrected concentration of the analyte in ng/mL, or parts-per-billion (ppb).
 y = The measured fluorescence for the sample.
 a = y intercept of the reference calibration curve in fluorescence.
 b = Slope of the reference calibration curve in fluorescence per ng mL⁻¹.

Samples with concentrations below MDL/MQL will be expressed appropriately, whereas over-range samples will be diluted and re-analyzed.

Once the data are transmitted to the central database, calculations may be performed to express the final results in ng of analyte per cm² of area sampled or ng of analyte per g of dust or any other form.

- 14.1 The following formula may be used to calculate the final results in ng of analyte per gram of dust sampled.

$$C_{x2} = (C_x + C_{MB}) \left(\frac{V_x}{DF} \right) \left(\frac{1}{Wt} \right)$$

where:

- C_{x2} = Corrected concentration of the analyte in ng/g of dust.
 C_x = Uncorrected concentration of the analyte in ng/mL or ppb.
 V_x = Final volume of the sample in mL.
DF = Dilution factor.
Wt = Mass of the dust sampled in g.
 C_{MB} = The average concentration of the analyte in method blanks in ng/mL or ppb.

TABLE 1. SUGGESTED CALIBRATION SOLUTIONS

Analyte	Concentration of the Analyte [ng/mL]						
	#1	#2	#3	#4	#5	#6	#7
As	0.0	0.05	0.1	0.3	0.5	0.7	1.0

TABLE 2. OPERATING PARAMETERS FOR THE INSTRUMENT

Hydride generation atomic fluorescence spectrometer (HGAF)	
<u>Experimental parameters:</u>	
Element	As
Source	BDHCL
Power-current (mA)	Iry - 27.5 Boost - 35
Sample volume (mL)	10-12
Measurement mode	Peak Ht.
Sensitivity	Range - 100 Fine - 10
PMT setting ^a	5.5
<u>Analysis cycle:</u>	
Delay (sec)	10
Rise (sec)	25
Analysis (min)	0.5
Memory (sec)	40
Reductant	1.3% NaBH ₄ in 0.1 M NaOH
Blank	25% HCl in 1% KI and 0.05% asorbic acid
Carrier gas flow (L/min)	0.3 (Ar)

^a Depends on the lamp intensity.

ACS - SAMPLE SUBMISSION FORM

Project: _____ Number of Samples: _____

Date: _____ Analyst: _____

Sample bar codes for all the samples submitted			

Analyte	Standard # or sample #	Concentration	Comments

Figure 1. Batch sample submission form (example).

**NATIONAL HUMAN EXPOSURE ASSESSMENT SURVEY
SAMPLE COLLECTION AND CUSTODY RECORD
RTI/EOHSI CONSORTIUM**

SAMPLE CODE:		PARTICIPANT I.D.:		COUNTY I.D.:					
<table border="1" style="margin: auto; border-collapse: collapse;"> <tr> <td style="padding: 5px;">SAMPLE TYPE:</td> <td style="padding: 5px;">House Dust</td> </tr> <tr> <td style="padding: 5px;">TO BE ANALYZED FOR:</td> <td style="padding: 5px;">As</td> </tr> </table>						SAMPLE TYPE:	House Dust	TO BE ANALYZED FOR:	As
SAMPLE TYPE:	House Dust								
TO BE ANALYZED FOR:	As								
<p>COLLECTION COMMENTS: _____</p> <p>_____</p> <p>_____</p> <p>_____</p>									
<p>PROCESSING COMMENTS: _____</p> <p>_____</p> <p>_____</p> <p>_____</p>									
CUSTODY RECORD									
--CUSTODY OF--				OPERATION PERFORMED					
ORG.	INITIALS	ID	DATE						
RTI				Sample Collected					
RTI				Sample Shipped to RTI					

Research Triangle Institute, Analytical and Chemical Sciences, P.O. Box 12194, RTP, NC 27709

Figure 2. Example of a printed custody record.