CHAIN OF CUSTODY FORM
(Original Stays With Samples)

US EPA Experimental Stream Facility
1003 US Highway 50, Milford, Ohio 45150
Phone # 513-965-4805
FAX # 513-965-4806

Originator: Carol Lepox
Ship Date: 11/26/2018 | Contact:
Destination: Cincinnati | Phone #:
Recipient: Sam Smith | E-Mail:
Phone #:

Collection Date YYYY-MM-DD	Collection Time HH:MM:SS	SiteID	LongID	CrossID	REP	Matrix Type*	TYPE (UKN DUP SPK BLK)	Preservative	Analyses Required
2018 11 19	07:00:00	FL	BTL	NA	1	DI	BLK		TNH4, TRP, TNO2-3, TNO3, TN, TP
20181119	07:00:00	FL	BTL	NA	2	DI	BLK		"
20181119	07:00:00	FL	FLD	NA	1	DI	BLK		"
20181119	07:00:00	FL	FLD	NA	2	DI	BLK		"
20181119	08:47:00	FL	47	0.1	1	SW	UKN		"
20181119	09:00:00	FL	43	0.1	1	SW	UKN		"
20181119	09:48:00	FL	41	0.1	1	SW	UKN		"
20181119	09:48:00	FL	41	0.1	2	SW	UKN		"
20181119	10:10:00	FL	45	0.1	1	SW	UKN		"
20181119	10:34:00	FL	04	0.1	1	SW	UKN		"
20181119	10:53:00	FL	03	0.1	1	SW	UKN		"
20181119	11:10:00	FL	14	0.1	1	SW	UKN		"
20181119	11:20:00	FL	15	0.1	1	SW	UKN		"
20181119	11:55:00	FL	07	0.1	1	SW	UKN		"
20181119	12:17:00	FL	01	0.1	1	SW	UKN		"
20181119	13:01:00	FL	05	0.1	1	SW	UKN		"
20181119	13:13:00	FL	02	0.1	1	SW	UKN		"
20181119	13:13:00	FL	02	0.1	2	SW	UKN		"
20181119	07:00:00	FL	LAB	NA	1	DI	BLK		"
20181119	07:00:00	FL	BTL	NA	1	DI	BLK	ZM HCL	TOC
20181119	07:00:00	FL	BTL	NA	2	DI	BLK	"	"
20181119	07:00:00	FL	FLD	NA	1	DI	BLK	"	"
20181119	07:00:00	FL	FLD	NA	2	DI	BLK	"	"
20181119	08:47:00	FL	47	0.1	1	SW	UKN	"	"
20181119	09:00:00	FL	43	0.1	1	SW	UKN	"	"
20181119	09:48:00	FL	41	0.1	1	SW	UKN	"	"
20181119	09:48:00	FL	41	0.1	2	SW	UKN	"	"
20181119	10:10:00	FL	45	0.1	1	SW	UKN	"	"

Relinquished By	Date	Time	Received By	Date	Time

CHAIN OF CUSTODY FORM
(Original Stays With Samples)

US EPA Experimental Stream Facility
1003 US Highway 50, Milford, Ohio 45150
Phone # 513-965-4805
FAX # 513-965-4806

Originator	
Ship Date	Contact
Destination	Phone #
Recipient	E-Mail
Phone #	

Collection Date YYYY-MM-DD	Collection Time HH:MM:SS	SiteID	LongID	CrossID	REP	Matrix Type*	TYPE (UKN DUP SPK BLK)	Preservative	Analyses Required
20181119	10:34:00	FL	04	0.1	1	SW	UKN	2M HCL	TOC
20181119	10:53:00	FL	03	0.1	1	SW	UKN	"	"
20181119	11:10:00	FL	14	0.1	1	SW	UKN	"	"
20181119	11:20:00	FL	15	0.1	1	SW	UKN	"	"
20181119	11:55:00	FL	07	0.1	1	SW	UKN	"	"
20181119	12:17:00	FL	01	0.1	1	SW	UKN	"	"
20181119	13:01:00	FL	05	0.1	1	SW	UKN	"	"
20181119	13:13:00	FL	02	0.1	1	SW	UKN	"	"
20181119	13:13:00	FL	02	0.1	2	SW	UKN	"	"
20181119	07:00:00	FL	LAB	NA	1	DI	BLK	"	"

Relinquished By	Date	Time	Received By	Date	Time

CHAIN OF CUSTODY FORM

Original Stays With Samples

US EPA Experimental Stream Facility
1003 US Highway 50, Milford, Ohio 45150
Phone # 513-965-4805
FAX # 513-965-4806

Originator:
Ship Date:
Destination:
Recipient:
Phone #:

Contact:
Phone #:
E-Mail:

Collection Date YYYY-MM-DD	Collection Time HH:MM:SS	SiteID	LongID	Cross ID	REP	Matrix Type*	TYPE (UKN, DUP, SPK, BLK)	Analyses Required	*Volume of liquid sample filtered (mL)	
20181119	07:00:00	FL	FLD	NA	1	DI	BLK	chlorophyll	384.72	
20181119	07:00:00	FL	FLD	NA	2	DI	BLK	"	406.61	
20181119	08:47:00	FL	47	0.1	1	SW	UKN	"	219.02	
20181119	09:00:00	FL	43	0.1	1	SW	UKN	"	200.29	
20181119	09:48:00	FL	41	0.1	1	SW	UKN	"	205.48	
20181119	09:48:00	FL	41	0.1	2	SW	UKN	"	191.37	
20181119	10:10:00	FL	45	0.1	1	SW	UKN	"	237.40	
20181119	10:34:00	FL	04	0.1	1	SW	UKN	"	226.08	
20181119	10:53:00	FL	03	0.1	1	SW	UKN	"	187.57	
20181119	11:10:00	FL	14	0.1	1	SW	UKN	"	217.20	
20181119	11:20:00	FL	15	0.1	1	SW	UKN	"	205.31	
20181119	11:55:00	FL	01	0.1	1	SW	UKN	"	204.00	
20181119	12:17:00	FL	01	0.1	1	SW	UKN	"	209.24	
20181119	13:01:00	FL	05	0.1	1	SW	UKN	"	215.52	
20181119	13:13:00	FL	02	0.1	1	SW	UKN	"	215.12	
20181119	13:13:00	FL	02	0.1	2	SW	UKN	"	218.20	

	Relinquished By	Date	Time	Received By			

GAS TRAP DATA SHEET

Lake: FALLS			11/19/18	11/20/18		
Site	**Lat**	**Long**	**Deployment time/date**	**Retrieval time/date**	**Volume recovered (mL)**	**Exetainer codes**
58	36.068722	78.775612	0825	0805	2	
52	36.074672	78.760379	0830	0813	0.5	
56	36.072740	78.766214	0840	0820	4.5	
47	36.068448	78.768527	0847	0825	6.0	
43	36.061445	78.773254	0900	0833	0	
59	36.062722	78.771151	0908	0844	0.5	
63	36.056180	78.770232	0917	0852	21	11/9 2018, FL.63.BT.1
54	36.057873	78.765119	0930	0900	0.5	
55	36.053247	78.761492	0935	0906	0.5	
61	36.053420	78.759431	0940	0912	0.5	
41	36.055640	78.758544	0948	0920	0.5	
49	36.051270	78.756747	0955	0929	0.5	
57	36.048533	78.752382	1004	0937	0.5	
45	36.044128	78.758154	1010	0945	0.5	
10	36.039807	78.756021	1020	0950	0.5	
4	36.012222	78.760996	1034	1007	0.5	
8	36.018958	78.748574	1044	1017	0.5	
3	36.032468	78.736721	1053	1026	1	
14	36.059610	78.717221	1110	1038	0.2	
15	36.022725	78.714674	1120	1050	0.1	
11	36.019282	78.716921	1135	1057	0	
7	36.005237	78.667806	1155	1113	0	
1	36.002138	78.634354	1217	1127	0.5	
21	36.008793	78.628131	1225	1135	0	
13	36.003605	78.646369	1237	1146	0	
12	35.965623	~~78.633851~~	78.633851 1254	1205	0.1	
5	35.970125	78.621254	1301	1214	0.1	
2	35.976262	78.588996	1313	1230	0	
16	35.956708	78.590517	1330	1244	6.0	
24	35.942307	78.588176	1342	1253	1.0	

DISSOLVED GAS AND AIR DATA SHEET

Dissolve Gas PreFix: 11192018.FL.

Lake:	Falls				Date: 11/20/18					
*Site	Dissolved gas Exetainer codes		Water Vol/ ~~Helium~~ Air Vol	Water Temp (degC)	Atmospheric Air Exetainer Codes PreFix: 11182018.FL.					BP (mm Hg)
43	..FL.43.DG.1	...FL.43.DG.2	115/25	10.4						
47	...FL.47.DG.1	..FL.47.DG.2	"	10.9						
45	..FL.45.DG.1	..FL.45.DG.2	"	11.1						
41	...EL.41.DG.1	.:.FL.41.DG.2	"	10.8	...41.AA.01	...41.AA.02	...41.AA.03	...41.AA.04	...41.AA.05	
4	...FL.04.DG.1	...FL.04.DG.2	"	11.2						
3	..FL.03.DG.1	...FL.03.DG.2	"	12.0						
14	...FL.14.DG.1	...FL.14.DG.2	"	12.2						
15	...FL.15.DG.1	...FL.15.DG.2	"	12.5						
7	..FL.07.DG.1	...FL.07.DG.2	"	12.5						
1	...FL.01.DG.1	...FL.01.DG.2	"	12.5						
5	...FL.05.DG.1	...FL.05.DG.2	"	14.0						
2	...FL.02.DG.1	...FL.02.DG.2	"	14.2	...02.AA.01	...02.AA.02	...02.AA.03	...02.AA.04	...02.AA.05	

*Site: all dissolved gas samples collected at 0.1m depth

Boat motor should be off for 2 minutes before sample collection

Goal is to inject 20mL of gas into exetainer

Signature

11/20/18

Date

GREENHOUSE GAS ANALYZER AND GPS DATA SHEET

FILE NAMES	
Lake:	FALLS
Sample Date:	11/20/18
GGA file	339-218-11-20-fa...txt
gps file name	

Time Sync GPS = 12:47:00 LGR = 12:57:13

DEPLOYMENT DATA

Sample site	Chamber start time	Bubbling observed (Y/N)	GPS logger started	Chamber volume (L)	CH_4 and CO_2 profile checked	Water Temp (degC)	
58	8:05:18	✓	✓	45	✓	10.3	
52	8:13:27	✓	✓	45	✓	10.9	
56	8:19:34	✓	✓	40	✓	10.9	
47	8:26:34	✓	✓	40	✓	10.4	CH4 from I-25
43	8:33:50	✓	✓	40	✓	10.4	
59	8:43:53	✓	✓	40	✓	10.4	
63	8:52:28	✓	✓	40	✓	10.3	
54	9:00:10	✓	✓	40	✓	10.4	
55	9:06:10	✓	✓	40	✓	10.4	
61	9:12:12	✓	✓	40	✓	10.6	
41	9:20:00	✓	✓	40	✓	10.8	
49	9:29:15	✓	✓	40	✓	10.8	
57	9:37:28	✓	✓	40	✓	10.8	
45	9:44:09	✓	✓	40	✓	11.1	
10	9:52:39	✓	✓	40	✓	11.1	
4	10:07:11	✓	✓	40	✓	11.2	
8	10:17:05	✓	✓	40	✓	11.2	
3	10:26:17	✓	✓	40	✓	12.0	
14	10:38:41	✓	✓	40	✓	12.2	
15	10:51:29	✓	✓	40	✓	12.5	
11	10:57:12	✓	✓	40	✓	12.5	
7	11:13:19	✓	✓	40	✓	12.5	
1	11:27:49	✓	✓	40	✓	12.5	
21	11:35:17	✓	✓	40	✓	13.1	
13	~~11:46:55~~	11:50:48 ✓	✓	40	✓	13.0	add p.f. to
12	12:05:08	Yes					
12	12:07:13	✓	✓	40	✓	13.5	
5	12:14:25	FAN	TURNED	OFF			
5	12:18:50	✓	✓	40	✓	14.0	
2	12:30:10	✓	✓	40	✓	14.2	
16	12:44:11	✓	✓	40	✓	14.4	
24	12:53:08	✓	✓	40	✓	14.7	

SONDE DATA SHEET

Lake: FALLS LAKE Sample Date: 11/19/18

Site	Site Depth (m)	Sample Depth (m)	Temp (C)	DO (mg/l)	DO (%sat)	sp.Cond (µs cm^{-1})	pH	Chl a (ug/L)	Turbidity (NTU)	Notes
58	4.80	0.1	10.07	8.85	78.6	0.063	7.60	6.0	24.3	
52	4.2	0.1	10.31	8.79	77.9	0.065	7.52	6.6	18.2	
56	5.3	0.1	10.11	9.07	80.6	0.064	7.60	5.8	19.9	
47	5.0	0.1	10.11	9.10	80.8	0.063	7.69	6.7	23.9	
47	5.0	4.5	9.67	8.34	72.8	0.066	7.64	6.5	47.0	
43	4.8	0.1	10.44	9.93	88.5	0.060	7.62	4.3	24.2	
43	4.8	4.3	9.86	8.37	73.7	0.065	7.63	5.5	37.3	
59	4.7	0.1	10.41	9.37	83.5	0.061	7.65	5.4	27.6	
63	5.1	0.1	10.54	9.33	83.2	0.062	7.73	7.2	25.1	
54	4.7	0.1	10.67	9.35	84.1	0.061	7.69	5.8	23.6	
55	4.9	0.1	10.96	9.80	88.5	0.061	7.74	5.8	23.7	
61	5.2	0.1	10.40	9.22	81.8	0.059	7.75	5.8	25.4	
41	3.7	0.1	10.48	9.74	86.0	0.059	7.68	5.1	25.9	
41	3.7	3.2	9.95	9.24	72.9	0.059	7.67	4.6	30.5	
49	5.6	0.1	10.71	9.36	83.6	0.059	7.59	6.8	25.4	
57	4.9	0.1	10.96	9.72	87.0	0.060	7.60	6.0	25.3	
45	5.1	0.1	10.44	9.35	83.1	0.060	7.57	4.5	23.3	
45	5.1	4.6	10.15	8.19	72.9	0.060	7.55	4.9	29.7	
10	4.4	0.1	10.86	8.92	80.7	0.060	7.54	5.4	24.6	
4	4.0	0.1	11.35	9.60	87.9	0.068	7.46	8.5	17.9	
4	4.0	3.5	10.88	7.34	62.4	0.061	7.50	6.5	48.0	
8	6.7	0.1	11.94	~~6.8~~ 9.70	89.3	0.062	7.46	6.3	20.8	
3	4.5	0.1	12.21	10.2	94.7	0.068	7.44	8.5	12.5	
3	4.5	4.0	10.61	8.5	76.2	0.064	7.49	6.2	44.0	
14	4.3	0.1	11.69	9.83	90.7	0.068	7.52	4.7	11.5	
14	4.3	3.8	9.95	8.03	70.4	0.048	7.59	7.1	35.6	
15	7.3	0.1	12.36	9.96	92.6	0.071	7.45	5.7	11.5	
15	7.3	6.8	10.87	8.5	76.6	0.066	7.46	5.2	24.7	

SONDE DATA SHEET

Lake: FALLS Sample Date: 11/19/18

Site	Site Depth (m)	Sample Depth (m)	Temp (C)	DO (mg/l)	DO (%sat)	sp.Cond (µs cm^{-1})	pH	Chl a (ug/L)	Turbidity (NTU)	Notes
11	6.1	0.1	12.46	10.15	94.9	0.071	7.43	10.8	11.8	
7	9.5	0.1	12.75	9.3	87.5	0.077	7.43	9.5	7.4	
7	9.5	1.0	12.18	9.02	83.5	0.071	7.43	9.0	7.8	
7	9.5	2.0	11.96	8.60	79.5	0.077	7.42	7.5	7.9	
7	9.5	3.0	11.92	8.39	77.7	0.077	7.41	7.2	8.1	
7	9.5	4.0	11.81	8.22	75.9	0.077	7.39	5.9	9.2	
7	9.5	5.0	11.76	8.10	74.7	0.076	7.38	5.7	11.2	
7	9.5	6.0	11.75	8.05	74.3	0.076	7.36	6.0	11.5	
7	9.5	7.0	11.74	8.05	74.3	0.076	7.36	6.0	11.8	
7	9.5	8.0	11.73	8.01	73.9	0.076	7.35	6.3	13.4	
7	9.5	9.0	11.72	7.95	73.3	0.076	7.34	11.8	47.3	
1	5.5	0.1	13.68	9.01	86.6	0.077	7.44	5.5	9.0	
1	5.5	5.0	12.47	8.27	77.1	0.076	7.42	7.1	13.5	
21	6.7	0.1	13.48	9.0	86.1	0.076	7.38	6.8	13.4	
13	8.4	0.1	13.95	8.9	86.1	0.078	7.39	5.3	8.1	
12	7.4	0.1	14.01	9.34	90.0	0.078	7.42	9.9	6.8	
5	8.1	0.1	14.60	9.31	91.4	0.078	7.39	5.6	6.0	
5	8.1	7.6	12.94	6.87	64.9	0.077	7.38	6.1	15.8	
2	6.4	0.1	14.36	10.48	102.5	0.077	7.44	13.8	6.7	
2	6.4	5.9	12.65	8.45	78.3	0.077	7.41	7.0	27.2	
16	9.4	0.1	15.23	8.18	81.4	0.078	7.41	4.7	5.6	
24	4.0	0.1	14.61	8.68	84.2	0.079	7.41	6.2	5.4	

Date 11/19/18

Cal cylinder serial # CC163816

Dilution system serial # EPA A20808

LGR file 559-2018-11-19_f0000.txt

Sample MFC gas correction factor 1.43

Zero Air Flow (Lpm)	Cal Gas Flow Readout (sccm)	CH4 cal (ppm)	CO2 cal (ppm)	CH4 resp (ppm)	CO2 resp (ppm)	Start Time	Stop Time
3.25	0.0	0.0	0.0	0.008	0.34	16:55:00	16:58:30
7	70.0	28.3	283.2				
6.5	70.0	30.5	304.7	27.6	281.9	16:58:30	17:03:36
6	70.0	33.0	329.8				
5.5	70.0	35.9	359.4	32.4	337.8	17:03:36	17:06:11
5	70.0	39.5	394.9				
4.5	70.0	43.8	438.1	39.20	412.6	17:06:11	17:08:26
4	70.0	49.2	491.9				
3.5	70.0	56.1	560.8	49.89	532.0	17:08:26	17:11:00
3	70.0	65.2	652.1				

Signature _____

Date 4/19/18

Cal cylinder serial # CC 498857

Dilution system serial # EPA A20808

LGR file SSA - 2018 - 10 - 23 - f0000.txt

Zero Air Flow (Lpm)	Cal Gas Flow Readout (sccm)	CH4 cal (ppm)	CO2 cal (ppm)	CH4 resp (ppm)	CO2 resp (ppm)	Start Time	Stop Time
3.25	0.0	0.0	0.0	0.014	0.50	17:12:25	17:16:32
9	60.0	2.0	20.2				
8.5	60.0	2.1	21.4	1.95	20.71	17:16:32	17:19:44
8	60.0	2.2	22.7				
7.5	60.0	2.4	24.2	2.19	23.1	17:19:44	17:22:22
7	60.0	2.6	25.9				
6.5	60.0	2.7	27.9	2.51	26.4	17:22:22	17:24:55
6	60.0	3.0	30.2				
5.5	60.0	3.2	32.9	2.96	31.2	17:24:55	17:27:36
5	60.0	3.6	36.1				

Signature _____

Date 11/20/19

Cal cylinder serial # CC14 3816

Dilution system serial # EPA A20808

LGR file 339-2019-11-20-f0001.txt


Sample MFC gas correction factor 1.43

Zero Air Flow (Lpm)	Cal Gas Flow Readout (sccm)	CH4 cal (ppm)	CO2 cal (ppm)	CH4 resp (ppm)	CO2 resp (ppm)	Start Time	Stop Time
3.25	0.0	0.0	0.0	0.007	0.11	16:26:00	16:31:10
7	70.0	28.3	283.2				
6.5	70.0	30.5	304.7	27.6	286.6	16:31:00	16:44:40
6	70.0	33.0	329.8				
5.5	70.0	35.9	359.4	32.3	339.3	16:44:40	16:47:55
5	70.0	39.5	394.9				
4.5	70.0	43.8	438.1	39.1	414.3	16:47:55	16:51:30
4	70.0	49.2	491.9				
3.5	70.0	56.1	560.8	49.7	531.8	16:51:30	16:54:50
3	70.0	65.2	652.1				

Signature _____

Date 11/20/18
Cal cylinder serial # CC498857
Dilution system serial # EPA A20808
LGR file 999-2018-11-20_f0001.txt

Zero Air Flow (Lpm)	Cal Gas Flow Readout (sccm)	CH4 cal (ppm)	CO2 cal (ppm)	CH4 resp (ppm)	CO2 resp (ppm)	Start Time	Stop Time
3.25	0.0	0.0	0.0	0.001	0.25	16:10:00	16:13:35
9	60.0	2.0	20.2				
8.5	60.0	2.1	21.4	1.890	19.50	16:13:35	16:17:18
8	60.0	2.2	22.7				
7.5	60.0	2.4	24.2	2.150	22.5	16:17:18	16:19:15
7	60.0	2.6	25.9				
6.5	60.0	2.7	27.9	2.48	25.78	16:19:15	16:21:52
6	60.0	3.0	30.2				
5.5	60.0	3.2	32.9	2.94	30.61	16:21:52	16:25:27
5	60.0	3.6	36.1				

Signature 

11/19/2018 Site List

	# Samples collected			
Site#	Bubble Trap	Dissolved Gas	Water	Air
~~53~~ 58	3			
~~42~~ 52	3	~~2~~	~~1~~	
~~50~~ 56	3			
~~44~~ 47	3	2 ~~2~~	~~1~~ 1	
~~52~~ 43	3	2	1	
~~56~~ 59	3			
~~47~~ 63	3			
~~43~~ 54	3	~~2~~ ~~2~~	~~1~~	
~~54~~ 55	3			
~~55~~ 61	3			
41	3	2	2	5
49	3			
57	3			
45	3	2	1	
10	3			
4	3	2	1	
8	3			
3	3	2	1	
14	3	2	1	
15	3	2	1	
11	3			
7	3	2	1	
1	3	2	1	
21	3			
13	3			
12	3			
5	3	2	1	
2	3	2	2	2
16	3			
24	3			

ATTACHMENT A
YSI 6600 SONDE CALIBRATION LOG SHEET

SONDE ID: STD 16 **DATE:** 11/19/18 **TECHNICIAN:** JTW

DATE / TIME OK? (YES) / NO

BATTERY VOLTS: 12.4 (Change if < 10.0 v) **CHANGED BATTERIES?** YES / (NO)

TURBIDITY WIPER: PARKS 180° FROM OPTICS? (YES) / NO **CHANGED?** YES / (NO)

CHLOROPHYLL WIPER: PARKS 180° FROM OPTICS? (YES) / NO **CHANGED?** YES / (NO)

DISSOLVED OXYGEN: "DO Sat %" **BAROMETRIC PRESSURE:** 755.7 mm Hg

SENSOR OUTPUT TEST: PASS / FAIL (For Rapid-Pulse Probes, change membrane if failed)

Record first 10 readings from Discrete Sampling:

DO MEMBRANE CHANGED? YES / NO **PROBE RE-SURFACED?** YES / NO **DATE/TIME:** _____

[~97] **INITIAL:** 99.4 **CALIBRATED:** 99.8 **POST-DEPLOYMENT:** 97.1
DO CHARGE [50 ± 25]: _____ **DO GAIN** [1.0 ± 0.35]: 1.04614

DEPTH: "Pressure" [~~m~~ ft] **PRESSURE OFFSET** [VENTED = 0 ± 6 / NON-VENTED = -14.7 ± 6]: -0.0684
[0] **INITIAL:** -0.144 **CALIBRATED:** 0.0 **POST-DEPLOYMENT:** -0.150

CONDUCTIVITY: "SpCond" [mS/cm] **CONDUCTIVITY CELL CONSTANT** [5.0 ± 0.45]: 5.19[illegible]014

1.413 [~~10~~] **INITIAL:** 1.431 **CALIBRATED:** 1.413 **POST-DEPLOYMENT:** 1.406 → T = 22.1 C

TURBIDITY: "Turb+NTU" [NTU]

[0] **INITIAL:** 0.3 **CALIBRATED:** 0.0 **POST-DEPLOYMENT:** 0.1

126 [~~1000~~] **INITIAL:** 123.9 **CALIBRATED:** 126.0 **POST-DEPLOYMENT:** 124.7

CHLOROPHYLL: "Chlor ug/L" [ug/L]

[0] **INITIAL:** 0.53 **CALIBRATED:** _____ **POST-DEPLOYMENT:** 0.0

[113] **INITIAL:** _____ **CALIBRATED:** _____ **POST-DEPLOYMENT:** _____

---

pH: "pH"

[7] **INITIAL:** 7.1 **CALIBRATED:** 7.0 **POST-DEPLOYMENT:** 7.046 **mV** [0 ± 40]: -31.8 pre

[10] **INITIAL:** 9.96 **CALIBRATED:** 10.0 **POST-DEPLOYMENT:** 9.98 **mV** [-177 ± 40]: -200.3 pre

---

ORP: "ORP mV" **ORP mV OFFSET [0 ± 100]:** _____

Printed copies of this document are uncontrolled. All users are responsible for confirming version status against the electronic version in the document control system.