CHAIN OF CUSTODY FORM
(Original Stays With Samples)

US EPA Experimental Stream Facility
1003 US Highway 50, Milford, Ohio 45150
Phone # 513-965-4805
FAX # 513-965-4806

Originator: CARL LENOX
Ship Date: 09272018 Contact:
Destination: Cincinnati Phone #:
Recipient: Sam Smith E-Mail:
Phone #:

Collection Date YYYY-MM-DD	Collection Time HH:MM:SS	SiteID	LongID	CrossID	REP	Matrix Type*	TYPE (UKN, DUP, SPK, BLK)	Preservative	Analyses Required
20180925	07:00:00	FL	BTL	NA	1	DI	BLK		TNH4, TRP, TNO2-3, TNO2, TN, TP
20180925	07:00:00	FL	BTL	NA	2	DI	BLK		"
20180925	07:00:00	FL	FLD	NA	1	DI	BLK		"
20180925	07:00:00	FL	FLD	NA	2	DI	BLK		"
20180925	07:42:00	FL	42	0.1	1	SW	UKN		"
20180925	08:03:00	FL	44	0.1	1	SW	UKN		"
20180925	08:39:00	FL	43	0.1	1	SW	UKN		"
20180925	09:02:00	FL	41	0.1	1	SW	UKN		"
20180925	09:02:00	FL	41	0.1	2	SW	UKN		"
20180925	09:49:00	FL	04	0.1	1	SW	UKN		"
20180925	10:13:00	FL	03	0.1	1	SW	UKN		"
20180925	10:27:00	FL	14	0.1	1	SW	UKN		"
20180925	10:41:00	FL	15	0.1	1	SW	UKN		"
20180925	11:04:00	FL	07	0.1	1	SW	UKN		"
20180925	11:27:00	FL	01	0.1	1	SW	UKN		"
20180925	12:17:00	FL	05	0.1	1	SW	UKN		"
20180925	12:30:00	FL	02	0.1	1	SW	UKN		"
20180925	12:30:00	FL	02	0.1	2	SW	UKN		"
20180925	07:00:00	FL	LAB	NA	1	DI	BLK		"
20180925	07:00:00	FL	BTL	NA	1	DI	BLK	2NHCL	TOC
20180925	07:00:00	FL	BTL	NA	2	DI	BLK	"	"
20180925	07:00:00	FL	FLD	NA	1	DI	BLK	"	"
20180925	07:00:00	FL	FLD	NA	2	DI	BLK	"	"
20180925	07:42:00	FL	42	0.1	1	SW	UKN	"	"
20180925	08:03:00	FL	44	0.1	1	SW	UKN	"	"
20180925	08:39:00	FL	43	0.1	1	SW	UKN	"	"
20180925	09:02:00	FL	41	0.1	1	SW	UKN	"	"
20180925	09:02:00	FL	41	0.1	2	SW	UKN	"	"

Relinquished By	Date:	Time:	Received By	Date:	Time:

CHAIN OF CUSTODY FORM
(Original Stays With Samples)

US EPA Experimental Stream Facility
1003 US Highway 50, Milford, Ohio 45150
Phone # 513-965-4805
FAX # 513-965-4806

Originator: CAROLENOX
Ship Date: 9/27/2018 Contact:
Destination: Cincinnati Phone #:
Recipient: Sam Smith E-Mail:
Phone #:

Collection Date YYYY-MM-DD	Collection Time HH:MM:SS	SiteID	LongID	CrossID	REP	Matrix Type*	TYPE (UKN, DUP, SPK, BLK)	Preservative	Analyses Required
20180925	09:49:00	FL	04	0.1	1	SW	UKN	2 MHCL	TOC
20180925	10:13:00	FL	03	0.1	1	SW	UKN	"	"
20180925	10:27:00	FL	14	0.1	1	SW	UKN	"	"
20180925	10:41:00	FL	15	0.1	1	SW	UKN	"	"
20180925	11:04:00	FL	07	0.1	1	SW	UKN	"	"
20180925	11:27:00	FL	01	0.1	1	SW	UKN	"	"
20180925	12:17:00	FL	05	0.1	1	SW	UKN	"	"
20180925	12:30:00	FL	02	0.1	1	SW	UKN	"	"
20180925	12:30:00	FL	02	0.1	2	SW	UKN	"	"
20180925	07:00:00	FL	LAB	NA	1	DI	BLK	"	"

Relinquished By	Date:	Time:	Received By	Date:	Time:

CHAIN OF CUSTODY FORM
(Original Stays With Samples)

US EPA Experimental Stream Facility
1003 US Highway 50, Milford, Ohio 45150
Phone # 513-965-4805
FAX # 513-965-4806

Originator:	Carol Lenox
Ship Date:	09272018
Destination:	Cincinnati
Recipient:	Sam Smith
Phone #:	
Contact:	
Phone #:	
E-Mail:	

Collection Date YYYY-MM-DD	Collection Time HH:MM:SS	SiteID	LongID	Cross ID	REP	Matrix Type*	TYPE (UKN, DUP, SPK, BLK)	Analyses Required	*Volume of liquid sample filtered (mL)	
20180925	07:00:00	FL	FLD	NA	1	DI	BLK	chlorophyll	392.86	
20180925	07:00:00	FL	FLD	NA	2	DI	BLK	"	368.43	
20180925	07:42:00	FL	42	0.1	1	SW	UKN	"	153.13	
20180925	08:03:00	FL	44	0.1	1	SW	UKN	"	156.46	
20180925	08:39:00	FL	43	0.1	1	SW	UKN	"	193.41	
20180925	09:02:00	FL	41	0.1	1	SW	UKN	"	184.54	
20180925	09:02:00	FL	41	0.1	2	SW	UKN	"	197.56	
20180925	09:49:00	FL	04	0.1	1	SW	UKN	"	168.37	
20180925	10:13:00	FL	03	0.1	1	SW	UKN	"	224.33	
20180925	10:27:00	FL	14	0.1	1	SW	UKN	"	214.09	
20180925	10:41:00	FL	15	0.1	1	SW	UKN	"	229.68	
20180925	11:04:00	FL	07	0.1	1	SW	UKN	"	222.55	
20180925	11:27:00	FL	01	0.1	1	SW	UKN	"	270.28	
20180925	12:17:00	FL	05	0.1	1	SW	UKN	"	249.78	
20180925	12:30:00	FL	02	0.1	1	SW	UKN	"	269.70	
20180925	12:30:00	FL	02	0.1	2	SW	UKN	"	275.97	

Relinquished By:	Date:	Time:	Received By:

GAS TRAP DATA SHEET

Lake: Falls Lake, NC			9/25/18	9/26/18		
Site	Lat	Long	Deployment time/date	Retrieval time/date	Volume recovered (mL)	Exetainer codes
53	36.069168	78.794424	0734	0757	6.5	
42	36.065205	78.795169	0742	0810	25.5	09252018.FL.42.BT.1
50	36.068100	78.786034	0754	0815	5	
44	36.072870	78.779331	0803	0827	0.5	
52	36.079713	78.760346	0816	0845	3	
56	36.072795	78.766252	0822	0854	19	09252018.FL.56.BT.1
47	36.068320	78.768639	0828	0905	7.5	
43	36.061375	78.773251	0838	0912	25, 6	09252018.FL.43.BT.1
54	36.057750	78.765169	0849	0921	2.5	
55	36.053133	78.761574	0855	0930	6	
41	36.055603	78.758479	0902	0937	7	
49	36.051462	78.756742	0911	0951	4	
57	36.048498	78.752314	0920	0959	20	09252018.FL.57.BT.1
45	36.043992	78.758199	0930	1009	8	
10	36.039647	78.756111	0938	1016	0.25	
4	36.012102	78.761086	0949	1033	25, 25, 25, 50	09252018.FL.04.BT.1/BT.2/BT.3
8	36.019005	78.748667	1000	1045	13.5	
3	36.032503	78.736871	1013	1056	8	
14	36.059452	78.717371	1027	1115	0.5	
15	36.022633	78.714627	1041	1133	6	
11	36.019168	78.717107	1051	1143	0	
7	36.005340	78.667687	1104	1158	0	
1	36.002060	78.634451	1127	1218	1	
21	36.008827	78.628179	1139	1231	5.5	
13	36.003590	78.646852	1149	1244	0.5	
12	35.965458	78.633861	1209	1301	0.5	
5	35.970072	78.621241	1217	1315	9	
2	35.976208	78.588941	1230	1327	9	
16	35.956780	78.590269	1242	1346	0.5	
24	35.941970	78.588152	1249	1358	0.5	

DISSOLVED GAS AND AIR DATA SHEET

Prefix: 09252018.FL.

Prefix for AA samples: 09252018.FL.

Lake:	Falls Lake, NC				Date:	09/26/18				
*Site	Dissolved gas Exetainer (DG) codes		Water Vol/ Helium Vol	Water Temp (degC)	Atmospheric Air Exetainer Codes (AA)					BP (mm Hg)
42	09252018.FL.42.DG.2	42.DG.1	114/25	25	09252018.FL.42.AA.1	42.AA.2	42.AA.3	42.AA.4	42.AA.5	
44	44.DG.1	44.DG.2		26	44.AA.1	44.AA.2	44.AA.3	44.AA.4	44.AA.5	
43	43.DG.1	43.DG.2		29	43.AA.1	43.AA.2	43.AA.3	43.AA.4	43.AA.5	
41	41.DG.1	41.DG.2		27	41.AA.1	41.AA.2	41.AA.3	41.AA.4	41.AA.5	
4	04.DG.1	04.DG.2		28	04.AA.1	04.AA.2	04.AA.3	04.AA.4	04.AA.5	
3	03.DG.1	03.DG.2		28	03.AA.1	03.AA.2	03.AA.3	03.AA.4	03.AA.5	
14	14.DG.1	14.DG.2		27	14.AA.1	14.AA.2	14.AA.3	14.AA.4	14.AA.5	
15	15.DG.1	15.DG.2		28	15.AA.1	15.AA.2	15.AA.3	15.AA.4	15.AA.5	
7	07.DG.1	07.DG.2		28	07.AA.1	07.AA.2	07.AA.3	07.AA.4	07.AA.5	
1	01.DG.1	01.DG.2		29	01.AA.1	01.AA.2	01.AA.3	01.AA.4	01.AA.5	
5	05.DG.1	05.DG.2		29	05.AA.1	05.AA.2	05.AA.3	05.AA.4	05.AA.5	
2	02.DG.1	02.DG.2		29	02.AA.1	02.AA.2	02.AA.3	02.AA.4	02.AA.5	

*Site: all dissolved gas samples collected at 0.1m depth

Boat motor should be off for 2 minutes before sample collection

Goal is to inject 20mL of gas into exetainer

Ambient Air Samples only collected @ Two Sites: 41 and 2

Signature: Ryan Daly

09/26/18

Date

GREENHOUSE GAS ANALYZER AND GPS DATA SHEET

FILE NAMES	
Lake:	FALLS
Sample Date:	9/26/18
GGA file	959-2018-09-26-f0000.txt
gps file name	

DEPLOYMENT DATA Clocks LGR = 14:25:20 GPS = 13:16:00

Sample site	Chamber start time	Bubbling observed (Y/N)	GPS logger started	Chamber volume (L)	CH_4 and CO_2 profile checked	Water Temp (degC)
53	9:07:10	N	✓	45	✓	24.0
42	9:17:34	N	✓	45	✓	25.0
50	9:25:57	N	✓	45	✓	25.0
44	9:36:55	N	✓	45	✓	25.0
52	9:54:17	Y	✓	45	✓	25.0
52	9:56:13	N	✓	45	✓	25.0
56	10:03:16	N	✓	50	✓	25.0
47	10:14:44	N	✓	50	✓	27.0
43	10:21:58	N	✓	50	✓	26.0
54	10:30:35	N	✓	50	✓	26.0
55	10:38:56	N	✓	45	✓	26.0
41	10:45:59	N	✓	45	✓	26.0
49	10:54:49	N	✓	50	✓	26.2
57	11:08:42	N	✓	50	✓	26.3
45	11:18:23	N	✓	50	✓	26.3
10	11:25:28	N	✓	45	✓	26.2
4	11:42:00	N	✓	45	✓	26.6
8	11:53:51	N	✓	45	✓	26.4
3	12:05:50	N	✓	45	✓	26.5
14	12:24:17	N	✓	45	✓	25.8
15	12:44:00	N	✓	45	✓	26.7
11	12:52:16	N	✓	45	✓	26.8
7	13:07:48	N	✓	45	✓	27.8
1	13:27:10	N	✓	45	✓	27.8
21	13:40:10	N	✓	45	✓	27.8
13	13:53:23	N	✓	45	✓	27.7
12	14:10:52	Y	✓	45	✓	28.3
12	14:12:27	N	✓	45	✓	28.3
5	14:23:57	N	✓	45	✓	28.6
2	14:36:30	N	✓	45	✓	28.1
16	14:55:05	N	✓	45	✓	28.2
24	15:06:55	Y	✓	45	✓	28.2
24	15:12:40	N	✓	45	✓	28.2

SONDE DATA SHEET

After Hurricane Florence

Lake: FALLS Sample Date: 9/25/18

Site	Site Depth (m)	Sample Depth (m)	Temp (C)	DO (mg/l)	DO (%sat)	sp.Cond (µs cm^{-1})	pH	Chl a (ug/L)	Turbidity (NTU)	Notes
53	~~3.5~~ 3.5	0.1	24.4	6.14	73.4	0.073	7.9	11.6	21.7	
43	3.5	0.1	24.3	6.23	74.3	0.068	8.0	15.1	21.9	
42	3.5	3.0	24.4	5.67	67.7	0.069	8.0	14.2	28.4	
50	3.2	0.1	24.3	5.53	65.7	0.065	~~8.25~~ 8.15	16.9	21.3	
441	3.5	0.1	24.5	6.51	77.4	0.066	8.04	16.3	15.7	
44	3.5	3.0	24.1	3.56	42.3	0.061	7.88	7.5	36.5	
52	3.6	0.1	24.1	2.40	27.1	0.069	7.81	7.0	19.0	
56	4.6	0.1	24.3	4.36	51.9	0.067	7.69	10.0	18.5	
47	4.4	0.1	24.5	5.04	60.0	0.068	7.60	12.8	16.7	
43	4.3	0.1	24.6	5.67	67.6	0.067	7.541	12.8	18.2	
43	~~3.8~~ 4.3	3.8	24.5	4.88	58.5	0.067	7.418	9.7	28.4	
54	4.1	0.1	24.6	6.44	77.1	0.068	7.45	16.6	16.4	
55	4.4	0.1	24.7	6.91	83.0	0.068	7.43	16.8	13.4	
41	3.2	0.1	24.7	7.0	84.1	0.068	7.45	15.0	12.5	
41	3.2	2.7	24.5	5.62	67.3	0.068	7.410	8.1	28.8	
49	4.9	0.1	24.8	7.15	86.1	0.069	7.41	17.9	12.0	
57	4.3	0.1	24.7	5.65	67.1	0.069	7.43	14.7	16.0	
45	4.6	0.1	24.7	6.23	75.1	0.068	7.36	23.2	16.2	
10	3.5	0.1	24.8	5.98	71.7	0.069	7.36	13.3	15.1	
4	3.2	0.1	25.4	7.38	89.7	0.078	7.35	15.7	14.3	
4	3.2	2.7	25.0	~~7.63~~ 6.50	78.9	0.078	7.33	12.3	23.0	
8	6.0	0.1	25.3	7.82	93.1	0.083	7.40	14.2	10.5	
3	3.9	0.1	25.2	6.00	72.0	0.088	7.36	12.0	8.5	
3	3.9	3.4	24.7	3.64	43.6	0.089	7.31	7.3	14.6	
14	3.7	0.1	25.4	4.73	56.6	0.080	7.42	9.1	11.5	
14	3.7	3.2	24.6	3.2	38.4	0.081	7.30	6.7	32.0	
15	6.6	0.1	25.9	7.93	110.0	0.086	7.28	20.8	8.1	
15	6.6	6.1	24.7	2.95	36.1	0.084	7.07	11.4	46	

SONDE DATA SHEET

Lake: FALLS Sample Date: 9/25 18

Site	Site Depth (m)	Sample Depth (m)	Temp (C)	DO (mg/l)	DO (%sat)	sp.Cond (µs cm^{-1})	pH	Chl a (ug/L)	Turbidity (NTU)	Notes
11	5.6	0.1	26.0	9.35	115.3	0.087	7.23	25.4	8.8	
7	8.0	0.1	26.0	9.49	117.3	0.103	7.58	17.4	4.5	
7	1	1.0	25.9	8.34	101.4	0.103	7.61	8.7	4.1	
7		2.0	25.8	7.26	88.5	0.103	7.59	7.2	4.0	
7		3.0	25.6	2.64	31.7	0.101	7.56	4.2	6.3	
7		4.0	25.3	2.67	33.0	0.101	7.44	5.0	9.7	
7		5.0	25.2	2.58	30.6	0.092	7.39	4.0	11.3	
7	1	6.0	25.0	1.55	18.4	0.088	7.35	4.4	14.1	
7		7.0	24.9	1.44	17.3	0.086	7.30	4.6	15.3	
7	8.0	7.5	24.9	1.29	15.5	0.086	~~7.17~~	~~5.0~~	26.5	
1	4.9	0.1	26.1	7.97	98.0	0101	7.22	19.0	4.0	
1	4.9	4.4	25.6	2.17	26.5	0.103	7.22	6.7	23.0	
21	6.9	0.1	26.8	6.14	76.1	0.099	7.25	11.0	3.4	
13	7.0	0.1	26.5	8.041	100.1	0.102	7.32	11.5	3.1	
12	6.1	0.1	26.9	7.03	87.8	0.100	7.41	13.5	2.8	
5	5.1	0.1	27.3	5.65	70.8	0.101	7.45	5.7	1.7	
5	~~4.6~~ 5.1	4.6	26.1	3.82	46.6	0.101	7.38	4.4	5.5	
2	5.8	0.1	26.9	5.76	71.5	0.100	7.36	10.0	3.1	
2	5.8	5.3	25.8	2.42	29.7	0.102	7.28	3.8	11.7	
16	7.3	0.1	27.3	~~5.30~~ 5.30	65.9	0.100	7.29	12.0	2.4	
24	4.3	0.1	27.54	6.91	87.1	0.097	7.34	7.0	1.6	

Date 9/20/18

Cal cylinder serial # CC163816

Dilution system serial # EPA A20808

LGR file 99-2018_09-20_f0000.txt

Sample MFC gas correction factor 1.43

RE

Zero Air Flow (Lpm)	Cal Gas Flow Readout (sccm)	CH4 cal (ppm)	CO2 cal (ppm)	CH4 resp (ppm)	CO2 resp (ppm)	Start Time	Stop Time
3.25	0.0	0.0	0.0	0.054	0.55	06:20:18	06:23:02
7	70.0	28.3	283.2				
6.5	70.0	30.5	304.7	24.1	248.5	06:27:13	06:31:08
6	70.0	33.0	329.8				
5.5	70.0	35.9	359.4	28.2	293.4	06:31:08	06:33:57
5	70.0	39.5	394.9				
4.5	70.0	43.8	438.1	34.2	358.8	06:33:57	06:37:10
4	70.0	49.2	491.9				
3.5	70.0	56.1	560.8	43.42	461.15	06:37:10	06:39:45
3	70.0	65.2	652.1				

Signature _____

note CAL GAS WAS Set to 60 sccm for all calibration points.

Need to re-calc dilution Concentrations

Date 9/26/18

Cal cylinder serial # ~~CC163816~~ CC498857

Dilution system serial # EPA A20808


LGR file 359-2018_09-26_focod.txt

	Zero Air Flow (Lpm)	Cal Gas Flow Readout (sccm)	CH4 cal (ppm)	CO2 cal (ppm)	CH4 resp (ppm)	CO2 resp (ppm)	Start Time	Stop Time
✓	3.25	0.0	0.0	0.0	~~0.054~~	0.55	~~06:22:18~~	~~06:25:02~~
	9	60.0	2.0	20.2	0.044			
✓	8.5	60.0	2.1	21.4	~~1.985~~	~~21.9~~	~~06:23:40~~	—
	8	60.0	2.2	22.7	1.94	20.7		
✓	7.5	60.0	2.4	24.2	2.19	23.1	06:46:51	06:49:16
	7	60.0	2.6	25.9				
✓	6.5	60.0	2.7	27.9	2.52	26.4	06:49:16	06:51:29
	6	60.0	3.0	30.2				
✓	5.5	60.0	3.2	32.9	2.98	31.0	06:51:29	06:53:36
	5	60.0	3.6	36.1				

START STOP

06:41:00 06:43:20

06:43:20 06:46:51

Signature 

Date 10/3/18
Cal cylinder serial # CC498857
Dilution system serial # EPA A20508
LGR file 994-2018-10-03_f0000.txt

Zero Air Flow (Lpm)	Cal Gas Flow Readout (sccm)	CH4 cal (ppm)	CO2 cal (ppm)	CH4 resp (ppm)	CO2 resp (ppm)	Start Time	Stop Time
3.25	0.0	0.0	0.0	2.006	0.34	15:54:20	15:57:20
9	60.0	2.0	20.2				
8.5	60.0	2.1	21.4	1.89	19.3	15:57:20	16:04:15
8	60.0	2.2	22.7				
7.5	60.0	2.4	24.2	2.15	22.0	16:04:15	16:11:00
7	60.0	2.6	25.9				
6.5	60.0	2.7	27.9	2.48	25.6	16:11:00	16:14:03
6	60.0	3.0	30.2				
5.5	60.0	3.2	32.9	2.93	30.3	16:14:03	16:17:40
5	60.0	3.6	36.1				

Signature _____

Date 10/31/18

POST

Cal cylinder serial # CC163816

Dilution system serial # EPA A26808

LGR file 999-2018-10-03-fuuuu.txt

Sample MFC gas correction factor 1.43

	Zero Air Flow (Lpm)	Cal Gas Flow Readout (sccm)	CH4 cal (ppm)	CO2 cal (ppm)	CH4 resp (ppm)	CO2 resp (ppm)	Start Time	Stop Time
✓	3.25	0.0	0.0	0.0	0.016	0.29	16:18:40	16:23:04
	7	70.0	28.3	283.2				
✓	6.5	70.0	30.5	304.7	27.49	285.0	16:23:04	16:28:21
	6	70.0	33.0	329.8				
✓	5.5	70.0	35.9	359.4	32.25	337.6	16:28:21	16:30:59
	5	70.0	39.5	394.9				
✓	4.5	70.0	43.8	438.1	38.99	412.0	16:30:59	16:33:28
	4	70.0	49.2	491.9				
✓	3.5	70.0	56.1	560.8	49.49	529.8	16:33:28	16:36:35
	3	70.0	65.2	652.1				

Signature _____

ATTACHMENT A
YSI 6600 SONDE CALIBRATION LOG SHEET

SONDE ID: STD 16 **DATE:** 9/25/18 **TECHNICIAN:** JTW

DATE / TIME OK? YES / NO

BATTERY VOLTS: _____ (Change if < 10.0 V) **CHANGED BATTERIES?** YES / NO

TURBIDITY WIPER: PARKS 180° FROM OPTICS? YES / NO **CHANGED?** YES / NO

CHLOROPHYLL WIPER: PARKS 180° FROM OPTICS? YES / NO **CHANGED?** YES / NO

DISSOLVED OXYGEN: "DO Sat %" **BAROMETRIC PRESSURE:** 756.8 mm Hg

SENSOR OUTPUT TEST: PASS / FAIL (For Rapid-Pulse Probes, change membrane if failed)

Record first 10 readings from Discrete Sampling:

_____ _____ _____ _____ _____

_____ _____ _____ _____ _____

DO MEMBRANE CHANGED? YES / NO **PROBE RE-SURFACED?** YES / NO **DATE/TIME:** _____

[~97] **INITIAL:** 99.6 **CALIBRATED:** 99.7 **POST-DEPLOYMENT:** 99.6

DO CHARGE [50 ± 25]: _____ **DO GAIN** [1.0 ± 0.35]: 1.04894

DEPTH: "Pressure" ~~[m]~~ ft **PRESSURE OFFSET** [VENTED = 0 ± 6 / NON-VENTED = -14.7 ± 6]: -0.067

[0] **INITIAL:** -0.146 **CALIBRATED:** 0.001 **POST-DEPLOYMENT:** -0.352

CONDUCTIVITY: "SpCond" [mS/cm] **CONDUCTIVITY CELL CONSTANT** [5.0 ± 0.45]: 5.0887

1.413 ~~[10]~~ **INITIAL:** 1.421 **CALIBRATED:** 1.413 **POST-DEPLOYMENT:** 1.408

Pre @ 22.36 °C
Post @ 22.22 °C

TURBIDITY: "Turb+NTU" [NTU]

[0] **INITIAL:** 0.2 **CALIBRATED:** 0.0 **POST-DEPLOYMENT:** -0.2

126 ~~[1000]~~ **INITIAL:** 128.1 **CALIBRATED:** 125.9 **POST-DEPLOYMENT:** 124.3

CHLOROPHYLL: "Chlor ug/L" [ug/L]

[0] **INITIAL:** 0.1 **CALIBRATED:** _____ **POST-DEPLOYMENT:** -0.2

5.7c 2 ~~[113]~~ **INITIAL:** _____ **CALIBRATED:** _____ **POST-DEPLOYMENT:** 8.6 T= 7.4°C

5.7c 41 11.6 T= 7.3 °C

PH: "pH"

[7] **INITIAL:** 6.81 **CALIBRATED:** 7.0 **POST-DEPLOYMENT:** 7.08 **MV** [0 ± 40]: -32.7 @ 7.0

[10] **INITIAL:** 9.95 **CALIBRATED:** 10.0 **POST-DEPLOYMENT:** 9.96 **MV** [-177 ± 40]: -203.4 @ 10.0

ORP: "ORP mV" **ORP MV OFFSET [0 ± 100]:** _____