

Medical Device Design: Design Control Methodologies

A Critical Analysis of the Design Control Process for a Medical Device Translation Guidance
Resource

Author: Ugonna Faithful Ogini
Medical Engineering

Contents

Abstract..... 2

1. Introduction..... 3

2. Methods: Design Control Methodology..... 3

 2.1 Design Study Framework..... 3

 2.2 Rationale for the Convergent Design Model.....4

 2.3 Design Control, Project Planning, and Regulatory Alignment.....5

3. Results and Discussion.....5

 3.1 Development of the Statement of Need.....5

 3.2 Design Specification (XDS).....6

 3.3 Concept Generation and Selection Methodology.....6

4. Conclusion and Future Work.....7

References..... 8

Appendices..... 9

Abstract

Medical device development requires the structured application of design control methodologies to ensure that user needs, regulatory expectations, and technical feasibility are aligned from project inception. This paper presents an individual critical analysis of a design control study for defining a medical device translation guidance resource that provides a pathway from concept through design, regulatory bodies, end-users, and disposal, undertaken within a multidisciplinary postgraduate cohort. The analysis focuses on the application of a convergent design model, the development of a robust Statement of Need informed by continuous end-user engagement, and the translation of those needs into a comprehensive design specification (XDS). From the full XDS, a subset of essential specifications was selected to enable systematic concept evaluation and final idea selection. International design control frameworks, including ISO 13485, FDA 21 CFR 820 principles, and EU/UK Medical Device Regulations, were applied as guiding constraints rather than claims of regulatory compliance. Design artefacts and decision-making processes are critically discussed to demonstrate methodological rigour, regulatory awareness, and coherence of early-phase design. The paper evaluates how design control decisions within a collaborative design study influence concept quality, clarity of purpose, and translational readiness in medical device development.

1. Introduction

Medical device development is characterised by stringent regulatory requirements, diverse stakeholder needs, and significant implications for patient safety. Decisions made during the early phases of design are widely recognised to exert a disproportionate influence on downstream usability, regulatory compliance, and overall product success (Cross, 2008). As a result, structured design control methodologies are essential to ensure that user needs are clearly articulated and translated into coherent design requirements.

The underlying design study was conducted collaboratively by a multidisciplinary postgraduate team comprising medical engineers, clinicians, psychologists, neuroscientists, and chemical engineers. The project was intentionally scoped to early-phase design activities, including need identification, specification development, and concept selection, rather than detailed engineering design, manufacturing, or validation testing.

This individual analysis places particular emphasis on the Statement of Need, which the author authored as a controlled document and which represents the most direct articulation of end-user requirements. The author coordinated the cross-functional team through the design process and compiled its collaborative input into the formal design control documentation. The clarity achieved at the needs stage provided a stable foundation for subsequent design activities, ensuring coherence across specification development and concept selection. While the ideas and design work were undertaken collaboratively, this paper presents an individual critical analysis of the methodologies, documentation, and decision-making processes employed during the team-based design study, with a focus on early-phase design control and methodological rigour.

2. Methods: Design Control Methodology

2.1 Design Study Framework

The work was structured as a design study, consistent with established methodologies in engineering design research (Blessing and Chakrabarti, 2009). Design studies emphasise systematic generation, refinement, and evaluation of design artefacts within realistic technical, regulatory, and user constraints, making them particularly appropriate for medical device development.

Within this framework, the collaborative project progressed through structured need identification, comprehensive specification development, and comparative concept evaluation, supported by documented

project planning and communication processes. Design activities were conducted in alignment with recognised medical device design control principles, ensuring that user needs, regulatory considerations, and technical feasibility were addressed coherently throughout the early stages of development. This approach enabled the synthesis of multidisciplinary inputs and informed decision-making consistent with advanced medical device design practice.

2.2 Rationale for the Convergent Design Model

A convergent design model was selected to guide the design study. This model emphasises iterative refinement, progressive narrowing of the solution space, and continuous incorporation of stakeholder input (Cross, 2008). Such characteristics are particularly advantageous in medical device design, where user needs may initially be broad or ambiguous and require iterative clarification.

The choice was also influenced by the exploratory nature of early-phase medical device design, where both the problem definition and potential solutions evolve through engagement with stakeholders. In healthcare contexts, user needs, regulatory expectations, and practical constraints are rarely fully understood at project outset and often require iterative clarification as understanding deepens. The convergent design model accommodates this uncertainty by allowing ideas, requirements, and concepts to be refined in parallel, supporting gradual alignment rather than premature commitment to a single solution (Ulrich and Eppinger, 2016).

From a design control perspective, this iterative convergence encourages reflective decision-making, ensuring that emerging concepts remain grounded in validated user needs. Compared with linear approaches, the convergent model supports early identification of misalignment between user expectations and proposed solutions, reducing the likelihood of downstream redesign, which is particularly relevant in medical device development where late-stage changes can have significant regulatory and usability implications (Otto and Wood, 2001).

In contrast, the Waterfall model enforces a rigid, linear progression between phases, limiting opportunities for iteration and feedback; this rigidity is poorly suited to early-phase development, where premature specification freeze can result in misalignment with user expectations. Pugh's Total Design model provides a comprehensive framework for engineering projects, but its breadth and structure can introduce unnecessary complexity during early conceptual phases where flexibility and rapid convergence are critical (Pugh, 1991). The convergent design model was therefore selected as the most appropriate methodology, as it enabled continuous end-user engagement, supported iterative refinement of the Statement of Need, and facilitated informed convergence toward a solution that balanced usability, regulatory awareness, and feasibility.




Figure 2.1. Convergent design model illustrating iterative narrowing from user needs to a selected concept.

2.3 Design Control, Project Planning, and Regulatory Alignment

Design activities were structured in accordance with recognised design control principles, including user need identification, translation into design inputs, documentation, and structured review (FDA, 1997; ISO 13485:2016). Although full regulatory compliance was beyond the scope of this early-phase study, regulatory awareness consistent with EU and UK Medical Device Regulations was embedded throughout the design process.

A comprehensive self-study was undertaken, reviewing relevant literature to identify research gaps and to map the current state of knowledge associated with the project topic and end-user needs. A formal project plan was developed collaboratively to coordinate tasks, allocate responsibilities, and manage timelines, with communication conducted through a combination of face-to-face meetings and Microsoft Teams to ensure transparency of decision-making and effective multidisciplinary collaboration. User-centred design principles aligned with ISO 62366-1 were applied through structured focus group engagement, ensuring that usability considerations informed early design decisions.

3. Results and Discussion

3.1 Development of the Statement of Need

The Statement of Need was developed through structured focus group engagement with representative end-users. Qualitative data were systematically analysed and synthesised into clear, auditable need statements, consistent with best practice in medical device development (Martin et al., 2006). Figure 3.1 shows an example segment, illustrating the translation of end-user input into structured requirements.

Project Title
Medical Device Design from Concept to Clinical Use
Description of Need
A resource to guide university academics and researchers through an optimal pathway from concept design through medical regulatory approval, clinical validation, patient access, and device disposal, to a level of accuracy consistent with medical device regulatory standards, to minimise time from lab to patient.

Figure 3.1. Example segment of the Statement of Need, illustrating translation of end-user input into structured requirements.

The Statement of Need played a central role in shaping the direction and quality of the overall design study. Rather than functioning solely as a descriptive summary of user requirements, it acted as a reference point against which subsequent design decisions were evaluated. Its development required careful interpretation and synthesis of qualitative end-user input, ensuring that expressed needs were translated into clear and actionable statements (Martin et al., 2006).

Particular emphasis was placed on accessibility, usability, and the provision of clear guidance within regulated medical device contexts, ensuring that later stages of the design process remained aligned with the practical challenges faced by users. As recognised in design control guidance, robust needs definition reduces ambiguity and limits rework in later development phases (FDA, 2014). The completed Statement of Need provided a clear articulation of the problem space and defined the scope and objectives of the project, ensuring that subsequent specification development and concept selection remained coherent and aligned, and directly contributing to the quality of the XDS and the robustness of the final selected concept.

3.2 Design Specification (XDS)

A comprehensive design specification (XDS) was developed directly from the Statement of Need as a formal controlled document, translating user needs into structured design requirements consistent with professional medical device design practice. The XDS addressed functional performance, usability, regulatory considerations, data protection, sustainability, and content structure. For the purpose of structured concept selection, twelve essential specifications were selected from the wider XDS; these represented the most critical requirements for differentiating between competing concepts while preserving alignment with end-user needs and regulatory considerations (Table 3.1).

Table 3.1. Summary of selected Essential XDS used for concept evaluation.

Essential XDS	Description
Performance 1.1	Interactive infographic or printed poster showing all stages
Performance 1.4	Documentation toolkit with templates for DHF, RMF, technical files
Environment 2.1	VR simulation or video walkthrough of hospital / home use
Environment 2.8	Sustainability report or interactive dashboard with lifecycle analysis
Environment 2.15	Responsive digital platform or interactive PDF
Transport 3.8	Secure cloud repository with encryption and version control
Standards & Regulations 4.1	GDPR-compliant portal with secure login and anonymised data
Standards & Regulations 4.5	Content review workflow tool with peer-review checkpoints
Standards & Regulations 4.6	Regulatory mapping chart or compliance dashboard
Standards & Regulations 4.8	Legal compliance guide with IP protection strategies
Standards & Regulations 4.13	Sustainability compliance checklist referencing ISO 14001
Packaging 5.1	Digital dashboard or flipbook organised by development stages

3.3 Concept Generation and Selection Methodology

Five individual concepts were generated independently by team members. Concept selection employed a structured, pairwise comparison method, in which each concept was systematically compared against the twelve selected Essential XDS. The selection of twelve essential specifications enabled focused comparison while preserving alignment with user-derived requirements, and the pairwise comparison method supported transparent evaluation consistent with established engineering design selection practice (Pugh, 1991).

This approach ensured that end-user considerations continued to influence decision-making beyond the initial needs analysis stage. By grounding concept evaluation in specifications derived from the Statement

of Need, the selection process remained traceable and defensible, which is particularly important within regulated medical device development environments (ISO 13485:2016). Concepts were compared pairwise, with scores assigned based on relative performance against each specification, and cumulative scores calculated to identify the most robust solution (Table 3.2).

Table 3.2. Summary of concept selection outcomes. Idea 1: structured digital multibook; Idea 2: offline-first app with accessibility modes and regulatory smart tools; Idea 3: subscription-based app with staged regulatory pathways; Idea 4: comprehensive web platform with dashboards and VR; Idea 5: ambient intelligence browser extension.

Essential XDS	Idea 1	Idea 2	Idea 3	Idea 4	Idea 5
Spec 1.1	1	4	1	2½	1½
Spec 1.4	2	4	1	1½	1½
Spec 2.1	1	3	1	4	1
Spec 2.8	1	4	2	2½	½
Spec 2.15	2	2	2	2	2
Spec 3.8	½	3	3	1½	2
Spec 4.1	½	1½	3½	3	1½
Spec 4.5	1	3½	2½	2	1
Spec 4.6	½	3½	2½	3½	½
Spec 4.8	1	4	2	2	1
Spec 4.13	1½	4	1½	1½	1½
Spec 5.1	½	4	3	2	½
Total	11½	42½	25	29	16½

The analysis demonstrated a clear distinction between concepts, with Idea 2, an offline-first application with accessibility modes and regulatory intelligence tools, achieving a substantially higher cumulative score (42½) than the alternatives, and it was therefore selected as the primary concept. To address identified gaps, selected elements from other proposals were integrated: interactive infographics for process quality (from Idea 4), physical hard-copy playback to support non-digital use (from Idea 1), and staged pathway control with expert review to strengthen compliance assurance (from Idea 3). The resulting concept is a platform that functions as an offline-capable, accessibility-driven regulatory and design companion supporting the full medical device development lifecycle, obtainable in a hard-copy book. This hybrid solution demonstrated superior performance across accessibility, regulatory guidance, user engagement, sustainability, and digital delivery criteria when compared with the alternative concepts generated during the study.

4. Conclusion and Future Work

This paper has provided an individual critical analysis of design control methodologies applied within a team-based early-phase medical device design study. Central to the success of the project was the development of a clear and robust Statement of Need, built from the team's continuous end-user engagement and captured in the controlled documentation that the author authored and coordinated. This clarity of purpose enabled the coherent development of a comprehensive XDS and facilitated structured, defensible concept selection.

The final selected concept, an offline-capable, accessibility-driven regulatory and design companion supporting the full medical device development lifecycle, was identified as the optimal solution because it most effectively satisfied the essential design specifications derived from end-user needs. Compared with alternative concepts, it demonstrated superior alignment with accessibility requirements, regulatory guidance provision, usability, and adaptability to diverse learning and clinical contexts. This outcome

illustrates how rigorous design control, anchored in a well-developed Statement of Need, directly enhances overall concept quality.

Completed design control artefacts included the Statement of Need, design specification, team skills matrix, literature review, idea generation and selection records, project plan, focus group report, and design review meeting records, together with a consolidated list of controlled documents. Documents requiring further development include the approved supplier list and supplier contract. Future work should focus on expanding traceability documentation, conducting formal risk management activities, and progressing toward detailed design, verification, and validation in accordance with ISO 13485 and EU/UK Medical Device Regulation requirements.

References

- Blessing, L.T.M. and Chakrabarti, A. (2009) *DRM: A Design Research Methodology*. London: Springer.
- Cross, N. (2008) *Engineering Design Methods: Strategies for Product Design*. 4th edn. Chichester: John Wiley & Sons.
- European Union (2017) 'Regulation (EU) 2017/745 of the European Parliament and of the Council of 5 April 2017 on medical devices', *Official Journal of the European Union*, L117, pp. 1–175.
- Food and Drug Administration (FDA) (1997) *Design Control Guidance for Medical Device Manufacturers*. Rockville, MD: U.S. FDA.
- Food and Drug Administration (FDA) (2014) *Design Control Guidance for Medical Device Manufacturers*. Rockville, MD: U.S. FDA.
- International Organization for Standardization (ISO) (2016) *ISO 13485:2016 Medical devices: Quality management systems: Requirements for regulatory purposes*. Geneva: ISO.
- International Organization for Standardization (ISO) (2015) *ISO 62366-1:2015 Medical devices: Application of usability engineering to medical devices*. Geneva: ISO.
- Martin, J.L., Murphy, E., Crowe, J.A. and Norris, B.J. (2006) 'Capturing user requirements in medical device development', *Applied Ergonomics*, 37(5), pp. 537–547.
- Medicines and Healthcare products Regulatory Agency (MHRA) (2023) *The Medical Devices Regulations 2002 (as amended)*. London: MHRA.
- Otto, K.N. and Wood, K.L. (2001) *Product Design: Techniques in Reverse Engineering and New Product Development*. Upper Saddle River, NJ: Prentice Hall.
- Pugh, S. (1991) *Total Design*. Wokingham: Addison-Wesley.
- Ulrich, K.T. and Eppinger, S.D. (2016) *Product Design and Development*. 6th edn. New York: McGraw-Hill Education.

Appendices

The following design control artefacts were produced during the study. As the project is at an early, ongoing phase, they are maintained together as a live Technical File (design and development records) rather than a closed Design History File:

- **Appendix A:** Team Skills Matrix
- **Appendix B:** Literature Review (Self-Study)
- **Appendix C:** Project Plan
- **Appendix D:** Focus Group Report
- **Appendix E:** Design Review Meeting (V1)

- **Appendix F1:** Statement of Need
- **Appendix F2:** Statement of Need Process Flowchart
- **Appendix G1:** XDS for Medical Devices Information Resource
- **Appendix G2:** Design Specification Flowchart
- **Appendix H:** Design Review Meeting (V2)
- **Appendix I:** Updated Team Skills Matrix
- **Appendix J:** Team Conceptual Idea Document (Individual)
- **Appendix K:** Final Selected Idea: Selection Criteria
- **Appendix L:** Example Idea Proforma

These are controlled documents produced within a team. They are listed here for completeness but are not reproduced, as they contain collaborative input and are treated as confidential, consistent with good data-protection and intellectual-property practice.