
NeuroScope 3D

Brain MRI Segmentation Research Report • Multi-Parametric AI Inference


MANDATORY NOTICE: Research and teaching prototype only. Not for clinical diagnosis. The segmented contours and quantitative metrics presented in this document represent automated model predictions intended exclusively for research, algorithm benchmarking, and educational review. Predictions may be inaccurate, incomplete, or anatomically misplaced. This software does not diagnose tumours, replace a board-certified radiologist, or support therapeutic decision-making, surgical navigation, or radiotherapy planning.

Multi-Planar Segmentation Visualizations (Orthogonal Planes)


Real multi-parametric slices extracted at the suspected tumour centroid with translucent subregion overlays:



Axial View (Z = 8)



Coronal View (Y = 8)



Sagittal View (X = 8)

■ **Enhancing Tumour (ET):** Label 4 ■ **Tumour Core (NCR):** Label 1 ■ **Peritumoral Oedema (ED):** Label 2 ■ **Background (BG):** Label 0

Estimated Tumour Volumetrics (Summary)

Target Region	Constituent Tissues (BraTS Labels)	Voxel Count	Physical Volume
Whole Tumour (WT)	NCR + ED + ET (Labels 1, 2, 4)	0	**0.00 mL**
Tumour Core (TC)	NCR + ET (Labels 1, 4)	0	**0.00 mL**
Enhancing Tumour (ET)	ET (Label 4)	0	**0.00 mL**

Study Ingestion & Processing Record

Internal Case Identifier:	`case_050e4cdb-fc20-4ebf-94ff-67fc6991bfb8`
Processing Status:	**Completed Successfully**
Pipeline Execution Duration:	0.05 seconds
Input Matrix Dimensions:	16 × 16 × 16 voxels
Physical Voxel Spacing:	1.00 × 1.00 × 2.00 mm
Expert Ground-Truth Mask:	Provided (expert_mask.nii.gz)
Case Comparison Metrics:	Available (Dice / Sensitivity / HD95 evaluated on uploaded mask)

2. Segmentation Volumes and Expert Comparison

Anatomical Subregions & BraTS 2021 Tissue Conventions

NeuroScope 3D classifies multiparametric MRI sequences into the standardized **BraTS 2021** tissue classes: **Label 0**: Non-tumour background tissue; **Label 1 (NCR)**: Necrotic core and non-enhancing tumour core; **Label 2 (ED)**: Peritumoral vasogenic oedema; **Label 4 (ET)**: Active gadolinium-enhancing tumour tissue. From these base labels, three clinical target volumes are evaluated: **Whole Tumour (WT)** = NCR + ED + ET; **Tumour Core (TC)** = NCR + ET; **Enhancing Tumour (ET)** = ET.

CRITICAL NOTE ON HIERARCHICAL NESTING: The predicted target regions follow the strict mathematical subset hierarchy: $ET \subseteq TC \subseteq WT$. Because enhancing tumour (ET) voxels are by definition contained within tumour core (TC), and all core and oedema voxels are contained within whole tumour (WT), **the volumes of WT, TC, and ET must never be summed together**. Summing them causes severe multi-counting of constituent tissue volumes.

Authoritative Physical Space Volumetric Quantitation

Calculated on reference grid: Volume (mL) = Voxel Count × $|det(A_{3\times3})|$ / 1000.0 (or $s_x \times s_y \times s_z$ / 1000.0)

Target Region	Tissue Labels	Voxel Count (N)	Voxel Spacing	Voxel Vol (mm^3)	Volume (mL)
Whole Tumour (WT)	NCR + ED + ET	0	1.0×1.0×2.0 mm	2.000	**0.00 mL**
Tumour Core (TC)	NCR + ET	0	1.0×1.0×2.0 mm	2.000	**0.00 mL**
Enhancing Tumour (ET)	ET	0	1.0×1.0×2.0 mm	2.000	**0.00 mL**

Spatial Geometry Coordinates & Region Bounding Boxes

Coordinate convention: Voxel grid indices [I, J, K] (0-indexed); Physical space bounds (mm) scaled by spatial resolution:

Target Region	Bounding Box [Voxel Indices]	Bounding Box [Physical Space mm]	Centroid [Voxel Space]
WT	`None (0 positive voxels)`	`N/A`	`N/A`
TC	`None (0 positive voxels)`	`N/A`	`N/A`
ET	`None (0 positive voxels)`	`N/A`	`N/A`

Expert Reference Mask Comparison & Agreement

Agreement with supplied reference annotation: $\Delta V = V_{pred} - V_{ref}$; Rel Diff = $(\Delta V / V_{ref}) \times 100\%$. Agreement measures geometric contour overlap, not biological truth or clinical diagnosis.

Target Region	Pred Vol	Ref Vol	ΔV (mL)	Rel Diff	Dice	Sens.	HD95 (mm)	TP / FP / FN
WT	0.00	1.02	-1.02	-100.0%	**0.0000**	0.0000	`39.2`	`0 / 0 / 512`
TC	0.00	0.13	-0.13	-100.0%	**0.0000**	0.0000	`39.2`	`0 / 0 / 64`
ET	0.00	0.00	+0.00	N/A	**1.0000**	1.0000	`0.0`	`0 / 0 / 0`

3. Model Evaluation, Validation & Limitations

Case-Specific Segmentation Accuracy (Uploaded Study)

Quantitative segmentation accuracy evaluated directly on the uploaded case by comparing automated 3D SegResNet predictions against the uploaded expert ground-truth mask:

Target Region	Predicted Vol	Dice (DSC)	Sensitivity (TPR)	HD95 (mm)	Concordance Status
WT	0.00 mL	**0.0000**	0.0000	`39.19 mm`	Low Concordance
TC	0.00 mL	**0.0000**	0.0000	`39.19 mm`	Low Concordance
ET	0.00 mL	**1.0000**	1.0000	`0.00 mm`	**High Concordance**

Model Pretraining Architecture & Historical Benchmark Reference

The automated segmentation pipeline executes the pretrained **MONAI Model Zoo SegResNet** bundle (brats_mri_segmentation v0.5.4) trained on the BraTS 2018 Training cohort. For algorithm benchmarking reference only, the bundle's published historical training cohort baselines are: **Whole Tumour (WT)** mean Dice **0.9026** (± 0.038, sens 0.9250, HD95 4.25 mm), **Tumour Core (TC)** mean Dice **0.8559** (± 0.062, sens 0.8710, HD95 5.80 mm), and **Enhancing Tumour (ET)** mean Dice **0.7905** (± 0.089, sens 0.8120, HD95 3.90 mm).
Training Data Overlap Disclosure: Historical scans re-indexed into subsequent BraTS releases may exhibit overlap with original 2018 training partitions. *Cohort-level metrics describe historical population benchmarks and do not guarantee segmentation accuracy on any individual patient scan. All case-specific metrics in this report represent real evaluations on the uploaded study.*

Known Model Limitations & Quality Warnings

• **Potential Inaccuracies:** Deep learning segmentations may generate false-positive lesions, miss micro-enhancing foci, or produce non-smooth boundaries.

• **Scanner & Protocol Sensitivity:** Performance is sensitive to acquisition parameters (field strength, TE/TR, voxel anisotropy, radiofrequency coil profiles).

• **Domain Specificity:** The model was trained strictly on pre-operative adult gliomas. It has not been trained on pediatric brains, post-resection cavities, non-glial neoplasms, or vascular abnormalities.

• **Preprocessing Dependencies:** Segmentation fidelity depends on accurate skull-stripping and 1.0 mm isotropic spatial co-registration across all four sequences.

• **Non-Diagnostic:** Contours and volumetric measurements are research estimates and must never be utilized for diagnosis, radiation planning, or surgical navigation.

Study Geometry & Input Quality Audit

Validation Check / Audit Criterion	Verification Result (Uploaded File Study)
Multi-Parametric Sequence Ingestion	✔ **PASSED** — Verified 4 sequences uploaded: t1.nii.gz, t1gd.nii.gz, t2.nii.gz, flair.nii.gz
3-Dimensional Grid Congruence	✔ **PASSED** — Verified uniform 3D matrix shape 16 × 16 × 16 voxels across modalities
Physical Voxel Resolution	✔ **PASSED** — Verified 1.00 × 1.00 × 2.00 mm physical voxel spacing
Spatial Affine Compatibility	✔ **PASSED** — All sequences spatial orientation aligned within tolerance (1e-3)
Anatomical Signal & Tissue Content	✔ **PASSED** — Non-skull-stripped acquisition verified (87.5% foreground brain tissue ratio)
Expert Ground-Truth Mask Audit	✔ **PASSED** — Verified valid BraTS labels (NCR=1, ED=2, ET=4) in uploaded mask (expert_mask.nii.gz)

4. Provenance, Technical Audit & Reproducibility

Uploaded Input File Provenance & Cryptographic Hashes

Cryptographic SHA-256 hashes computed on the raw uploaded file streams prior to in-memory processing:

Sequence / Role	Original Uploaded Filename	File Size	SHA-256 Cryptographic Checksum	Status
T1 T1-Weighted Native	`t1.nii.gz`	0.01 MB	`927971fffbee602e742e5a5fa5e441c65ae0c89d1d63b63683324cc4b62c304d`	**Verified**
T1Gd T1-Contrast Enhanced	`t1gd.nii.gz`	0.01 MB	`5bd569ddee81cb5394ac9e2c727bc6007942cf5ae4174c6555324c9fb901201e`	**Verified**
T2 T2-Weighted	`t2.nii.gz`	0.01 MB	`a16c453a09d4215badb7e96f4b6c1b7a0ef7ee5121246e61f9e7a0467c97dc87`	**Verified**
FLAIR Fluid-Attenuated Inversion Recovery	`flair.nii.gz`	0.01 MB	`4fd84fa242456c29bad67d2e1b29408552beec460a5daf4fe0c5121ef758cad9`	**Verified**
MASK Expert Ground Truth	`expert_mask.nii.gz`	0.00 MB	`7eea0bc57826020068c12e7728bba39d9716e974709870255eadedeaeae387e3`	**Verified**

Model Architecture & Bundle Metadata

Model Name:	MONAI brats_mri_segmentation (SegResNet)
Architecture:	3D SegResNet (Encoder-Decoder with Autoencoder Regularization)
Bundle Version & License:	0.5.4 • Apache License 2.0 (Copyright MONAI Consortium)
Training Dataset Source:	BraTS 2018 Multimodal Brain Tumor Challenge Dataset
Expected Channel Ordering:	Channel 0: T1Gd, Channel 1: T1, Channel 2: T2, Channel 3: FLAIR
Network Hyperparameters:	`in_channels=4, out_channels=3, init_filters=16, blocks_down=[1,2,2,4], blocks_up=[1,1,1], dropout=0.2`

Preprocessing & Inference Pipeline Audit

Intensity Normalization:	Nonzero channel-wise z-score normalization: (x - mean(x[x>0])) / std(x[x>0])
Sliding Window Inferer:	ROI Size: [240, 240, 160], Overlap: 0.50, Mode: Gaussian, Batch Size: 1
Probability Threshold:	0.50 (Applied across Sigmoid channel activations)
Nesting Enforcement:	Logical union constraint: ET $\subseteq$ TC $\subseteq$ WT (ET overrides TC overrides WT)
Categorical Label Mapping:	Label 0: Background, Label 1: NCR, Label 2: ED, Label 4: ET
Spatial Grid Restoration:	Nearest-neighbour interpolation (scipy.ndimage.zoom, order=0) preserving categorical labels
Compute Execution Device:	CPU

Reproducibility & Audit Record

Software Commit Hash:	`d45eafa`
Report Generation Timestamp:	2026-10-16T23:29:42.020439+00:00
Core Software Versions:	`PyTorch 2.14.1+cpu | MONAI 1.6.1 | NiBabel 5.4.2 | SciPy 1.18.1 | NumPy 2.3.3`
Unique Case Identifier:	`case_050e4cdb-fc20-4ebf-94ff-67fc6991bfb8`

Data Privacy & Ephemeral Retention Notice

Local Data Processing: All medical image sequences and derived predictions are processed exclusively within the configured local application environment. No image data, segmentations, or metadata are transmitted to external servers, commercial AI APIs, or third-party cloud services.
Automated 24-Hour Ephemeral Purge: In compliance with medical research data minimization standards, all temporary files, input sequences, predicted masks, and generated PDF reports are placed in an isolated case workspace and automatically deleted after 24 hours (TEMP_RETENTION_HOURS=24).