

## CERTIFICADO PATRON DE CND


**TEND** 

NDT Reference Standard Certificate



2c43





Valores Pedidos / Requested Values

Valores Medidos / Measured Values

Posicion Position	Tipo de Notch Notch Type	Ancho / Diam Max [mm] Max Width / Diam	Profundidad Nominal [%] Depth	Tolerancia [%] Tolerance	Profundidad Maxima [mm] Max Depth	Profundidad Minima [mm] Min Depth	Profundidad Nominal [mm] Nom Depth	Longitud Maxima [mm] Max Length	[mm] Length	Profundida d A [mm] Depth A	Profundida d B [mm] Depth B	Ancho / Diam [mm] Width / Diam	Responsable Responsible	Fecha Date
Ι Δ	Agujero Fondo Plano Flat bottom hole	1,50	0	±15	0,000	0,000	0,000	-	-	0,000	-	1,50	Sebastian Monteros	27/08/04
В	Longitudinal Externo longitudinal OD	1,00	5	±15	0,322	0,238	0,280	38,10	38,10	5,000	5,000	1,00	Sebastian Monteros	27/08/04
С	Longitudinal Externo longitudinal OD	1,00	10	±15	0,644	0,476	0,560	38,10	38,10	10,000	10,000	1,00	Sebastian Monteros	27/08/04
D	Longitudinal Interno Longitudinal ID	1,00	13	±15	0,805	0,595	0,700	38,10	38,10	12,500	12,500	1,00	Sebastian Monteros	27/08/04

## Instrumentos de Medicion Utilizados:

## Notas:

1.-Todas las dimensiones son en milimetros excepto que se indique lo contrario.

2.-Minima Profundidad Nominal de las ranuras 0.3 mm.

CERTIFICADO CON REPLICA NO

## Notes:

- 1.-All dimensions in milimeters unless stated.
- 2.-Minimum notch nominal depth 0.3 mm.

CERTIFIED WITH PLASTIC REPLICA NO