


CERTIFICADO PATRON DE CND

TEND

NDT Reference Standard Certificate







Valores Pedidos / Requested Values

Valores Medidos / Measured Values

Posicion Position	Notch Type	Ancho / Diam Max [mm] Max Width / Diam	Profundidad Nominal [%] Depth				Profundidad Nominal [mm] Nom Depth	Longitud Maxima [mm] Max Length	Longitud [mm] Length	Profundida d A [mm] Depth A	Profundida d B [mm] Depth B	Ancho / Diam [mm] Width / Diam	Responsable Responsible	Fecha Date
Δ	Longitudinal Interno Longitudinal ID	1,00	10	±15	0,460	0,340	0,400	38,00	38,00	0,458	0,452	0,97	Daniel Carcacha	11/12/15

Instrumentos de Medicion Utilizados:

Proyector de Perfiles Mitutoyo PJ-A3000 Codigo: 302-711A

Calibre Pie de Rey Codigo:

Notas:

1.-Todas las dimensiones son en milimetros excepto que se indique lo contrario.

2.-Minima Profundidad Nominal de las ranuras 0.3 mm.

CERTIFICADO CON REPLICA SI

Notes:

- 1.-All dimensions in milimeters unless stated.
- 2.-Minimum notch nominal depth 0.3 mm.

CERTIFIED WITH PLASTIC REPLICA YES

1/2