

## CERTIFICADO PATRON DE CND


**TEND** 

NDT Reference Standard Certificate



1C1936





Valores Pedidos / Requested Values

Valores Medidos / Measured Values

Posicion Position	Tipo de Notch Notch Type	Ancho / Diam Max [mm] Max Width / Diam	Nominal [%]		Profundidad Maxima [mm] Max Depth	Profundidad Minima [mm] Min Depth	Nom Depth	Longitud Maxima [mm] Max Length	[mm] Length	Profundida d A [mm] Depth A	Profundida d B [mm] Depth B	Ancho / Diam [mm] Width / Diam	Responsable Responsible	Fecha Date
В	Longitudinal Externo longitudinal OD	1,00	5	±15	0,460	0,340	0,400	50,80	47,50	0,459	0,456	0,84	Daniel Carcacha	15/02/10
С	Transversal Externo Transverse OD	1,00	5	±15	0,460	0,340	0,400	50,80	0,40	0,460	0,448	0,86	Daniel Carcacha	15/02/10

## Instrumentos de Medicion Utilizados:

## Notas:

1.-Todas las dimensiones son en milimetros excepto que se indique lo contrario.

2.-Minima Profundidad Nominal de las ranuras 0.3 mm.

CERTIFICADO CON REPLICA NO

## Notes:

- 1.-All dimensions in milimeters unless stated.
- 2.-Minimum notch nominal depth 0.3 mm.

CERTIFIED WITH PLASTIC REPLICA NO