

## CERTIFICADO PATRON DE CND


**TEND** 

NDT Reference Standard Certificate

Tubo patrón número
Reference standard #

2c129





Valores Pedidos / Requested Values

Valores Medidos / Measured Values

Posicion Position	Tipo de Notch Notch Type	Ancho / Diam Max [mm] Max Width / Diam	Profundidad Nominal [%] Depth		Profundidad Maxima [mm] Max Depth	Profundidad Minima [mm] Min Depth	Profundidad Nominal [mm] Nom Depth	Longitud Maxima [mm] Max Length	Longitud [mm] Length	Profundida d A [mm] Depth A		Ancho / Diam [mm] Width / Diam	Responsable Responsible	Fecha Date
Α	Transversal Externo Transverse OD	1,00	5	±15	0,869	0,642	0,756	38,10	38,00	0,765	0,760	0,99	Sebastian Monteros	31/08/04
В	Longitudinal Externo longitudinal OD	1,00	5	±15	0,869	0,642	0,756	38,10	38,00	0,752	0,748	0,99	Sebastian Monteros	31/08/04
С	Transversal Interno Transverse ID	1,00	10	±15	1,738	1,284	1,511	38,10	38,50	1,568	1,489	0,99	Sebastian Monteros	31/08/04

## Instrumentos de Medicion Utilizados:

## Notas:

1.-Todas las dimensiones son en milimetros excepto que se indique lo contrario.

2.-Minima Profundidad Nominal de las ranuras 0.3 mm.

CERTIFICADO CON REPLICA NO

## Notes:

- 1.-All dimensions in milimeters unless stated.
- 2.-Minimum notch nominal depth 0.3 mm.

CERTIFIED WITH PLASTIC REPLICA NO