

CERTIFICADO PATRON DE CND


TEND

NDT Reference Standard Certificate

Tubo patrón número
Reference standard #

2u186





Valores Pedidos / Requested Values

Valores Medidos / Measured Values

Posicion Position	Tipo de Notch Notch Type	Ancho / Diam Max [mm] Max Width / Diam	Profundidad Nominal [%] Depth	Tolerancia [%] Tolerance	Profundidad Maxima [mm] Max Depth	Profundidad Minima [mm] Min Depth	Profundidad Nominal [mm] Nom Depth	Longitud Maxima [mm] Max Length	Longitud [mm] Length	Profundida d A [mm] Depth A	Profundida d B [mm] Depth B	Ancho / Diam [mm] Width / Diam	Responsable Responsible	Fecha Date
Α	Transversal Interno Transverse ID	1,00	13	±15	0,920	0,680	0,800	50,80	46,00	0,907	0,880	0,96	Sebastian Monteros	23/06/04
В	Longitudinal Externo longitudinal OD	1,00	13	±15	0,920	0,680	0,800	50,80	46,00	0,879	0,867	0,95	Sebastian Monteros	23/06/04
С	Transversal Externo Transverse OD	1,00	13	±15	0,920	0,680	0,800	50,80	47,00	0,894	0,881	0,99	Sebastian Monteros	23/06/04
D	Longitudinal Interno Longitudinal ID	1,00	13	±15	0,920	0,680	0,800	50,80	49,00	0,920	0,915	0,95	Sebastian Monteros	23/06/04
F	Chavetero Grove	10,00	50	±15	3,680	2,720	3,200	10,00	10,00	3,000	3,000	10,00	Sebastian Monteros	23/06/04
G	Longitudinal Externo longitudinal OD	1,00	13	±15	0,920	0,680	0,800	50,80	48,00	0,863	0,890	1,00	Sebastian Monteros	23/06/04
Н	Transversal Interno Transverse ID	1,00	13	±15	0,920	0,680	0,800	50,80	45,00	0,815	0,890	0,96	Sebastian Monteros	23/06/04
I	Agujero Fondo Plano Flat bottom hole	6,35	50	±15	3,680	2,720	3,200		•	3,000	-	6,00	Sebastian Monteros	23/06/04

Notas:

1.-Todas las dimensiones son en milimetros excepto que se indique lo contrario.

2.-Minima Profundidad Nominal de las ranuras 0.3 mm.

CERTIFICADO CON REPLICA NO

Notes:

- 1.-All dimensions in milimeters unless stated.
- 2.-Minimum notch nominal depth 0.3 mm.

CERTIFIED WITH PLASTIC REPLICA NO

OT-PMANT | Fecha | 23/06/2004 | Hoja | 1/2 | Documento | TAND-001-05 | Date | 11:36:49 a.m. | Page | Vigencia | 17/04/2006



CERTIFICADO PATRON DE CND

TEND

TAND-001-05

17/04/2006

NDT Reference Standard Certificate

Tubo patrón número
Reference standard #

2u186

	Datos del Producto Product Information									
Diametro [mm] Diameter	219,1	Cliente Client	Ciclo Product	72578						
Espesor [mm] Wall Thickness	6,4	Tratamiento Termico Heat Treatment	Colada _{Heat}	0						
Trazabilidad Traceability	0		Expediente Manufacturing Order							

Instrumentos de Medicion Utilizados:

Notas:

1.-Todas las dimensiones son en milimetros excepto que se indique lo contrario.

2.-Minima Profundidad Nominal de las ranuras 0.3 mm.

CERTIFICADO CON REPLICA NO

Notes:

1.-All dimensions in milimeters unless stated.

2.-Minimum notch nominal depth 0.3 mm.

CERTIFIED WITH PLASTIC REPLICA NO