


CERTIFICADO PATRON DE CND

TEND

NDT Reference Standard Certificate



Datos del Producto Product Information Ciclo Product Cliente Diametro [mm] 33,4 0 Diameter Espesor [mm]
Wall Thickness Tratamiento Termico Colada 3,75 0 **Heat Treatment** Heat Trazabilidad Expediente Manufacturing Order 0 Traceability



Valores Pedidos / Requested Values

Valores Medidos / Measured Values

sicion sition	Notch Type	Ancho / Diam Max [mm] Max Width / Diam	Profundidad Nominal [%] Depth			Profundidad Minima [mm] Min Depth	Profundidad Nominal [mm] Nom Depth	Longitud Maxima [mm] Max Length	Longitud [mm] Length	Profundida d A [mm] Depth A	Profundida d B [mm] Depth B	Ancho / Diam [mm] Width / Diam	Responsable Responsible	Fecha Date
Δ	Longitudinal Externo longitudinal OD	0,60	13	±15	0,539	0,398	0,469	100,00	100,00	12,500	12,500	0,60	Sebastian Monteros	15/12/06

Instrumentos de Medicion Utilizados:

Notas:

1.-Todas las dimensiones son en milimetros excepto que se indique lo contrario.

2.-Minima Profundidad Nominal de las ranuras 0.3 mm.

CERTIFICADO CON REPLICA SI

Notes:

- 1.-All dimensions in milimeters unless stated.
- 2.-Minimum notch nominal depth 0.3 mm.

CERTIFIED WITH PLASTIC REPLICA YES

1/2