

CERTIFICADO PATRON DE CND


TEND

NDT Reference Standard Certificate



1E68

Datos del Producto Product Information								
Diametro [mm]	33,4	Cliente Client	Ciclo Product	0				
Espesor [mm] Wall Thickness	3,75	Tratamiento Termico Heat Treatment	Colada _{Heat}	0				
Trazabilidad Traceability	0		Expediente Manufacturing Order					



Valores Pedidos / Requested Values

Valores Medidos / Measured Values

Posicion Position	· ·	Ancho / Diam Max [mm] Max Width / Diam	Nominal [%]		Profundidad Maxima [mm] Max Depth	Profundidad Minima [mm] Min Depth	Profundidad Nominal [mm] Nom Depth	Longitud Maxima [mm] Max Length	Longitud [mm] Length	Profundida d A [mm] Depth A	Profundida d B [mm] Depth B	Ancho / Diam [mm] Width / Diam	Responsable Responsible	Fecha Date
Α	Longitudinal Externo longitudinal OD	10,00	1	±15	0,043	0,032	0,038	100,00	100,00	1,000	1,000	10,00	Sebastian Monteros	15/12/06
В	Longitudinal Externo longitudinal OD	1,00	10	±15	0,431	0,319	0,375	100,00	100,00	10,000	10,000	1,00	Sebastian Monteros	15/12/06

Instrumentos de Medicion Utilizados:

Notas:

1.-Todas las dimensiones son en milimetros excepto que se indique lo contrario.

2.-Minima Profundidad Nominal de las ranuras 0.3 mm.

CERTIFICADO CON REPLICA NO

Notes:

1.-All dimensions in milimeters unless stated.

1/2

2.-Minimum notch nominal depth 0.3 mm.

CERTIFIED WITH PLASTIC REPLICA NO