

## CERTIFICADO PATRON DE CND

**TEND** 

TAND-001-05


17/04/2006

NDT Reference Standard Certificate

Tubo patrón número
Reference standard #

2c208

Datos del Producto Product Information								
Diametro [mm]	273,1	Cliente Client	Ciclo Product	73320				
Espesor [mm] Wall Thickness	11,1	Tratamiento Termico	Colada <sub>Heat</sub>	0				
Trazabilidad Traceability	0		Expediente Manufacturing Order					



Valores Pedidos / Requested Values

Valores Medidos / Measured Values

Posicion Position	Tipo de Notch Notch Type	Ancho / Diam Max [mm] Max Width / Diam	Profundidad Nominal [%] Depth	Tolerancia [%] Tolerance	Profundidad Maxima [mm] Max Depth	Profundidad Minima [mm] Min Depth	Profundidad Nominal [mm] Nom Depth	Longitud Maxima [mm] Max Length	Longitud [mm] Length	Profundida d A [mm] Depth A	Profundida d B [mm] Depth B	Ancho / Diam [mm] Width / Diam	Responsable Responsible	Fecha Date
Α	Longitudinal Interno Longitudinal ID	1,00	5	±15	0,638	0,472	0,555	50,00	49,00	0,600	0,550	0,94	Sebastian Monteros	02/09/04
В	Longitudinal Externo longitudinal OD	1,00	5	±15	0,638	0,472	0,555	50,00	50,00	0,500	0,530	0,93	Sebastian Monteros	02/09/04
С	Transversal Externo Transverse OD	1,00	5	±15	0,638	0,472	0,555	50,00	46,00	0,560	0,580	0,92	Sebastian Monteros	02/09/04
D	Transversal Interno Transverse ID	1,00	5	±15	0,638	0,472	0,555	50,00	45,00	0,630	0,620	0,94	Sebastian Monteros	02/09/04

## Instrumentos de Medicion Utilizados:

## Notas:

1.-Todas las dimensiones son en milimetros excepto que se indique lo contrario.

2.-Minima Profundidad Nominal de las ranuras 0.3 mm.

CERTIFICADO CON REPLICA NO

## Notes:

- 1.-All dimensions in milimeters unless stated.
- 2.-Minimum notch nominal depth 0.3 mm.

CERTIFIED WITH PLASTIC REPLICA NO