

## CERTIFICADO PATRON DE CND


**TEND** 

NDT Reference Standard Certificate

Tubo patrón número
Reference standard #

2c148

Datos del Producto Product Information Ciclo Product Cliente Client Diametro [mm] 244.5 0 Diameter Espesor [mm]
Wall Thickness Tratamiento Termico Colada Heat 7,92 0 **Heat Treatment** Trazabilidad Expediente Manufacturing Order 0 Traceability



Valores Pedidos / Requested Values

Valores Medidos / Measured Values

Posicion Position	Tipo de Notch Notch Type	Ancho / Diam Max [mm] Max Width / Diam	Profundidad Nominal [%] Depth	Tolerancia [%] Tolerance	Profundidad Maxima [mm] Max Depth	Profundidad Minima [mm] Min Depth	Profundidad Nominal [mm] Nom Depth	Longitud Maxima [mm] Max Length	Longitud [mm] Length	Profundida d A [mm] Depth A	Profundida d B [mm] Depth B	Ancho / Diam [mm] Width / Diam	Responsable Responsible	Fecha Date
Α	Longitudinal Interno Longitudinal ID	1,00	5	±15	0,455	0,337	0,396	38,10	37,60	0,398	1,000	0,99	Sebastian Monteros	01/09/04
В	Transversal Interno Transverse ID	1,00	5	±15	0,455	0,337	0,396	38,10	37,90	0,388	0,399	0,99	Sebastian Monteros	01/09/04
С	Longitudinal Externo longitudinal OD	1,00	5	±15	0,455	0,337	0,396	38,10	38,00	0,415	0,409	0,99	Sebastian Monteros	01/09/04
D	Transversal Externo Transverse OD	1,00	5	±15	0,455	0,337	0,396	38,10	38,00	0,375	0,367	0,99	Sebastian Monteros	01/09/04
Е	Longitudinal Externo longitudinal OD	1,00	13	±15	1,139	0,842	0,990	38,10	38,10	1,056	1,068	0,99	Sebastian Monteros	01/09/04
F	Longitudinal Interno Longitudinal ID	1,00	13	±15	1,139	0,842	0,990	38,10	38,10	1,068	1,079	0,99	Sebastian Monteros	01/09/04
G	Longitudinal Externo longitudinal OD	1,00	10	±15	0,911	0,673	0,792	38,00	37,00	0,910	0,907	0,98	Default Default	20/06/11
Н	Longitudinal Interno Longitudinal ID	1,00	10	±15	0,911	0,673	0,792	38,00	37,00	0,904	0,901	0,97	Default Default	20/06/11

#### Notas:

1.-Todas las dimensiones son en milimetros excepto que se indique lo contrario.

2.-Minima Profundidad Nominal de las ranuras 0.3 mm.

CERTIFICADO CON REPLICA NO

#### Notes:

- 1.-All dimensions in milimeters unless stated.
- 2.-Minimum notch nominal depth 0.3 mm.

CERTIFIED WITH PLASTIC REPLICA NO



# CERTIFICADO PATRON DE CND

**TEND** 

NDT Reference Standard Certificate

Tubo patrón número
Reference standard #

2c148

Datos del Producto Product Information									
Diametro [mm] Diameter	244,5	Cliente Client	Ciclo Product	0					
Espesor [mm] Wall Thickness	7,92	Tratamiento Termico Heat Treatment	Colada <sub>Heat</sub>	0					
Trazabilidad Traceability	0		Expediente Manufacturing Order						

### Instrumentos de Medicion Utilizados:

Notas:

1.-Todas las dimensiones son en milimetros excepto que se indique lo contrario.

2.-Minima Profundidad Nominal de las ranuras 0.3 mm.

CERTIFICADO CON REPLICA NO

Notes:

1.-All dimensions in milimeters unless stated.

2.-Minimum notch nominal depth 0.3 mm.

CERTIFIED WITH PLASTIC REPLICA NO