

## CERTIFICADO PATRON DE CND


**TEND** 

NDT Reference Standard Certificate

Tubo patrón número Reference standard #

1U0437

Datos del Producto Product Information									
Diametro [mm] Diameter	93,17	Cliente Client	DEMO	Ciclo Product	27080				
Espesor [mm] Wall Thickness	12,45	Tratamiento Termico	DEMO	Colada <sub>Heat</sub>	38478				
Trazabilidad Traceability	0			Expediente Manufacturing Order	DEMO				



Valores Pedidos / Requested Values

Valores Medidos / Measured Values

Posicion Position	Tipo de Notch Notch Type	Ancho / Diam Max [mm] Max Width / Diam	Profundidad Nominal [%] Depth			Profundidad Minima [mm] Min Depth	Profundidad Nominal [mm] Nom Depth	Longitud Maxima [mm] Max Length	Longitud [mm] Length	Profundida d A [mm] Depth A		Ancho / Diam [mm] Width / Diam	Responsable Responsible	Fecha Date
ι Δ	Longitudinal Interno Longitudinal ID	1,00	5	±15	0,716	0,529	0,623	38,10	-	-	-	-	Daniel Carcacha	10/05/10
В	Transversal Interno Transverse ID	1,00	5	±15	0,716	0,529	0,623	38,10	-	-	-	-	Daniel Carcacha	10/05/10
D	Longitudinal Externo Iongitudinal OD	1,00	5	±15	0,716	0,529	0,623	25,40	-	-	-	-	Daniel Carcacha	10/05/10

## Instrumentos de Medicion Utilizados:

## Notas:

1.-Todas las dimensiones son en milimetros excepto que se indique lo contrario.

2.-Minima Profundidad Nominal de las ranuras 0.3 mm.

CERTIFICADO CON REPLICA SI

## Notes:

- 1.-All dimensions in milimeters unless stated.
- 2.-Minimum notch nominal depth 0.3 mm.

CERTIFIED WITH PLASTIC REPLICA YES