

CERTIFICADO PATRON DE CND


TEND

NDT Reference Standard Certificate

Tubo patrón número
Reference standard #

1C3086





Valores Pedidos / Requested Values

Valores Medidos / Measured Values

Posicion Position	Tipo de Notch Notch Type	Ancho / Diam Max [mm] Max Width / Diam	Profundidad Nominal [%] Depth	Tolerancia [%] Tolerance	Profundidad Maxima [mm] Max Depth	Profundidad Minima [mm] Min Depth	Profundidad Nominal [mm] Nom Depth	Longitud Maxima [mm] Max Length	Longitud [mm] Length	Profundida d A [mm] Depth A	Profundida d B [mm] Depth B	Ancho / Diam [mm] Width / Diam	Responsable Responsible	Fecha Date
Α	Longitudinal Externo longitudinal OD	1,00	10	±15	1,633	1,207	1,420	50,00	50,00	1,295	1,397	1,00	Agustin Torres	09/06/20
В	Transversal Externo Transverse OD	1,00	10	±15	1,633	1,207	1,420	50,00	35,00	1,295	1,219	1,00	Agustin Torres	09/06/20
1 (-	Longitudinal Externo longitudinal OD	1,00	10	±15	1,633	1,207	1,420	50,00	50,00	1,371	1,397	1,00	Agustin Torres	09/06/20
D	Transversal Externo Transverse OD	1,00	10	±15	1,633	1,207	1,420	50,00	30,00	1,219	1,346	1,00	Agustin Torres	09/06/20
Е	Longitudinal Externo longitudinal OD	1,00	10	±15	1,633	1,207	1,420	50,00	50,00	1,244	1,320	1,00	Agustin Torres	09/06/20
F	Transversal Externo Transverse OD	1,00	10	±15	1,633	1,207	1,420	50,00	35,00	1,625	1,207	1,00	Agustin Torres	09/06/20

Notas:

- 1.-Todas las dimensiones son en milimetros excepto que se indique lo contrario.
- 2.-Minima Profundidad Nominal de las ranuras 0.3 mm.

CERTIFICADO CON REPLICA NO

Notes:

- 1.-All dimensions in milimeters unless stated.
- 2.-Minimum notch nominal depth 0.3 mm.
- CERTIFIED WITH PLASTIC REPLICA NO

OT-PMANT

Fecha Date 09/06/2020 08:04:09 a.m.

Hoja Page 1/2 Documento Vigencia

TAND-001-05 17/04/2006



CERTIFICADO PATRON DE CND

TEND

NDT Reference Standard Certificate

Tubo patrón número Reference standard #

1C3086

	Datos del Producto Product Information									
Diametro [mm]	63,5	Cliente Client	Metalmecanic a	Ciclo Product	27673					
Espesor [mm] Wall Thickness	14,2	Tratamiento Termico	SI	Colada _{Heat}	28712					
Trazabilidad Traceability	0			Expediente Manufacturing Order	5/1704.06					

Instrumentos de Medicion Utilizados:

Comparador a cuchilla Codigo:

Notas:

1.-Todas las dimensiones son en milimetros excepto que se indique lo contrario.

2.-Minima Profundidad Nominal de las ranuras 0.3 mm.

CERTIFICADO CON REPLICA NO

Notes:

1.-All dimensions in milimeters unless stated.

2.-Minimum notch nominal depth 0.3 mm.

CERTIFIED WITH PLASTIC REPLICA NO