

CERTIFICADO PATRON DE CND


TEND

NDT Reference Standard Certificate



1c590

Datos del Producto Product Information								
Diametro [mm]	114,3	Cliente Client	Ciclo Product	72627				
Espesor [mm] Wall Thickness	8,56	Tratamiento Termico Heat Treatment	Colada _{Heat}	0				
Trazabilidad Traceability	0		Expediente Manufacturing Order					



Valores Pedidos / Requested Values

Valores Medidos / Measured Values

	·													
Posicion Position	Tipo de Notch Notch Type	Ancho / Diam Max [mm] Max Width / Diam	Profundidad Nominal [%] Depth		Profundidad Maxima [mm] Max Depth	Profundidad Minima [mm] Min Depth		Longitud Maxima [mm] Max Length	Longitud [mm] Length	Profundida d A [mm] Depth A		Ancho / Diam [mm] Width / Diam	Responsable Responsible	Fecha Date
ι Δ	Agujero Pasante Drilled hole	1,60	-	-	-	-	-	-	-	-	-	1,60	Sebastian Monteros	05/11/04
В	Longitudinal Externo longitudinal OD	1,00	5	±15	0,492	0,364	0,428	50,00	50,00	5,000	5,000	1,00	Sebastian Monteros	05/11/04
С	Transversal Externo Transverse OD	1,00	5	±15	0,492	0,364	0,428	50,00	50,00	5,000	5,000	1,00	Sebastian Monteros	05/11/04

Instrumentos de Medicion Utilizados:

Notas:

1.-Todas las dimensiones son en milimetros excepto que se indique lo contrario.

2.-Minima Profundidad Nominal de las ranuras 0.3 mm.

CERTIFICADO CON REPLICA NO

Notes:

1.-All dimensions in milimeters unless stated.

1/2

2.-Minimum notch nominal depth 0.3 mm.

CERTIFIED WITH PLASTIC REPLICA NO