

CERTIFICADO PATRON DE CND


TEND

NDT Reference Standard Certificate



1c588

Datos del Producto Product Information									
Diametro [mm]	60,3	Cliente Client		Ciclo Product	0				
Espesor [mm] Wall Thickness	8,75	Tratamiento Termico Heat Treatment	NO	Colada _{Heat}	0				
Trazabilidad Traceability	0			Expediente Manufacturing Order					



Valores Pedidos / Requested Values

Valores Medidos / Measured Values

Posicion Position	Tipo de Notch Notch Type	Ancho / Diam Max [mm] Max Width / Diam	Profundidad Nominal [%] Depth			Profundidad Minima [mm] Min Depth	Profundidad Nominal [mm] Nom Depth	Longitud Maxima [mm] Max Length	Longitud [mm] Length	Profundida d A [mm] Depth A		Ancho / Diam [mm] Width / Diam	Responsable Responsible	Fecha Date
ι Δ	Agujero Pasante Drilled hole	1,60	-	-	-	-	-	-	-	-	-	1,60	Sebastian Monteros	05/11/04
В	Longitudinal Externo longitudinal OD	1,00	6	±15	0,604	0,446	0,525	50,00	50,00	6,000	6,000	1,00	Sebastian Monteros	05/11/04
С	Transversal Externo Transverse OD	1,00	5	±15	0,503	0,372	0,438	50,00	50,00	5,000	5,000	1,00	Sebastian Monteros	05/11/04

Instrumentos de Medicion Utilizados:

Notas:

1.-Todas las dimensiones son en milimetros excepto que se indique lo contrario.

2.-Minima Profundidad Nominal de las ranuras 0.3 mm.

CERTIFICADO CON REPLICA NO

Notes:

- 1.-All dimensions in milimeters unless stated.
- 2.-Minimum notch nominal depth 0.3 mm.

CERTIFIED WITH PLASTIC REPLICA NO