

## CERTIFICADO PATRON DE CND


**TEND** 

NDT Reference Standard Certificate

Tubo patrón número
Reference standard #

**1T068** 





Valores Pedidos / Requested Values

Valores Medidos / Measured Values

Posicion Position	Tipo de Notch Notch Type	Ancho / Diam Max [mm] Max Width / Diam	Profundidad Nominal [%] Depth	Tolerancia [%] Tolerance	Profundidad Maxima [mm] Max Depth	Profundidad Minima [mm] Min Depth	Profundidad Nominal [mm] Nom Depth	Longitud Maxima [mm] Max Length	Longitud [mm] Length	Profundida d A [mm] Depth A	Profundida d B [mm] Depth B	Ancho / Diam [mm] Width / Diam	Responsable Responsible	Fecha Date
Α	Transversal Interno Transverse ID	1,00	5	±15	0,424	0,313	0,369	50,80	40,00	0,365	0,362	0,93	Default Default	24/02/11
В	Longitudinal Interno Longitudinal ID	-	-	-	-	-	-	-	-	-	-	-	Default Default	24/02/11
С	Longitudinal Interno Longitudinal ID	1,00	5	±15	0,424	0,313	0,369	50,80	46,00	0,392	0,410	0,96	Default Default	24/02/11
D	Transversal Externo Transverse OD	1,00	5	±15	0,424	0,313	0,369	50,80	41,00	0,410	0,405	0,92	Default Default	24/02/11
Е	Longitudinal Externo longitudinal OD	1,00	5	±15	0,424	0,313	0,369	50,80	47,00	0,387	0,386	0,96	Default Default	24/02/11
F	Longitudinal Externo longitudinal OD	1,00	10	±15	0,848	0,626	0,737	50,80	46,00	0,790	0,822	0,96	Default Default	24/02/11
G	Transversal Externo Transverse OD	1,00	10	±15	0,848	0,626	0,737	50,80	42,00	0,800	0,762	0,95	Default Default	24/02/11
Н	Longitudinal Interno Longitudinal ID	1,00	10	±15	0,848	0,626	0,737	50,80	47,00	0,784	0,777	0,93	Default Default	24/02/11

#### Notas:

- 1.-Todas las dimensiones son en milimetros excepto que se indique lo contrario.
- 2.-Minima Profundidad Nominal de las ranuras 0.3 mm.

CERTIFICADO CON REPLICA SI

#### Notes:

- 1.-All dimensions in milimeters unless stated.
- 2.-Minimum notch nominal depth 0.3 mm.
- CERTIFIED WITH PLASTIC REPLICA YES

OT-PMANT

Fecha Date

24/02/2011 07:53:15 a.m. Hoja Page 1/2 Documento Vigencia

TAND-001-05 17/04/2006



# CERTIFICADO PATRON DE CND

**TEND** 

TAND-001-05

17/04/2006

NDT Reference Standard Certificate

Tubo patrón número
Reference standard #

**1T068** 

Datos del Producto Product Information									
Diametro [mm]	114,3	Cliente Client	JUSEPIN	Ciclo Product	25147				
Espesor [mm] Wall Thickness	7,37	Tratamiento Termico	SI	Colada <sub>Heat</sub>	0				
Trazabilidad Traceability	0			Expediente Manufacturing Order					

Posicion Position	Tipo de Notch Notch Type	Ancho / Diam Max [mm] Max Width / Diam	Profundidad Nominal [%] Depth			Profundidad Minima [mm] Min Depth	Profundidad Nominal [mm] Nom Depth	Longitud Maxima [mm] Max Length	Longitud [mm] Length	Profundida d A [mm] Depth A	Profundida d B [mm] Depth B	Ancho / Diam [mm] Width / Diam	Responsable Responsible	Fecha Date
I	Transversal Interno Transverse ID	1,00	10	±15	0,848	0,626	0,737	50,80	43,00	0,783	0,710	0,98	Default Default	24/02/11
J	Defecto Natural Natural defect	-	-	-	-	-	-	-	-	-	-	-	Default Default	24/02/11

### Instrumentos de Medicion Utilizados:

Notas:

1.-Todas las dimensiones son en milimetros excepto que se indique lo contrario.

2.-Minima Profundidad Nominal de las ranuras 0.3 mm.

CERTIFICADO CON REPLICA SI

Notes:

1.-All dimensions in milimeters unless stated.

2.-Minimum notch nominal depth 0.3 mm.

CERTIFIED WITH PLASTIC REPLICA YES